
R A C V

R E S P O N S E A C T I O N C O N T R A C T F O R

Remedial, Enforcement Oversight, and
Non-Time Critical Removal Activities at Sites of Release
or Threatened Release of Hazardous Substances in Region V

**ADDENDUM I, REVISION I
QUALITY ASSURANCE PROJECT PLAN**

**ONALASKA MUNICIPAL LANDFILL
Onalaska, Wisconsin**

Long-Term Response Action

WA No. 103-RALR-05L5 / Contract No. 68-W6-0025

April 1, 2002

PREPARED FOR

U.S. Environmental Protection Agency



PREPARED BY

CH2M HILL

Ecology and Environment, Inc.

TN & Associates, Inc.

Tucker, Young, Jackson, Tull, Inc.

**ADDENDUM I, REVISION I
QUALITY ASSURANCE PROJECT PLAN**

**ONALASKA MUNICIPAL LANDFILL
Onalaska, Wisconsin**

Long-Term Response Action

WA No. 103-RALR-05L5 / Contract No. 68-W6-0025

April 1, 2002

Section 1
Title Page
Remedial Planning Activities
(RAC V)
Contract No. 68-W6-0025
Addendum I-Revision I
to the
Quality Assurance Project Plan (QAPP)

Project Title: Long-Term Response Action
Onalaska Municipal Landfill
Onalaska, Wisconsin

EPA No: WA 103-RALR-05L5

EPA Remedial Project Officer: Kevin Adler *Tim Paensville*

Prepared by: CH2M HILL Date: April 2002

Approved by: *Jim Fisher* Date: 4/1/02
Jim Fisher

Approved by: *Ike Johnson* Date: 4/1/02
Ike Johnson
CH2M HILL Site Manager

Approved by: *[Signature]* Date: 4/24/02
EPA Region 5
Remedial Project Manager

Approved by: _____ Date: _____
EPA Region 5
Quality Assurance Manager

Contents

1. Title Page	iii
2. Acronyms	ix
3. Project Description	1
3.1 Introduction	1
3.2 Site Description	1
3.3 Site History and Background	1
3.3.1 Site History	1
3.3.2 Background.....	1
3.4 Target Compounds	1
3.5 Project Objectives	2
3.5.1 Intended Data Usage.....	2
3.5.2 Data Quality Objectives	2
3.6 Groundwater Monitoring Network	4
3.6.1 Piezometers.....	4
3.6.2 Monitoring Wells	4
3.6.3 Extraction Wells	4
3.6.4 Residential Wells	4
3.6.5 Surface Water and Sediments	4
3.6.6 Background/Baseline Monitoring	4
3.6.7 Sampling	5
3.7 Project Schedule	5
4. Project Organization and Responsibility	7
4.1 Management Responsibilities	7
4.1.1 USEPA Region 5 Remedial Project Manager	9
4.1.2 State Project Manager.....	9
4.1.3 CH2M HILL Program Manager	9
4.1.4 CH2M HILL Site Manager	9
4.1.5 USEPA Region 5 Quality Assurance Manager	9
4.2 Quality Assurance Organization	9
4.3 Field Operations.....	9
4.3.1 Field Team Leaders	9
4.3.2 Field Team Members.....	9
4.4 Laboratories	9
5. Quality Assurance Objectives for Measurement Data	11
5.1 Level of Quality Control Effort	11
5.1.1 Accuracy, Precision, and Sensitivity of Analysis	11
5.1.2 Completeness, Representativeness, and Comparability	11
5.2 Method Detection Limits	11
6. Sampling Procedures	13
7. Sample Custody Procedures	15
7.1 Introduction	15
7.2 Field Custody Procedures.....	15

7.2.1	Field Procedures	15
7.2.2	Sample Documentation Procedures.....	15
7.2.3	Transfer of Custody Procedures.....	15
7.2.4	Field Logbook.....	15
7.2.5	Corrections to Documentation.....	15
7.2.6	Distribution of Completed Documents	15
7.2.7	Site Manager’s Responsibility.....	15
7.3	Laboratory Custody Procedures for the Contract Laboratory	15
7.4	Final Evidence Files Custody Procedures	16
8.	Calibration Procedures and Frequency	17
8.1	Special Analytical Services	17
8.2	Field Instruments	17
9.	Analytical Procedures	19
9.1	Special Analytical Services	19
9.2	Field Instruments	19
10.	Internal Quality Control Checks	21
10.1	Special Analytical Services	21
10.2	Field Instruments	21
11.	Data Reduction, Validation, and Reporting	23
11.1	Data Reduction.....	23
11.1.1	Laboratory Analysis	23
11.1.2	Field Measurements	23
11.2	Data Validation	23
11.2.1	Laboratory Analysis	23
11.2.2	Field Measurements	23
11.3	Data Reporting	23
11.3.1	Laboratory Analysis	23
11.3.2	Field Measurements	24
11.4	Annual Evaluation.....	24
12.	Performance and System Audits.....	25
12.1	External Audits.....	25
12.1.1	Laboratories	25
12.1.2	Field Audits	25
12.2	Internal Audits	25
12.2.1	Field Audits	25
13.	Preventive Maintenance	27
13.1	Laboratory Instruments	27
13.2	Field Instruments	27
14.	Specific Routine Procedures to Assess Data Precision, Accuracy, and Completeness..	29
14.1	Field Measurements.....	29
14.2	Laboratory Data	29
14.2.1	Precision.....	29
14.2.2	Accuracy	29
14.2.3	Completeness	29
14.3	Project Assessment.....	29
15.	Corrective Actions	31
15.1	Sample Collection/Field Measurements.....	31

15.2 Laboratory Analyses—Laboratory Corrective Actions.....	31
16. Quality Assurance Reports to Management.....	33

Tables

3 Parameters, Sample Containers, Preservatives, Sample Matrix, Storage Methods, and Holding Times.....	13
4 Sampling and Analysis Summary	14

Figures

4 Project Organization.....	8
-----------------------------	---

Attachments

A Standard Operating Procedures	
---------------------------------	--

Acronyms

ARARs	applicable or relevant and appropriate requirements
ASTM	American Society for Testing and Materials
CERCLA	Comprehensive Environmental Response, Compensation, and Liability Act
COC	chain of custody
CRL	Central Regional Laboratory
DI	deionized
DMZ	Design Management Zone
DNR	Wisconsin Department of Natural Resources
DOT	Department of Transportation
DQO	data quality objective
EM	electromagnetic
ES	Enforcement Standard
EPA	U.S. Environmental Protection Agency
FAA	furnace atomic adsorption
FS	Feasibility Study
FSP	Field Sampling Plan
GC/MS	gas chromatograph/mass spectrometer
HPLC	high-pressure liquid chromatography
ICP	inductively coupled plasma
I.D.	inner diameter
LCS	laboratory control sample
LEL	lower explosive limit
LOD	level of detection
LSSS	Laboratory Scientific Support Section
LTRA	Long-Term Response Action
MCL	maximum contaminant level
MDL	method detection limits
MS/MSD	matrix spike/matrix spike duplicate
NEIC	National Enforcement Investigations Center
NGVD	National Geodetic Vertical Datum of 1929
NIST	National Institute of Standards and Technology
NPL	National Priorities List
O.D.	outer diameter
OVA	organic vapor analyzer
PAH	polycyclic aromatic hydrocarbon
PAL	Preventative Action Limits
PARCC	precision, accuracy, representativeness, completeness, and comparability
PCB	polychlorinated biphenyl
PCE	tetrachloroethene
PE	performance evaluation
PIC	pressurized ion chamber
%R	percent recovery

PRP	potentially responsible party
PVC	polyvinyl chloride
QA	quality assurance
QAM	Quality Assurance Manager
QAPP	Quality Assurance Project Plan
QAR	Quality Assurance Reviewer
QAS	Quality Assurance Section
QC	quality control
RA	Remedial Action
RAS	Routine Analytical Services
RCRA	Resource Conservation and Recovery Act
RI	Remedial Investigation
RI/FS	Remedial Investigation/Feasibility Study
RL	reporting limits
ROD	Record of Decision
RPD	relative percent difference
SAS	Special Analytical Services
SOP	Standard Operating Procedure
SOW	statement of work
SVOC	semivolatile organic compound
TAL	Target Analyte List
TCA	trichloroethane
TCE	trichloroethene
TCL	Target Compound List
TCLP	Toxicity Characteristic Leaching Procedure
TDS	total dissolved solids
TIC	total inorganic carbon
TOC	total organic carbon
USGS	U.S. Geological Survey
VOC	volatile organic compound
WA	work assignment

SECTION 3

Project Description

3.1 Introduction

No modifications have been made.

3.2 Site Description

No modifications have been made.

3.3 Site History and Background

No modifications have been made.

3.3.1 Site History

No modifications have been made.

3.3.2 Background

No modifications have been made.

3.4 Target Compounds

Delete existing text and replace with the following:

The list of target compounds have changed to match the requirements of the *Natural Attenuation Plan*¹. Sampling requirements for the natural attenuation study, including sample containers, preservatives, sample matrix, storage methods, and holding times, are provided in Table 3.

The parameters to be monitored, monitoring locations, and sampling frequency have been adjusted each year since the groundwater monitoring began in 1995. The rationale for the changes made prior to the natural attenuation study are documented in the 1995–2000 *Annual Groundwater Quality and Capture Reports*. The recommended changes documented in these reports were reviewed and approved by the USEPA and the Wisconsin Department of Natural Resources (WDNR) prior to implementation. The parameters monitored and the monitoring locations were revised in support of the natural attenuation study initiated in the fall of 2001. The parameters to be monitored during the natural attenuation study were based on the needs of this study and on the list of previously detected parameters as provided in the *Natural Attenuation Plan*. This list includes:

- VOCs detected during the Remedial Investigation (RI);

¹ CH2M HILL. *Natural Attenuation Plan, Onalaska Municipal Landfill Site, Onalaska, Wisconsin*. September 2000.

- Parameters that are useful in evaluating the natural attenuation processes; and
- A one-time dual sampling and analysis for dissolved metals, direct field filtered using the sample pump, and total metals for a baseline comparison. Thereafter, all metals sampling and analysis will be total metals (per the recommendation of the USEPA).

Oil and grease analyses and odor analyses were removed as target compounds, with the approval of the USEPA and WDNR, prior to the spring 2001 sampling event. Oil and grease were dropped for the following reasons:

- The analysis method recommended by the USEPA changed, and the detection limit for the new method is greater than the oil and grease concentrations that have been detected at the site in recent years.
- There are no groundwater standards for oil and grease.
- Oil and grease monitoring is not needed to evaluate the effectiveness of the remedial action.

Odor was dropped because of safety concerns expressed by prospective subcontract laboratories during the subcontract bidding process (the method calls for the physical inhalation of gases coming from the water samples).

Turbidity, total dissolved solids (TDS), and color were eliminated from the list when the *Natural Attenuation Plan* was developed. These parameters were eliminated because sufficient data had been collected from 1995 through spring 2001, and based on an evaluation of the data, it was concluded that they are not needed to evaluate the effectiveness of the remedial action. Additional information regarding the rationale for parameters in the current sampling program can be found in the *Natural Attenuation Plan*.

3.5 Project Objectives

Project objectives are as described in the *Natural Attenuation Plan*.

3.5.1 Intended Data Usage

Project objectives are as described in the *Natural Attenuation Plan*.

3.5.2 Data Quality Objectives

Delete existing text for this section and replace with the following:

The overall QA objective for the project is to collect data of known and acceptable quality to support project decisions. To achieve that objective, a plan was established for representative characterization of site groundwater. The objective of this QAPP is to develop and implement procedures for sampling, decontamination, COC, laboratory analyses, instrument calibration, data reduction and reporting, internal QC, audits, preventive maintenance, and corrective action to help generate data of known and acceptable quality.

3.5.2 Data Quality Objectives

Data quality objectives (DQOs) specify the quality of data required to support decisions made during or after site-related activities. Project-specific DQOs are developed with a seven-step process:

Step 1: State the Problem

The 1997 QAPP was developed in conjunction with the remedial action, which consisted of a groundwater collection and treatment system and a bioventing system. Revisions to the 1997 QAPP are necessary to implement a natural attenuation study as described in the *Natural Attenuation Plan* developed for the site.

Step 2: Identify the Decision

The objective of this investigation is to gather the information necessary to determine if natural attenuation is an effective remedy as described in the *Natural Attenuation Plan*.

Step 3: Identify Inputs to the Decision

Additional inputs are required to supplement those collected from previous site investigations. Laboratory and field measurements will provide the additional information required to assess the site. The inputs that will be applied to the decision process include:

- Chemical analytical results from offsite laboratory analysis of site media, including VOCs, select metals, and general chemistry constituents.
- Field measurement of selected parameters described in the *Natural Attenuation Plan*.

Step 4: Define the Boundaries of the Study

The investigation areas and proposed sampling locations are shown and discussed in the *Natural Attenuation Plan*.

Step 5: Develop a Decision Rule

The analytical results, field monitoring results, and groundwater level measurements acquired during this investigation will be used to monitor the extent of contamination and to assess the effectiveness of natural attenuation for VOCs.

This information will serve as the basis for the decisionmaking process.

Step 6: Specify Limits on Decision Errors

The potential for significant decision errors is considered unlikely because the scope of the investigation is a result of extensive review of the existing data. The sampling locations are spatially and conservatively set so that sampling should be adequate to satisfactorily assess current site conditions. Sampling and monitoring results will be used to determine if additional sampling locations are appropriate as discussed in the *Natural Attenuation Plan*.

The analytical methods and laboratories to be used will provide the level of QA and defensibility to match the existing data and meet the criteria of validation as outlined in

Laboratory Data Validation Functional Guidelines for Evaluating Organic Analyses (USEPA October 1999), *Laboratory Data Validation Functional Guidelines for Evaluating Inorganic Analyses* (USEPA February 1994), and supporting documents.

Step 7: Optimize the Design

The groundwater monitoring locations will be analyzed annually, at a minimum, as described in the *Natural Attenuation Plan*. The need for or benefit of modifying the plan to improve the design will be assessed and documented in the annual report.

3.6 Groundwater Monitoring Network

Table 4 of this addendum replaces the information provided in Figure 3, referenced in the last sentence in Section 3.6.

3.6.1 Piezometers

Delete existing text from this section and replace with the following:

Piezometer information is described in the *Natural Attenuation Plan*.

3.6.2 Monitoring Wells

Delete existing text from this section and replace with the following:

Monitoring well information is described in the *Natural Attenuation Plan*.

3.6.3 Extraction Wells

Delete existing text from this section and replace with the following:

The extraction wells are no longer in operation and are not included as part of the *Natural Attenuation Plan*.

3.6.4 Residential Wells

Delete existing text from this section and replace with the following:

Residential well information is described in the *Natural Attenuation Plan*.

3.6.5 Surface Water and Sediments

No modifications have been made.

3.6.6 Background/Baseline Monitoring

The text in Section 3.6.6 should be deleted and the following paragraph shall be added to the end of Section 3.6.6:

The rationale for the changes in the *Groundwater Monitoring Plan*, including piezometers, monitoring wells, extraction wells, and residential wells, can be found in the *Annual Groundwater Quality and Capture Reports* and the *Natural Attenuation Plan*. The revised tables

showing the monitoring wells and analytes that will be analyzed are provided in the *Natural Attenuation Plan*. Sampling frequency is described in the *Natural Attenuation Plan*.

3.6.7 Sampling

Delete existing text in Section 3.6.7.

The sampling plans provided in Tables 3 and 4 replace previous sample plans.

3.7 Project Schedule

Delete the text in Section 3.7 and replace with:

Schedule related information is provided in the *Natural Attenuation Plan*.

SECTION 4

Project Organization and Responsibility

No modifications have been made.

4.1 Management Responsibilities

Delete:

Quality Assurance Manager (QAM)
Willie Harris (U.S. EPA Region 5)

Replace with:

Quality Assurance Reviewer (QAR)




FIGURE 4
PROJECT ORGANIZATION
 ONALASKA LANDFILL
 QAPP

4.1.1 USEPA Region 5 Remedial Project Manager

No modifications have been made.

4.1.2 State Project Manager

No modifications have been made.

4.1.3 CH2M HILL Program Manager

No modifications have been made.

4.1.4 CH2M HILL Site Manager

No modifications have been made.

4.1.5 USEPA Region 5 Quality Assurance Manager

Delete "SAS" and replace with "SOP."

4.2 Quality Assurance Organization

Delete:

Willie Harris, USEPA Region 5 QA Manager (QAM)

Replace with:

USEPA Region 5 Quality Assurance Reviewer (QAR)

4.3 Field Operations

Delete:

External Field Audits USEPA Region 5 CRL

Replace with:

External Field Audits USEPA Region 5

4.3.1 Field Team Leaders

No modifications have been made.

4.3.2 Field Team Members

No modifications have been made.

4.4 Laboratories

Delete all references to "SAS" and replace with "SOP."

SECTION 5

Quality Assurance Objectives for Measurement Data

No modifications have been made.

5.1 Level of Quality Control Effort

The following notation shall be added to Section 5.1:

The SOPs provided in Attachment A to this addendum replace the previously provided SASs. The QC stated in the SOPs supercedes and replaces previously listed QC regulations.

5.1.1 Accuracy, Precision, and Sensitivity of Analysis

Delete all references to "SAS" and replace with "SOP."

5.1.2 Completeness, Representativeness, and Comparability

No additional modifications have been made.

5.2 Method Detection Limits

Delete existing text from Section 5.2 and replace with:

Method detection limits can be found in SOPs.

SECTION 6

Sampling Procedures

Delete existing text in Section 6 and replace it with:

Detailed sampling procedures are provided in the *Natural Attenuation Plan*. Table 3 is a summary of sample matrixes and parameters. Table 4 is a sampling and analysis plan summary.

TABLE 3
Parameters, Sample Containers, Preservatives, Sample Matrix, Storage Methods, and Holding Times

Analysis	Method	Matrix	Container	Preservation and Storage	Maximum Hold Time
VOCs					
Benzene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Ethylbenzene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Toluene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Total Xylenes	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
1,1-Dichloroethane	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
1,1-Dichloroethene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
1,1,1-Trichloroethane	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Cis-1,2-Dichloroethene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Trans-1,2-Dichloroethene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Trichloroethene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Tetrachloroethene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Methylene Chloride	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Vinyl Chloride (Chloroethene)	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
1, 2, 4-Trimethylbenzene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
1, 3, 5-Trimethylbenzene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Naphthalene	GC/MS SW826 8260B	Water	Three 40-mL vials	HCl to pH \leq 2, 4°C	14 days
Inorganic Constituents					
Arsenic	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH \leq 2	180 days
Barium	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH \leq 2	180 days
Iron	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH \leq 2	180 days
Lead	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH \leq 2	180 days
Manganese	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH \leq 2	180 days
Cadmium	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH \leq 2	180 days
Cobalt	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH \leq 2	180 days

TABLE 3
Parameters, Sample Containers, Preservatives, Sample Matrix, Storage Methods, and Holding Times

Analysis	Method	Matrix	Container	Preservation and Storage	Maximum Hold Time
Mercury	GC/MS SW846 7470A	Water	250-mL bottles	HNO ₃ to pH _≤ 2	180 days
Vanadium	GC/MS SW846 6010B	Water	250-mL bottles	HNO ₃ to pH _≤ 2	180 days
Dissolved Oxygen	Field measurement	Water	—	—	—
ORP	Field measurement	Water	—	—	—
pH	Field measurement	Water	—	—	—
Temperature	Field measurement	Water	—	—	—
Specific Conductance	Field measurement	Water	—	—	—
Nitrate	E300	Water	100-mL polyethylene bottle	4°C	48 hours
Sulfate	E300	Water	250-mL polyethylene bottle	4°C	28 days
Methane, Ethane, Ethene	RSK-175	Water	Two 40-mL vials	HCl to pH _≤ 2, 4°C	Not determined
Alkalinity	E310.1	Water	100-mL polyethylene bottle	4°C	14 days
Chloride	E300	Water	250-mL polyethylene bottle	4°C	28 days
Total Organic Carbon	E415.1	Water	Two 40-mL vials	H ₂ SO ₄ to pH _≤ 2, 4°C	28 days

— = Not applicable

TABLE 4
Sampling and Analysis Summary

	Number of Field Samples				Number of Lab QC Samples		Total Samples	Sampling Frequency
	FS	TB	FB	FD	MS	MSD		
Group I Wells								
VOCS	15	3	2	2	1	1	24	Semiannual
SVOCS	15	0	2	2	1	1	21	Semiannual
Metals	15	3	2	2	1	1	21	Semiannual
Natural attenuation	15	3	2	2	1	1	21	Semiannual
Group II Wells								
VOCS	11	2	1	1	1	1	17	Annual
SVOCS	11	0	1	1	1	1	15	Annual
Metals	11	0	1	1	1	1	15	Annual
Natural attenuation	11	0	1	1	1	1	15	Annual

Sample Custody Procedures

7.1 Introduction

No modifications have been made.

7.2 Field Custody Procedures

No modifications have been made.

7.2.1 Field Procedures

Delete all references to "SAS" and replace with "SOP."

7.2.2 Sample Documentation Procedures

Delete the text contained in the parentheses in Step 12.

7.2.3 Transfer of Custody Procedures

No modifications have been made.

7.2.4 Field Logbook

No modifications have been made.

7.2.5 Corrections to Documentation

No modifications have been made.

7.2.6 Distribution of Completed Documents

No modifications have been made.

7.2.7 Site Manager's Responsibility

No modifications have been made.

7.3 Laboratory Custody Procedures for the Contract Laboratory

Delete the following from Section 7.3:

The laboratory chain-of-custody procedures for organic and inorganic compound SAS analyses are described in the *Contracted Laboratories Quality Assurance Plan*.

Replace with:

The contracted laboratory's chain-of-custody SOP outlines the sample custody procedures.

7.4 Final Evidence Files Custody Procedures

No modifications have been made.

SECTION 8

Calibration Procedures and Frequency

No modifications have been made.

8.1 Special Analytical Services

Delete the existing section title and replace with the following:

8.1 Standard Operating Procedures

For the laboratory SOP analysis, the calibration procedures and frequency are presented in the SOPs, provided in Attachment A to this addendum.

8.2 Field Instruments

No modifications have been made.

SECTION 9

Analytical Procedures

No modifications have been made.

9.1 Special Analytical Services

Delete the existing section title and text and replace with the following:

9.1 Standard Operating Procedures

Severn Trent Laboratory of North Canton, Ohio, will perform the analytical work according to the SOPs it has submitted. The laboratory will follow the QA/QC procedures and limits stated in the supplied SOPs. The attached SOPs will replace SASs. SASs no longer will be used in support of the project. Attachment A to this addendum contains the following laboratory SOPs:

- Gas Chromatograph/Mass Spectrometry (GC/MS) Analysis of Volatile Organics—Method 524.2 Compounds by Ion Trap
- Determination of Volatile Organics by GC/MS Based on Methods SW-846 8260B, 8260A, and 624
- Preparation and Analysis of Mercury in Aqueous Samples by Cold Vapor Atomic Absorption SW846 7470A and MCAWW 245.1
- Acid Digestion of Aqueous Samples by SW846 and MCAWW 200 Series Methods
- Analysis of Dissolved Gas in Water (RSK-176)
- Gas Chromatographic Analysis based on Methods 8000B, 8021B, 8081A, 8082, 8151A, 8310, 8141A, 8015B, 608, and 610 and on Wisconsin DNR Modified DRO Method
- Trace Element Analysis, SW-846 Method 6010B and EPA Method 200.7
- Inductively Coupled Plasma-Mass Spectrometry, EPA Methods 6020 and 200.8
- Alkalinity (Total)
- Total Organic Carbon (TOC) and Total Inorganic Carbon (TIC)
- Determination of Inorganic Anions by Ion Chromatography
- Sample Receipt and Chain of Custody (STL-QA-0006, Revision 2)
- Data Review and Verification (STL-QA-0011, Revision 2)

9.2 Field Instruments

No modifications have been made.

SECTION 10

Internal Quality Control Checks

10.1 Special Analytical Services

Delete the existing title and text for Section 10.1 and replace with the following:

10.1 Standard Operating Procedures

The analytical QC procedures are provided in the attached SOPs.

10.2 Field Instruments

No modifications have been made.

Data Reduction, Validation, and Reporting

11.1 Data Reduction

No modifications have been made.

11.1.1 Laboratory Analysis

Delete the following sentence:

Data reduction, evaluation, and reporting for samples analyzed by this laboratory will be performed according to specifications outlined in the SASs.

Replace with:

Data reduction, evaluation, and reporting for samples analyzed by this laboratory will be performed according to specifications outlined in the SOPs.

11.1.2 Field Measurements

No modifications have been made.

11.2 Data Validation

11.2.1 Laboratory Analysis

Delete the references to the Laboratory Data Validation Functional Guidelines and replace them with:

Laboratory Data Validation Functional Guidelines for Evaluating Organic Analyses—USEPA October 1999.

Laboratory Data Validation Functional Guidelines for Evaluating Inorganic Analyses—USEPA February 1994.

11.2.2 Field Measurements

No modifications have been made.

11.3 Data Reporting

11.3.1 Laboratory Analysis

Delete the existing text in section 11.3.1 and replace with the following:

The contracted laboratory's SOP outlines data reporting procedures.

11.3.2 Field Measurements

No additional modifications have been made.

11.4 Annual Evaluation

No modifications have been made.

SECTION 12

Performance and System Audits

No modifications have been made.

12.1 External Audits

No modifications have been made.

12.1.1 Laboratories

Add to the end of Section 12.1.1:

USEPA Region 5 may perform external laboratory audits.

12.1.2 Field Audits

Delete:

All field activities conducted by CH2M HILL may be subject to onsite audit by USEPA's Region 5 Central District office or CRL.

Replace with:

All field activities conducted by CH2M HILL may be subject to onsite audit by USEPA Region 5.

12.2 Internal Audits

Add the following text:

Internal audits will be performed as described in the contracted laboratory's SOPs.

12.2.1 Field Audits

No modifications have been made.

SECTION 13

Preventive Maintenance

13.1 Laboratory Instruments

Delete the existing text in Section 13.1 and replace with the following:

Information regarding laboratory instruments is specified in the contracted laboratory's SOPs.

13.2 Field Instruments

No modifications have been made.

SECTION 14

Specific Routine Procedures to Assess Data Precision, Accuracy, and Completeness

14.1 Field Measurements

No modifications have been made.

14.2 Laboratory Data

No modifications have been made.

14.2.1 Precision

Delete section title and replace with:

14.2.1 Precision and Accuracy

Add the text from Section 14.2.2 to the end of existing text in Section 14.2.1. Add the following text to the end of Section 14.2.1:

14.2.1.1 Quality Control-Accuracy and Precision

The accuracy and precision limits of the matrix spike/matrix spike duplicate (MS/MSD) samples for Method 8260 are determined with the use of control charts, which are modified almost daily. Therefore, these limits cannot be included in the SOP.

Regarding the accuracy and precision limits for a laboratory control sample (LCS) duplicate for Method 8260, if an LCS duplicate is performed (traditionally done only when a sample batch is analyzed without the inclusion of a MS/MSD), it will be evaluated using the laboratory MSD precision limits from the time period it was analyzed.

14.2.2 Accuracy

Delete section and add text to Section 14.2.1.

14.2.3 Completeness

No modifications have been made.

14.3 Project Assessment

No modifications have been made.

SECTION 15

Corrective Actions

No modifications have been made.

15.1 Sample Collection/Field Measurements

No modifications have been made.

15.2 Laboratory Analyses—Laboratory Corrective Actions

Delete the first sentence of this section.

SECTION 16

Quality Assurance Reports to Management

No modifications have been made.

Controlled Copy
Copy No: _____

SOP No: NC-WC-0084
Revision No: 3
Revision Date: 10/03/00
Page 1 of 22

Implementation Date: _____

STL STANDARD OPERATING PROCEDURE

**TITLE: DETERMINATION OF INORGANIC ANIONS BY ION
CHROMATOGRAPHY**

(SUPERSEDES: REVISION (2))

Intranet Controlled

Prepared by: _____ Date

Reviewed by: _____ Date
Technology Specialist

Approved by: _____ Date
Quality Assurance Manager

Approved by: _____ Date
Environmental Health and Safety Coordinator

Approved by: _____ Date
Laboratory Director

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use it for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

1. SCOPE AND APPLICATION.....3

2. SUMMARY OF METHOD3

3. DEFINITIONS3

4. INTERFERENCES.....3

5. SAFETY.....4

6. EQUIPMENT AND SUPPLIES4

7. REAGENTS AND STANDARDS.....5

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE..... 11

9. QUALITY CONTROL..... 12

10. CALIBRATION AND STANDARDIZATION 14

11. PROCEDURE 15

12. DATA ANALYSIS AND CALCULATIONS..... 16

13. METHOD PERFORMANCE..... 16

14. POLLUTION PREVENTION 17

15. WASTE MANAGEMENT 17

16. REFERENCES..... 17

17. MISCELLANEOUS (TABLES, APPENDICES, ETC...)..... 18

1. SCOPE AND APPLICATION

- 1.1. This method covers the determination of fluoride, chloride, nitrite, bromide, nitrate, ortho-phosphate and sulfate in drinking water, surface water, mixed domestic and industrial wastewaters, groundwater, reagent waters, solids (after extraction 11.7) and leachates (when no acetic acid is used).
- 1.2. A listing of associated LIMs method codes is located in Section 8.2.
- 1.3. This document accurately reflects current laboratory standard operating procedures (SOP) as of the date above. All facility SOPs are maintained and updated as necessary.

2. SUMMARY OF METHOD

- 2.1. A 25 uL volume of sample is introduced into the ion chromatograph. The sample is pumped through three different ion exchange columns and into a conductivity detector. The first two columns, a precolumn or guard column and a separator column, are packed with low-capacity, strongly basic anion exchange resin. Ions are separated into discrete bands based on their affinity for the exchange sites of the resin. The last column is a suppresser column that reduces the background conductivity of the eluent to a low or negligible level and converts the anions in the sample to their corresponding acids. The separated anions in their acid form are measured using an electrical conductivity cell. Anions are identified based on their retention times compared to known standards. Quantitation is accomplished by measuring the peak height or area and comparing it to a calibration curve generated from known standards.

3. DEFINITIONS

- 3.1. Refer to the glossary in the Laboratory Quality Manual (LQM).

4. INTERFERENCES

- 4.1. Interferences can be caused by substances with retention times that are similar to and overlap those of the anion of interest. Large amounts of an anion can interfere with the peak resolution of an adjacent anion. Sample dilution and/or fortification can be used to solve most interference problems associated with retention times.
- 4.2. The water dip or negative peak that elutes near, and can interfere with, the fluoride peak can usually be eliminated by the addition of concentrated eluent to each standard and sample.

- 4.3. Method interferences may be caused by contaminants in the reagent water, reagents, glassware and other sample processing apparatus that lead to discrete artifacts or an elevated baseline in the ion chromatograms.
- 4.4. Any anion that is not retained by the column or only slightly retained will elute in the area of fluoride and interfere. Known coelution is caused by carbonate and other small organic anions. At concentrations of fluoride above 1.5 mg/L, this interference may not be significant; however, it is the responsibility of the user to generate precision and accuracy information in each sample matrix.
- 4.5. The acetate anion elutes early during the chromatographic run. The retention times of the anions also seem to differ when large amounts of acetate are present. Therefore, this method is not recommended for leachates of solid samples when acetic acid is used for pH adjustment.

5. SAFETY

- 5.1. Procedures shall be carried out in a manner that protects the health and safety of all STL associates.
- 5.2. Eye protection that satisfied ANSI Z87.1 (as per the Chemical Hygiene Plan), laboratory coat, and appropriate gloves must be worn while samples, standards, solvents, and reagents are being handled. Disposable gloves that have been contaminated will be removed and discarded; other gloves will be cleaned immediately.
- 5.3. The health and safety hazards of many of the chemicals used in this procedure have not been fully defined. Additional health and safety information can be obtained from the Material Safety Data Sheets (MSDS) maintained in the laboratory.
- 5.4. Exposure to chemicals must be maintained **as low as reasonably achievable**, therefore; unless they are known to be non-hazardous, all samples must be opened, transferred and prepared in a fume hood, or under other means of mechanical ventilation when possible. Solvent and waste containers will be kept closed unless transfers are being made.
- 5.5. The preparation of standards and reagents will be conducted in a fume hood with the sash closed as far as the operation will permit.
- 5.6. All work must be stopped in the event of a known or potential compromise to the health and safety of a STL associate. The situation must be reported **immediately** to a laboratory supervisor.

6. EQUIPMENT AND SUPPLIES

- 6.1. Balance -- Analytical, capable of accurately weighing to the nearest 0.0001 g.
- 6.2. Ion Chromatograph -- Analytical system complete with ion chromatograph and all required accessories including analytical columns, compressed gases and detectors.
 - 6.2.1. Anion guard column: A protector of the separator column. If omitted from the system the retention times will be shorter. Usually packed with same substrate as the separator column. 4 x 50 mm, Dionex IonPac AG14 P/N 46134, or equivalent.
 - 6.2.2. Anion separator column: The separation shown in Figure 1 was generated using a Dionex IonPac AS14 column (P/N 46134). Equivalent column may be used if comparable resolution is obtained, and the requirements of Sect. 9.2 can be met.
 - 6.2.3. Anion suppresser device: Dionex anion micro membrane suppresser (P/N 37106) or ASRS-Ultra Self-Regenerating Suppressor (4mm) P/N 53946 or equivalent.
 - 6.2.4. Detector -- Conductivity cell: approximately 1.25 uL internal volume, Dionex, or equivalent.
 - 6.2.5. Dionex --PeakNet 5.1Data Chromatography Software or equivalent.
- 6.3. Assorted laboratory glassware (pipettes, volumetric flasks, etc.).

7. REAGENTS AND STANDARDS

- 7.1. Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available. Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.
- 7.2. Reagent water: Distilled or deionized water, free of the anions of interest. Water should contain particles no larger than 0.20 microns.
- 7.3. Eluent solution: sodium bicarbonate (CASRN 144-55-8) 1.0 mM, sodium carbonate (CASRN 497-19-8) 3.5 mM. Dissolve 1.680 g sodium bicarbonate (NaHCO_3) and 7.417 g of sodium carbonate (Na_2CO_3) in reagent water (7.2) and dilute to 100 mL in a volumetric flask. Take 10 mL of this concentrated eluent solution and dilute to 2 L for

use as the working eluent solution or dissolve the entire bicarbonate/carbonate amount in 20 L of reagent water.

- 7.4. Stock solutions (1,000 mg/L): All stocks may be prepared as described below or purchased from commercial sources. Primary and secondary sources are required for each target analyte.
- 7.4.1. Fluoride stock solution (1.00 mL = 1.00 mg F⁻): In a 1 liter volumetric flask, dissolve 2.2100 g of sodium fluoride (NaF) in reagent water, and dilute to volume with reagent water. Store in chemical- resistant glass or polyethylene.
- 7.4.2. Chloride stock solution (1.00 mL = 1.00 mg Cl⁻): Dry sodium chloride (NaCl) for 12 hours at 105°C, and cool in a desiccator. In a 1 liter volumetric flask, dissolve 1.6485 g of the dry salt in reagent water and dilute to volume with reagent water.
- 7.4.3. Nitrite stock solution (1.00 mL = 1.00 mg NO₂⁻ - N): Place approximately 10.0 g of sodium nitrite (KNO₂) in a 125 mL beaker and dry to constant weight (about 24 hours) in a desiccator. In a 1 liter volumetric flask, dissolve 6.0790 g of the dried salt in reagent water and dilute to volume with reagent water. Store in a sterilized glass bottle. Refrigerate and prepare monthly.
- Nitrite is easily oxidized, especially in the presence of moisture, and only fresh reagents are to be used.
 - Prepare sterile bottles for storing nitrite solutions by heating for 1 hour at 170°C in an air oven.
- 7.4.4. Bromide stock solution (1.00 mL = 1.00 mg Br⁻): Dry approximately 5.0 g of sodium bromide (NaBr) for 12 hours at 105°C, and cool in a desiccator. In a 1 liter volumetric flask, dissolve 1.2876 g of the dried salt in reagent water and dilute to volume with reagent water.
- 7.4.5. Nitrate stock solution (1.00 mL = 1.00 mg NO₃⁻ - N): Dry approximately 10.00 g of sodium nitrate (KNO₃) at 105°C for 24 hours. In a 1 liter volumetric flask, dissolve 7.2200 g of the dried salt in reagent water and dilute to volume with reagent water.
- 7.4.6. Phosphate stock solution (1.00 mL = 1.00 mg PO₄ - P): Dry approximately 10.00 g of potassium dihydrogen phosphate (KH₂PO₄) for 1 hour at 105°C and cool in a desiccator. In a 1 liter volumetric flask, dissolve 4.3937 g of the dry salt in reagent water and dilute to volume with reagent water.

- 7.4.7. Sulfate stock solution (1.00 mL = 1.00 mg SO_4^{--2}): Dry approximately 5.00 g of potassium sulfate (K_2SO_4) at 105°C for 1 hour and cool in a desiccator. In a 1 liter volumetric flask, dissolve 1.8141 g of the dried salt in reagent water and dilute to volume with reagent water.
- 7.4.8. Commercial stock solution A: F^- - 25 mg/L, Cl^- - 500 mg/L, Br^- - 100 mg/L, NO_3^- - N- 25 mg/L, $\text{PO}_4 - \text{P}$ - 25 mg/L, SO_4^{--2} - 500 mg/L
- 7.4.9. Commercial stock solution B: NO_2^- - N- 25 mg/L
- 7.4.10. Commercial IC Spike solution A: : F^- - 125 mg/L, Cl^- - 2500 mg/L, Br^- - 500 mg/L, NO_3^- - N- 125 mg/L, $\text{PO}_4 - \text{P}$ - 125 mg/L, SO_4^{--2} - 2500 mg/L
- 7.4.11. Commercial IC Spike solution B: NO_2^- - N- 125 mg/L
- 7.5. Working standards: Prepare calibration standard #5 in a 10 mL volumetric flask and transfer to a vial. Adjust the amount of stock solution used to prepare the working standards if the stock concentration differs from 1000 mg/L as assumed Alternatively prepare Cal standard #5 by mixing 4.0 mL commercial stock A, 4.0 mL commercial stock B and 2.0 mL of reagent water.

Calibration Standard #5

Analyte	mL of Stock	Final Conc.
Fluoride	0.10mL	10.0 mg/L
Chloride	2.0 mL	200. mg/L
Nitrite	0.10 mL	10.0 mg/L
Bromide	0.40 mL	40.0 mg/L
Nitrate	0.10 mL	10.0 mg/L
Ortho-Phosphate	0.10 mL	10.0 mg/L
Sulfate	2.0 mL	200. mg/L

- 7.5.1. In 5 mL PolyVials prepare the following calibration standards in reagent grade water. Final concentrations of working standards are shown below.

Calibration Standard #4: take 2.50 mL of calibration standard #5 and add 2.50 mL of reagent water.

Calibration Standard #2: take 250 μ L of calibration standard #5 and add 4.75 mL of reagent water.

Calibration Standard #1: take 25.0 μ L of calibration standard #5 and add 4.95 mL of reagent water.

Calibration Standard #3: take 1.25 mL of calibration standard #5 and add 3.75 mL of reagent water.

Calibration Standard #1

Analyte	25.0 μ L of Cal Std #5	Final Conc
Fluoride		0.05 mg/L
Chloride		1.0 mg/L
Nitrite		0.05 mg/L
Bromide		0.20 mg/L
Nitrate		0.05 mg/L
Ortho-Phosphate		0.05 mg/L
Sulfate		1.0 mg/L

Calibration Standard #2

Analyte	250 μ L of Cal Std #5	Final Conc.
Fluoride		0.5 mg/L
Chloride		10. mg/L
Nitrite		0.5 mg/L
Bromide		2.0 mg/L
Nitrate		0.5 mg/L
Ortho-Phosphate		0.5 mg/L
Sulfate		10. mg/L

Calibration Standard #3

Analyte	1.25 mL of Cal Std #5	Final Conc.
Fluoride		2.5 mg/L
Chloride		50. mg/L
Nitrite		2.5 mg/L
Bromide		10. mg/L
Nitrate		2.5 mg/L
Ortho-Phosphate		2.5 mg/L
Sulfate		50. mg/L

Calibration Standard #4

Analyte	2.5 mL of Cal Std #5	Final Conc.
---------	----------------------	-------------

Fluoride		5.0 mg/L
Chloride		100 mg/L
Nitrite		5.0 mg/L
Bromide		40. mg/L
Nitrate		5.0 mg/L
Ortho-Phosphate		5.0 mg/L
Sulfate		100 mg/L

7.5.2. Prepare or purchase a secondary stock standard(s) using a standards source other than that used for the primary standards as described in Section 7.5. Dilute these stock standards to as indicated in the table below to prepare the mixture to be used for the LCS and CCV solution. Alternatively prepare this solution by mixing 0.50 mL commercial stock A, 0.50 mL commercial stock B and 4.0 mL of reagent water.

LCS & Continuing Calibration Verification Solution

Analyte	Final Conc. (V _f =5ml)
Fluoride	2.5 mg/L
Chloride	50. mg/L
Nitrite	2.5 mg/L
Bromide	10. mg/L
Nitrate	2.5 mg/L
Ortho Phosphate	2.5 mg/L
Sulfate	50. mg/L

7.5.3. Prepare or purchase a secondary stock standard(s) using a standards source other than that used for the primary standards as described in Section 7.5. Dilute these stock standards to prepare the mixture to be used for the Matrix Spike solution. Alternatively purchase these mixes (ready to use) from a commercial source. Add 100 uL of each IC Spike solution to 5 mL of sample when preparing the MS/MSD. Dilute as needed after spiking the sample.

Matrix Spike "True" Values

Analyte	Final Conc.
Fluoride	2.5 mg/L
Chloride	50. mg/L
Nitrite	2.5 mg/L
Bromide	10. mg/L
Nitrate	2.5 mg/L
Ortho Phosphate	2.5 mg/L
Sulfate	50. mg/L

NOTE: Stock standards, calibration standard #5 and LCS standard should be stored in the dark at $4^{\circ} \pm 2^{\circ}\text{C}$. Replace these standards when instrument response indicates target analyte degradation may have occurred or after the standard has expired (12 months commercial mix or 6 months in house mix), which ever occurs first. Nitrite and ortho-phosphate are particularly light and oxygen sensitive.

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE

- 8.1. Samples should be collected in plastic or glass bottles. All bottles must be thoroughly cleaned and rinsed with reagent water. The volume collected should be sufficient to ensure a representative sample, allow for replicate analysis, if required, and minimize waste disposal.
- 8.2. Sample preservation and holding times for the anions that can be determined by this method for water samples are as follows:

QuanTIMs Method Code		Analyte	Preservation	Holding Time
EPA 300.0A	SW846 9056A	Fluoride	4° ± 2°C	28 days
C8	3C			
CX	3D	Chloride	4° ± 2°C	28 days
GO	3E	Nitrite	4° ± 2°C	48 hours
GM	3F	Bromide	4° ± 2°C	28 days
C9	3G	Nitrate	4° ± 2°C	48 hours
DO	3H	Ortho Phosphate	4° ± 2°C	48 hours
CY	3I	Sulfate	4° ± 2°C	28 days

Note: Soil leachates will follow the same preservation and holding times as the water samples; starting from the time of extraction.

9. QUALITY CONTROL

9.1. The STL QC Program document provides further details of the QC and corrective action guidelines presented in this SOP. Refer to this document if additional guidance is required.

9.2. Table I provides a summary of quality control requirements including type, frequency, acceptance criteria and corrective action.

9.3. Initial Demonstration of Capability

9.3.1. Prior to the analysis of any samples by Ion Chromatography, the following requirements must be met:

9.3.1.1. Method Detection Limit (MDL): An MDL must be determined prior to analysis of any samples. The MDL is determined using seven replicates of reagent water spiked with the anions of interest that has been carried through the entire analytical procedure. MDLs must be redetermined on an annual basis. The spike level must be greater than the calculated MDL

but less than or equal to 10x the MDL. The result of the MDL determination must be below the STL reporting limit.

- 9.4. Batch definition: Preparation and QC batch definitions are provided in the STL QC Policy.
- 9.5. Method Blank (MB): One method blank must be processed with each preparation batch. The method blank consists of reagent grade water that has been taken through the entire preparation and analytical process. The method blank is used to identify any system and process interferences or contamination of the analytical system that may lead to the reporting of elevated analyte concentrations or false positive data. The method blank should not contain any analyte of interest above the reporting limit.
- 9.6. Laboratory Control Sample (LCS): One LCS must be processed with each preparation batch and must be carried through the entire analytical procedure. The LCS is used to monitor the accuracy of the analytical process. On-going monitoring of the LCS results provides evidence that the laboratory is performing the method within acceptable accuracy and precision guidelines. If the result is outside established control limits the system is out of control and corrective action must occur. Until in-house limits are established, a control limit of 90 - 110% recovery must be applied. Corrective action will be repreparation and reanalysis of the batch unless the client agrees that other corrective action is acceptable. The LCS consists of reagent grade water containing a known amount of target analytes that has been injected into the ion chromatography system. The LCS is prepared from a separate stock standard, or neat material, of a different manufacturer than the stock, or neat material, used to prepare the calibration standard.
- 9.7. Matrix Spike/Matrix Spike Duplicate (MS/MSD): One MS/MSD pair must be processed for each QC batch. A matrix spike (MS) is a field sample to which a known concentration of target analyte has been added. A matrix spike duplicate (MSD) is a second aliquot of the same sample (spiked identically as the MS) prepared and analyzed along with the sample and matrix spike. Some client specific DQO's may require the use of sample duplicates in place of or in addition to MS/MSD's. The MS/MSD results are used to determine the effect of a matrix on the precision and accuracy of the analytical process. Due to the potential variability of the matrix of each sample, these results may have immediate bearing only on the specific sample spiked. Spiking levels will be the same as the LCS values.
- If the MS/MSD recovery or RPD falls outside the acceptance range, the recovery of the analyte must be in control for the LCS. Until in-house control limits are established, a control limit of 90-110% recovery and 20% RPD must be applied to the MS/MSD.

- If the native analyte concentration in the MS/MSD exceeds 4x the spike level for that analyte, the recovery data are reported as NC (i.e. not calculated).
 - If the LCS recovery is within limits, then the laboratory operation is in control and the results may be accepted.
 - If the recovery of the LCS is outside the limits, corrective action must be taken. Corrective action will include reparation and reanalysis of the batch.
 - If a MS/MSD is not possible due to limited sample volume then a LCS duplicate must be analyzed. The RPD of the LCS and LCSD must be compared to the matrix spike limits.
- 9.8. Continuing Calibration Verification (CCV/CCB): Continuing calibration is verified by analyzing the calibration standard after every ten (10) samples. The CCV must fall within +/- 10% of the true value for each target analyte. A CCB is analyzed immediately following the CCV to monitor low level accuracy and system cleanliness. The CCB result must be below the reporting limit for that analyte. If either the CCV or CCB fail to meet criteria, the analysis must be terminated, the problem corrected and reparation and analysis of all samples following the last CCV and CCB which were in control.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Establish ion chromatographic operating parameters equivalent to those indicated in table 2. Refer to Table 3 for typical standard run retention times. Other than the presence of the analytical column the instrument conditions are the same.
- 10.2. For each analyte of interest, prepare a **minimum** of 3 calibration standards and a blank by adding accurately measured volumes of one or more stock standards to a volumetric flask and dilution to volume with reagent water. If a sample analyte concentration exceeds the calibration range the sample may be diluted to fall within the range. If this is not possible then three new calibration concentrations must be chosen, two of which must bracket the concentration of the sample analyte of interest. Each attenuation range of the instrument used to analyze a sample must be calibrated individually.
- 10.3. Using an injection volume of 25 uL of each calibration standard, tabulate peak height or area responses against the concentration. The results are used to prepare a calibration curve for each analyte. During this procedure, retention times must be recorded. All

analytes will be calibrated using a quadratic regression forced through the origin.
Correlation coefficients (R^2) must be 0.995 or better.

11. PROCEDURE

- 11.1. One time procedural variations are allowed only if deemed necessary in the professional judgment of supervision to accommodate variation in sample matrix, radioactivity, chemistry, sample size, or other parameters. Any variation in procedure shall be completely documented using a Nonconformance Memo and is approved by a Technical Specialist and QA Manager. If contractually required, the client shall be notified. The Nonconformance Memo shall be filed in the project file.
- 11.2. Any unauthorized deviations from this procedure must also be documented as a nonconformance, with a cause and corrective action described.
- 11.3. Table 2 summarizes the recommended operating conditions for the ion chromatograph. Included in this table are estimated retention times that can be achieved by this method. Other columns, chromatographic conditions, or detectors may be used if the requirements of Sect. 9.2 are met.
- 11.4. Check system calibration daily as outlined in Table 1 and, if required, recalibrate as described in Sect 10.
- 11.5. Load and inject a fixed amount (25 uL) of settled & filtered sample. If the sample is cloudy then it should be filtered prior to loading into the autosampler polyvial. Flush injection loop thoroughly, using each new sample. Use the same size loop for standards and samples. Record the resulting peak size in area or peak height units. An automated constant volume injection system may also be used.
- 11.6. The width of the retention time window used to make identifications should be based upon measurements of actual retention time variations of various concentration. Three times the standard deviation of a retention time can be used to calculate a suggested window size for each analyte. However, the experience of the analyst should weigh heavily in the interpretation of chromatograms since retention time is concentration dependent for most analytes..
- 11.7. If the response for the peak exceeds the working range of the system, dilute the sample with an appropriate amount of reagent water and reanalyze.
- 11.8. If the resulting chromatogram fails to produce adequate resolution, or if identification of specific anions is questionable, fortify the sample with an appropriate amount of standard and reanalyze.

NOTE: Retention time is affected by concentration. Nitrate and sulfate exhibit the greatest amount of change, although all anions are affected to some degree. See Table 3. In some cases this peak migration may produce poor resolution or identification.

- 11.9. The following extraction should be used for solid materials: Add an amount of reagent water equal to ten times the weight of dry solid material taken as a sample. This slurry is mixed for one hour using a magnetic stirring device or tumbler. Filter the resulting slurry before injecting using a 0.45 um membrane type filter. This can be the type that attaches directly to the end of the syringe
- 11.10. Should more complete resolution be needed between peaks the eluent (7.3) can be diluted. This will spread out the run but will also cause the later eluting anions to be retained longer. The analyst must determine to what extent the eluent is diluted. This dilution should not be considered a deviation from the method.

12. DATA ANALYSIS AND CALCULATIONS

- 12.1. Prepare a calibration curve for each analyte by plotting instrument response against standard concentration. Compute sample concentration by comparing sample response with the standard curve. Multiply answer by appropriate dilution factor.
- 12.2. Report only those values that fall between the lowest and the highest calibration standards. Samples exceeding the highest standard should be diluted and reanalyzed.
- 12.3. Report results in mg/L for aqueous samples and mg/Kg for 1 hour leachates and mg/L for 18 hour leachates of solid samples.
- 12.4. Report NO_2^- as N
 NO_3^- as N
 $\text{HPO}_4^{=}$ as P

13. METHOD PERFORMANCE

13.1. The reporting limits for the following analytes are based on a 25 uL injection volume:

Analyte	Water RL	Soil RL
Fluoride	1.0 mg/L	10 mg/kg
Chloride	1.0 mg/L	10 mg/kg
Nitrite	0.5 mg/L	5 mg/kg
Bromide	0.5 mg/L	5 mg/kg
Nitrate	0.05 mg/L	0.5 mg/kg
O-Phosphate	0.5 mg/L	5 mg/kg
Sulfate	1.0 mg/L	10 mg/kg

13.2. The group/team leader has the responsibility to ensure that this procedure is performed by an analyst who has been properly trained in its use and has the required experience. The analyst must be given two blind performance samples to analyze or process for analysis. Upon successful completion of the performance evaluation (PE) samples, these analyses will be documented as initial qualification,. Requalification must be performed annually thereafter for this procedure. The group/team leader must document the training and PE performance and submit the results to the QA Manager for inclusion in the associate's training files.

14. POLLUTION PREVENTION

14.1. This method does not contain any specific modifications that serve to minimize or prevent pollution.

15. WASTE MANAGEMENT

15.1. Waste generated in this procedure must be segregated, and disposed of according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

16.1. Method 300.0, "Determination of Inorganic Anions by Ion Chromatography", Environmental Monitoring Systems Laboratory, Office of Research and Development, U.S. Environmental Protection Agency, Cincinnati, Ohio, Revision 2.1, August 1993.

16.2. Method 9056A, "Determination of Inorganic Anions by Ion Chromatography", SW846, Test Methods for Evaluating Solid Waste, Third Edition, Draft Revision 1, September 1999.

16.3. STL North Canton Laboratory Quality Manual (LQM), current version.

17. MISCELLANEOUS (TABLES, APPENDICES, ETC...)

17.1. Attachment #1, method Flow Chart


17.2. Table 1, Quality Control Samples

17.3. Table 2, Standard Instrument Operating Parameters

17.4. Table 3, Retention Time Matrix

17.5. Figure 1, Example Chromatogram

Determination Of Inorganic Anions By Ion Chromatography



ATTACHMENT #1

TABLE 1
QUALITY CONTROL SAMPLES

QC Samples	Frequency	Acceptance Criteria	Corrective Action
Initial Calibration Verification (ICV)	At the start of each day following calibrating prior to sample analysis	+/- 10% of true value	Recalibrate and reanalyze
Initial Calibration Blank (ICB)	After Initial Calibration Verification and prior to sample analysis	< the Reporting Limit	Reprepare and reanalyze
Laboratory Control Sample (LCS)	1 per batch of 20 samples	Meets laboratory historical limits	Reanalyze all samples associated with unacceptable LCS
Matrix Spike Sample (MS/MSD)	1 MS/MSD pair per batch or 20 samples	Meets laboratory historical limits	Supervisor's technical judgment
Continuing Calibration Verification (CCV)	Between each group of 10 injections and at the end of the analytical sequence	+/- 10% of true value	Recalibrate and reanalyze all samples since the last acceptable CCV
Continuing Calibration Blank (CCB)	Between each group of 10 injections and at the end of the analytical sequence	< the Reporting Limit	Recalibrate and reanalyze all samples since the last acceptable CCB

TABLE 2

Standard Instrument Operating Parameters

Standard Conditions:

Eluent Pump Rate: 1.20 mL/min (DX-120 and DX-320)
 Sample Loop: 25 uL
 Eluent: 1.0mM sodium bicarbonate, 3.5mM sodium carbonate
 Detector output Baseline conductivity should be 15 - 20 uS prior to sample analysis.

TABLE 3

Standard Run Retention Time Matrix (minutes)*

Analyte	Concentration (mg/L)												RT window	
	0.05	0.2	0.5	1	2	2.5	5	10	20	40	50	100		200
F ⁻	2.75	2.75				2.75	2.75	2.75						
Cl ⁻				3.97					3.98		4.03	4.08	4.17	
NO ₂ ⁻	4.80	4.80				4.78	4.78	4.80						
Br ⁻		6.15			6.13			6.10	6.08	6.07				
NO ₃ ⁻	7.33	7.27				7.17	7.13	7.07						
o-PO ₄ ²⁻	9.53	9.53				9.52	9.50	9.48						
SO ₄ ²⁻				11.50				11.48			11.43	11.38	11.27	

* Analyte retention time is concentration dependent for most anions. Retention time increases with increasing concentration for chloride. Retention time decreases with increasing concentration for bromide, nitrate, ortho-phosphate and sulfate.

EXAMPLE ION CHROMATOGRAM

cal std 4 IC stds 9001/9002




Figure #1

Controlled Copy
Copy No: _____
Implementation Date _____

SOP No: CORP-GC-0001NC
Revision No: 5.6
Revision Date: 05/25/01
Page 1 of 92

STL NORTH CANTON STANDARD OPERATING PROCEDURE

**TITLE: GAS CHROMATOGRAPHIC ANALYSIS BASED ON METHOD 8000B,
8021B, 8081A, 8082, 8151A, 8310, 8141A, 8015B, 608, 610 and Wisconsin DNR Modified
DRO Method**

(SUPERSEDES: Revision 5.5 Dated 03/16/01)

Prepared by: _____ Date _____
Approved by: _____
Technology Specialist Date _____
Approved by: _____
Quality Assurance Manager Date _____
Approved by: _____
Environmental Health and Safety Date _____
Approved by: _____
Laboratory Director Date _____
Approved by: *[Signature]* 6/4/01
Corporate Technology Date _____

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use it for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

TABLE OF CONTENTS

1.	Scope and Application	4
2.	Summary of Method	4
3.	Definitions	4
4.	Interferences	4
5.	Safety	4
6.	Equipment and Supplies	5
7.	Reagents and Standards	5
8.	Sample Preservation and Storage	5
9.	Quality Control	6
10.	Calibration and Standardization	9
11.	Procedure	14
12.	Data Analysis and Calculations	16
13.	Method Performance	19
14.	Pollution Prevention	20
15.	Waste Management	20
16.	References	20
17.	Miscellaneous	20

List of Appendices:

Appendix A	Analysis of Volatile Organics based on Method 8021B
Appendix B	Analysis of Organochlorine Pesticides based on Method 8081A
Appendix C	Analysis of Organochlorine Pesticides and PCBs based on Method 8082
Appendix D	Analysis of Phenoxy Acid Herbicides based on Method 8151A
Appendix E	Analysis of Polynuclear Hydrocarbons based on Method 8310
Appendix F	Analysis of Organophosphorus Pesticides Based on Method 8141A
Appendix G	Analysis of Total Petroleum Hydrocarbons Based on Method 8015B and Wisconsin DRO
Appendix H	Analysis of Non-Halogenated Organic Compounds by Method 8015B, Direct Injection
Appendix I	Analysis of Phillips 66 Compounds

List of Tables

Table A1	Standard analyte lists for 8021B
Table A2-A4	Recommended conditions for method 8021B
Table A5	Surrogate and Internal standard concentrations for aqueous and low level soil samples, method 8021B
Table A6	Concentrations for LCS and MS/MSD compounds, low level soil and aqueous, method 8021B.
Table B1	Standard analyte and reporting limits, method 8081A
Table B2	Recommended conditions, method 8081A
Table B3	Calibration levels, method 8081A
Table B4	Column degradation evaluation mix, method 8081A
Table B5	LCS/matrix spike and surrogate levels, method 8081A
Table B6	LCS/matrix spike and surrogate levels for TCLP, method 8081A
Table B7	Suggested analytical sequence, method 8081A
Table B8	Performance limits, method 8081A
Table C1	Standard analyte list, method 8082
Table C2	Instrumental conditions, method 8082
Table C3	Calibration standards, method 8082
Table C4	LCS/Matrix spike and surrogate levels, method 8082
Table C5	Suggested Analytical Sequence, Method 8082
Table D1	Standard Analyte List, method 8151A
Table D2	Instrumental conditions, method 8151A
Table D3	LCS/Matrix spike and surrogate levels, method 8151B
Table D4	Performance limits, method 8151B
Table E1	Standard Analyte List for Method 8310
Table E2	Instrumental Conditions for Method 8310
Table E3	LCS/Matrix spike and surrogate compounds, Method 8310
Table E4	Calibration Standards, Method 8310
Table E5	Compound/Detector List, Method 8310
Table F1	Analyte and Reporting Limit List, Method 8141A
Table F2	Instrument Conditions, Method 8141A
Table F3	Initial Calibration, Method 8141A
Table F4	LCS/Matrix spike and surrogate compounds, Method 8141A
Table G1	Suggested GC Temperature Program for TPH Analysis
Table G2	Reporting Limits for TPH Analysis
Table H1	Non-Halogenated Organic Compounds Reporting Limits
Table H2	Non-Halogenated Organic Compounds Working Standards
Table I1	Phillips 66 Compounds Reporting Limits

1. SCOPE AND APPLICATION

This SOP describes procedures for analysis of organic analytes by Gas Chromatography (GC). The procedures are based on SW-846 methodology and are applicable for measurements made to comply with the Resource Conservation and Recovery Act (RCRA). Individual analytes and methods are described in the appendices. Appendices B, C, and E include criteria for the analysis of wastewater by Methods 608 and 610. Appendix G describes procedures for the analysis of petroleum hydrocarbons by SW-846 methodology and the Wisconsin DNR Modified DRO method. Appendix H includes criteria for the analysis of non-halogenated organic compounds by Method 8015B, Direct Injection. Appendix I describes the analysis of Phillips 66 analytes by Method 8015B.

2. SUMMARY OF METHOD

In general, semivolatile analytes in aqueous samples are prepared for analysis using continuous or separatory funnel liquid / liquid extraction or solid phase extraction (SOP # CORP-OP-0001NCNC) Solid samples are prepared using sonication, Soxhlet or pressurized fluid extraction (SOP # CORP-OP-0001NC). Volatile analytes are prepared for analysis using purge and trap methodology (Appendix A).

After the initial preparation step, the sample is introduced to the GC and concentrations of target analytes are measured by the detector response within a defined retention time window, relative to the response to standard concentrations. Internal or external standardization procedures are used as specified in the method appendices.

3. DEFINITIONS

Definitions of terms used in this SOP may be found in the glossary of the STL North Canton Laboratory Quality Manual (LQM), current version.

4. INTERFERENCES

Contamination by carryover can occur when a low concentration sample is analyzed after a high concentration sample. In addition, some purge and trap autosamplers are susceptible to port specific contamination. Co-elution of target analytes with non-targets can occur, resulting in false positives or biased high results. In particular, this is a problem with non-selective detectors such as the Flame Ionization Detector (FID). See the appendices for interferences specific to individual tests and suggested corrective actions.

5. SAFETY

- 5.1. Procedures shall be carried out in a manner that protects the health and safety of all STL North Canton associates. The following requirements must be met:

Eye protection that prevents splash, laboratory coat, and appropriate gloves must be worn while samples, standards, solvents, and reagents are being handled. Disposable gloves that have become contaminated will be removed and discarded; other gloves will be cleaned immediately. Refer to the STL North Canton Chemical Hygiene plan for a complete description of personal protection equipment.

The health and safety hazards of many of the chemicals used in this procedure have not been fully defined. Additional health and safety information can be obtained from the MSDS files maintained in the laboratory. Specific hazards are covered in the appendices.

- 5.1.1. Opened containers of neat standards will be handled in a fume hood.

- 5.2. Sample extracts and standards, which are in a flammable solvent, shall be stored in an explosion-proof refrigerator.

5.3. When using hydrogen gas as a carrier, all precautions listed in the CHP shall be observed.

5.4. Standard preparation and dilution shall be performed inside an operating fume hood.

6. EQUIPMENT AND SUPPLIES

An analytical system complete with a gas chromatograph is required. A data system capable of measuring peak area and/or height is required. Recommended equipment and supplies for individual methods are listed in each method appendix.

7. REAGENTS AND STANDARDS

7.1. Stock Standards

Stock standards are purchased as certified solutions or prepared from pure solutions. Stock standards for method 8021B are stored at -10 to -20°C. Other stock standard solutions are stored as recommended by the manufacturer. All stock standards must be protected from light. Stock standard solutions should be brought to room temperature before using.

Semivolatile stock standard solutions must be replaced after one year. Stock standards of gases must be replaced at least every week, unless the acceptability of the standard is demonstrated (Less than 20% drift from the initial calibration is an acceptable demonstration). Other volatile stock standards must be replaced every 6 months or sooner if comparison with check standards prepared from an independent source indicates a problem.

7.1.1. Expiration times for all standards are measured from the time the standard is prepared or from the time that the standard ampoule is opened, if the standard is supplied in a sealed ampoule. If a vendor-supplied standard has an earlier expiration date then that date is used.

7.2. Calibration Standards

7.2.1. Volatile Calibration Standards

The procedure for preparation of volatile standards is given in Appendix A.

7.2.2. Semivolatile Calibration Standards

Semivolatile calibration standards are prepared as dilutions of the stock standards. Surrogates and internal standards are used as specified in the method appendices. Semivolatile calibration solutions must be refrigerated at $\leq 6^{\circ}\text{C}$ and protected from light. The standards must be replaced at least every six months or sooner if comparison with check standards indicates a problem.

7.3. Gases for carrier and make-up: Hydrogen, Helium, Nitrogen, Argon/Methane.

7.4. Quality control (QC) Standards

QC standards (matrix spiking and LCS standards) are prepared and stored in the same way as calibration standards. They must be made from a stock independent from the calibration standards.

8. SAMPLE PRESERVATION AND STORAGE

Semivolatile extracts must be refrigerated at $\leq 6^{\circ}\text{C}$ and analyzed within 40 days of the end of the extraction. Volatile sample storage conditions and holding times are given in Appendix A.

9. QUALITY CONTROL

9.1. Initial Demonstration of Capability

- 9.1.1. For the standard analyte list, the initial demonstration and method detection limit (MDL) studies described in Section 13 must be acceptable before analysis of samples may begin.
- 9.1.2. For non-standard analytes, a MDL study must be performed and calibration curve generated before analyzing any samples, unless lesser requirements are previously agreed to with the client. In any event the minimum initial demonstration required is analysis of an extracted standard at the reporting limit and a single point calibration.

9.2. Batch Definition

Batches are defined at the sample preparation stage. Batches should be kept together through the whole analytical process as far as possible, but it is not mandatory to analyze prepared extracts on the same instrument or in the same sequence. Refer to the STL North Canton QC Program document (QA-003) for further details of the batch definition.

9.2.1. Quality Control Batch

The batch is a set of up to 20 samples of the same matrix processed using the same procedures and reagents within the same time period. The Quality Control batch must contain a matrix spike / spike duplicate (MS/MSD), a Laboratory Control Sample (LCS), and a method blank. Laboratory generated QC samples (Blank, LCS, MS/MSD) do not count towards the maximum 20 samples in a batch. Field QC samples are included in the batch count. In some cases, at client request, the MS/MSD may be replaced with a matrix spike and sample duplicate. If insufficient sample is available for an MS/MSD a LCSD may be substituted.

9.3. Control Limits

In-house historical control limits must be determined for surrogates, matrix spikes, and laboratory control samples (LCS). These limits must be determined at least annually. The recovery limits are mean recovery +/- 3 standard deviations, unless that limit is tighter than the calibration criteria, in which case limits may be widened. Refer to policy QA-003 for more details.

- 9.3.1. These limits do not apply to dilutions (except for tests without a separate extraction), but surrogate and matrix spike recoveries will be reported unless the dilution is more than 5X.
- 9.3.2. All surrogate, LCS, and MS recoveries (except for dilutions) must be entered into QuantIMS (when available) or other database so that accurate historical control limits can be generated. For tests without a separate extraction, surrogates and matrix spikes will be reported for all dilutions.
- 9.3.3. Refer to the QC Program document (QA-003) for further details of control limits.

9.4. Surrogates

All methods must use surrogates to the extent possible. Surrogate recoveries in samples and QC samples must be assessed to ensure that recoveries are within established limits. If any surrogates are outside limits, the following corrective actions must take place (except for dilutions):

- Check all calculations for error.
- Ensure that instrument performance is acceptable.
- Recalculate the data and/or reanalyze the extract if either of the above checks reveal a problem.

- Reprepare and reanalyze the sample or flag the data as "Estimated Concentration" if neither of the above resolves the problem. Repreparation is not necessary if there is obvious chromatographic interference.
 - The decision to reanalyze or flag the data should be made in consultation with the client. It is only necessary to reprepare / reanalyze a sample once to demonstrate that poor surrogate recovery is due to matrix effect, unless the analyst believes that the repeated out of control results are not due to matrix effect.
- 9.4.1. If dual column analysis is used the choice of which result to report is made in the same way as for samples (Section 12.1.2) unless one column is out of control, in which case the in-control result is reported.
- 9.4.2. If the surrogates are out of control for the sample, matrix spike, and matrix spike duplicate, then matrix effect has been demonstrated for that sample and reparation is not necessary. If the sample is out of control and the MS and/or MSD is in control, then reparation or flagging of the data is required.
- 9.4.3. Refer to the STL North Canton QC Program document (QA-003) for further details of the corrective actions.
- 9.5. Method Blanks
- For each batch of samples, analyze a method blank. The method blank consists of reagent water for aqueous semivolatile samples, and sodium sulfate for semivolatile soils tests (Refer to SOP No. CORP-OP-0001NC for details). For low level volatiles, the method blank consists of reagent water. For medium level volatiles, the method blank consists of methanol as described in Appendix A. Surrogates are added and the method blank is carried through the entire analytical procedure. The method blank must not contain any analyte of interest at or above the reporting limit (except common laboratory contaminants, see below) or at or above 5% of the measured concentration of that analyte in the associated samples, whichever is higher.
- If the analyte is a common laboratory contaminant (methylene chloride, acetone, 2-butanone, phthalate esters) the data may be reported with qualifiers if the concentration of the analyte is less than five times the reporting limit. Such action must be taken in consultation with the client.
- Re-extraction and reanalysis of samples associated with an unacceptable method blank is required when reportable concentrations are determined in the samples.
- If there is no target analyte greater than the RL in the samples associated with an unacceptable method blank, the data may be reported with qualifiers. Such action should be taken in consultation with the client.
- 9.5.1. Refer to the STL North Canton QC Program document (QA-003) for further details of the corrective actions.
- 9.6. Instrument Blanks
- 9.6.1. An instrument blank must be analyzed during any 12-hour period of analysis that does not contain a method blank.
- 9.6.2. An instrument blank consists of the appropriate solvent with internal standards added. If internal standards are not used the surrogates should be added.
- 9.6.3. Control criteria are the same as for the method blank, except that only reanalysis of affected samples would be required, not re-extraction.

9.7. Laboratory Control Samples (LCS)

For each batch of samples, analyze a LCS. The LCS contains a representative subset of the analytes of interest, and must contain the same analytes as the matrix spike. The LCS may also contain the full set of analytes with a subset of control analytes. If any control analyte or surrogate is outside established control limits, the system is out of control and corrective action must occur. Corrective action will normally be reparation and reanalysis of the batch; however, if the matrix spike and matrix spike duplicate are within limits; the batch may be acceptable.

9.7.1. Refer to the STL North Canton QC Program document (QA-003) for further details of the corrective action.

9.7.2. If dual column analysis is used the choice of which result to report is made in the same way as for samples (Section 12.1.2) unless one column is out of control, in which case the in control result is reported.

9.7.3. LCS compound lists are included in the appendices.

9.8. Matrix Spikes

For each QC batch, analyze a matrix spike and matrix spike duplicate. Spiking compounds and levels are given in the appendices. Compare the percent recovery and relative percent difference (RPD) to those in the laboratory specific historically generated limits.

- If any individual recovery or RPD falls outside the acceptable range, corrective action must occur. The initial corrective action will be to check the recovery of that analyte in the Laboratory Control Sample (LCS). Generally, if the recovery of the analyte in the LCS is within limits, then the laboratory operation is in control and analysis may proceed.
- If the recovery for any component is outside QC limits for both the Matrix spike / spike duplicate and the LCS, the laboratory is out of control and corrective action must be taken. Corrective action will normally include reparation and reanalysis of the batch.
- If a MS/MSD is not possible due to limited sample, then a LCS duplicate should be analyzed.
- The matrix spike / duplicate must be analyzed at the same dilution as the unspiked sample, unless the matrix spike components would then be above the calibration range.

9.8.1. If dual column analysis is used the choice of which result to report is made in the same way as for samples (Section 12.1.2) unless one column is out of control, in which case the in control result is reported.

9.9. Quality Assurance Summaries

Certain clients may require specific project or program QC that may supersede these method requirements. Quality Assurance Summaries should be developed to address these requirements.

9.10. STL North Canton QC Program

Further details of QC and corrective action guidelines are presented in the STL QC Program document (QA-003). Refer to this document if in doubt regarding corrective actions.

10. CALIBRATION AND STANDARDIZATION

Internal or external calibration may be used. Internal calibration is recommended unless the sample matrix is likely to interfere with the quantitation of the internal standard. In either event prepare standards containing each analyte of interest at a minimum of five concentration levels. The low level standard should be at or below the reporting limit. The other standards define the working range of the detector. Recommended calibration levels are given in the appendices.

- 10.1. A new calibration curve must be generated after major changes to the system or when the continuing calibration criteria cannot be met. Major changes include new columns, changing PID lamps or FID jets or replacing the ECD detector. A new calibration is not required after clipping the column, replacing the septum or syringe, or other minor maintenance.
- 10.2. With the exception of 10.3 below, it is NOT acceptable to remove points from a calibration curve for the purpose of meeting criteria, unless the points are the highest or lowest on the curve AND the reporting limit and/or linear range is adjusted accordingly. In any event, at least 5 points must be included in the calibration curve. Quadratic (second order) calibrations require at least six points. Third order calibrations require at least seven points.
- 10.3. A level may be removed from the calibration if the reason can be clearly documented, for example a broken vial or no purge run. A minimum of five levels must remain in the calibration. The documentation must be retained with the initial calibration. Alternatively, if the analyst believes that a point on the curve is inaccurate, the point may be reanalyzed and the reanalysis used for the calibration. All initial calibration points must be analyzed without any changes to instrument conditions, and all points must be analyzed within 24 hours.
- 10.4. External standard calibration
- Quantitation by the external standard method assumes a proportional relationship between the calibration run and the analyte in the sample. To use this approach, introduce each calibration standard into the GC using the technique that will be used for samples. The ratio of the peak height or area response to the mass or concentration injected may be used to prepare a calibration curve.

$$\text{Calibration Factor (CF)} = \frac{\text{Area or Height of Peak}}{\text{Mass Injected (ng)}}$$

Some data systems may use the inverse of this formula. This is acceptable so long as the same formula is used for standards and samples. It is also possible to use the concentration of the standard rather than the mass injected. (This would require changes in the equations used to calculate the sample concentrations). Use of peak area or height must be consistent. However, if matrix interferences would make quantitation using peak area inaccurate for a particular sample, then peak height may be used as a substitute.

- 10.5. Internal standard calibration
- 10.5.1. The internal standard approach assumes that variations in instrument sensitivity, amount injected etc. can be corrected by determining the ratio of the response of the analyte to the response of an internal standard that has been added to the extract. To use this approach, select one or more internal standard(s) that are similar in analytical behavior to the compounds of interest. Recommended internal standards are given in the appendices. The analyst must demonstrate that the measurement of the internal standard is not affected by method or matrix interferences. If the sample matrix interferes with quantitation of the internal standard, then the external standard approach must be used instead. In this event use the response factors from the previous continuing calibration to quantitate the analytes in the sample with the interference (applies only to the sample with the interference).
- 10.5.2. Introduce each calibration standard into the GC using the technique that will be used for samples. Response factors (RF) for each compound are calculated as follows:

$$RF = \frac{A_s \times C_{is}}{A_{is} \times C_s}$$

Where:

A_s = Response for the analyte to be measured

A_{is} = Response for the internal standard

C_{is} = Concentration of internal standard

C_s = Concentration of the analyte to be determined in the standard

10.6. Calibration curve fits

Average response factor, linear regression, or quadratic curves may be used to fit the data. Average response factor may be used if the average % RSD of the response factors or calibration factors of all the analytes in the calibration standard taken together is $\leq 20\%$. The average %RSD is calculated by summing the RSD value for each analyte and dividing by the total number of analytes.

10.6.1. In general, for environmental analysis, average response factors are the most appropriate calibration model. Linear or curved regression fits should only be used if the analyst has reason to believe that the average RF model does not fit the normal concentration/response behavior of the detector.

10.6.2. Average response factor

The average response factor may be used if the average percent relative standard deviation (%RSD) of all the response factors taken together is $\leq 20\%$.

The equation for average response factor is:

$$\text{Average response factor} = \overline{RF} = \frac{\sum_{i=1}^n RF_i}{n}$$

Where: n = Number of calibration levels

$\sum_{i=1}^n RF_i$ = Sum of response factors for each calibration level

10.6.3. Linear regression

The linear fit uses the following functions:

10.6.3.1. External Standard

$$y = ax + b$$

or

$$x = \frac{(y - b)}{a}$$

Where: y = Instrument response

x = Concentration

a = Slope

b = Intercept

10.6.3.2. Internal Standard

$$C_s = \frac{\left[\frac{A_s C_{is}}{A_{is}} - b \right]}{a}$$

Where: C_s = Concentration in the sample

A_s = Area of target peak in the sample

A_{is} = Area of internal standard in the sample

C_{is} = Concentration of the internal standard

10.6.4. Quadratic curve

The quadratic curve uses the following functions:

10.6.4.1. External standard

$$y = ax + cx^2 + b$$

Where c is the curvature

10.6.4.2. Internal Standard

$$y = a\left(\frac{A_s \times C_{is}}{A_{is}}\right) + c\left(\frac{A_s \times C_{is}}{A_{is}}\right)^2 + b$$

10.7. Evaluation of calibration curves

- 10.7.1. The percent relative standard error (%RSE) from the calibration curve is used to evaluate the initial calibration. This provides a measure of how much error is associated with using the calibration curve for quantitation.
- 10.7.2. The least squares regression line is calculated and used to calculate the predicted concentration for each level. The percent relative standard error is calculated as follows:

$$\% RSE = 100\% \times \sqrt{\frac{\sum_{i=1}^N \left[\frac{C_i - PC_i}{C_i} \right]^2}{(N - P)}}$$

Where:

N = Number of points in the curve

P = Number of parameters in the curve (= 1 for average response factor, 2 for linear, 3 for quadratic)

C_i = True concentration for level i

PC_i = Predicted concentration for level i

Note that when average response factors are used, %RSE is equivalent to %RSD.

10.8. The following requirements must be met for any calibration to be used:

- Response must increase with increasing concentration.
- If a curve is used, the intercept of the curve at zero response must be less than \pm the reporting limit for the analyte.
- The average Relative Standard Error (RSD for average response factors) of the calibration points from the curve used must be $\leq 20\%$.
- Some data systems will not measure the %RSE from a linear or quadratic fit. For the linear case, the correlation coefficient may be used as an alternative to the %RSE, and must be greater than or equal to 0.990. For the quadratic case the Coefficient of Determination may be used, and must be greater or equal to 0.990.

Note: The Relative Standard Error (RSE) is superior to the Correlation Coefficient (r) and Coefficient of Determination (r^2) for testing the fit of a set of calibration points to a line. The lower points on a curve have little effect on r . As a result a curve may have a very good correlation coefficient (>0.995), while also having $> 100\%$ error at the low point.

10.9. Weighting of data points

10.9.1. In linear and quadratic calibration fits, the points at the lower end of the calibration curve have less absolute variance than points at the high concentration end of the curve. This can cause severe errors in quantitation at the low end of the calibration. However, in environmental analysis, accuracy at the low end of the curve is very important. For this reason it is preferable to increase the weighting of the lower concentration points. $1/\text{Concentration}^2$ weighting (often called $1/X^2$ weighting) will improve accuracy at the low end of the curve and should be used if the data system has this capability.

10.10. Non-standard analytes are sometimes requested. For these analytes, it may be acceptable to analyze a single standard at the reporting limit with each continuing calibration rather than a five point initial calibration. This action must be with client approval. If the analyte is detected in any of the samples, a five point initial calibration must be generated and the sample(s) reanalyzed for quantitation.

10.11. Calibration Verification

10.11.1. 12 hour Calibration

The working calibration curve or RF must be verified by the analysis of a mid point calibration standard at the beginning, after every 12 hours, and at the end of the analysis sequence. The center of each retention time window is updated with each 12-hour calibration or calibration verification.

10.11.2. Calibration Verification

It may be appropriate to analyze a mid point standard more frequently than every 12 hours. If these calibration verification standards are analyzed, requirements are the same as the 12-hour calibration with the exception that retention times are not updated.

10.11.3. Any individual compounds with %D < 15% meet the calibration criteria. The calibration verification is also acceptable if the average of the %D for all the analytes is < 15%. This average is calculated by summing the entire absolute %D results in the calibration (including surrogates) and dividing by the number of analytes. Any analyte that is reportable as found must have a % difference of < 15% in the calibration verification or 12 hour calibration, on the column used for quantitation. Refer to section 12.1.2 for which result to report.

10.11.4. It is not necessary to run a calibration verification standard at the beginning of the sequence if samples are analyzed immediately after the completion of the initial calibration.

10.11.5. Samples quantitated by external standard methods must be bracketed by calibration verification standards that meet the criteria listed above. The bracketing standards on the column used for calibration must meet the same criteria as the opening standards. Bracketing is not necessary for internal standard methods.

10.11.6. If the analyst notes that a CCV has failed and can document the reason for failure (e.g. no purge, broken vial, carryover from the previous sample etc.) then a second CCV may be analyzed without any adjustments to the instrument. If this CCV meets criteria then the preceding samples have been successfully bracketed. If adjustments to the instrument are performed before the repeat CCV then the preceding samples have not been successfully bracketed but analysis may continue.

10.11.7. In general, it is not advisable to analyze repeat CCVs on unattended runs. If repeat CCVs are analyzed then the first will serve as the bracketing standard for the preceding samples and the last will serve as the CCV for the following samples.

10.11.8. If highly contaminated samples are expected it is acceptable to analyze blanks or primers at any point in the run.

10.11.9. % Difference calculation

% Difference for internal and external methods is calculated as follows

Internal Standard:

External standard:

$$\%D = \frac{RF_c - \overline{RF}}{\overline{RF}} \times 100$$

$$\%D = \frac{CF_c - \overline{CF}}{\overline{CF}} \times 100$$

Where RF_c and CF_c are the response and calibration factors from the continuing calibration

\overline{RF} and \overline{CF} are the average response and calibration factors from the initial calibration

10.11.10.% Drift calculation

% Drift is used for comparing the continuing calibration to a linear or quadratic curve. The criteria for % drift are the same as for % difference

$$\% \text{ Drift} = \frac{\text{Calculated Conc.} - \text{Theoretical Conc.}}{\text{Theoretical Conc.}} \times 100\%$$

10.11.11. Corrective Actions for Continuing Calibration

If the overall average %D of all analytes is greater than $\pm 15\%$ corrective action must be taken. This may include clipping the column, changing the liner or other minor instrument adjustments, followed by reanalyzing the standard. If the overall average %D still varies by more than $\pm 15\%$, a new calibration curve must be prepared.

10.11.12. Corrective Action for Samples

For internal standard methods, any samples injected after a standard not meeting the calibration criteria must be re-injected.

For external standard methods, any samples injected after the last good continuing calibration standard must be re-injected.

If the average %D for all the analytes in the calibration is over 15%, but all of the analytes requested for a particular sample have %D $\leq 15\%$, then the analysis is acceptable for that sample.

11. PROCEDURE

11.1. Extraction

Extraction procedures are referenced in the appendices.

11.2. Cleanup

Cleanup procedures are referenced in the appendices.

11.3. Gas Chromatography

Chromatographic conditions for individual methods are presented in the appendices.

11.4. Sample Introduction

In general, volatiles analytes are introduced using purge and trap as described in Appendix A. Semivolatile analytes are introduced by direct injection of the extract. Samples, standards, and QC must be introduced using the same procedure.

11.5. Analytical Sequence

An analytical sequence starts with an initial calibration or a daily calibration. Refer to the individual method appendices for method specific details of daily calibrations and analytical sequences.

11.5.1. The daily calibration includes analysis of standards containing all single response analytes and updating the retention time windows.

11.5.2. If there is a break in the analytical sequence of greater than 12 hours, a new analytical sequence must be started with a daily calibration.

11.6. Retention Time Windows

11.6.1. Retention time windows must be determined for all analytes. Make an injection of all analytes of interest each day over a three-day period. Calculate the standard deviation of the three retention times for each analyte (relative retention times may also be used). For multi-response analytes (e.g., Aroclors) use the retention time of major peaks. Plus or minus three times the standard deviation of the retention times of each analyte defines the retention time window.

11.6.2. The center of the retention time window is the retention time from the last of the three standards. The centers of the windows are updated with the mid- point of the initial calibration and each 12-hour calibration. The widths of the windows will remain the same until new windows are generated following the installation of a new column.

11.6.3. If the retention time window as calculated above is less than +/- 0.05 minutes, use +/- 0.05 minutes as the retention time window. This allows for slight variations in retention times caused by sample matrix.

11.6.4. The laboratory must calculate new retention time windows each time a new column is installed. The new windows must be generated within one week of the installation of the new column. Until these standards have been run on the new column, the retention time windows from the old column may be used, updated with the retention times from the new initial calibration.

11.6.5. Corrective Action for Retention Times

The retention times of all compounds in the 12 hour calibration or calibration verification standard must be within the retention time window. If this condition is not met, all samples analyzed after the last compliant standard must be reanalyzed unless the following conditions are met for any compound that elutes outside the retention time window:

The retention time of that compound in the standard must be within a retention time range equal to twice the original window.

No peak that would be reportable may be present on the sample chromatogram within an elution time range equal to three times the original retention time window.

11.7. Daily Retention Time Windows

The center of the retention time windows determined in Section 11.6 are adjusted to the retention time of each analyte as determined in the 12 hour calibration standards or continuing calibration verification standards. (See the method 8081A and 8082 appendices for exceptions for multi-response

components.) The retention time windows must be updated at the beginning of each analytical sequence and with each 12-hour calibration or continuing calibration verification.

11.8. Percent Moisture

Analytical results may be reported as dry or wet weight, as required by the client. Percent moisture must be determined if results will be reported as dry weight. Refer to SOP CORP-OP-0001NC for determination of percent moisture.

11.9. Procedural Variations

Procedural variations are allowed only if deemed necessary in the professional judgment of the supervisor to accommodate variation in sample matrix, radioactivity, chemistry, sample size, or other parameters. Any variation in procedure shall be completely documented using a Nonconformance Memo and approved by a supervisor and QA/QC manager. If contractually required, the client shall be notified. The Nonconformance Memo shall be filed in the project file. The nonconformance is also addressed in the case narrative. Any unauthorized deviations from this procedure must also be documented as a nonconformance, with a cause and corrective action described.

12. DATA ANALYSIS AND CALCULATIONS

12.1. Qualitative Identification

12.1.1. Tentative identification occurs when a peak is found within the retention time window for an analyte, at a concentration above the reporting limit, or above the MDL if J flags are required. Normally confirmation is required on a second column, but if the detector is sufficiently specific or if the sample matrix is well enough defined, single column analysis may be adequate. In some cases GC/MS confirmation may be required. Client specific requirements may also define the need for second column confirmation and / or GC/MS confirmation. Refer to the appendices for test specific requirements for confirmation. Identification is confirmed if a peak is also present in the retention time window for that analyte on the confirmatory column, at a concentration greater than the reporting limit (MDL if J flag confirmation required).

12.1.2. Dual column quantitation

For confirmed results, two approaches are available to the analyst.

A) The primary column approach

Or

B) The better result approach

Both are acceptable to avoid the reporting of erroneous or unconfirmed data.

12.1.2.1. Primary column approach:

The result from the primary column is normally reported. The result from the secondary column is reported if any of the following three bulleted possibilities are true.

- There is obvious chromatographic interference on the primary column
- The result on the primary column is > 40% greater than the result on the secondary column
- Continuing or bracketing standard fails on the primary column but is acceptable on the secondary column. (If the primary column result is > 40% higher than the secondary and the primary column calibration fails, then the sample must be evaluated for reanalysis.)

12.1.2.2. Better result approach

The lower of the two results is normally reported. The lower result is considered better because the higher result is generally higher because of chromatographic interference. The higher result is reported if any of the following two bulleted possibilities are true.

- There is obvious chromatographic interference on the column with the lower result
- The continuing or bracketing calibration on the column with the lower result fails. (If the higher result is > 40% higher and the calibration on the column with the lower result fails, then the sample must be evaluated for reanalysis.)

12.1.3. If the Relative percent difference (RPD) between the response on the two columns is greater than 40%, or if the opinion of an experienced analyst is that the complexity of the matrix is resulting in false positives, the confirmation is suspect and the results are qualified. RPD is calculated using the following formula:

$$RPD = \frac{R_1 - R_2}{\frac{1}{2}(R_1 + R_2)}$$

Where R=Result

12.1.4. Multi-response Analytes

For multi-response analytes, the analyst should use the retention time window, but should rely primarily on pattern recognition. The pattern of peaks will normally serve as confirmation.

12.1.5. The experience of the analyst should weigh heavily in the interpretation of the chromatogram. For example, sample matrix or laboratory temperature fluctuation may result in variation of retention times.

12.2. Calibration Range

If concentrations of any analytes exceed the working range as defined by the calibration standards, then the sample must be diluted and reanalyzed. Dilutions should target the most concentrated analyte in the upper half (over 50% of the high level standard) of the calibration range. It may be necessary to dilute samples due to matrix.

12.3. Dilutions

Samples may be screened to determine the appropriate dilution for the initial run. If the initial diluted run has no hits or hits below 20% of the calibration range and the matrix allows for analysis at a lesser dilution, then the sample must be reanalyzed at a dilution targeted to bring the largest hit above 50% of the calibration range.

12.3.1. Guidance for Dilutions Due to Matrix

If the sample is initially run at a dilution and only minor matrix peaks are, then the sample should be reanalyzed at a more concentrated dilution. Analyst judgement is required to determine the most concentrated dilution that will not result in instrument contamination.

12.3.2. Reporting Dilutions

The most concentrated dilution with no target compounds above the calibration range will be reported. Other dilutions may be reported at client request, if the lower dilutions will not cause detector saturation, column overload, or carryover. Analyst judgement and client site history will factors in the reporting of dual dilutions.

12.4. Interferences

If peak detection is prevented by interferences, further cleanup should be attempted. If no further cleanup is reasonable, then elevation of reporting levels and/or lack of positive identification must be addressed in the case narrative.

12.5. Internal Standard Criteria for Continuing Calibration

If internal standard calibration is used, then the internal standard response in a continuing calibration standard must be within 50 to 150% of the response in the mid level of the initial calibration.

12.6. Calculations

Capabilities of individual data systems may require the use of different formulas than those presented here. When this is the case, the calculations used must be shown to be equivalent and must be documented in an appendix attached to this document.

12.6.1. External Standard Calculations

12.6.1.1. Aqueous samples

$$\text{Concentration (mg / L)} = \frac{(A_x \times V_i \times D_f)}{(CF \times V_i \times V_s)}$$

Where:

A_x = Response for the analyte in the sample

V_i = Volume of extract injected, μL

D_f = Dilution factor

V_i = Volume of total extract, μL

V_s = Volume of sample extracted or purged, mL

CF = Calibration factor, area or height/ng, Section 10.1

12.6.1.2. Non-aqueous Samples

$$\text{Concentration (mg / kg)} = \frac{(A_x \times V_i \times D_f)}{(CF \times V_i \times W \times D)}$$

Where:

W = Weight of sample extracted or purged, g

$$D = \frac{100 - \% \text{Moisture}}{100} \quad (D = 1 \text{ if wet weight is required})$$

12.6.2. Internal Standard Calculations

12.6.2.1. Aqueous Samples

$$\text{Concentration (mg / L)} = \frac{(A_x \times C_{is} \times D_f)}{(A_{is} \times RF \times V_s)}$$

Where:

C_{is} = Amount of internal standard added, ng

A_{is} = Response of the internal standard

RF = Response factor for analyte

12.6.2.2. Non-aqueous Samples

$$\text{Concentration (mg / kg)} = \frac{(A_x \times C_{is} \times D_f)}{(A_{is} \times RF \times W \times D)}$$

12.6.3. Surrogate Recovery

Concentrations of surrogate compounds are calculated using the same equations as for the target compounds. The response factor from the initial calibration is used. Surrogate recovery is calculated using the following equation:

$$\% \text{ Recovery} = \frac{\text{Concentration (or amount) found}}{\text{Concentration (or amount) spiked}} \times 100$$

13. METHOD PERFORMANCE

13.1. Method Detection Limit

Each laboratory must generate a valid method detection limit for each analyte of interest. The MDL must be below the reporting limit for each analyte. The procedure for determination of the method detection limit is given in 40 CFR Part 136, Appendix B, and further defined in QA Policy #: QA-005.

13.2. Initial Demonstration

Each laboratory must make a one time initial demonstration of capability for each individual method. Demonstration of capability for both soils and water matrices is required. This requires analysis of QC check samples containing all of the standard analytes for the method. For some tests it may be necessary to use more than one QC check mix to cover all analytes of interest.

13.2.1. Four aliquots of the QC check sample are analyzed using the same procedures used to analyze samples, including sample preparation. The concentration of the QC check sample should be equivalent to a mid-level calibration.

13.2.2. Calculate the average recovery and standard deviation of the recovery for each analyte of interest. Compare these results with the acceptance criteria given in each appendix.

13.2.3. If any analyte does not meet the acceptance criteria, the test must be repeated. Only those analytes that did not meet criteria in the first test need to be evaluated. Repeated failure for any analyte indicates the need for the laboratory to evaluate the analytical procedure and take corrective action.

13.3. Training Qualification

The group/team leader has the responsibility to ensure that an analyst who has been properly trained in its use and has the required experience performs this procedure.

14. POLLUTION PREVENTION

This method does not contain any specific modifications that serve to minimize or prevent pollution.

15. WASTE MANAGEMENT

Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition, Final Update III, December 1996, and Section 8000B

17. MISCELLANEOUS

17.1. Modifications from Reference Method

17.1.1. Chapter 1 of SW-846 states that the method blank should not contain any analyte of interest at or above the Method Detection Limit. This SOP states that the Method Blank must not contain any analyte of interest at or above the reporting limit. Common lab contaminants are allowed to be up to 5 times the reporting limit in the blank following consultation with the client.

17.2. Modifications from Previous Revision


The calibration criteria in section 10.11 have been rewritten to improve consistency with SW-846 and to improve clarity.

17.3. Facility Specific SOPs

Each facility shall attach a list of facility specific SOPs or approved attachments (if applicable) which are required to implement this SOP or which are used in conjunction with this SOP. If no facility specific SOPs or amendments are to be attached, a statement must be attached specifying that there are none.


17.4. Flow Diagrams

17.4.1. Initial demonstration and MDL¹



¹ This flow diagram is for guidance and cannot cover all eventualities. Consult the SOP text and a supervisor if in doubt.

17.4.2. Sample Analysis¹



¹ This flow diagram is for guidance and cannot cover all eventualities. Consult the SOP text and a supervisor if in doubt.

1. SCOPE AND APPLICATION

- 1.1. This method describes sample preparation and extraction for the analysis of volatile organics by a purge and trap procedure, following method 8021B. However, where required by a client QAPP this section may also be used to analyze aromatic volatiles by discontinued methods 8020A and 8010B. All requirements of the 8000B section of this SOP must be met except when superseded by this Appendix. Refer to Table A-1 for the individual analytes normally determined by these procedures.
- 1.2. Compounds within the scope of this method have boiling points below 200°C and are soluble or slightly soluble in water. Classes of compounds best suited to purge-and-trap analysis include low molecular weight halogenated hydrocarbons, aromatics, ketones, nitriles, acetates, acrylates, ethers, and sulfides.
- 1.3. Water samples and soils samples with low levels of contamination may be analyzed directly by purge-and-trap extraction and gas chromatography. Higher concentrations of these analytes in soil may be determined by the medium level methanol extraction procedure.
- 1.4. This method also describes the preparation of water-miscible liquids, non-water-miscible liquids, solids, wastes, and soils/sediments for analysis by the purge-and-trap procedure.
- 1.5. The associated LIMS method code is QR.

2. SUMMARY OF METHOD

- 2.1. An inert gas is bubbled through the sample at ambient temperature or at 40°C (40°C required for low-level soils), and the volatile components are transferred from the aqueous phase to the vapor phase. The vapor is swept through a sorbent column where the volatile components are adsorbed. After purging is completed, the sorbent column is heated and back-flushed with inert gas to desorb the components onto a gas chromatographic column. Analytes are detected using a photoionization Detector, an electrolytic conductivity detector or a combination of both.
- 2.2. For soil samples, a portion of the sample is dispersed in methanol to dissolve the volatile organic constituents. A portion of the methanolic solution is combined with water. It is then analyzed by purge-and-trap GC following the normal water method. If very low detection limits are needed for soil samples then direct purge using sodium bisulfate preservation may be necessary.

3. DEFINITIONS

Refer to the STL North Canton Laboratory Quality Manual (LQM), current version, for definitions of terms used in this SOP.

4. INTERFERENCES

- 4.1. Refer to section 4 of the method 8000B part of this SOP for general information on chromatographic interferences.
- 4.2. Impurities in the purge gas, and from organic compounds out-gassing from the plumbing ahead of the trap, account for the majority of contamination problems. The analytical system must be demonstrated to be free from contamination under the conditions of the analysis by running laboratory reagent blanks. The use of non-TFE plastic tubing, non-TFE thread sealants, or flow controllers with rubber components in the purging device should be avoided.
- 4.3. Samples can be contaminated by diffusion of volatile organics (particularly methylene chloride and fluorocarbons) through the septum seal of the sample vial during shipment and storage. A trip blank

**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B**

prepared from organic-free reagent water and carried through sampling and handling protocols serves as a check on such contamination.

- 4.4. Contamination by carryover can occur whenever high-concentration and low-concentration samples are analyzed sequentially. Whenever an unusually concentrated sample is analyzed, it should be followed by an analysis of organic-free reagent water to check for cross-contamination. The trap and other parts of the system are subject to contamination. Therefore, frequent bake-out and purging of the system may be required.
- 4.5. When utilizing an autosampler system, which has multiple ports for sample analysis, it is likely that only a single stage or port may be contaminated by a highly concentrated sample. If a port is suspect, a water blank should be analyzed to verify lack of contamination. If the water blank and subsequent blanks on that port show contamination consistent with the concentrated sample, further maintenance is required. This may include replacing or cleaning the multi-port valve, transfer lines, etc.
- 4.6. A holding blank is kept in the sample refrigerator. This is analyzed and replaced every 14 days. If the holding blank does not meet the method blank criteria, the source of contamination must be found and corrected. Evaluation of all samples analyzed in the 14-day period prior to the analysis of the contaminated holding blank is required.
- 4.7. Acidification of samples may result in hydrolysis of 2-chloroethyl-vinyl ether.

5. SAFETY

- 5.1. Refer to section 5 of the Method 8000B section of this SOP for general safety requirements.
- 5.2. Often, purge vessels on purge-and-trap instrumentation are pressurized by the time analysis is completed. Therefore, vent the pressure prior to removal of these vessels to prevent the contents from spraying out.
- 5.3. The toxicity or carcinogenicity of each chemical used in this procedure has not been fully defined. Additional health and safety information can be obtained from the MSDS files maintained in the laboratory. The following specific hazards are known:

Methanol -- Flammable and toxic
- 5.4. The following method analytes have been tentatively classified as known or suspected human or mammalian carcinogens: Benzene, Carbon Tetrachloride, 1,4-Dichlorobenzene, 1,2-Dichloroethane, Hexachlorobutadiene, 1,1,2,2-Tetrachloroethane, 1,1,2-Trichloroethane, Chloroform, 1,2-Dibromoethane, Tetrachloroethene, Trichloroethene, Vinyl Chloride. Pure standard materials and stock standard solutions of these compounds should be handled in a hood.
- 5.5. Methanol shall not be used in a CaptAir hood.

6. EQUIPMENT AND SUPPLIES

- 6.1. Microsyringes -- 10 μ L, 25 μ L, 100 μ L, 250 μ L, 500 μ L, and 1000 μ L. These should be equipped with a 20 gauge (0.006" ID) needle. These will be used to measure and dispense methanolic solutions and aqueous samples.
- 6.2. Gas tight syringes -- 5 mL and 25 mL. Used for measuring sample volumes.
- 6.3. Purge and Trap Apparatus -- A device capable of extracting volatile compounds, trapping on a sorbent trap, and introducing onto a gas chromatograph.

**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B**

- 6.4. Purge and Trap Autosampler -- In order to maintain high sample throughput, an autosampler is highly recommended.
- 6.5. Trap -- The trap used is dependent on the class of compound to be analyzed. Refer to Table A-2 for suggested traps for specific tests.
- 6.6. Purge Vessels -- These are dependent on the purge and trap unit/autosampler used. Both disposable culture tubes (needle sparge units) and specially designed vessels with fritted bottoms may be used. Follow the manufacturer's suggestions for configuration.
- 6.7. Columns - Refer to Table A-2 for details of columns.
- 6.8. Volumetric flasks, Class A: 5 mL to 250 mL
- 6.9. pH paper
- 6.10. Balance capable of weighing to 0.01g for samples.

7. REAGENTS AND SUPPLIES

- 7.1. Refer to the method 8000B section of this SOP for general requirements for reagents and supplies.

- 7.2. Organic Free Water

Organic free water is defined as water in which an interferent is not observed at the reporting limit of the compounds of interest. Suggested methods for generating organic free water include:

- Filtration through a carbon bed.
- Continuously sparging water with helium or nitrogen.
- Use of commercial water purification systems.

Other methods may be used, so long as the requirement that the water show no interference is met. The procedure used should be documented in a lab specific attachment.

- 7.3. Sodium Bisulfate

- 7.4. Methanol -- Purge and Trap Grade

- 7.5. Standards

Refer to tables A-5 and A-6 for details of surrogate, matrix spiking and internal standards. Calibration standard levels are not specified, since they may depend on the sensitivity and linear range of specific detectors. However, the low level standard must be equivalent to the reporting limits specified in Table A-1.

- 7.5.1. Volatile standards are prepared by injecting a measured volume of the stock standard into a syringe containing the appropriate volume of organic free water. The calibration standard is then loaded into the purge device.

8. SAMPLE COLLECTION, PRESERVATION, AND STORAGE

- 8.1. Holding times for all volatile analysis are 14 days from sample collection.
- 8.2. Water samples are normally preserved at pH < 2 with 1:1 hydrochloric acid. If residual chlorine is present, 2 drops of 10% sodium thiosulfate are added.

**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B**

- 8.3. Solid samples are field preserved with sodium bisulfate solution for low level analysis, or with methanol for medium level analysis. Soil samples can also be taken using the EnCore™ sampler and preserved in the lab within 48 hours of sampling. At specific client request, unpreserved soil samples may be accepted.
- 8.4. There are several methods of sampling soil. The recommended method, which provides the minimum of field difficulties, is to take an EnCore sample. (The 5 g or 25 g sampler can be used, depending on client preference). Following shipment back to the lab the soil is preserved in methanol. This is the medium level procedure. If very low detection limits are needed (< 50 µg/kg for most analytes) then it will be necessary to use two additional 5 g EnCore samplers or to use field preservation.
- 8.5. Sample collection for medium level analysis using EnCore samplers.
- 8.5.1. Ship one 5 g (or 25 g) EnCore sampler per field sample position.
- 8.5.2. An additional bottle must be shipped for percent moisture determination.
- 8.5.3. When the samples are returned to the lab, extrude the (nominal) 5g (or 25 g) sample into a tared VOA vial containing 5 mL methanol (25 mL methanol for the 25 g sampler). Obtain the weight of the soil added to the vial and note on the label.
- 8.5.4. Add the correct amount of surrogate spiking mixture. (Add 100 µL of 250 µg/mL solution for a nominal 25 g sample, 20µL for a nominal 5 g sample.)
- 8.5.5. Add the correct amount of matrix spiking solution to the matrix spike and matrix spike duplicate samples. (Add 100 µL of 250 µg/mL solution for a nominal 25 g sample, 20µL for a nominal 5 g sample.) The addition of spike introduces a slight error, (0.4%) which can be neglected, into the calculations.
- 8.5.6. Prepare an LCS for each batch by adding the correct amount of matrix spiking solution to clean methanol. (100 µL of spike to 25 mL methanol or 20 µL spike to 5 mL methanol).
- 8.5.7. Shake the samples for two minutes to distribute the methanol throughout the soil.
- 8.5.8. Allow to settle, then remove a portion of methanol and store in a clean Teflon capped vial at 4+2°C until analysis.
- 8.6. Sample collection for medium level analysis using field methanol preservation
- 8.6.1. Prepare a VOA vial by adding 5 mL purge and trap grade methanol. (If a 25 g sample is to be used, add 25 mL methanol to the VOA vial).
- 8.6.2. Seal the bottle and attach a label.
- 8.6.3. Weigh the bottle to the nearest 0.01g and note the weight on the label.
- 8.6.4. Ship with appropriate sampling instructions.
- 8.6.5. Each sample will require an additional bottle with no preservative for percent moisture determination.
- 8.6.6. At client request, the methanol addition and weighing may also be performed in the field.
- 8.6.7. When the samples are returned to the lab, obtain the weight of the soil added to the vial and note on the label.

**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B**

- 8.6.8. Add the correct amount of surrogate spiking mixture. (Add 100 μL of 250 $\mu\text{g}/\text{mL}$ solution for a nominal 25 g sample, 20 μL for a nominal 5 g sample.)
- 8.6.9. Add the correct amount of matrix spiking solution to the matrix spike and matrix spike duplicate samples. (Add 100 μL of 250 $\mu\text{g}/\text{mL}$ solution for a nominal 25 g sample, 20 μL for a nominal 5 g sample.) The addition of spike introduces a slight error, (0.4%) which can be neglected, into the calculations.
- 8.6.10. Prepare an LCS for each batch by adding the correct amount of matrix spiking solution to clean methanol. (100 μL of spike to 25 mL methanol or 20 μL spike to 5 mL methanol).
- 8.6.11. Shake the samples for two minutes to distribute the methanol throughout the soil.
- 8.6.12. Allow to settle, then remove a portion of methanol and store in a clean Teflon capped vial at 4 \pm 2 $^{\circ}\text{C}$ until analysis.
- 8.7. Low level procedure
- 8.7.1. If low detection limits are required (typically < 50 $\mu\text{g}/\text{kg}$) sodium bisulfate preservation must be used. However, it is also necessary to take a sample for the medium level (methanol preserved) procedure, in case the concentration of analytes in the soil is above the calibration range of the low-level procedure.
- 8.7.2. A purge and trap autosampler capable of sampling from a sealed vial is required for analysis of samples collected using this method. (Varian Archon or O.I. 4552).
- 8.7.3. The soil sample is taken using a 5g EnCore sampling device and returned to the lab. It is recommended that two EnCore samplers be used for each field sample position, to allow for any reruns than may be necessary. A separate sample for % moisture determination is also necessary.
- 8.7.4. Prepare VOA vials by adding a magnetic stir bar, approximately 1 g of sodium bisulfate and 5 mL of reagent water.
- 8.7.5. Seal the vial and attach a label. The label must not cover the neck of the vial or the autosampler will malfunction.
- 8.7.6. Weigh the vial to the nearest 0.01g and note the weight on the label.
- 8.7.7. Extrude the soil sample from the EnCore sampler into the prepared VOA vial. Reweigh the vial to obtain the weight of soil and note on the label.
- 8.7.8. **Note:** Soils containing carbonates may effervesce when added to the sodium bisulfate solution. If this is the case at a specific site, add 5 mL of water instead, and freeze at $\geq -10^{\circ}\text{C}$ until analysis.
- 8.7.9. Alternatively the sodium bisulfate preservation may be performed in the field. Ship at least two vials per sample. The field samplers must determine the weight of soil sampled. Each sample will require an additional bottle with no preservative for percent moisture determination, and an additional bottle preserved with methanol for the medium level procedure.
- 8.8. Aqueous samples are stored in glass containers with Teflon lined septa at 4 $^{\circ}\text{C}$ \pm 2 $^{\circ}\text{C}$, with minimum headspace.
- 8.9. Medium level solid extracts are aliquoted into 2 - 5 mL glass vials with Teflon lined caps and stored at 4 $^{\circ}\text{C}$ \pm 2 $^{\circ}\text{C}$. The extracts are stored with minimum headspace.
- 8.10. The maximum holding time is 14 days from sampling until the sample is analyzed. (Samples that are found to be unpreserved still have a 14 day holding time. However they should be analyzed as soon as

**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B**


possible. The lack of preservation should be addressed in the case narrative). Maximum holding time for the EnCore sampler (before the sample is added to methanol or sodium bisulfate) is 48 hours.

- 8.11. A holding blank is stored with the samples. This is analyzed and replaced if any of the trip blanks show any contamination. Otherwise it is replaced every 14 days.


Note: Freezing is not allowed for Ohio VAP solids.

ANALYSIS OF VOLATILE ORGANICS BASED ON METHOD 8021B


EnCore procedure when low level is not required (field steps in gray)



EnCore procedure when low level is required



Field methanol extraction procedure (field steps in gray)



**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B****9. QUALITY CONTROL**

- 9.1. Refer to the method 8000B section of this SOP, section 9, for general quality control procedures, including batch definition, requirements for method blanks, LCS, matrix spikes, surrogates, and control limits.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Refer to the method 8000B section of this SOP, section 10, for general calibration procedures.
- 10.2. Gas Chromatograph Operating Conditions

Various column configurations are possible. If dual column confirmation is necessary, the sample may be split using a Y splitter at the injector end to direct the sample to two columns and two detectors. For simultaneous determination of aromatic and halogenated volatiles, a single column is used and the PID and ELCD detectors are connected in series.

- 10.2.1. Refer to Table A-2, A-3 and A-4 for GC operating conditions.

10.3. Initial Calibration

- 10.3.1. Refer to Section 10 of the 8000B section of this SOP for details of initial calibration criteria.
- 10.3.2. Low level soil samples must be purged at 40°C; therefore the calibration curve must also be purged at 40°C. In addition, the low level soil calibration solutions should contain approximately the same amount of sodium bisulfate as the samples.
- 10.3.3. The low level calibration must be at the reporting limit or below. The remaining standards encompass the working range of the detector.
- 10.3.4. Calibrate the instrument using the same volume that will be used during sample analysis.

10.4. Calibration Verification

- 10.4.1. A mid level calibration standard is used for the calibration verification. The gases have 20 % D criteria rather than the 15% used for other analytes.
- 10.4.2. A calibration verification run is performed after every 10 samples for this method.
- 10.4.3. Bracketing of samples with calibration verification runs is only necessary for external standard analysis.

11. PROCEDURE

- 11.1. Refer to the method 8000B section of this SOP for general procedural requirements.
- 11.2. Analytical Sequence
- The analytical sequence starts with an initial calibration of at least five points, or a 12 hour calibration that meets % difference criteria from an existing initial calibration.
- 11.3. Confirmation
- The PID and ELCD detectors are sufficiently selective that second column confirmation is not always necessary. Requirements for second column confirmation should be decided in consultation with the client. If the PID and ELCD are used in series confirmatory information for many analytes can be gained by comparing the relative response from the two detectors.

- 11.4. Aqueous Sample Analysis (Purge and Trap units using sparge vessels)
- 11.4.1. Depending on the sensitivity of the instrument and capabilities of the purge and trap device, 5, 10, 20, or 25 mL sample volumes may be analyzed. A 5 mL sample volume is recommended.
 - 11.4.2. Rinse a 5 mL (or 25 mL for larger sample volumes) gas-tight syringe with organic free water. Fill the syringe with the sample to be analyzed, and compress to volume.
 - 11.4.3. Check and document the pH of the sample remaining in the VOA vial after loading the syringe.
 - 11.4.4. This procedure invalidates the contents of the VOA vial for further analysis, unless an aliquot is transferred to a smaller VOA vial with no headspace (e.g., 20 mL) at the same time the analysis aliquot is removed.
 - 11.4.5. Spike with the appropriate volume of surrogate/internal standard solution and spike solution (if required) through the barrel of the syringe. The method blank is spiked with surrogates only, the LCS and matrix spikes with the surrogate and matrix spiking solutions. Refer to Tables A-5 and A-6 for volumes and concentrations of spiking solutions.
 - 11.4.6. Load onto the purge and trap device and start the run.
 - 11.4.7. If the initial analysis of a sample or a dilution of the sample has a concentration of analytes that exceeds the initial calibration range, the sample must be reanalyzed at a higher dilution. When a sample has a high response for a compound, an organic free water blank should follow analysis. It is recognized that during automated unattended analysis, this may not occur. If any potential carryover hits are present in samples following highly contaminated samples, the sample must be reanalyzed to determine if any of these hits are a result of carryover or are actually present in the sample.
 - 11.4.8. Dilutions may be made in gas tight syringes unless the volume of sample used is less than 5 μ L, in which case dilution in volumetric flasks will be necessary.
 - 11.4.8.1. Spike with the same volume of surrogate/internal standard solution as used for undiluted samples prior to loading onto the purge and trap device.
 - 11.4.8.2. For Matrix spike / matrix spike duplicates where the sample requires dilution, the sample is spiked after the dilution is performed.
- 11.5. Aqueous and Soil Sample Analysis (Purge and Trap units that sample directly from the VOA vial)
- 11.5.1. Units, which sample from the VOA vial, should be equipped with a module, which automatically adds surrogate and internal standard solution to the sample prior to purging the sample.
 - 11.5.2. If the autosampler uses automatic IS/SS injection, no further preparation of the VOA vial is needed. Otherwise the internal and surrogate standards must be added to the vial. *Note:* Aqueous samples with high amounts of sediment present in the vial may not be suitable for analysis on this instrumentation, or they may need to be analyzed as soils.
 - 11.5.3. Sample remaining in the vial after sampling with one of these mechanisms is no longer valid for further analysis. A fresh VOA vial must be used for further sample analysis.
 - 11.5.4. Check the pH of the sample remaining in the VOA vial after analysis is completed.

**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B****11.6. Low-Level Solids Analysis using discrete autosamplers**

Note: This technique may seriously underestimate analyte concentration and must not be used except at specific client request for the purpose of comparability with previous data. It is no longer part of SW-846.

This method is based on purging a heated sediment/soil sample mixed with reagent water containing the surrogate and, if applicable, internal and matrix spiking standards. Analyze all reagent blanks and standards under the same conditions as the samples (e.g., heated). The calibration curve is also heated during analysis. Purge temperature is 40°C.

11.6.1. *Do not discard any supernatant liquids. Mix the contents of the container with a narrow metal spatula.*

11.6.2. *Weigh out 5 g (or other appropriate aliquot) of sample into a disposable culture tube or other purge vessel. Record the weight to the nearest 0.1 g. If method sensitivity is demonstrated, a smaller aliquot may be used. Do not use aliquots less than 1.0 g. If the sample is contaminated with analytes such that a purge amount less than 1.0 g is appropriate, use the medium level method described in section 11.7.*

11.6.3. *Connect the purge vessel to the purge and trap device.*

11.6.4. *Rinse a 5 mL gas-tight syringe with organic free water, and fill. Compress to 5 mL. Add surrogate/internal standard (and matrix spike solutions if required.) (See Tables A-5, A-6, A-7 and A-8.) Add directly to the sample from 11.6.2.*

11.6.5. *The above steps should be performed rapidly and without interruption to avoid loss of volatile organics.*

11.6.6. *Add the heater jacket or other heating device and start the purge and trap unit.*

11.6.7. *Soil samples that have low IS recovery when analyzed (<50%) should be reanalyzed once to confirm matrix effect. If external standard calibration is used, samples with surrogate recovery below the control limit should be reanalyzed once to confirm matrix effect.*

11.7. Methanol Extract Soils

11.7.1. *Rinse a gas-tight syringe with organic free water. Fill the syringe with the same volume of organic free water as used in the calibrations. Add no more than 2% (v/v) (100 µL for a 5 mL purge) methanolic extract (from Section 8.5 or 8.6) to the syringe. Add internal standard (if used). Load the sample onto the purge and trap device and analyze as for aqueous samples. If less than 5µL of methanolic extract is to be added to the water, dilute the methanolic extract such that a volume greater than 5µL will be added to the water in the syringe.*

12. DATA ANALYSIS AND CALCULATIONS

Refer to section 12 of the 8000B section of this SOP.

13. METHOD PERFORMANCE

13.1. *Performance limits for the four replicate initial demonstration of capability required under Section 13.1 of the 8000B section of this SOP.*

14. POLLUTION PREVENTION

This method does not contain any specific modifications that serve to minimize or prevent pollution.

**ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B****15. WASTE MANAGEMENT**

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

- 16.1. Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition, Final Update III, December 1996, Sections 5000, 5030B, 5035 and 8021B

17. MISCELLANEOUS

- 17.1. Modifications from Reference Method

- 17.2. Modifications from previous revision

17.2.1. No revisions were made to this appendix.

- 17.3. Facility Specific SOPs

Each facility shall attach a list of facility specific SOPs or approved attachments (if applicable) which are required to implement this SOP or which are used in conjunction with this SOP. If no facility specific SOPs or amendments are to be attached, a statement must be attached specifying that there are none.

ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B

Revision No. 5.6

Revision Date: 05/25/01

Page A13 of A16

17.4. Tables

Table A-1					
Standard Analyte List					
Test	Compound	CAS number	Reporting Limit, µg/L or µg/kg		
			Aqueous	Low Soil	Medium Soil
Halogenated volatiles by 8021B	Bromodichloromethane	75-27-4	1.0	1.0	50
	Bromoform	75-25-2	1.0	1.0	50
	Bromomethane	74-83-9	1.0	1.0	50
	Carbon Tetrachloride	56-23-5	1.0	1.0	50
	Chlorobenzene	108-90-7	1.0	1.0	50
	Chloroethane	70-00-3	1.0	1.0	50
	2-Chloroethyl vinyl ether	110-75-8	5.0	5.0	250
	Chloroform	67-66-3	1.0	1.0	50
	Chloromethane	74-87-3	1.0	1.0	50
	Dibromochloromethane	124-48-1	1.0	1.0	50
	1,2-Dichlorobenzene	95-50-1	1.0	1.0	50
	1,3-Dichlorobenzene	541-73-1	1.0	1.0	50
	1,4-Dichlorobenzene	106-46-7	1.0	1.0	50
	Dichlorodifluoromethane	75-71-8	1.0	1.0	50
	1,1-Dichloroethane	75-34-3	1.0	1.0	50
	1,2-Dichloroethane	107-06-2	1.0	1.0	50
	1,1-Dichloroethene	75-45-4	1.0	1.0	50
	cis-1,2 Dichloroethene	156-59-4	1.0	1.0	50
	trans-1,2-Dichloroethene	156-60-5	1.0	1.0	50
	Dichloromethane(DCM)	75-09-2	5.0	5.0	250
	1,2-Dichloropropane	78-87-5	1.0	1.0	50
	cis-1,3-Dichloropropene	10061-01-5	1.0	1.0	50
	trans-1,3-Dichloropropene	10061-02-6	1.0	1.0	50
	1,1,2,2-Tetrachloroethane	79-34-5	1.0	1.0	50
	Tetrachloroethene	127-18-4	1.0	1.0	50
	1,1,1-Trichloroethane	71-55-6	1.0	1.0	50
	1,1,2-Trichloroethane	79-00-5	1.0	1.0	50
Trichloroethene	79-01-6	1.0	1.0	50	
Trichlorofluoromethane	75-69-4	1.0	1.0	50	
Vinyl Chloride	75-01-4	1.0	1.0	50	
Additional halogenated volatiles	Benzyl Chloride	100-44-7	5.0	5.0	250
	Bromobenzene	108-86-1	1.0	1.0	50
	Dibromomethane	74-95-3	1.0	1.0	50
	1,1,1,2-Tetrachloroethane	630-20-6	1.0	1.0	50
	Freon 113	76-13-1	1.0	1.0	50
	1,2,3-Trichloropropane	96-18-4	1.0	1.0	50
BTEX by 8021B	Benzene	71-43-2	1.0	1.0	50
	Ethyl Benzene	100-41-4	1.0	1.0	50
	Toluene	108-88-3	1.0	1.0	50
	Xylenes (total)	1330-20-7	1.0	1.0	50
Aromatic volatiles by 8021B	Benzene	71-43-2	1.0	1.0	50

10/11/01

ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B

Revision No. 5.6

Revision Date: 05/25/01

Page A14 of A16

Table A-1					
Standard Analyte List					
Test	Compound	CAS number	Reporting Limit, µg/L or µg/kg		
			Aqueous	Low Soil	Medium Soil
	Chlorobenzene	108-90-7	1.0	1.0	50
	1,2-Dichlorobenzene	75-34-3	1.0	1.0	50
	1,3-Dichlorobenzene	107-06-2	1.0	1.0	50
	1,4-Dichlorobenzene	75-45-4	1.0	1.0	50
	Ethyl Benzene	100-41-4	1.0	1.0	50
	Toluene	108-88-3	1.0	1.0	50
	Xylenes (total)	1330-20-7	1.0	1.0	50
Additional aromatic and unsaturated volatiles	1,2,4 Trimethylbenzene	95-63-6	1.0	1.0	50
	1,3,5 Trimethylbenzene	108-67-8	1.0	1.0	50
	Acetone	67-64-1	10	10	500
	MEK (2-butanone)	78-93-3	5.0	5.0	250
	MIBK (4-methyl-2-pentanone)	108-10-1	5.0	5.0	250
	Naphthalene	91-20-3	2.0	2.0	250
	Styrene	100-42-5	1.0	1.0	50
Methyl tert-butyl ether (MTBE)	1634-04-4	1.0	1.0	50	
Combined halogenated and aromatic volatiles by 8021B	Benzene	71-43-2	1.0	1.0	50
	Bromobenzene	108-86-1	1.0	1.0	50
	Bromochloromethane	74-97-5	1.0	1.0	50
	Bromodichloromethane	75-27-4	1.0	1.0	50
	Bromoform	75-25-2	1.0	1.0	50
	Bromomethane	74-83-9	1.0	1.0	50
	n-butylbenzene	104-51-8	1.0	1.0	50
	sec-Butylbenzene	135-98-8	1.0	1.0	50
	tert-Butylbenzene	98-06-6	1.0	1.0	50
	Carbon Tetrachloride	56-23-5	1.0	1.0	50
	Chlorobenzene	108-90-7	1.0	1.0	50
	Chlorodibromomethane	124-48-1	1.0	1.0	50
	Chloroethane	75-00-3	1.0	1.0	50
	Chloroform	67-66-3	1.0	1.0	50
	Chloromethane	74-87-3	1.0	1.0	50
	2-Chlorotoluene	95-49-8	1.0	1.0	50
	4-Chlorotoluene	106-43-4	1.0	1.0	50
	1,2-Dibromo-3-Chloropropane(DBCP)	96-12-8	1.0	1.0	50
	1,2-Dibromoethane(EDB)	106-93-4	1.0	1.0	50
	Dibromomethane	74-95-3	1.0	1.0	50
	1,2-Dichlorobenzene	95-50-1	1.0	1.0	50
	1,3-Dichlorobenzene	541-73-1	1.0	1.0	50
	1,4-Dichlorobenzene	106-46-7	1.0	1.0	50
	Dichlorodifluoromethane	75-71-8	1.0	1.0	50
	1,1-Dichloroethane	75-34-3	1.0	1.0	50
	1,2-Dichloroethane	107-06-2	1.0	1.0	50
	1,1-Dichloroethene	75-35-4	1.0	1.0	50
cis-1,2-Dichloroethene	156-59-4	1.0	1.0	50	
trans-1,2-Dichloroethene	156-60-5	1.0	1.0	50	
1,2-Dichloropropane	78-87-5	1.0	1.0	50	

ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B

Revision No. 5.6

Revision Date: 05/25/01

Page A15 of A16

Table A-1					
Standard Analyte List					
Test	Compound	CAS number	Reporting Limit, µg/L or µg/kg		
			Aqueous	Low Soil	Medium Soil
	1,3-Dichloropropane	142-28-9	1.0	1.0	50
	2,2-Dichloropropane	590-20-7	1.0	1.0	50
	1,1-Dichloropropene	563-58-6	1.0	1.0	50
	cis-1,3-Dichloropropene	10061-01-5	1.0	1.0	50
	trans-1,3-Dichloropropene	10061-02-6	1.0	1.0	50
	Ethylbenzene	100-41-4	1.0	1.0	50
	Hexachlorobutadiene	87-68-3	1.0	1.0	50
	Isopropylbenzene	98-82-8	1.0	1.0	50
	p-Isopropyltoluene	99-87-6	1.0	1.0	50
	Methylene Chloride	75-09-2	5.0	5.0	250
	Naphthalene	91-20-3	2.0	2.0	250
	n-Propylbenzene	10306501	1.0	1.0	50
	Styrene	100-42-5	1.0	1.0	50
	1,1,1,2-Tetrachloroethane	630-20-6	1.0	1.0	50
	1,1,2,2-Tetrachloroethane	79-34-5	1.0	1.0	50
	Tetrachloroethene	127-18-4	1.0	1.0	50
	Toluene	108-88-3	1.0	1.0	50
	1,2,3-Trichlorobenzene	87-61-6	1.0	1.0	50
	1,2,4-Trichlorobenzene	120-82-1	1.0	1.0	50
	1,1,1-Trichloroethane	71-55-6	1.0	1.0	50
	1,1,2-Trichloroethane	79-00-5	1.0	1.0	50
	Trichloroethene	79-01-6	1.0	1.0	50
	Trichlorofluoromethane	75-69-4	1.0	1.0	50
	1,2,3-Trichloropropane	96-18-4	1.0	1.0	50
	1,2,4-Trimethylbenzene	95-63-6	1.0	1.0	50
	1,3,5-Trimethylbenzene	108-67-8	1.0	1.0	50
	Vinyl Chloride	75-01-4	1.0	1.0	50
	Xylenes (total)	1330-20-7	1.0	1.0	50

Table A-2	
Recommended Conditions for Aromatic Volatiles	
Parameter	Recommended Conditions
Temperature program	50°C, 1min, 10°C/min to 200°C, 1min
Column 1	Rtx-502.2 or DB-502.2 60m x 0.53mm 3.0µm
Column 2	Rtx-1 or DB-1 60m x 0.53mm 3.0 µm
Carrier gas	Helium or hydrogen
Purge Flow / time	40 mL/min, 11 minutes
Desorb Temp / time	180°C, 2 minutes (220°C for Vocarb 3000)
Bake Time / temp	200°C, 12 minutes (230°C for Vocarb 3000)
Transfer line / valve temp	115°C

ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B

Revision No. 5.6

Revision Date: 05/25/01

Page A16 of A16

Table A-3	
Recommended Conditions for Method Halogenated Volatiles	
Parameter	Recommended Conditions
Temperature program	35°C, 12 min, then 4°C/min to 200°C, hold for 5 min
Column 1	DB-VRX or RTX-502.2 105m x 0.53 mm id df = 3.0um
Column 2	DB-1 or RTX-1 105m x 0.53 mm ID df = 3.0um
Column 3	Rtx - Volatiles 120m x 0.53mm ID df=2.0um
Carrier gas	Helium
Purge Flow / time	40 mL/min, 11 minutes
Desorb Temp / time	180°C, 2 minutes (220°C for Vocarb 3000)
Bake Time / temp	200°C, 12 minutes (230°C for Vocarb 3000)
Transfer line / valve temp	115°C

Table A-4	
Recommended Conditions for Method Combined Aromatic and Halogenated Volatiles	
Parameter	Recommended Conditions
Temperature program	35°C, 12 min, then 4°C/min to 200°C, hold for 5 min
Column 1	DB-VRX or RTX-502.2 105m x 0.53 mm id df = 3.0um
Column 2	DB-1 or RTX-1 105m x 0.53 mm ID df = 3.0um
Column 3	Rtx - Volatiles 120m x 0.53mm ID df=2.0um
Carrier gas	Helium
Purge Flow / time	40 mL/min, 11 minutes
Desorb Temp / time	180°C, 2 minutes (220°C for Vocarb 3000)
Bake Time / temp	200°C, 12 minutes (230°C for Vocarb 3000)
Transfer line / valve temp	115°C

ANALYSIS OF VOLATILE ORGANICS BASED ON
METHOD 8021B

Table A-5 Surrogate and Internal Standard Concentrations				
Standard	Components	Working Solution µg/mL	Spike amount µL (for 5 mL purge)	Final concentration µg/L (µg/kg)
Aromatic volatiles IS/SS	4-Chlorotoluene (SS)	20	5	20
	1-Chloro-4-fluorobenzene (IS)	40		40
Halogenated volatiles IS/SS	4-chlorotoluene (SS)	20	5	20
	1-Chloro-4-fluorobenzene (IS)	40		40
Combined Aromatic and halogenated volatiles IS/SS	Fluorobenzene (SS)	20	5	20
	1,4-Dichlorobutane (SS)	20	5	20
	1-Chloro-4-fluorobenzene (IS)	40	10	40

It may be necessary to select different surrogates in order to minimize sample interferences. 1-chloro-4-fluorobenzene and 4-chlorotoluene are fairly well resolved from analytes listed in this SOP. However 4-chlorotoluene may sometimes be requested as a target analyte. Other surrogates that may be considered, and issues associated with their use are:

Bromochloromethane:	Elutes very close to chloroform and cis-1, 2-dichloroethene on the 502.2 column. May be a target analyte.
1,2-Bromochloroethane:	
1-Chloro-2-fluorobenzene:	Elutes close to ethylbenzene on DB-1 or Rtx-1 and close to m,p-xylene on 502.2
a,a,a-Trifluorotoluene:	Good for aromatic volatiles, coelutes or very close to trichloroethene
Bromofluorobenzene:	Close to 1,1,2,2-trichloroethane and 1,2,3-trichloropropane on the 502.2 column. Good on DB-1 or Rtx-1.
2-Bromo-1-chloropropane:	May coelute with 1,1,2-trichloroethane

Table A-6 Concentrations for LCS and MS/MSD compounds				
Standard	Components	Working Solution µg/mL	Spike amount µL (5 mL purge)	Final concentration µg/L (µg/kg)
Aromatic	Benzene	20	5	20
	Toluene	20		20
	Chlorobenzene	20		20
Halogenated	Chlorobenzene	20	5	20
	1,1-Dichloroethene	20		20
	Trichloroethene	20		20
Combination aromatic / halogenated	Benzene	20	5	20
	Toluene	20		20
	Chlorobenzene	20		20
	1,1-Dichloroethene	20		20
	Trichloroethene	20		20

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608****1. SCOPE AND APPLICATION**

This SOP Appendix describes procedures to be used when SW-846 Method 8081A is applied to the analysis of organochlorine pesticides by GC/ECD. This Appendix may also be applied when discontinued SW-846 Method 8080A is requested, and is applicable to extracts derived from any matrix which are prepared according to the appropriate STL North Canton sample extraction SOPs. (CORP-OP-0001NC) Criteria for the analysis of wastewater by Method 608 is also included in this appendix.

Table B-1 lists compounds, which are routinely determined by this method, and gives the Reporting Limits (RL) for each matrix. RLs given are based on the low level standard and the sample preparation concentration factors. Matrix interferences may result in higher RLs than those listed.

At client request, this method may also be used for the analysis of PCBs (Arochlors) in combination with pesticides, although these are normally analyzed following method 8082, as described in Appendix C of this SOP. In any event, if samples for PCB analysis do not need the acid clean up procedure, then the same injection may be used for method 8081B and 8082, assuming all calibration and QC requirements for both methods are met. Extracts that have been acid cleaned may not be analyzed for pesticides, since several of the pesticides will be degraded.

- 1.1. The associated LIMS method codes are QJ (8081A) and DM (608).

2. SUMMARY OF METHOD

This method presents conditions for the analysis of prepared extracts of organochlorine pesticides. The pesticides are injected onto the column and separated and detected by electron capture detection. Quantitation may be by internal or external standard methods.

3. DEFINITIONS

Refer to the STL North Canton Laboratory Quality Manual (LQM), current version, for definitions of terms used in this document.

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for information regarding chromatographic interferences.
- 4.2. Interferences in the GC analysis arise from many compounds amenable to gas chromatography that give a measurable response on the electron capture detector. Phthalate esters, which are common plasticizers, can pose a major problem in the determinations. Avoiding contact with any plastic materials minimizes interferences from phthalates.
- 4.3. Sulfur will interfere and can be removed using procedures described in SOP CORP-OP-0001NC.
- 4.4. Interferences co-extracted from samples will vary considerably from source to source. The presence of interferences may raise quantitation limits for individual samples. Specific cleanups may be performed on the sample extracts, including florisil cleanup (Method 3620), Gel Permeation Chromatography (Method 3640), and Sulfur cleanup (Method 3660). These cleanup procedures are included in SOP # CORP-OP-0001NC. Using hexane / acetone as the extraction solvent (rather than hexane / methylene chloride) will reduce the amount of interferences extracted.

5. SAFETY

- 5.1. Refer to section 5 of the Method 8000B SOP for general safety requirements.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

- 5.2. Aroclors have been classified as a potential carcinogen under OSHA. Concentrated solutions of Aroclors must be handled with extreme care to avoid excess exposure. Contaminated gloves and clothing must be removed immediately. Contaminated skin surfaces must be washed thoroughly.
- 5.3. The following parameters covered by this method have been tentatively classified as known or suspected, human or mammalian carcinogens: 4,4'-DDT, 4,4'-DDD, and the BHCs. Primary standards of these toxic compounds should be prepared in a hood.
- 5.4. All ⁶³Ni sources shall be leak tested every six months, or in accordance with the manufacturer's general radioactive material license.
- 5.5. All ⁶³Ni sources shall be inventoried every six months. If a detector is missing, the Director, EH&S shall be immediately notified and a letter sent to the NRC or local state agency.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP. A ⁶³Ni electron capture detector is required.
- 6.2. Refer to Table B-2 for analytical columns.
- 6.3. Microsyringes, various sizes, for standards preparation, sample injection, and extract dilution.

7. REAGENTS AND STANDARDS

- 7.1. Refer to the method 8000B section of this SOP for general requirements for reagents and supplies.
- 7.2. Refer to Table B-3 for details of calibration standards.
- 7.3. Surrogate Standards
Tetrachloro-m-xylene and decachlorobiphenyl are the surrogate standards. Refer to tables B-5 and B-6 for details of surrogate standards.
- 7.4. Column Degradation Evaluation Mix
A mid-level standard containing 4,4'-DDT and Endrin and not containing any of their breakdown products must be prepared for evaluation of degradation of these compounds by the GC column and injection port. This mix must be replaced after one year, or whenever corrective action to columns fails to eliminate the breakdown of the compounds, whichever is shorter. This solution also contains the surrogates. Refer to Table B-4 for details of the column degradation evaluation mix.

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE

Refer to Section 8 of the 8000B section of this SOP.

9. QUALITY CONTROL

Refer to Section 9 of the 8000B section of this SOP.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

- 10.2. Refer to Table B-2 for details of GC operating conditions. The conditions listed should result in resolution of all analytes listed in Table B-1 on both columns. Closely eluting pairs are DDE and Dieldrin on the Rtx-5 or DB-5 column and Endosulfan II and DDD on the 1701 column.

Note: Method 608 requires a minimum of three concentration levels for initial calibrations. The relative standard deviation, RSD must be less than 10%.

10.3. Column Degradation Evaluation

Before any calibration runs, either initial or 12 hour, The column evaluation mix must be injected before each initial or daily calibration. The degradation of DDT and endrin must be calculated (see equations 9 and 10) and each shown to be less than 15% before calibration can proceed. This is only necessary if the target compound list includes DDT, Endrin, or any of their degradation products.

If the breakdown of DDT and/or endrin exceeds the limits given above, corrective action must be taken. This action may include:

- Replacement of the injection port liner or the glass wool.
- Cutting off a portion of the injection end of a capillary column.
- Replacing the GC column.

10.4. Initial Calibration

Refer to Section 10 of the 8000B section of this SOP for details of calibration procedures.

- 10.4.1. Refer to Table B-7 for the initial calibration analytical sequence.

- 10.4.2. The response for each single-peak analyte will be calculated by the procedures described in the general method for GC analysis.

- 10.4.3. The surrogate calibration curve is calculated from the Individual AB mix. Surrogates in the other calibration standards are used only as retention time markers. If there are resolution problems, then the A and B mixes may be analyzed separately.

- 10.4.4. For multi-component pesticides:

Single point calibration is used for multi-component pesticides (typically toxaphene and technical chlordane). Two options are possible; the same quantitation option must be used for standards and samples. Refer to section 12.3 for guidance on which option to use.

- 10.4.5. For multi-component analytes, the mid level standard must be analyzed as part of the initial calibration. This single point calibration is used to quantitate multi-component analytes.

- 10.4.6. The analyst may include a full 5 point calibration for any of the multi-component analytes with the initial calibration.

10.5. 12 hour Calibration Verification

The 12 hour calibration verification sequence must be analyzed within 12 hours of the start of the initial calibration and at least once every 12 hours thereafter if samples are being analyzed. If more than 12 hours have elapsed since the injection of the last sample in the analytical sequence, a new analytical sequence must be started with a 12 hour calibration. A mid level calibration standard is used for the 12 hour calibration. Refer to the 8000B section of this SOP for acceptance criteria.

Note : Method 608 requires that the working calibration curve must be verified on each working day by the measurement of one or more calibration standards. If the response for any parameter varies from the predicted response by more than $\pm 15\%$, the test must be repeated using a fresh calibration standard.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**Revision No: 5.6
Revision Date: 05/25/01
Page B4 of B12

10.5.1. At a minimum, the 12 hour calibration includes analysis of the breakdown mix followed by mid level standards of any single and multi-component analytes.

10.5.2. The retention time windows for any analytes included in the 12 hour calibration are updated.

10.6. Continuing Calibration

The AB calibration mix is analyzed as the continuing calibration standard. At a minimum, this is analyzed after every 20 samples, including matrix spikes, LCS, and method blanks. If 12 hours elapse analyze the 12 hour standard sequence instead. The continuing calibration standard need not include multi-component analytes. If instrument drift is expected due to sample matrix or other factors, it may be advisable to analyze the continuing calibration standard more frequently.

10.6.1. A mid level calibration standard is used for the continuing calibration.

11. PROCEDURE

11.1. Refer to the method 8000B section of this SOP for general procedural requirements.

11.2. Extraction

The extraction procedure is described in SOP No. CORP-OP-0001NC.

11.3. Cleanup

Cleanup procedures are described in SOP No. CORP-OP-0001NC.

11.4. Suggested gas chromatographic conditions are given in Table B-2.

11.5. Allow extracts to warm to ambient temperature before injection.

11.6. The suggested analytical sequence is given in Table B-7.

12. DATA ANALYSIS AND CALCULATIONS

12.1. Refer to the 8000B section of this SOP for identification and quantitation of single component analytes.

12.2. Identification of Multi-component Analytes

Retention time windows are also used for identification of multi-component analytes, but the "fingerprint" produced by major peaks of those compounds in the standard is used in tandem with the retention times to identify the compounds. The ratios of the areas of the major peaks are also taken into consideration. Identification of these compounds may be made even if the retention times of the peaks in the sample fall outside of the retention time windows of the standard, if in the analyst's judgment the fingerprint (retention time and peak ratios) resembles the standard chromatogram.

12.3. Quantitation of Multi-component Analytes

Use 3-10 major peaks or total area for quantitation as described in section 10.4.4, initial calibration of multi-component analytes.

12.3.1. If there are no interfering peaks within the envelope of the multi-component analyte, the total area of the standards and samples may be used for quantitation. Any surrogate or extraneous peaks within the envelope must be subtracted from the total area.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

12.3.1.1. Multiple peak option

This option is particularly valuable if toxaphene is identified but interferences make quantitation based on total area difficult. Select 3-10 major peaks in the analyte pattern. Calculate the response using the total area or total height of these peaks. Alternatively, find the response of each of the 3-10 peaks per multi-peak pesticide, and use these responses independently, averaging the resultant concentrations found in samples for a final concentration result. When using this option, it is appropriate to remove peaks that appear to be coeluting with contaminant peaks from the quantitation. (i.e. peaks which are significantly larger than would be expected from the rest of the pattern.)

Chlordane may be quantitated either using the multiple peak option (12.3.1.1) total area option (12.3.1.2.) or by quantitation of the major components, α -chlordane, γ -chlordane and heptachlor.

12.3.1.2. Total area option

The total area of the standards and samples may be used for quantitation of multi-component analytes. Any surrogate or extraneous peaks within the envelope must be subtracted from the total area. This option should not be used if there are significant interference peaks within the multi-component pattern in the samples. The retention time window for total area measurement must contain at least 90% of the area of the analyte.

- 12.4. Second column confirmation multi-component analytes will only be performed when requested by the client, because the appearance of the multiple peaks in the sample usually serves as a confirmation of analyte presence.
- 12.5. Surrogate recovery results are calculated and reported for decachlorobiphenyl (DCB) and tetrachloro-m-xylene (TCMX). Corrective action is only necessary if DCB and TCMX are both outside of acceptance limits.
- 12.6. Calculation of Column Degradation/% Breakdown (%B)

Equation 9

$$DDT \%B = \frac{A_{DDD} + A_{DDE}}{A_{DDD} + A_{DDE} + A_{DDT}} \times 100$$

where:

A_{DDD} , A_{DDE} , and A_{DDT} = the response of the peaks for 4,4'-DDD, 4,4'-DDE, and 4,4'-DDT in the column degradation evaluation mix.

Equation 10

$$Endrin \%B = \frac{A_{EK} + A_{EA}}{A_{EK} + A_{EA} + A_E} \times 100$$

where:

A_{EK} , A_{EA} , and A_E = the response of endrin ketone, endrin aldehyde, and endrin in the column degradation evaluation mix.

13. METHOD PERFORMANCE

- 13.1. Performance limits for the four replicate initial demonstration of capability required under Section 13.1 of the main body of this SOP. Example performance limits are listed in Table B-8. The spiking level should be equivalent to a mid level calibration.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

14. POLLUTION PREVENTION

Refer to section 14 of the 8000B section of this SOP.

15. WASTE MANAGEMENT

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

- 16.1. SW846, Update III, December 1996, Method 8081A
- 16.2. CFR136A, Method 608 – Organochlorine Pesticides and PCBs

17. MISCELLANEOUS

- 17.1. Modifications from Reference Method
None
- 17.2. Modifications from Previous Revisions
 - 17.2.1. No revisions were made to this appendix.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

Revision No: 5.6

Revision Date: 05/25/01

Page B7 of B12

17.3. Tables

Table B-1 Standard Analyte list and Reporting Limits			
Compound	Reporting Limit, µg/L or µg/kg		
	water	soil	waste
Aldrin	0.05	1.7	50
α-BHC	0.05	1.7	50
β-BHC	0.05	1.7	50
δ-BHC	0.05	1.7	50
γ-BHC (Lindane)	0.05	1.7	50
α-Chlordane	0.05	1.7	50
γ-Chlordane	0.05	1.7	50
Chlordane (technical)	0.5	17	500
4,4'-DDD	0.05	1.7	50
4,4'-DDE	0.05	1.7	50
4,4'-DDT	0.05	1.7	50
Dieldrin	0.05	1.7	50
Endosulfan I	0.05	1.7	50
Endosulfan II	0.05	1.7	50
Endosulfan Sulfate	0.05	1.7	50
Endrin	0.05	1.7	50
Endrin Aldehyde	0.05	1.7	50
Heptachlor	0.05	1.7	50
Heptachlor Epoxide	0.05	1.7	50
Methoxychlor	0.1	3.3	100
Toxaphene	2.0	67	2000
APPENDIX IX ADD ONS			
Diallate	1.0	33	1000
Isodrin	0.1	3.3	100
Chlorobenzilate	0.1	3.3	100
<i>Kepone</i> ¹	1.0	33	1000

¹ Kepone is sometimes requested for analysis by method 8081A. However kepone may produce peaks with broad tails that elute later than the standard by up to a minute (presumably due to hemi-acetal formation). As a result kepone analysis by 8081A is unreliable and not recommended. Analysis by method 8270C is a possible alternative. Note: alpha chlordane, gamma chlordane, and endrin ketone are not required for some projects.

The following concentration factors are assumed in calculating the Reporting Limits:

	<u>Extraction Vol.</u>	<u>Final Vol.</u>
Ground water	1000 mL	10 mL
Low-level Soil	30 g	10 mL
High-level soil / waste	1 g	10 mL

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**Revision No: 5.6Revision Date: 05/25/01

Page B8 of B12

Table B-2	
Parameter	Recommended Conditions
Injection port temp	220°C
Detector temp	325°C
Temperature program	120°C for 1 min, 8.5°C/min to 285°C, , 6 min hold
Column 1	Rtx-CLPesticides 30m x 0.32mm id, 0.5µm
Column 2	Rtx-35 30m x 0.32 mm id, 0.5µm
Column 3	DB-608, 30m X 0.32 mm, 0.25µm
Injection	2µL
Carrier gas	Helium or Hydrogen
Make up gas	Nitrogen
Y splitter	Restek or J&W or Supelco glass tee

ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608

Revision No: 5.6

Revision Date: 05/25/01

Page B9 of B12

Table B-3						
Calibration Levels ng/mL						
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6 ²
Individual Mix AB¹						
Aldrin	5	10	25	50	100	200
g-BHC (Lindane)	5	10	25	50	100	200
Heptachlor	5	10	25	50	100	200
Methoxychlor	10	20	50	100	200	400
Dieldrin	5	10	25	50	100	200
Endosulfan I	5	10	25	50	100	200
Endosulfan II	5	10	25	50	100	200
4,4'-DDT	5	10	25	50	100	200
Endrin Aldehyde	5	10	25	50	100	200
Endrin Ketone	5	10	25	50	100	200
β-BHC	5	10	25	50	100	200
δ-BHC	5	10	25	50	100	200
α-BHC	5	10	25	50	100	200
4,4'-DDD	5	10	25	50	100	200
4,4'-DDE	5	10	25	50	100	200
Endosulfan Sulfate	5	10	25	50	100	200
Endrin	5	10	25	50	100	200
α-Chlordane ³	5	10	25	50	100	200
γ-Chlordane ³	5	10	25	50	100	200
Multi-component Standards						
Chlordane (Technical)			250 ⁴			
Toxaphene			1000 ⁵			
Surrogates are included with all the calibration mixes at the following levels:						
Tetrachloro-m-xylene	5	10	25	50	100	200
Decachlorobiphenyl	5	10	25	50	100	200

¹ Standards may be split into an A and B mix if resolution of all compounds on both columns is not obtained.

² Level 6 is optional and should only be used if linearity can be maintained on the instrument to this level.

³ Compounds may be used in lieu of running a daily technical Chlordane standard for samples that are non-detect for technical Chlordane.

⁴ This standard may be used for quantitation of technical chlordane between 50 and 1000 ng/mL. If the chlordane is more concentrated, the extract must be diluted and reanalyzed.

⁵ This standard may be used for quantitation of toxaphene between 200 and 4000 ng/mL. If the toxaphene is more concentrated, the extract must be diluted and reanalyzed.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

Revision No: 5.6
Revision Date: 05/25/01
Page B10 of B12

Table B-4	
Column Degradation Evaluation Mix ng/mL	
Component	Concentration
4,4'-DDT	25
Endrin	25
Tetrachloro-m-xylene (Surrogate)	20
Decachlorobiphenyl (Surrogate)	20

Table B-5			
LCS/Matrix Spike and Surrogate Spike levels µg/L or µg/kg			
	Aqueous	Soil	Waste
gamma BHC (Lindane)	0.20	33.3	200
Aldrin	0.20	33.3	200
Heptachlor	0.20	33.3	200
Dieldrin	0.50	33.3	500
Endrin	0.50	33.3	500
4,4' DDT	0.50	33.3	500
Tetrachloro-m-xylene (Surrogate)	0.20	33.3	200
Decachlorobiphenyl (Surrogate)	0.20	33.3	200

Table B-6		
LCS/Matrix Spike and Surrogate Spike levels for TCLP µg/L or µg/kg		
	Aqueous	Waste
Heptachlor	5	500
Heptachlor epoxide	5	500
Lindane	5	500
Endrin	5	500
Methoxychlor	10	1000

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

Revision No: 5.6Revision Date: 05/25/01

Page B11 of B12

**Table B-7
Suggested Analytical Sequence**

Initial Calibration

Solvent blank (optional)	
Breakdown Mix	
Individual mix AB	All levels
Technical Chlordane	Level 3 ¹
Toxaphene	Level 3 ¹
Up to 20 samples unless 12 hours comes first)	
Solvent blank (optional)	
Individual mix AB	Mid level (Continuing calibration)
Samples	
After 12 hours:	
Breakdown mix	
Individual mix AB	
Any other single component analytes	

Any multi-component analytes

¹ A five point curve for any of the multi-component analytes may be included. If Arochlors are included, a 5 point calibration for Arochlor 1016/1260 should be included with the initial calibration and a single point for the other Arochlors. The mid point 1016/1260 mix is included with the daily calibration (every 12 hours).

Note: A solvent blank or primer may be analyzed at any time during the sequence when highly contaminated samples are expected. A solvent blank or primer may not be analyzed as routine immediately prior to standards.

12 hour Calibration

At least every 12 hours, counting from the start of the initial calibration, or from the start of the last daily calibration, the retention time windows must be updated using the Individual mix AB, and the breakdown mix must be run before the continuing calibration.

**ANALYSIS OF ORGANOCHLORINE PESTICIDES
BASED ON METHOD 8081A and Method 608**

Revision No: 5.6Revision Date: 05/25/01

Page B12 of B12

Table B-8		
Example Performance limits, four replicate initial demonstration of capability		
Compound	Initial demonstration, mean recovery limits	Initial demonstration, RSD limits
Aldrin	46-112	21
alpha-BHC	51-122	24
beta-BHC	61-120	32
delta-BHC	49.5-118.5	36
gamma-BHC	57-116	23
Chlordane	44.8-108.6	20
4,4'-DDD	52-126	28
4,4'-DDE	46-120	27.5
4,4'-DDT	54-137	36
Dieldrin	42.5-124.5	38
Endosulfan I	43-141	24.5
Endosulfan II	78-171	61
Endosulfan Sulfate	62-132	27
Endrin	49-126	37
Heptachlor	57-100	20
Heptachlor Epoxide	43.5-131.5	25.4
Toxaphene	44.4-111.2	20

1. SCOPE AND APPLICATION

- 1.1. This SOP Appendix describes procedures to be used when SW-846 Method 8000B is applied to the analysis of polychlorinated biphenyls (PCB) by GC/ECD. This Appendix is to be applied when SW-846 Method 8082 is requested, and is applicable to extracts derived from any matrix which are prepared according to the appropriate STL sample extraction SOPs. (CORP-OP-0001NC). The PCBs are determined and quantitated as Arochlor mixes. Criteria for the analysis of wastewater by Method 608 is also included in this appendix.

Table C-1 lists compounds, which are routinely determined by this method, and gives the Reporting Limits (RL) for each matrix. RLs given are based on the low level standard and the sample preparation concentration factors. Matrix interferences may result in higher RLs than those listed.

Note: SW-846 method 8082 provides incomplete guidance for determination of individual PCB congeners. This SOP does not include directions for congener specific analysis.

- 1.2. The associated LIMS method codes are QH (8082) and DM (608).

2. SUMMARY OF METHOD

This method presents conditions for the analysis of prepared extracts of PCBs. The PCBs are injected onto the column and separated and detected by electron capture detection. Quantitation is by the external standard method.

3. DEFINITIONS

Refer to the STL North Canton Laboratory Quality Manual (LQM), current version, for definitions of terms used in this document.

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for information regarding chromatographic interferences.
- 4.2. Interferences in the GC analysis arise from many compounds amenable to gas chromatography that give a measurable response on the electron capture detector. Phthalate esters, which are common plasticizers, can pose a major problem in the determinations. Avoiding contact with any plastic materials minimizes interferences from phthalates.
- 4.3. Sulfur will interfere and can be removed using procedures described in SOP CORP-OP-0001NC.
- 4.4. Interferences co-extracted from samples will vary considerably from source to source. The presence of interferences may raise quantitation limits for individual samples. Specific cleanups may be performed on the sample extracts, including florisil cleanup (Method 3620), Gel Permeation Chromatography (Method 3640), and Sulfur cleanup (Method 3660). These cleanup procedures are included in SOP # CORP-OP-0001NC.

5. SAFETY

- 5.1. Refer to section 5 of the Method 8000B SOP for general safety requirements.
- 5.2. Aroclors have been classified as a potential carcinogen under OSHA. Concentrated solutions of Aroclors must be handled with extreme care to avoid excess exposure. Contaminated gloves and clothing must be removed immediately. Contaminated skin surfaces must be washed thoroughly.

ANALYSIS OF PCBs BASED ON METHOD 8082 and Method 608

- 5.3. All ^{63}Ni sources shall be leak tested every six months, or in accordance with the manufacturer's general radioactive material license.
- 5.4. All ^{63}Ni sources shall be inventoried every six months. If a detector is missing, the Director, EH&S shall be immediately notified and a letter sent to the NRC or local state agency.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP. A ^{63}Ni electron capture detector is required.
- 6.2. Refer to Table C-2 for analytical columns.
- 6.3. Microsyringes, various sizes, for standards preparation, sample injection, and extract dilution.

7. REAGENTS AND STANDARDS

- 7.1. Refer to the method 8000B section of this SOP for general requirements for reagents and supplies. The standards must be replaced at least every six months or sooner if comparison with check standards indicates a problem.
- 7.2. Refer to Table C-3 for details of calibration standards.
- 7.3. Surrogate Standards
Tetrachloro-m-xylene and decachlorobiphenyl are the surrogate standards. Other surrogates may be used at client request. Refer to Table C-4 for details of surrogate standards.

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE

Refer to Section 8 of the 8000B section of this SOP.

9. QUALITY CONTROL

Refer to Section 9 of the 8000B section of this SOP.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.
- 10.2. Initial Calibration
 - 10.2.1. Refer to Table C-5 for the initial calibration analytical sequence.
 - 10.2.2. The response for each Arochlor will be calculated by the procedures described in the general method for GC analysis, with the following modifications.
 - 10.2.3. A five point calibration of the Arochlor 1016/1260 mix is generated with at least mid level single points for the other Arochlor mixes. The average response factor is used to quantitate Arochlors 1260 and 1016, other Arochlors are quantitated from the mid level single point.
Note: Method 608 requires a minimum of three concentration levels for initial calibrations. The relative standard deviation, RSD must be less than 10%.
 - 10.2.4. The analyst may include a full 5 point calibration for any of the Arochlors with the initial calibration.

- 10.2.5. The high and low standards for the initial 5 point calibration of 1016 / 1260 define the acceptable quantitation range for the other Arochlors. If any Arochlor is determined above this concentration the extract must be diluted and reanalyzed.
- 10.2.6. If the analyst knows that a specific Arochlor is of interest for a particular project, that Arochlor may be used for the five point calibration rather than the 1016 / 1260 mix.
- 10.2.7. The surrogate calibration curve is calculated from the Arochlor 1016/1260 mix. Surrogates in the other calibration standards are used only as retention time markers.
- 10.2.8. Two options are possible for quantitation of Arochlors. The same quantitation option must be used for standards and samples.

- 10.2.8.1. Multiple peak option

Select 3-10 major peaks in the analyte pattern. Calculate the response using the total area or total height of these peaks. Alternatively, find the response of each of the 3-10 peaks per Arochlor, and use these responses independently, averaging the resultant concentrations found in samples for a final concentration result. When using this option, it is appropriate to remove peaks that appear to be coeluting with contaminant peaks from the quantitation. (i.e. peaks which are significantly larger than would be expected from the rest of the pattern.)

- 10.2.8.2. Total area option

The total area of the standards and samples may be used for quantitation of multi-component analytes. Any surrogate or extraneous peaks within the envelope must be subtracted from the total area. This option should not be used if there are significant interference peaks within the multi-component pattern in the samples. The retention time window for total area measurement must contain at least 90% of the area of the analyte.

- 10.3. 12 hour Calibration

The 12 hour calibration verification must be analyzed within 12 hours of the start of the initial calibration and at least once every 12 hours thereafter if samples are being analyzed. If there is a break in the analytical sequence of greater than 12 hours, then a new continuing calibration run must be analyzed before proceeding with the sequence. If more than 12 hours have elapsed since the injection of the last sample in the analytical sequence, a new analytical sequence must be started with a 12 hour calibration.

- 10.3.1. At a minimum, the 12 hour calibration includes analysis of the Arochlor 1260 / 1016 mix.
 - 10.3.2. It is adequate to verify calibration with a mixture of Arochlors 1016 and 1260. If a specific Arochlor is expected, it may be included in the daily calibration check.
 - 10.3.3. The retention time windows for any analytes included in the daily calibration and CCVs are updated.
 - 10.3.4. For this method samples must be bracketed with successful calibration verification runs.

- 10.4. Calibration verification

The Arochlor 1260/1016 calibration mix is analyzed as the calibration verification standard. This is analyzed after every 20 samples, including matrix spikes, LCS, and method blanks. (Depending on the type of samples, it may be advisable to analyze verifications more frequently in order to minimize reruns.).

- 10.4.1. A mid level standard is used for the calibration verification.

Note : Method 608 requires that the working calibration curve must be verified on each working day by the measurement of one or more calibration standards. If the response for any parameter varies from the predicted response by more than $\pm 15\%$, the test must be repeated using a fresh calibration standard.

11. PROCEDURE

- 11.1. Refer to the method 8000B section of this SOP for general procedural requirements.
- 11.2. Extraction
The extraction procedure is described in SOP No. CORP-OP-0001NC.
- 11.3. Cleanup
Cleanup procedures are described in SOP No. CORP-OP-0001NC.
- 11.4. Suggested gas chromatographic conditions are given in Table C-2.
- 11.5. Allow extracts to warm to ambient temperature before injection.
- 11.6. The suggested analytical sequence is given in Table C-5.

12. DATA ANALYSIS AND CALCULATIONS

- 12.1. Identification of Arochlors
Retention time windows are used for identification of Arochlors, but the "fingerprint" produced by major peaks of those analytes in the standard is used in tandem with the retention times for identification. The ratios of the areas of the major peaks are also taken into consideration. Identification may be made even if the retention times of the peaks in the sample fall outside of the retention time windows of the standard, if in the analyst's judgment the fingerprint (retention time and peak ratios) resembles the standard chromatogram.
A clearly identifiable Arochlor pattern serves as confirmation of single column GC analysis. Dual column confirmation may be used for specific program requirements or by client request.
- 12.2. Quantitation of Arochlors
Use 3-10 major peaks or total area for quantitation
If the analyst believes that a combination of Aroclor 1254 and 1260, or a combination of 1242, 1248 and 1232 is present, then only the predominant Arochlor is quantitated and reported, but the suspicion of multiple Arochlors is discussed in the narrative. If well separated Arochlor patterns are present, and then both Arochlors are quantitated and reported.
- 12.3. If there are no interfering peaks within the envelope of the Arochlor, the total area of the standards and samples may be used for quantitation. Any surrogate or extraneous peaks within the envelope must be subtracted from the total area.
- 12.4. Second column confirmation of Arochlors will only be performed when requested by the client. The appearance of the multiple peaks in the sample usually serves as a confirmation of Arochlor presence.
- 12.5. Surrogate recovery results are calculated and reported for decachlorobiphenyl (DCB) and tetrachloro-m-xylene (TCMX). Corrective action is only necessary if DCB and TCMX are both outside of acceptance limits.

ANALYSIS OF PCBs BASED ON METHOD 8082 and Method 608

13. METHOD PERFORMANCE

- 13.1. Performance limits for the four replicate initial demonstration of capability are required as referenced under Section 13.1 of the main body of this SOP.
- 13.2. Method detection limits (MDL) are determined for all Arochlors.

14. POLLUTION PREVENTION

Refer to section 14 of the 8000B section of this SOP.

15. WASTE MANAGEMENT

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

- 16.1. SW846, Update III, December 1996, Method 8082
- 16.2. CFR136A, Method 608, Organochlorine Pesticides and PCBs.

17. MISCELLANEOUS

- 17.1. Modifications from Reference Method
 - 17.1.1. Method 8082 includes limited direction for congener specific quantitation. This is outside the scope of this SOP.
- 17.2. Modifications from Previous Revisions
 - No changes were made to this Appendix

17.3. Tables

Table C-1 Standard Analyte list and Reporting Limits			
Compound	Reporting Limit, $\mu\text{g/L}$ or $\mu\text{g/kg}$		
	water	soil	waste
Aroclor-1016	1.0	33	1000
Aroclor-1221	1.0	33	1000
Aroclor-1232	1.0	33	1000
Aroclor 1242	1.0	33	1000
Aroclor-1248	1.0	33	1000
Aroclor-1254	1.0	33	1000
Aroclor-1260	1.0	33	1000

The following concentration factors are assumed in calculating the Reporting Limits:

	<u>Extraction Vol.</u>	<u>Final Vol.</u>
Ground water	1000 mL	10 mL
Low-level Soil	30 g	10 mL
High-level soil / waste	1 g	10 mL

Table C-2	
Parameter	Recommended Conditions
Injection port temp	220°C
Detector temp	325°C
Temperature program	70°C for 0.5min, 30°C/min to 190°C, 2.5°C/min to 225, 18°C/min to 280°C, 3 min hold
Column 1	DB-5 or Rtx-5 30m x 0.32mm id, 0.5 μm
Column 2	DB-1701 or Rtx 1701 30m x 0.32 mm id, 0.25 μm
Column 3	DB-608, 30m X 0.32 mm, 0.25 μm
Injection	1-2 μL
Carrier gas	Helium or Hydrogen
Make up gas	Nitrogen
Y splitter	Restek or J&W or Supelco glass tee

Table C-3						
Calibration Levels ng/mL						
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6 ¹
Aroclor 1016/1260	100	200	500	1000	2000	4000
Aroclor 1242 ²			500			
Aroclor 1221 +1254 ²			500			
Aroclor 1232 ²			500			
Aroclor 1248 ²			500			
Surrogates are included with all the calibration mixes at the following levels:						
Tetrachloro-m-xylene	5	10	25	50	100	200
Decachlorobiphenyl	5	10	25	50	100	200

¹ Level 6 is optional and should only be used if linearity can be maintained on the instrument to this level.

² Aroclors may be quantitated within the range 100 to 2000 ng/mL (4000ng/mL if the level 6 1016/1260 standard is included). If the Aroclor is more concentrated, it must be reanalyzed at a dilution.

Table C-4			
LCS/Matrix Spike and Surrogate Spike levels for Aroclor analysis with Acid Cleanup			
µg/L or µg/kg			
	Aqueous	Soil	Waste
Aroclor 1016/1260	10	333	10,000
Tetrachloro-m-xylene (Surrogate)	0.20	6.67	200
Decachlorobiphenyl (Surrogate)	0.20	6.67	200

Table C-5			
Michigan Analyte List and Reporting Limits ¹			
Compound	Reporting Limit		
	water (µg/L)	soil (µg/Kg)	
Aroclor-1016	0.2	330	
Aroclor-1221	0.2	330	
Aroclor-1232	0.4	330	
Aroclor 1242	0.2	330	
Aroclor-1248	0.2	330	
Aroclor-1254	0.2	330	
Aroclor-1260	0.2	330	

¹ Reporting Limits are only for samples performed under the Michigan program

Table C-5
Suggested Analytical Sequence

Initial Calibration

Injection

1	Solvent blank (optional)	
2	Aroclor 1016/1260	Level 1
3	Aroclor 1016/1260	Level 2
4	Aroclor 1016/1260	Level 3
5	Aroclor 1016/1260	Level 4
6	Aroclor 1016/1260	Level 5
7	Aroclor 1232	Level 3
8	Aroclor 1242	Level 3
9	Aroclor 1248	Level 3
10	Aroclor 1221/1254	Level 3
11-30	Sample 1-20 (or as many samples as can be analyzed in 12 hours)	
	Solvent blank (optional)	
32	Aroclor 1016/1260	Level 3

etc

Note: A solvent blank or primer may be analyzed at any time during the sequence when highly contaminated samples are expected. A solvent blank or primer may not be analyzed as routine immediately prior to standards.

12 hour Calibration

At least every 12 hours, counting from the start of the initial calibration, or from the start of the last daily calibration, the retention time windows must be updated using the Aroclor 1260 / 1016 mix. Mid level standards of any other Aroclors expected to be present in the samples are also injected.

**ANALYSIS OF PHENOXY ACID HERBICIDES BASED ON
SW-846 METHOD 8151A****1. SCOPE AND APPLICATION**

This method is applicable to the gas chromatographic determination of Chlorinated phenoxy acid herbicides in extracts prepared by SOP CORP-OP-0001NC. The herbicides listed in Table D1 are routinely analyzed.

Other chlorinated acids may be analyzed by this method if the quality control criteria in Section 9 and the initial demonstration of method performance in Section 13 are met.

- 1.1. The associated LIMS method code is QS.

2. SUMMARY OF METHOD

This method presents conditions for the analysis of prepared extracts of phenoxy acid herbicides by gas chromatography. The herbicides, as their methyl esters, are injected onto the column, separated, and detected by electron capture detectors. Quantitation is by the external standard method.

3. DEFINITIONS

Refer to the STL North Canton Laboratory Quality Manual (LQM), current version, for definitions of terms used in this document.

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for general information regarding chromatographic interferences.
- 4.2. Chlorinated acids and phenols cause the most direct interference with this method.
- 4.3. Sulfur may interfere and may be removed by the procedure described in SOP#CORP-OP-0001NC.

5. SAFETY

- 5.1. Refer to section 5 of the Method 8000B SOP for general safety requirements.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP. A Ni₆₃ electron capture detector is required.
- 6.2. Refer to Table D2 for analytical columns.
- 6.3. Microsyringes, various sizes, for standards preparation, sample injection, and extract dilution.

7. REAGENTS AND STANDARDS

- 7.1. Refer to section 7 of the 8000B section of this SOP for general information on reagents and standards.

8. SAMPLE PREPARATION, PRESERVATION AND STORAGE

Refer to Section 8 of the 8000B section of this SOP.

9. QUALITY CONTROL

- 9.1. Refer to Section 9 of the 8000B section of this SOP for quality control requirements, including the initial demonstration of capability, definition of a batch, surrogate limits, method blanks, laboratory control samples (LCS), and matrix spikes (MS).

- 9.2. Refer to Table D-3 for the components and levels of the LCS and MS mixes.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.
- 10.2. Calibration standards are prepared from purchased standards in the methyl ester form.
- 10.3. The low level standard must be at or below the laboratory reporting limit. Other standards are chosen to bracket the expected range of concentrations found in samples, without saturating the detector or leading to excessive carryover.
- 10.4. Refer to Table D-2, for details of GC operating conditions.

11. PROCEDURE

- 11.1. Refer to the method 8000B section of this SOP for procedural requirements.
- 11.2. Extraction
The extraction procedure is described in SOP #CORP-OP-0001NC.
- 11.3. Cleanup
The alkaline hydrolysis and subsequent extraction of the basic solution described in the extraction procedure provides an effective cleanup.
- 11.4. Analytical Sequence
The analytical sequence starts with an initial calibration of at least five points, or a daily calibration that meets % difference criteria from an existing initial calibration.
- 11.4.1. The daily calibration must be analyzed at least once every 24 hours when samples are being analyzed. If there is a break in the analytical sequence of greater than 12 hours, then a new continuing calibration run must be analyzed before proceeding with the sequence. If more than 24 hours have elapsed since the injection of the last sample in the analytical sequence, a new analytical sequence must be started with a daily calibration.
- 11.4.2. The daily calibration consists of mid level standards of all analytes of interest. Retention time windows must be updated with the daily calibration.
- 11.4.3. After every 12 hours a continuing calibration is analyzed. The continuing calibration consists of mid level standards of all analytes of interest. Retention time windows are updated with continuing calibrations.
- 11.5. Gas Chromatography
Chromatographic conditions are listed in Table D-2.

12. DATA ANALYSIS AND CALCULATIONS

- 12.1. Refer to the 8000B section of this SOP for identification and quantitation of single component analytes.
- 12.2. The herbicides are analyzed as their methyl esters, but reported as the free acid. For this reason it is necessary to correct the results for the molecular weight of the ester versus the free acid. This is achieved through the concentrations of the calibration standards. For example the 20µg/L calibration standard for 2,4-D contains 21.3 µg/L of the methyl ester. No further correction is necessary.

**ANALYSIS OF PHENOXY ACID HERBICIDES BASED ON
SW-846 METHOD 8151A**

- 12.3. A routine 10X dilution occurs on final extracts for all samples. Due to a QuantIMS limitation, the dilution factor field in QuantIMS cannot be used when a dilution is routine, because the dilution factor is automatically applied to all reference values creating reporting problems. For the herbicide analysis, the extract volume will be 10mL and an aliquot at 10X dilution will be analyzed. The final extract volume recorded on the laboratory bench sheet will be recorded as 100mL to avoid using the dilution factor field in QuantIMS.

13. METHOD PERFORMANCE

- 13.1. The EPA for this method has not published multiple laboratory performance data. Performance limits for the four replicate initial demonstration of capability are required as referenced under Section 13.1 of the main body of this SOP.

14. POLLUTION PREVENTION

This method does not contain any specific modifications that serve to minimize or prevent pollution.

15. WASTE MANAGEMENT

Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

Method 8151A, SW-846, Update III, December 1996

17. MISCELLANEOUS

- 17.1. Modifications from Reference Method
Refer to the method 8000B section of this SOP for modifications from the reference method.
- 17.2. Modifications from Previous Revision
The calibration procedure has been changed to require esterification of the calibration standards

ANALYSIS OF PHENOXY ACID HERBICIDES BASED ON
SW-846 METHOD 8151A

Revision No: 5.6

Revision Date: 05/25/01

Page D4 of D5

17.3. Tables

Table D-1				
Standard Analyte list				
Compound	CAS Number	Reporting Limit, µg/L or µg/kg		
		Aqueous	Soil	Waste
2,4-D	94-75-7	4	80	4000
2,4-DB	94-82-6	4	80	4000
2,4,5-TP (Silvex)	93-72-1	1	20	1000
2,4,5-T	93-76-5	1	20	1000
Dalapon	75-99-0	2	40	2000
Dicamba	1918-00-9	2	40	2000
Dichloroprop	120-36-5	4	80	4000
Dinoseb	88-85-7	0.6	12	600
MCPA	94-74-6	400	8000	400,000
MCPP	93-65-2	400	8000	400,000

The following concentration factors are assumed in calculating the Reporting Limits:

	Extraction Vol.	Final Vol.	Dilution Factor
Ground water	1000 mL	10 mL	20
Low-level Soil without GPC	50 g	10 mL	20
High-level soil / waste	1 g	10 mL	20

Specific reporting limits are highly matrix dependent. The reporting limits listed above are provided for guidance only and may not always be achievable. For special projects, the extracts may be analyzed without any dilution, resulting in reporting limits 20 times lower than those in Table D-1.

Table D-2	
Instrumental Conditions	
PARAMETER	Recommended conditions
Injection port temp	220°C
Detector temp	325°C
Temperature program	80,2/30/170,0/1/180,1
Column 1	DB-5MS or RTX 5 30x0.32, 0.5µm
Column 2	DB-1701 or Rtx-1701
Injection	1-2µL
Carrier gas	Helium / Hydrogen
Make up gas	Nitrogen

Recommended conditions should result in resolution of all analytes listed in Table D-1.

The reporting limits listed in Table D-1 will be achieved with these calibration levels and a 20 fold dilution of the sample extract. Lower reporting limits can be achieved with lesser dilutions of the sample extract.

ANALYSIS OF PHENOXY ACID HERBICIDES BASED ON
SW-846 METHOD 8151A

Revision No: 5.6

Revision Date: 05/25/01

Page D5 of D5

Table D-3			
LCS/Matrix Spike and Surrogate Spike levels $\mu\text{g/L}$ or $\mu\text{g/kg}$ ¹			
	Aqueous	Soil	Waste
2,4-D	16	800	16000
Silvex	4	200	4000
2,4,5-T	4	200	4000
2,4-DB	16	800	16000
Dalapon	8	400	8000
DCAA (surrogate)	16	800	16000

¹ LCS, MS and SS spikes are as the free acid.

1. SCOPE AND APPLICATION

- 1.1. This SOP Appendix describes procedures to be used when SW-846 Method 8000B is applied to the analysis of polynuclear aromatic hydrocarbons (PAH) by High Performance Liquid Chromatography. This Appendix is to be applied when SW-846 Method 8310 or EPA Method 610 is requested, and is applicable to extracts derived from any matrix which are prepared according to the appropriate STL North Canton sample extraction SOPs. (CORP-OP-0001NC).

Table E-1 lists compounds, which are routinely determined by this method, and gives the Reporting Limits (RL) for each matrix. RLs given are based on the low level standard and the sample preparation concentration factors. Matrix interferences may result in higher RLs than those listed. Note: Reporting limits are subject to change due to compound sensitivity, calibration levels, MDLs, and other factors.

- 1.2. The associated LIMS method codes are SG (8310) and VT (610).

2. SUMMARY OF METHOD

This method presents conditions for the analysis of prepared extracts of polynuclear aromatic hydrocarbons by high performance liquid chromatography. The compounds are injected onto a column, separated, and detected by ultraviolet and fluorescence detectors. Quantitation is by the external standard method.

3. DEFINITIONS

Refer to the STL North Canton Laboratory Quality Manual (LQM), current version, for definitions of terms used in this document.

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for general information regarding chromatographic interferences.

5. SAFETY

- 5.1. Refer to section 5 of the Method 8000B SOP for general safety requirements.
- 5.2. The following parameters covered by this method have been tentatively classified as known or suspected, human or mammalian carcinogens: Benzo(a)anthracene, Benzo(a)pyrene, and Dibenzo(a,h)anthracene. Primary standards of these toxic compounds should be prepared in a hood.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP. An Ultraviolet detector and Fluorescence detector is required.
- 6.2. Refer to Table E2 for analytical columns.
- 6.3. Microsyringes, various sizes, for standards preparation, sample injection, and extract dilution.

7. REAGENTS AND STANDARDS

- 7.1. Refer to section 7 of the 8000B section of this SOP for general information on reagents and standards.
- 7.2. Refer to Table E-3 and E-4 for details of calibration and other standards.

7.3. Surrogate Standards

Terphenyl-d14 and Benzo(e)pyrene are the surrogate standards. Refer to table E-3 for details of surrogate standards.

8. SAMPLE PREPARATION, PRESERVATION AND STORAGE

Refer to section 8 of the 8000B section of this SOP.

9. QUALITY CONTROL

9.1. Refer to Section 9 of the 8000B section of this SOP for quality control requirements, including the initial demonstration of capability, definition of a batch, surrogate limits, method blanks, laboratory control samples (LCS), and matrix spikes (MS).

9.2. Refer to Table E-3 for the components and levels of the LCS and MS mixes.

10. CALIBRATION AND STANDARDIZATION

10.1. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.

Note: The Initial Calibration requirement for Method 610 is as follows: If the ratio of response to amount injected (calibration factor) is a constant over the working range (<10% relative standard deviation, RSD), linearity through the origin can be assumed and the average ratio or calibration factor can be used in place of a calibration curve.

10.2. Calibration standards are prepared from purchased solutions. Table E-4 lists the calibration levels.

10.3. The low level standard must be at or below the laboratory reporting limit. Other standards are chosen to bracket the expected range of concentrations found in samples, without saturating the detector or leading to excessive carryover.

10.4. Refer to Table E-2, for details of LC operating conditions.

11. PROCEDURE

11.1. Refer to the method 8000B section of this SOP for procedural requirements.

11.2. Extraction

The extraction procedure is described in SOP #CORP-OP-0001NC.

11.3. Cleanup

11.4. Analytical Sequence

The analytical sequence starts with an initial calibration of at least five points, or a daily calibration that meets % difference criteria from an existing initial calibration.

Note: Method 610 requires that continuing calibration verifications may not vary from the predicted response for any parameter by more than $\pm 15\%$. The average response may not be used for Method 610.

11.4.1. The daily calibration must be analyzed at least once every 24 hours when samples are being analyzed. If there is a break in the analytical sequence of greater than 12 hours, then a new continuing calibration run must be analyzed before proceeding with the sequence. If more than 24

hours have elapsed since the injection of the last sample in the analytical sequence, a new analytical sequence must be started with a daily calibration.

11.4.2. The daily calibration consists of mid level standards of all analytes of interest. Retention time windows must be updated with the daily calibration.

11.4.3. After every 12 hours a continuing calibration is analyzed. The continuing calibration consists of mid level standards of all analytes of interest. Retention time windows are updated with continuing calibrations.

11.5. High Performance Liquid Chromatography
Chromatographic conditions are listed in Table E-2.

12. DATA ANALYSIS AND CALCULATIONS

12.1. Refer to the 8000B section of this SOP for identification and quantitation of single component analytes.

12.2. Surrogate recovery results are calculated and reported for Terphenyl-d14 and Benzo(e)pyrene unless it is determined that sample interference has adversely affected the quantitation of one of the surrogates. One surrogate must be within QC criteria. Corrective action is only necessary if Terphenyl-d14 and Benzo(e)pyrene are both outside of acceptance limits.

13. METHOD PERFORMANCE

13.1. Performance limits for the four replicate initial demonstration of capability required as referenced under Section 13.1

14. POLLUTION PREVENTION

Refer to section 14 of the 8000B section of this SOP.

15. WASTE MANAGEMENT

15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

16.1. SW846, Update III, December 1996, Method 8310.

16.2. CFR136 Appendix A Method 610-Polynuclear Aromatic Hydrocarbons.

17. MISCELLANEOUS

17.1. Reporting limits

17.1.1. The lower standard reporting limits are listed in Table

17.1.2. If samples require dilution or smaller volumes than specified in this method, the RL will be elevated.

17.2. Elution time

17.2.1. For any positive detection in a sample, the chromatogram is overlaid on screen with the nearest standard for elution time matching and pattern recognition using the Target system.

17.3. Analyte detection

17.3.1. Some compounds are detected only on the ultraviolet detector. For other compounds, the lower of the concentrations calculated from both detectors is reported, the second detector serving as confirmation of the analytes presence.

17.3.2. For any compound that is detected only on the ultraviolet detector, confirmation analysis is not possible. These detected compounds must be narrated that the quantitation is not confirmed by a second detector (Fluorescence).

17.3.3. Some compounds, due to calibration levels and detector limitations, can only be confirmed using the Fluorescence detector. The result will always be reported from the UV detector. These compounds will be narrated, when applicable. Table E-5 lists the compounds and respective detectors.

17.4. Troubleshooting guide

17.4.1. Consult the instrument manufacturer's operating manual for guidance.

17.5. Tables

Table E-1			
Standard Analyte list			
Compound	CAS Number	Reporting Limit, µg/L or µg/kg	
		Aqueous	Soil
Acenaphthene	83-32-9	1	100
Acenaphthylene	208-96-8	1	100
Anthracene	120-12-7	2	100
Benzo(a)anthracene	56-55-3	0.1	5
Benzo(a)pyrene	50-32-8	0.1	5
Benzo(b)fluoranthene	205-99-2	0.1	5
Benzo(g,h,i)perylene	191-24-2	0.1	10
Benzo(k)fluoranthene	207-08-9	0.1	1.7
Chrysene	218-01-9	0.1	5
Dibenzo(a,h)anthracene	53-70-3	0.1	5
Fluoranthene	206-44-0	0.1	10
Fluorene	86-73-7	1	100
Indeno(1,2,3-cd)pyrene	193-39-5	0.1	5
Naphthalene	91-20-3	2	100
Phenanthrene	85-01-8	1	100
Pyrene	129-00-0	0.1	5
1-Methylnaphthalene*	90-12-0	2	100
2-Methylnaphthalene*	91-57-6	2	100

Table E-2				
Instrumental Conditions				
PARAMETER	Recommended conditions			
Injection volume	20 µL			
Detector	HP Series 1100 Fluorescence Detector			
Detector	HP Series 1100 Variable Wavelength Detector (UV)			
Solvent Program	Time	% Water	% Acetonitrile	Flow
	0 min.	50	50	1.5 mL/m
	5 min.	45	55	1.5 mL/m
	10 min.	35	65	2.0 mL/m
	15 min.	25	75	2.0 mL/m
	20 min.	15	85	2.0 mL/m
	25 min.	5	95	2.0 mL/m
30 min.	45	55	2.0 mL/m	
Column 1	PAH Hypersil 5 micron			
Injection	20 µL			
Solvent	Acetonitrile, HPLC grade			
Solvent	Reagent Water, HPLC grade			

Compound	Concentration	Compound	Concentration
Acenaphthene	10 ug/mL	Dibenzo(a,h)anthracene	2 ug/mL
Acenaphthylene	10 ug/mL	Fluoranthene	2 ug/mL
Anthracene	10 ug/mL	Fluorene	2 ug/mL
Benzo(a)anthracene	2 ug/mL	Indeno(1,2,3-cd)pyrene	2 ug/mL
Benzo(a)pyrene	2 ug/mL	Naphthalene	10 ug/mL
Benzo(b)fluoranthene	2 ug/mL	Phenanthrene	2 ug/mL
Benzo(g,h,i)perylene	2 ug/mL	Pyrene	2 ug/mL
Benzo(k)fluoranthene	2 ug/mL	Terphenyl-d14*	1 ug/mL
Chrysene	2 ug/mL	Benzo(e)pyrene*	5 ug/mL

Bold print denotes control compounds in the LCS/LCSD and MS/MSD.

* Denotes surrogate compounds

Component	ST	1:2	1:5	1:10	1:20	1:50	1:100
Naphthalene	50	25	10	5	2.5	1	0.5
Acenaphthylene	50	25	10	5	2.5	1	0.5
Terphenyl d14 (surrogate)	50	25	10	5	2.5	1	0.5
Acenaphthene	10	5	2	1	0.5	0.2	0.1
Fluorene	10	5	2	1	0.5	0.2	0.1
Phenanthrene	10	5	2	1	0.5	0.2	0.1
Anthracene	10	5	2	1	0.5	0.2	0.1
Chrysene	10	5	2	1	0.5	0.2	0.1
1-Methylnaphthalene	50	25	10	5	2.5	1	0.5
Fluoranthene	10	5	2	1	0.5	0.2	0.1
Pyrene	10	5	2	1	0.5	0.2	0.1
Benzo(a)anthracene	10	5	2	1	0.5	0.2	0.1
Benzo(e)pyrene (surrogate)	50	25	10	5	2.5	1	0.5
Benzo(b)fluoranthene	10	5	2	1	0.5	0.2	0.1
Benzo(k)fluoranthene	10	5	2	1	0.5	0.2	0.1
Benzo(a)pyrene	10	5	2	1	0.5	0.2	0.1
Dibenzo(a,h)anthracene	10	5	2	1	0.5	0.2	0.1
Benzo(g,h,i)perylene	10	5	2	1	0.5	0.2	0.1
Indeno(1,2,3-cd)pyrene	10	5	2	1	0.5	0.2	0.1
2-Methylnaphthalene	50	25	10	5	2.5	1	0.5

Table E-5 Compound/Detector		
Compound	Ultraviolet 254 nm	Fluorescence 280 nm
Acenaphthene	X	X
Acenaphthylene	X	
Anthracene	X	Retention Time confirmation only
Benzo(a)anthracene	X	X
Benzo(a)pyrene	X	X
Benzo(b)fluoranthene	X	X
Benzo(g,h,i)perylene	X	X
Benzo(k)fluoranthene	X	X
Chrysene	X	X
Dibenzo(a,h)anthracene	X	X
Fluoranthene	X	X
Fluorene	X	X
Indeno(1,2,3-cd)pyrene	X	X
Naphthalene	X	X
Phenanthrene	X	X
Pyrene	X	X
1-Methylnaphthalene	X	X
2-Methylnaphthalene	X	X

1. SCOPE AND APPLICATION

- 1.1. This method is applicable to the determination of the concentration of certain Organophosphorus Pesticides in waters, wastewaters, oils, soils, and sludges. It is based on SW846 Method 8141A. Table F1 shows reporting limits for compounds routinely analyzed by this method. The compounds include, but are not limited to, those shown in Table F1.
- 1.2. This document accurately reflects current laboratory standard operating procedures (SOP) as of the date above. All facility SOPs are maintained and updated as necessary by the laboratory QA department.
- 1.3. The associated LIMS method code is P2.

2. SUMMARY OF METHOD

- 2.1. An aliquot of prepared sample is injected in a gas chromatograph (GC) and compounds in the effluent are detected by a flame photometric detector. Appropriate preparation techniques are described in SOP CORP-OP-0001NC. Ultrasonic Extraction (Method 3550) is **NOT** an appropriate sample preparation for Method 8141 and should not be used because of the potential for destruction of target analytes during the ultrasonic extraction process.

3. DEFINITIONS

Refer to the STL North Canton Laboratory Quality Manual (LQM), current version, for definitions of terms used in this document.

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for general information regarding chromatographic interferences.
- 4.2. Analytical difficulties encountered for target analysis include:
 - 4.2.1. The water solubility of Dichlorvos (DDVP) is 10 g/L at 20°C, and recovery is poor from aqueous solution.
 - 4.2.2. Naled is converted to Dichlorvos (DDVP) on column by debromination. This reaction may also occur during sample workup.
 - 4.2.3. Trichlorfon rearranges and is dehydrochlorinated in acidic, neutral, or basic media to form Dichlorvos (DDVP) and hydrochloric acid. If this method is to be used for the determination of organophosphates in the presence of Trichlorfon, the analyst should be aware of the possibility of rearrangement to Dichlorvos to prevent misidentification.
 - 4.2.4. Merphos is a single component pesticide that is readily oxidized to Merphos oxone. Chromatographic analysis of Merphos almost always results in two peaks.

5. SAFETY

- 5.1. Refer to Section 5 of the Method 8000B SOP for general safety requirements.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP.
- 6.2. Refer to Table F2 for Instrument settings.
- 6.3. Microsyringes, various sizes, for standards preparation, sample injection, and extract dilution.

7. REAGENTS AND STANDARDS

- 7.1. Refer to Section 7 of the 8000B section of this SOP for general information on reagents and standards.
- 7.2. Refer to Table F-3 for details of calibration and other standards.
- 7.3. Surrogate Standards
Triphenyl phosphate is the surrogate standard. Refer to table F-4 for details of the surrogate standard.

8. SAMPLE PREPARATION, PRESERVATION AND STORAGE

Refer to section 8 of the 8000B section of this SOP.

9. QUALITY CONTROL

- 9.1. Refer to Section 9 of the 8000B section of this SOP for quality control requirements, including the initial demonstration of capability, definition of a batch, surrogate limits, method blanks, laboratory control samples (LCS), and matrix spikes (MS).
- 9.2. Refer to Table F-4 for the components and levels of the LCS and MS mixes.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.
- 10.2. Calibration standards are made up from purchased solutions. Table F-3 lists the calibration levels.
- 10.3. The low level standard must be at or below the laboratory reporting limit. Other standards are chosen to bracket the expected range of concentrations found in samples, without saturating the detector or leading to excessive carryover.

11. PROCEDURE

- 11.1. Refer to the method 8000B section of this SOP for procedural requirements.
- 11.2. Extraction
The extraction procedure is described in SOP #CORP-OP-0001NC.
- 11.3. Cleanup
- 11.4. Analytical Sequence
The analytical sequence starts with an initial calibration of at least five points, or a daily calibration that meets % difference criteria from an existing initial calibration.

- 11.4.1. The daily calibration must be analyzed at least once every 24 hours when samples are being analyzed. If there is a break in the analytical sequence of greater than 12 hours, then a new continuing calibration run must be analyzed before proceeding with the sequence. If more than 24 hours have elapsed since the injection of the last sample in the analytical sequence, a new analytical sequence must be started with a daily calibration.
- 11.4.2. The daily calibration consists of mid level standards of all analytes of interest. Retention time windows must be updated with the daily calibration.
- 11.4.3. After every 12 hours a continuing calibration is analyzed. The continuing calibration consists of mid level standards of all analytes of interest. Retention time windows are updated with continuing calibrations.

12. DATA ANALYSIS AND CALCULATIONS

- 12.1. Refer to the 8000B section of this SOP for identification and quantitation of single component analytes.
- 12.2. Surrogate recovery results are calculated and reported for Triphenylphosphate unless it is determined that sample interference has adversely affected the quantitation of one of the surrogates. The surrogate must be within QC criteria. Corrective action is only necessary if Triphenylphosphate is outside of acceptance limits.

13. METHOD PERFORMANCE

- 13.1. Performance limits for the four replicate initial demonstration of capability are required as referenced under Section 13.1 of the main body of this SOP.

14. POLLUTION PREVENTION

Refer to section 14 of the 8000B section of this SOP.

15. WASTE MANAGEMENT

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

SW846, Update III, December 1996, Method 8141A.

17. MISCELLANEOUS

- 17.1. Reporting limits
 - 17.1.1. The lower standard reporting limits are listed in Table F-1
 - 17.1.2. If samples require dilution or smaller volumes than specified in this method, the RL will be elevated.
 - 17.1.2.1. The nature of the FPD detector contributes to high dilutions for Method 8141A. There is a phenomenon known as quenching that occurs. This happens when light absorption occurs in the flame of the FPD due to hydrocarbons, sulfur, and certain light absorbing compounds.

When this happens the analytes of interest do not reach the photomultiplier tube and are not detected even though they may be present.

17.2. Elution time

17.2.1. For any positive detection in a sample, the chromatogram is overlaid on screen with the nearest standard for elution time matching and pattern recognition using the Target system.

17.3. Troubleshooting guide

17.3.1. Consult the instrument manufacturer's operating manual for guidance.

Table F1: Organophosphorus Pesticides Routinely Analyzed and Reporting Limits

Compound	CAS Number	Reporting Limits	
		Water, µg/L	Solid, µg/kg
Azinphos methyl	86-50-0	1.0	33
Bolstar (Suprofos)	35400-43-2	1.0	33
Chlorpyrifos	2921-88-2	1.0	33
Coumaphos	56-72-4	1.0	33
Demeton, O and S	8065-48-3	1.0	33
Diazinon	333-41-5	1.0	33
Dichlorvos	62-73-7	1.0	33
Disulfoton	298-04-4	1.0	33
Ethoprop	13194-48-4	1.0	33
Fensulfothion	115-90-2	1.0	33
Fenthion	55-38-9	1.0	33
Malathion	121-75-5	1.0	33
Merphos	150-50-5	1.0	33
Methyl Parathion	298-00-0	1.0	33
Mevinphos	7786-34-7	1.0	33
Naled	300-76-5	1.0	33
Phorate	298-02-2	1.0	33
Ronnel	299-84-3	1.0	33
Stirophos	22248-79-9	1.0	33
Tokuthion	34643-46-4	1.0	33
Trichloronate	327-98-0	1.0	33
O,O,O-Trientyl phosphorothioate	126-68-1	1.0	33
Thionazin	297-97-2	1.0	33
Sulfotepp	3689-24-5	1.0	33
Dimethoate	60-51-5	1.0	33
Parathion	56-38-2	1.0	33
Famphur	52-85-7	1.0	33

Table F2: Instrumental Conditions	
PARAMETER	Recommended conditions
Injection port temp	175°C
Detector temp	230°C
Initial temp	135°C
Temperature program	(A) 5°C/minute (B) 20°C/minute
Final Temp	(A) 245°C (B) 295°C
Final Hold Time	(A) 0 minutes (B) 7 minutes
Column 1	RTX-OPP, 30 meter, 0.32 mm, 0.5 µm film
Column 2	RTX-1, 30 meter, 0.32 mm, 0.5 µm film
Injection	1-2µL
Carrier gas	Helium / Hydrogen
Make up gas	Nitrogen

Table F3 Initial Calibration Concentrations (ng/uL)

COMPOUND	LEVEL1	LEVEL2	LEVEL3	LEVEL4	LEVEL5	LEVEL6	LEVEL7
o,o,o-Triethylphosphate	0.2	0.5	1	2	5	10	20
Dichlorvos	0.2	0.5	1	2	5	10	20
Mevinphos	0.2	0.5	1	2	5	10	20
Thionazin	0.2	0.5	1	2	5	10	20
Ethoprop	0.2	0.5	1	2	5	10	20
Naled	0.2	0.5	1	2	5	10	20
Sulfotepp	0.2	0.5	1	2	5	10	20
Phorate	0.2	0.5	1	2	5	10	20
Demeton	0.2	0.5	1	2	5	10	20
Dimethoate	0.2	0.5	1	2	5	10	20
Diazinon	0.2	0.5	1	2	5	10	20
Disulfoton	0.2	0.5	1	2	5	10	20
Methyl Parathion	0.2	0.5	1	2	5	10	20
Ronnel	0.2	0.5	1	2	5	10	20
Fenthion	0.2	0.5	1	2	5	10	20
Chlorpyrifos	0.2	0.5	1	2	5	10	20
Parathion	0.2	0.5	1	2	5	10	20
Malathion	0.2	0.5	1	2	5	10	20
Trichloronate	0.2	0.5	1	2	5	10	20
Merphos	0.2	0.5	1	2	5	10	20
Stirophos	0.2	0.5	1	2	5	10	20
Tokuthion	0.2	0.5	1	2	5	10	20
Fensulfothion	0.2	0.5	1	2	5	10	20
Bolstar	0.2	0.5	1	2	5	10	20
Famphur	0.2	0.5	1	2	5	10	20
Azinphos methyl	0.2	0.5	1	2	5	10	20
Coumaphos	0.2	0.5	1	2	5	10	20

Table F4: LCS/Matrix Spike and Surrogate Spike Compounds – 20 ug/mL

Compound	Compound
Thinazin	Sulfotepp
Phorate	Disulfoton
Methyl Parathion	Parathion
Famphur	O,O,O-Triethylphosphate
Dimethoate	Triphenyl Phosphate - Surrogate

1. Scope and Application

- 1.1. This method is applicable to the determination of the concentration and **tentative** identification of petroleum hydrocarbon mixes in waters, wastewaters, soils, and sludges.
- 1.2. This SOP is based on SW-846 Method 8015B, Modified, Revision 3, December 1996 and Wisconsin DNR Modified DRO method.
- 1.3. The associated LIMS method codes are HS (8015 MOD), KI (8015B), and C6 (Wisconsin DRO).

2. Summary of Method

- 2.1. This method provides gas chromatographic conditions for detection and identification of total petroleum hydrocarbons. Prior to the use of this method, appropriate sample preparation techniques are used.
- 2.2. Wisconsin DRO Method is designed to measure the concentration of diesel range organics in water and soil. This corresponds to a hydrocarbon range of C₁₀-C₂₈ and a boiling point range between approximately 170 °C and 430 °C.
- 2.3. An aliquot of the prepared sample is injected into a gas chromatograph (GC) and compounds in the effluent are detected by a flame ionization detector (FID).

3. Definitions

- 3.1. Refer to the glossary in the STL North Canton Laboratory Quality Manual (LQM), current version.
- 3.2. Wisconsin DNR DRO Method Definition - Diesel Range Organics (DRO) – All the chromatographic response falling between the onset of the n-decane (n-C₁₀) peak and the conclusion of the n-octacosane (n-C₂₈) peak. Quantitation is based on a direct comparison of the total area within this range to the total area of the Diesel Component Standard.
- 3.3. Wisconsin DNR DRO Method Definition –Diesel Component Standard: A ten component blend of typical diesel compounds. This standard serves as a quantitation standard and is used to establish a retention time window for diesel range organics.

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for general information regarding chromatographic interferences.

5. SAFETY

- 5.1. Refer to Section 5 of the Method 8000B SOP for general safety requirements.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP.
- 6.2. Microsyringes, various sizes, for standards preparation, sample injection, and extract dilution.

7. REAGENTS AND STANDARDS

- 7.1. Refer to Section 7 of the 8000B section of this SOP.
- 7.2. The petroleum hydrocarbons are purchased from a chemical supplier when available. When no chemical supplier is available, the fuels are purchased from public sources.
- 7.3. The Diesel Component Stock Standard for the Wisconsin DRO method may be commercially prepared standards if the standards are certified by the manufacturer or by an independent source.

8. SAMPLE PREPARATION, PRESERVATION AND STORAGE

- 8.1. Refer to Section 8 of the 8000B section of this SOP.
- 8.2. Wisconsin DNR DRO aqueous samples must be preserved with HCL such that the pH is < 2.

9. QUALITY CONTROL

- 9.1. Refer to Section 9 of the 8000B section of this SOP for quality control requirements, including the initial demonstration of capability, definition of a batch, surrogate limits, method blanks, laboratory control samples (LCS), and matrix spikes (MS).
- 9.2. LCS recoveries are calculated from a Diesel Calibration. The Wisconsin DNR DRO method specifies the recovery of the LCS-water spike must be between 75%-115% and the RPD <20%. LCS Soil recoveries must fall between 70%-120% of the known concentration and the RSD must be <20%.
 - 9.2.1. The Wisconsin DNR DRO Method requires a Duplicate Laboratory Control Spike with each batch. MS/MSD samples are not required for this method. One LCS must be run at the beginning of the batch of samples and the other at the end.
- 9.3. MS/MSD recoveries are calculated from a Diesel calibration.
- 9.4. Surrogates
 - 9.4.1. Because of the nature of the TPH analysis, whereas certain petroleum mixtures can override the C9 surrogate, the C9 surrogate recoveries are advisory. Re-extraction due to surrogate recoveries is determined by analyst judgement.

NOTE: Ohio VAP rules require reanalysis when surrogate recoveries are outside of control limits.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Recommended Instrument Conditions
 - 10.1.1. Hydrogen carrier gas - flow rate 5 - 6 mL/min
 - 10.1.2. Detector gas mixture - air hydrogen mixture in a 10:1 ratio, air 80 - 120 mL/min, hydrogen 8 -12 mL/min
 - 10.1.3. Temperature Program - refer to Appendix
 - 10.1.4. Injection volume - 1 µL
- 10.2. Initial Calibration

-
- 10.2.1. Analyze a five point Diesel calibration standard and a five point DRO calibration standard referring to the recommended instrument conditions. The calibration concentrations are 100, 200, 500, 1000, and 2000 ng/uL. A 5000ng/uL standard may be analyzed if needed. The retention time window of C10-C32 shall be used for the Diesel calibration.
 - 10.2.2. The Wisconsin DNR DRO method requires the retention time window be defined as beginning 0.1 minutes before the onset of the n-decane peak and ending 0.1 minutes after the conclusion of the n-octacosane peak in the calibration run.
 - 10.2.3. Wisconsin DNR DRO quantitation is based on a direct comparison of the total area within the retention time window to the total area of the Diesel Component Standard.
 - 10.2.4. Wisconsin DNR DRO integration must be "baseline to baseline" as opposed to a "valley to valley". Baseline to baseline is defined here as a flat baseline drawn parallel to the x-axis of chromatographic graph that includes all responses within the retention time window.
 - 10.2.5. A linear regression calibration curve must be prepared for Wisconsin DNR DRO analysis. The curve must have a correlation coefficient of at least 99%.
- 10.3. Continuing Calibration
 - 10.3.1. The calibration verification must bracket the LCS, MS, MSD extracts and must meet passing criteria. The Wisconsin DNR DRO method requires the calibration verification must not vary from the calibration curve by more than 20%.
 - 10.4. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.

11. PROCEDURE

- 11.1. Refer to the method 8000B section of this SOP for procedural requirements.
- 11.2. Extraction

The extraction procedure is described in SOP #CORP-OP-0001NC.
- 11.3. Cleanup
- 11.4. Analytical Sequence – Refer to Section 11 in the 8000B Section of this SOP.
- 11.5. Petroleum Hydrocarbon Identification and/or Fingerprinting
 - 11.5.1. To identify the type of petroleum hydrocarbon, compare the chromatographic peak pattern to the patterns of known petroleum hydrocarbons analyzed under identical chromatographic conditions. Samples are quantified against diesel, but fingerprinting may be done when client requested.
 - 11.5.2. Positive matching may not be possible, even using site-specific hydrocarbons. Degradation of the pattern can occur during environmental exposure of the fuel. See Table 2 for possible fingerprints.
- 11.6. Sample Quantification
 - 11.6.1. Samples are quantified against the initial calibration of diesel or DRO on a single column.
 - 11.6.2. The total height or area of the hydrocarbon is determined in the same manner used for the hydrocarbon standard.

11.6.3. If the amount of sample injected into the GC exceeds the working range of the calibration curve, an appropriate dilution is performed before reanalysis.

12. DATA ANALYSIS AND CALCULATIONS

- 12.1. Refer to the 8000B section of this SOP for identification and quantitation of single component analytes.
- 12.2. Surrogate recovery results are calculated and reported for Nonane (C-9) unless it is determined that sample interference has adversely affected the quantitation of one of the surrogates. The surrogate must be within QC criteria. Corrective action is only necessary if Nonane (C-9) is outside of acceptance limits.

13. METHOD PERFORMANCE

- 13.1. Performance limits for the four replicate initial demonstration of capability are required as referenced under Section 13.1 of the main body of this SOP.

14. POLLUTION PREVENTION

Refer to section 14 of the 8000B section of this SOP.

15. WASTE MANAGEMENT

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

- 16.1. SW846, Method 8015B, Nonhalogenated Organics Using GC/FID, Test Methods for Evaluating Solid Waste, Third Edition, USEPA
- 16.2. Method for Determining Diesel Range Organics, Modified DRO, Wisconsin DNR, Public Hearing Draft, September 1995.
- 16.3. Related SOP
 - 16.3.1. CORP-OP-0001NC, Extraction and Cleanup of Organic Compounds from Waters and Soils, Based on SW846 3500 Series, 3600 Series, 8150, 8151, and 600 Series Methods

Table G1: Suggested GC Temperature Program for TPH analysis

Initial Temperature	40°C
Initial Hold Time	4 minutes
Temperature Program	10°C/minute
Final Temperature	280°C
Final Hold Time	10 minutes

Table G2: Reporting Limits for TPH Analysis

Analyte	Reporting Limits		
	Water (µg/L)	Solids (mg/kg)	Waste Dilution (mg/kg)
TPH (as Diesel) or DRO	100	3.3	200
C10-C32 (Ohio VAP only)	100	3.3	200
Fingerprint Compounds¹			
Mineral Spirits	Kerosene	Motor Oil	
Hydraulic Oil	Jet Fuel	Stoddard Solvent	
DRO Spiking Solution			
Decane	Dodecane	Tetradecane	
Hexadecane	Octobecane	Eicosane	
Docosane	Tetracosane	Hexacosane	
Octacosane			

¹ This list represents most of the common petroleum hydrocarbons. The list may be expanded to include other petroleum hydrocarbons.

**MODIFIED NON-HALOGENATED ORGANIC COMPOUNDS
BY 8015B, DIRECT INJECTION****1. Scope and Application**

- 1.1. This method is applicable to the determination of the concentration of various Non-halogenated Organic Compounds in waters, wastes, sludges, and solids. It is based on SW-846 Method 8015B.
- 1.2. The applicable LIMs method codes are J7 (GC/FID 8015) and QU (Semivolatile Organics, 8015B). The preparation code is 88.

2. Summary of Method

- 2.1. This method provides gas chromatographic conditions for the detection of various nonhalogenated organic compounds. Samples are introduced to the GC by direct injection. Detection is achieved by a flame ionization detector (FID).

3. Definitions

- 3.1. Refer to the glossary in the STL North Canton Laboratory Quality Manual (LQM).

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for general information regarding chromatographic interferences.

5. SAFETY

- 5.1. Refer to section 5 of the Method 8000B SOP for general safety requirements.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP.
- 6.2. Recommended Columns:
 - 6.2.1. RTX Stabilwax, fused silica, 60 m x 0.53, 0.5 µm film thickness, or equivalent column.
 - 6.2.2. Stabilwax-DA, fused silica, 60 m x 0.32, 0.5 µm film thickness, or equivalent column.
- 6.3. Detectors: Flame ionization (FID)
- 6.4. Microsyringes, various sizes, for standards preparation, sample injection, and extract dilution.

7. REAGENTS AND STANDARDS

- 7.1. Refer to Section 7 of the 8000B section of this SOP.
 - 7.1.1. Reagent water

- 7.2. Standards

**MODIFIED NON-HALOGENATED ORGANIC COMPOUNDS
BY 8015B, DIRECT INJECTION**

7.2.1. Refer to Table H2.

8. SAMPLE PREPARATION, PRESERVATION AND STORAGE

8.1. Refer to section 8 of the 8000B section of this SOP.

9. QUALITY CONTROL

9.1. Refer to Section 9 of the 8000B section of this SOP for quality control requirements, including the initial demonstration of capability, definition of a batch, surrogate limits, method blanks, laboratory control samples (LCS), and matrix spikes (MS).

10. CALIBRATION AND STANDARDIZATION

10.1. Recommended Instrument Conditions

10.1.1. The following conditions are recommended. The following table lists specific information for various compounds.

10.1.2. RTX- Stabilwax and Stabilwax-DA Columns:

Carrier Gas:	Hydrogen
Initial Temp:	45 °C
Initial Hold:	3 mins
Ramp Rate A:	5 °C/min
Ramp Rate B:	30 °C/min
Final Hold A:	0 min
Final Hold B:	3 mins
Analysis Time:	17.83 mins
Injector Temp:	275 °C
FID Temp:	300 °C
Injection Vol:	1 µL

10.2. Initial Calibration

10.2.1. For each non-halogenated organic compound and surrogate standard, analyze five or more calibration standards referring to the recommended GC conditions in Section 10.1. One of the standards analyzed should be at or near the concentration which corresponds to the calibration range.

10.2.2. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.

11. PROCEDURE

11.1. Refer to the method 8000B section of this SOP for procedural requirements.

11.2. Sample Preparation Summary

11.2.1. Samples received fall into three general categories: waters, soils, or wastes.

**MODIFIED NON-HALOGENATED ORGANIC COMPOUNDS
BY 8015B, DIRECT INJECTION****11.3. Sample Preparation Procedure****11.3.1. Waters**

11.3.1.1. Add surrogate solution to the sample to achieve a concentration of 100 mg/L.

11.3.2. Sediments/soils and waste

11.3.2.1. Mix the contents of the sample container with a narrow metal or wood spatula. Weigh 5 g (wet weight) into a tared culture tube. Use a top-loading balance. Record the weight to 0.01 gram.

11.3.2.2. Quickly add 5 mL of reagent water. Add surrogate standard. Cap the vial and vortex to mix for two minutes.

11.3.2.3. If extract is cloudy or has suspended sediment particles, refrigerate and allow sample to sit for a maximum of 24 hours. Filter sample if necessary.

11.4. Sample Analysis**11.4.1. Preliminary Evaluation**

11.4.1.1. The sample or sample extract is introduced to the GC column by direct inject techniques. The concentration of the sample components is then calculated from the resulting chromatograms.

11.4.2. Analytical Sequence – Refer to Section 11 in the 8000B Section of this SOP.

11.4.3. Inject 1 μ L of the sample extract or diluted sample into the GC using the same operating conditions and techniques as those used in the calibration of the instrument.

11.5. Analytical Documentation

11.5.1. Record all analytical information in the analytical logbook/logsheets, including the analytical data from standards, blanks, LCSs, MS/MSDs, and any corrective actions or modifications to the method.

11.5.2. All standards are logged into a department standard logbook. All standards are assigned a unique number for identification. Logbooks are reviewed by the supervisor or designee.

11.5.3. Documentation such as all associated instrument printouts (final runs, screens, reruns, QC samples, etc.) and daily calibration data corresponding to all final runs is available for each data file.

12. DATA ANALYSIS AND CALCULATIONS

12.1. Refer to the 8000B section of this SOP for identification and quantitation of single component analytes.

12.2. Surrogate recovery results are calculated and reported unless it is determined that sample interference has adversely affected the quantitation of the surrogate. The surrogate must be within QC criteria. Corrective action is only necessary if the surrogate is outside of acceptance limits.

13. METHOD PERFORMANCE

13.1. Performance limits for the four replicate initial demonstration of capability are required as referenced under Section 13.1 of the main body of this SOP.

10/11/01

**MODIFIED NON-HALOGENATED ORGANIC COMPOUNDS
BY 8015B, DIRECT INJECTION**

Revision No: 5.6

Revision Date: 05/25/01

Page H4 of 5

14. POLLUTION PREVENTION

Refer to section 14 of the 8000B section of this SOP.

15. WASTE MANAGEMENT

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

- 16.1. SW846, Method 8015B, Nonhalogenated Organics Using GC/FID, Test Methods for Evaluating Solid Waste, Third Edition, USEPA

17. Miscellaneous (Tables, Appendices, Etc...)

Table H1 Non-halogenated Organic Compounds, Reporting Limits¹

Compound	CAS Number	Reporting Limits	
		Water, mg/L	Solid, mg/kg
2-Methoxyethanol	109-86-4	1.0	1.0
Methanol	67-56-1	1.0	0.5
Isopropyl alcohol	67-63-0	1.0	0.5
n-Propyl alcohol	71-23-8	1.0	0.5
Ethanol	64-17-5	1.0	0.5
n-Butanol	71-36-3	1.0	0.5
1,4-Dioxane	123-91-1	1.0	0.5
Ethylene oxide	75-21-8	1.0	0.5
iso-Butanol	78-83-1	1.0	0.5

¹ If samples require dilution or smaller volumes than specified in this method, the RL will be elevated.

Table H2 Non-halogenated Organic Compounds Working Standards

**MODIFIED NON-HALOGENATED ORGANIC COMPOUNDS
BY 8015B, DIRECT INJECTION**Revision No: 5.6Revision Date: 05/25/01

Page H5 of 5

COMPOUND	LEVEL1	LEVEL2	LEVEL3	LEVEL4	LEVEL5	LEVEL 6
2-Methoxyethanol	1.0	2.0	5.0	10.0	20.0	50.0
Methanol	1.0	2.0	5.0	10.0	20.0	50.0
Isopropyl alcohol	1.0	2.0	5.0	10.0	20.0	50.0
n-Propyl alcohol	1.0	2.0	5.0	10.0	20.0	50.0
Ethanol	1.0	2.0	5.0	10.0	20.0	50.0
n-Butanol	1.0	2.0	5.0	10.0	20.0	50.0
1,4-Dioxane	1.0	2.0	5.0	10.0	20.0	50.0
Ethylene oxide	1.0	2.0	5.0	10.0	20.0	50.0
iso-Butanol	1.0	2.0	5.0	10.0	20.0	50.0

1. Scope and Application

- 1.1. This method is applicable to the determination of the concentration of Sulfolane and N-Methyl-2-pyrrolidone in water and solid samples. It is based on SW846 Method 8015B. The working linear range is 50 to 1000 µg/L. Table I1 lists the reporting limits associated with this method.
- 1.2. The applicable LIMS method code is KU.

2. Summary of Method

- 2.1. This method provides gas chromatographic conditions for the detection of mg/L levels of Phillips 66 compounds in water. Prior to use of this method, appropriate sample extraction techniques must be used.

3. Definitions

- 3.1. Refer to the glossary in the STL North Canton Laboratory Quality Manual (LQM), current version.

4. INTERFERENCES

- 4.1. Refer to the method 8000B section of this SOP for general information regarding chromatographic interferences.

5. SAFETY

- 5.1. Refer to section 5 of the Method 8000B SOP for general safety requirements.
- 5.2. The health and safety hazards of many of the chemicals used in this procedure have not been fully defined. Additional health and safety information can be obtained from the Material Safety Data Sheets (MSDS) maintained in the laboratory. The following specific hazards are known:
 - 5.2.1. Chemicals that have been classified as **carcinogens**, or **potential carcinogens**, under OSHA include: **Methylene Chloride**.

6. EQUIPMENT AND SUPPLIES

- 6.1. Refer to Section 6 of the 8000B section of this SOP.
- 6.2. Gas Chromatograph
 - 6.2.1. Gas Chromatograph: Modified to accept capillary columns
 - 6.2.2. Data System: Capable of peak integration
 - 6.2.3. Gas Chromatographic Column: 30 m x 0.32 mm ID RTX-5 fused silica capillary column
 - 6.2.4. Autosampler: Capable of reproducible injections
 - 6.2.5. Carrier Gas: Hydrogen
 - 6.2.6. Detector: Flame ionization (FID)

- 6.3. Volumetric Flasks: 10, 50, and 100 mL
- 6.4. Microsyringe: 10 μ L
- 6.5. Pipettes: Disposable μ L, Pasteur
- 6.6. Autosampler Vials: 1 mL with 11 mm crimp cap, Teflon[®]/silicone septum liner.

7. REAGENTS AND STANDARDS

- 7.1. Refer to Section 7 of the 8000B section of this SOP.
- 7.2. Reagents
 - 7.2.1. Methylene Chloride: Pesticide grade or equivalent
- 7.3. Standards
 - 7.3.1. Refer to Section 7 of the 8000B section of this SOP.

8. SAMPLE PREPARATION, PRESERVATION AND STORAGE

- 8.1. Refer to Section 8 of the 8000B section of this SOP.

9. QUALITY CONTROL

- 9.1. Refer to Section 9 of the 8000B section of this SOP for quality control requirements, including the initial demonstration of capability, definition of a batch, surrogate limits, method blanks, laboratory control samples (LCS), and matrix spikes (MS).
- 9.2. Surrogates are not used for this analysis.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Initial Calibration
 - 10.1.1. Refer to Section 10 of the 8000B section of this SOP for general calibration requirements.

11. PROCEDURE

- 11.1. Refer to the method 8000B section of this SOP for procedural requirements.
- 11.2. Sample Analysis
 - 11.2.1. Summary
 - 11.2.1.1. The sample extract is injected onto the GC column. The compounds are then identified and quantitated.
 - 11.2.2. Recommended Instrument Conditions

11.2.2.1. GC Conditions

Initial Temperature:	45°C
Initial Hold Time:	4 minutes
Temperature Program:	15°C/minute
Final Temperature:	300°C, hold 2 minutes
Final Time:	23 minutes
Carrier Gas:	Hydrogen
Injection Volume:	1 µL

11.2.3. Sample Analysis Procedure

11.2.3.1. Preliminary Evaluation

The sample extracts may be screened to determine the level of analyte present. If the level of analyte exceeds the working range of the calibration curve, an appropriate dilution is performed to bring the level within the calibration range.

11.2.4. Inject 1 µL of the sample extract or diluted sample into the GC using the same conditions as those used in calibration.

11.2.5. Identification

11.2.5.1. Analytes of interest are identified by comparing retention times with known standards.

11.2.5.2. A single column is used for identification.

11.2.6. Sample Quantification

11.2.6.1. Refer to Section 11 in the 8000B Section of this SOP

11.3. Analytical Documentation

11.3.1. Record all analytical information in the analytical logbook/logsheets, including the analytical data from standards, blanks, LCSS, MS/MSDs, and any corrective actions or modifications to the method.

11.3.2. All standards are logged into a department standard logbook. All standards are assigned a unique number for identification. Logbooks are reviewed by the supervisor or designee.

11.3.3. Documentation such as all associated instrument printouts (final runs, screens, reruns, QC samples, etc.) and daily calibration data corresponding to all final runs is available for each data file.

12. DATA ANALYSIS AND CALCULATIONS

12.1. Refer to the 8000B section of this SOP for identification and quantitation of single component analytes.

13. METHOD PERFORMANCE

- 13.1. Performance limits for the four replicate initial demonstration of capability are required as referenced under Section 13.1 of the main body of this SOP.

14. POLLUTION PREVENTION

- 14.1. Refer to section 14 of the 8000B section of this SOP.

15. WASTE MANAGEMENT

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

- 16.1. SW846, Method 8015B, Nonhalogenated Organics Using GC/FID, Test Methods for Evaluating Solid Waste, Third Edition, USEPA

17. Miscellaneous (Tables, Appendices, Etc...)

- 17.1. Reporting limits

17.1.1. The lower reporting limits are shown in Table II

17.1.2. If samples require dilution or smaller volumes than specified in this method, the RL will be elevated.

TABLE II
PHILLIPS 66 REPORTING LIMITS

Compound	Reporting Limits, $\mu\text{g/L}$
Tetramethylene sulfone (Sulfolane)	50
N-Methyl-2-pyrrolidone	50

Controlled Copy
Copy No: _____

SOP No. CORP-MS-0002NC
Revision No. 2.3
Revision Date: 05/23/01
Page 1 of 67

Implementation Date: 06/05/01

STL STANDARD OPERATING PROCEDURE

**TITLE: DETERMINATION OF VOLATILE ORGANICS BY GC/MS BASED ON
METHOD 8260B, 8260A, AND 624**

(SUPERSEDES: REVISION 2.2, DATED 11/28/00)

Prepared by: Thomas E. Stiller 5/25/01
Date

Approved by: Ramona Evans 5/25/01
Technology Specialist Date

Approved by: Paul A. Cook 5/25/01
Quality Assurance Manager Date

Approved by: [Signature] 5-29-01
Environmental Health and Safety Coordinator Date

Approved by: Christy R. Ormond 5-25-01
Lab Manager Date

Approved by: [Signature] 6/4/01
Corporate Technology and/or Corporate Quality Assurance Date

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use it for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

Controlled Copy
Copy No: _____

SOP No. CORP-MS-0002NC
Revision No. 2.3
Revision Date: 05/23/01
Page 1 of 67

Implementation Date: _____

STL STANDARD OPERATING PROCEDURE

**TITLE: DETERMINATION OF VOLATILE ORGANICS BY GC/MS BASED ON
METHOD 8260B, 8260A, AND 624**

(SUPERSEDES: REVISION 2.2, DATED 11/28/00)

Prepared by: _____ Date _____
Intranet Controlled

Approved by: _____ Date _____
Technology Specialist

Approved by: _____ Date _____
Quality Assurance Manager

Approved by: _____ Date _____
Environmental Health and Safety Coordinator

Approved by: _____ Date _____
Lab Manager

Approved by: *Phil Beal* _____ Date 6/4/01
Corporate Technology and/or Corporate Quality Assurance

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use it for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

TABLE OF CONTENTS

1. SCOPE AND APPLICATION.....	4
2. SUMMARY OF METHOD	4
3. DEFINITIONS	5
4. INTERFERENCES	7
5. SAFETY	7
6. EQUIPMENT AND SUPPLIES	8
7. REAGENTS AND STANDARDS	11
8. SAMPLE COLLECTION, PRESERVATION AND STORAGE.....	12
9. QUALITY CONTROL	19
10. CALIBRATION AND STANDARDIZATION	23
11. PROCEDURE	27
12. DATA ANALYSIS AND CALCULATIONS	32
13. METHOD PERFORMANCE.....	39
14. POLLUTION PREVENTION.....	40
15. WASTE MANAGEMENT.....	40
16. REFERENCES	40
17. MISCELLANEOUS	40
APPENDIX A	
ANALYSIS OF VOLATILE ORGANICS BY METHOD 624	55

LIST OF TABLES

Table 1	STL Primary Standard and Reporting Limits
Table 2	STL Primary Standard Calibration Levels
Table 2A	STL Primary Standard Calibration Levels- Low Level
Table 3	STL Appendix IX Standard and Reporting Limits
Table 4	STL Appendix IX Standard Calibration Levels
Table 5	Reportable Analytes for STL Standard Tests, Primary Standard
Table 6	Reportable Analytes for STL Standard Tests, Appendix IX Standard
Table 7	Internal Standards
Table 8	Surrogate Standards
Table 9	Matrix Spike / LCS Control Standards
Table 10	BFB Tune Criteria
Table 11	SPCC Compounds and Minimum Response Factors
Table 12	CCC Compounds
Table 13	Characteristic Ions
Table A-1	Method 624 Analytes and Reporting Limits
Table A-2	Method 624 QC Acceptance Criteria

1.0 SCOPE AND APPLICATION

- 1.1. This method is applicable to the determination of Volatile Organic Compounds in waters, wastewater, soils, sludges and other solid matrices. Standard analytes are listed in Tables 5 and 6.
- 1.2. This SOP is applicable to method 8260B. It may also be used for analysis following method 8260A. Appendix A presents modifications to the procedures in the main SOP that are necessary for analysis of wastewater by method 624. The associated LIMS method codes are QK (8260B), DN (624), and MZ (8260A). Ohio VAP projects are distinguished by Program Code 2J. The following Prep Codes are used: 15 (5 mL purge), 25 (25 mL purge), 4B (Methanol preservation, EnCore™), 4D (Sodium Bisulfate preservation, EnCore™), 4P (Frozen, EnCore™), and 73 (5030A Methanol Prep).
- 1.3. This method can be used to quantify most volatile organic compounds that have boiling points below 200°C and are insoluble or slightly soluble in water. Volatile water soluble compounds can be included in this analytical technique; however, for more soluble compounds, quantitation limits are approximately ten times higher because of poor purging efficiency.
- 1.4. The method is based upon a purge and trap, gas chromatograph/mass spectrometric (GC/MS) procedure. The approximate working range is 5 to 200 µg/L for 5 mL waters, 1 to 40 µg/L for 25 mL purge waters, 5 to 200 µg/kg for low-level soils, and 250 to 25,000 µg/kg for medium-level soils. Reporting limits are listed in Tables 1 and 3.
- 1.5. Method performance is monitored through the use of surrogate compounds, matrix spike/matrix spike duplicates, and laboratory control spike samples.

2. SUMMARY OF METHOD

- 2.1. Volatile compounds are introduced into the gas chromatograph by the purge and trap method. The components are separated via the chromatograph and detected using a mass spectrometer, which is used to provide both qualitative and quantitative information.
- 2.2. Aqueous samples are purged directly. Generally, soils are preserved by extracting the volatile analytes into methanol. If especially low detection limits are required, soil samples may be preserved with sodium bisulfate and purged directly.
- 2.3. In the purge and trap process, an inert gas is bubbled through the solution at ambient temperature or at 40°C (40°C required for low level soils) and the volatile components are efficiently transferred from the aqueous phase to the vapor phase. The vapor is swept

through a sorbant column where the volatile components are trapped. After purging is completed, the sorbant column (trap) is heated and backflushed with inert gas to desorb the components onto a gas chromatographic column. The gas chromatographic column is then heated to elute the components which are detected with a mass spectrometer.

- 2.4. Qualitative identifications are confirmed by analyzing standards under the same conditions used for samples and comparing the resultant mass spectra and GC retention times. Each identified component is quantified by relating the MS response for an appropriate selected ion produced by that compound to the MS response for another ion produced by an internal standard.

3. DEFINITIONS

3.1. Batch

The batch is a set of up to 20 samples of the same matrix processed using the same procedures and reagents within the same time period. Using this method, each BFB analysis will normally start a new batch. Batches for medium level soils are defined at the sample preparation stage and may be analyzed on multiple instruments over multiple days, although reasonable effort should be made to keep the samples together.

- 3.1.1. The Quality Control batch must contain a matrix spike/spike duplicate (MS/MSD), a Laboratory Control Sample (LCS), and a method blank. In some cases, at client request, the MS/MSD may be replaced with a matrix spike and sample duplicate. Refer to the STL QC Program document (QA-003) for further details of the batch definition.

3.2. Method Blank

- 3.2.1. A method blank consisting of all reagents added to the samples must be analyzed with each batch of samples. The method blank is used to identify any background interference or contamination of the analytical system which may lead to the reporting of elevated concentration levels or false positive data.

3.3. Laboratory Control Sample (LCS)

- 3.3.1. Laboratory Control Samples are well characterized, laboratory generated samples used to monitor the laboratory's day-to-day performance of routine analytical methods. The LCS, spiked with a group of target compounds representative of the method analytes, is used to monitor the accuracy of the analytical process, independent of matrix effects. Ongoing monitoring of the LCS results provides evidence that the laboratory is performing the method within accepted QC guidelines for accuracy and precision.

3.4. Surrogates

- 3.4.1. Surrogates are organic compounds which are similar to the target analyte(s) in chemical composition and behavior in the analytical process, but which are not normally found in environmental samples. Each sample, blank, LCS, and MS/MSD is spiked with surrogate standards. Surrogate spike recoveries must be evaluated by determining whether the concentration (measured as percent recovery) falls within the required recovery limits.

3.5. Matrix Spike/Matrix Spike Duplicate (MS/MSD)

- 3.5.1. A matrix spike is an environmental sample to which known concentrations of target analytes have been added. A matrix spike duplicate is a second aliquot of the same sample which is prepared and analyzed along with the sample and matrix spike. Matrix spikes and duplicates are used to evaluate accuracy and precision in the actual sample matrix.

3.6. Calibration Check Compound (CCC)

- 3.6.1. CCCs are a representative group of compounds which are used to evaluate initial calibrations and continuing calibrations. Relative percent difference for the initial calibration and % drift for the continuing calibration response factors are calculated and compared to the specified method criteria.

3.7. System Performance Check Compounds (SPCC)

SPCCs are compounds which are sensitive to system performance problems and are used to evaluate system performance and sensitivity. A response factor from the continuing calibration is calculated for the SPCC compounds and compared to the specified method criteria.

4. INTERFERENCES

- 4.1. Method interferences may be caused by contaminants in solvents, reagents, glassware, and other processing apparatus that lead to discrete artifacts. All of these materials must be routinely demonstrated to be free from interferences under conditions of the analysis by running laboratory method blanks as described in the Quality Control section. The use of ultra high purity gases, pre-purged purified reagent water, and approved lots of purge and trap grade methanol will greatly reduce introduction of contaminants. In extreme cases the purging vessels may be pre-purged to isolate the instrument from laboratory air contaminated by solvents used in other parts of the laboratory.
- 4.2. Samples can be contaminated by diffusion of volatile organics (particularly methylene chloride and fluorocarbons) into the sample through the septum seal during shipment and storage. A field blank prepared from reagent water and carried through the sampling and handling protocol can serve as a check on such contamination.
- 4.3. Matrix interferences may be caused by non-target contaminants that are coextracted from the sample. The extent of matrix interferences will vary considerably from source to source depending upon the nature and diversity of the site being sampled.
- 4.4. Cross-contamination can occur whenever high-level and low-level samples are analyzed sequentially or in the same purge position on an autosampler. Whenever an unusually concentrated sample is analyzed, it should be followed by one or more blanks to check for cross-contamination. The purge and trap system may require extensive bake-out and cleaning after a high-level sample.
- 4.5. Some samples may foam when purged due to surfactants present in the sample. When this kind of sample is encountered an antifoaming agent (e.g., J.T. Baker's Antifoam B silicone emulsion) can be used. A blank spiked with this agent must be analyzed with the sample because of the non-target interferences associated with the agent.

5. SAFETY

- 5.1. Procedures shall be carried out in a manner that protects the health and safety of all STL associates.
- 5.2. The Chemical Hygiene Plan (CHP) gives details about the specific health and safety practices which are to be followed in the laboratory area. Personnel must receive training in

the CHP, including the written Hazard Communication plan, prior to working in the laboratory. Consult the CHP, the STL Health and Safety Policies and Procedures Manual, and available Material Safety Data Sheets (MSDS) prior to using the chemicals in the method.

- 5.3. Consult the STL Health and Safety Policies and Procedures Manual for information on Personal Protective Equipment. Eye protection that protects against splash and a laboratory coat must be worn in the lab. Appropriate gloves must be worn while samples, standards, solvents, and reagents are being handled. Disposable gloves that have been contaminated will be removed and discarded; other gloves will be cleaned immediately. Disposable gloves shall not be reused.
- 5.4. The health and safety hazards of many of the chemicals used in this procedure have not been fully defined, therefore each chemical compound should be treated as a potential health hazard. Additional health and safety information can be obtained from the MSDS files maintained in the laboratory. The following specific hazards are known:
 - 5.4.1. Chemicals that have been classified as carcinogens, or potential carcinogens, under OSHA include: Acrylonitrile, benzene, carbon tetrachloride, chloroform, 1,2-dibromo-3-chloropropane, 1,4-dichlorobenzene, and vinyl chloride.
 - 5.4.2. Chemicals known to be flammable are: **Methanol**.
- 5.5. Exposure to chemicals must be maintained **as low as reasonably achievable**, therefore, unless they are known to be non-hazardous, all samples should be opened, transferred, and prepared in a fume hood, or under other means of mechanical ventilation. Solvent and waste containers will be kept closed unless transfers are being made.
- 5.6. The preparation of standards and reagents will be conducted in a fume hood with the sash closed as far as the operations will permit.
- 5.7. All work must be stopped in the event of a known or potential compromise to the health and safety of a STL associate. The situation must be reported **immediately** to a laboratory supervisor.
- 5.8. Laboratory personnel assigned to perform hazardous waste disposal procedures must have a working knowledge of the established procedures and practices outlined in the STL Health and Safety Manual. These employees must have training on the hazardous waste disposal practices initially upon assignment of these tasks, followed by an annual refresher training.

6. EQUIPMENT AND SUPPLIES

- 6.1. Microsyringes: 10 µL and larger, 0.006 inch ID needle.

- 6.2. Syringe: 5 or 25 mL glass with luerlok tip, if applicable to the purging device.
- 6.3. Balance: Analytical, capable of accurately weighing 0.0001 g, and a top-loading balance capable of weighing 0.1 g
- 6.4. Glassware:
 - 6.4.1. Vials: 20 mL with screw caps and Teflon liners.
 - 6.4.2. Volumetric flasks: 10 mL and 100 mL, class A with ground-glass stoppers.
- 6.5. Spatula: Stainless steel.
- 6.6. Disposable pipets: Pasteur.
- 6.7. pH paper: Wide range.
- 6.8. Gases:
 - 6.8.1. Helium: Ultra high purity, gr. 5, 99.999%.
 - 6.8.2. Nitrogen: Ultra high purity, from cylinders of gas generators, may be used as an alternative to helium for purge gas.
 - 6.8.3. Compressed air: Used for instrument pneumatics.
 - 6.8.4. Liquid nitrogen: Used for cryogenic cooling if necessary.
- 6.9. Purge and Trap Device: The purge and trap device consists of the sample purger, the trap, and the desorber.
 - 6.9.1. Sample Purger: The recommended purging chamber is designed to accept 5 mL samples with a water column at least 3 cm deep. The purge gas must pass through the water column as finely divided bubbles, each with a diameter of less than 3 mm at the origin. The purge gas must be introduced no more than 5 mm from the base of the water column. Alternative sample purge devices may be used provided equivalent performance is demonstrated. Low level soils are purged directly from a VOA vial.
 - 6.9.2. Trap: A variety of traps may be used, depending on the target analytes required. For most purposes the Vocarb 3000 trap is suitable. Other traps, such as Vocarb 4000, or Tenax / Silica gel / Charcoal may be used if the Quality Control criteria are met.

6.9.3. Desorber: The desorber should be capable of rapidly heating the trap to 180°C. Many such devices are commercially available.

6.9.4. Sample Heater: A heater capable of maintaining the purge device at 40°C is necessary for low level soil analysis.

6.10. Gas Chromatograph/Mass Spectrometer System:

6.10.1. Gas Chromatograph: The gas chromatograph (GC) system must be capable of temperature programming.

6.10.2. Gas Chromatographic Columns: Capillary columns are used. Some typical columns are listed below:

6.10.2.1. Column 1: 105m x 0.53 ID Rtx-624 with 3 µm film thickness.

6.10.2.2. Column 2: 75 m x 0.53 ID DB-624 widebore with 3 µm film thickness.

6.10.2.3. Mass Spectrometer: The mass spectrometer must be capable of scanning 35-300 AMU every two seconds or less, using 70 volts electron energy in the electron impact mode and capable of producing a mass spectrum that meets the required criteria when 50 ng of 4-Bromofluorobenzene (BFB) are injected onto the gas chromatograph column inlet.

6.10.3. GC/MS interface: In general glass jet separators are used but any interface (including direct introduction to the mass spectrometer) that achieves all acceptance criteria may be used.

6.10.4. Data System: A computer system that allows the continuous acquisition and storage on machine readable media of all mass spectra obtained throughout the duration of the chromatographic program. The computer must have software that allows searching any GC/MS data file for ions of a specified mass and plotting such ion abundances versus time or scan number. This type of plot is defined as an Extracted Ion Current Profile (EICP). Software must also be available that allows integrating the abundances in any EICP between the specified time or scan-number limits. Also, for the non-target compounds, software must be available that allows for the comparison of sample spectra against reference library spectra. The most recent release of the NIST/EPA mass spectral library should be used as the reference library. The computer system must also be capable of backing up data for long-term off-line storage.

6.10.5. Cryogenic Cooling: Some columns require the use of liquid nitrogen to achieve the subambient temperature required for the proper separation of the gases.

7. REAGENTS AND STANDARDS

7.1. Reagents

7.1.1. Methanol: Purge and Trap Grade, High Purity

7.1.2. Reagent Water: High purity water that meets the requirements for a method blank when analyzed. (See section 9.4) Reagent water may be purchased as commercial distilled water and prepared by purging with an inert gas overnight. Other methods of preparing reagent water are acceptable.

7.2. Standards

7.2.1. Calibration Standard

7.2.1.1. Stock Solutions: Stock solutions may be purchased as certified solutions from commercial sources or prepared from pure standard materials as appropriate. These standards are prepared in methanol and stored in Teflon-sealed screw-cap bottles with minimal headspace at -10° to -20°C .

7.2.1.2. Working standards: A working solution containing the compounds of interest prepared from the stock solution(s) in methanol. These standards are stored in the freezer or as recommended by the manufacturer. Working standards are monitored by comparison to the initial calibration curve. If any of the calibration check compounds drift in response from the initial calibration by more than 20% then corrective action is necessary. This may include steps such as instrument maintenance, preparing a new calibration verification standard or tuning the instrument. If the corrective actions do not correct the problem then a new initial calibration must be performed.

7.2.1.3. Aqueous Calibration Standards are prepared in reagent water using the secondary dilution standards. These aqueous standards must be prepared daily.

7.2.1.4. If stock or secondary dilution standards are purchased in sealed ampoules they may be used up to the manufacturers expiration date.

7.2.2. Internal Standards: Internal standards are added to all samples, standards, and blank analyses. Refer to Table 7 for internal standard components.

7.2.3. Surrogate Standards: Refer to Table 8 for surrogate standard components and spiking levels.

-
- 7.2.4. Laboratory Control Sample Spiking Solutions: Refer to Table 9 for LCS components and spiking levels.
- 7.2.5. Matrix Spiking Solutions: The matrix spike contains the same components as the LCS. Refer to Table 9.
- 7.2.6. Tuning Standard: A standard is made up that will deliver 50 ng on column upon injection. A recommended concentration of 25 ng/ μ L of 4-Bromofluorobenzene in methanol is prepared as described in Sections 7.2.1.1 and 7.2.1.2.

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE

- 8.1. Holding times for all volatile analysis are 14 days from sample collection.
- 8.2. Water samples are normally preserved at $\text{pH} \leq 2$ with 1:1 hydrochloric acid. If residual chlorine is present, 2 drops of 10% sodium thiosulfate are added.
- 8.3. Solid samples are field preserved with sodium bisulfate solution for low level analysis, or with methanol for medium level analysis. Soil samples can also be taken using the EnCore™ sampler and preserved in the lab within 48 hours of sampling. At specific client request, unpreserved soil samples may be accepted.
- 8.4. There are several methods of sampling soil. The recommended method, which provides the minimum of field difficulties, is to take an EnCore™ sample. (The 5 g or 25 g sampler can be used, depending on client preference). Following shipment back to the lab the soil is preserved in methanol. This is the medium level procedure. If very low detection limits are needed ($< 50 \mu\text{g}/\text{kg}$ for most analytes) then it will be necessary to use two additional 5 g EnCore™ samplers or to use field preservation.
- 8.5. Sample collection for medium level analysis using EnCore™ samplers.
- 8.5.1. Ship one 5 g (or 25 g) EnCore™ sampler per field sample position.
- 8.5.2. An additional bottle must be shipped for percent moisture determination.
- 8.5.3. When the samples are returned to the lab, extrude the (nominal) 5g (or 25 g) sample into a tared VOA vial containing 5 mL methanol (25 mL methanol for the 25 g sampler). Obtain the weight of the soil added to the vial and note on the label.
- 8.5.4. Add the correct amount of surrogate spiking mixture. (Add 25 μ L of 2500 $\mu\text{g}/\text{mL}$ solution for a nominal 25 g sample, 5 μ L for a nominal 5 g sample.) Refer to Section 17.8 for Michigan project criteria.

-
- 8.5.5. Add the correct amount of matrix spiking solution to the matrix spike and matrix spike duplicate samples. (Add 500 μL of 50 $\mu\text{g}/\text{mL}$ solution for a nominal 25 g sample, 100 μL for a nominal 5 g sample.) Reduce the volume of methanol added to ensure the final volume is 25 mL for nominal 25 g sample or 5 mL methanol for a nominal 5 g sample. Refer to Section 17.8 for Michigan project criteria.
 - 8.5.6. Prepare an LCS for each batch by adding the correct amount of matrix spiking solution to clean methanol. (50 μL of spike to 25 mL methanol or 10 μL spike to 5 mL methanol). Refer to Section 17.8 for Michigan project criteria.
 - 8.5.7. Shake the samples for two minutes to distribute the methanol throughout the soil.
 - 8.5.8. Allow to settle, then remove a portion of methanol and store in a clean Teflon capped vial at $4 \pm 2^\circ\text{C}$ until analysis.
- 8.6. Sample collection for medium level analysis using field methanol preservation
 - 8.6.1. Prepare a 2 oz sample container by adding 25 mL purge and trap grade methanol. (If a 5 g sample is to be used, add 5 mL methanol to a 2 oz container or VOA vial).
 - 8.6.2. Seal the bottle and attach a label.
 - 8.6.3. Weigh the bottle to the nearest 0.01g and note the weight on the label.
 - 8.6.4. Ship with appropriate sampling instructions.
 - 8.6.5. Each sample will require an additional bottle with no preservative for percent moisture determination.
 - 8.6.6. At client request, the methanol addition and weighing may also be performed in the field.
 - 8.6.7. When the samples are returned to the lab, obtain the weight of the soil added to the vial and note on the label.
 - 8.6.8. Add the correct amount of surrogate spiking mixture. (Add 25 μL of 2500 $\mu\text{g}/\text{mL}$ solution for a nominal 25 g sample, 5 μL for a nominal 5 g sample.) Refer to Section 17.8 for Michigan project criteria.
 - 8.6.9. Add the correct amount of matrix spiking solution to the matrix spike and matrix spike duplicate samples. (Add 25 μL of 50 $\mu\text{g}/\text{mL}$ solution for a nominal 25 g sample, 100 μL for a nominal 5 g sample.) Reduce the volume of methanol added to ensure

the final volume is 25 mL for nominal 25 g sample or 5 mL methanol for a nominal 5 g sample. Refer to Section 17.8 for Michigan project criteria.

8.6.10. Prepare an LCS for each batch by adding the correct amount of matrix spiking solution to clean methanol. (500 μ L of spike to 25 mL methanol or 100 μ L spike to 5 mL methanol). Refer to Section 17.8 for Michigan project criteria.

8.6.11. Shake the samples for two minutes to distribute the methanol throughout the soil.

8.6.12. Allow to settle, then remove a portion of methanol and store in a clean Teflon capped vial at 4+2°C until analysis.

8.7. Low level procedure

8.7.1. If low detection limits are required (typically < 50 μ g/kg) sodium bisulfate preservation must be used. However, it is also necessary to take a sample for the medium level (field methanol preserved or using the EnCore™ sampler) procedure, in case the concentration of analytes in the soil is above the calibration range of the low level procedure.

8.7.2. A purge and trap autosampler capable of sampling from a sealed vial is required for analysis of samples collected using this method. (Varian Archon or O.I. 4552).

8.7.3. The soil sample is taken using a 5g EnCore™ sampling device and returned to the lab. It is recommended that two EnCore™ samplers be used for each field sample position, to allow for any reruns than may be necessary. A separate sample for % moisture determination is also necessary.

8.7.4. Prepare VOA vials by adding a magnetic stir bar, approximately 1 g of sodium bisulfate and 5 mL of reagent water.

8.7.5. Seal and label the vial. It is strongly recommended that the vial is labeled with an indelible marker rather than a paper label, since paper labels may cause the autosampler to bind and malfunction. The label absolutely must not cover the neck of the vial or the autosampler will malfunction.

8.7.6. Weigh the vial to the nearest 0.1g and note the weight on the label.

8.7.7. Extrude the soil sample from the EnCore™ sampler into the prepared VOA vial. Reweigh the vial to obtain the weight of soil and note on the label.

Note: Soils containing carbonates may effervesce when added to the sodium bisulfate solution. If this is the case at a specific site, add 5 mL of water instead, and freeze at

<-10°C within 48 hours, analyzed within 12 days after preserving with water, and stored at a 45 degree angle in the freezer.

Note: Freezing is not allowed for Ohio VAP soil samples.

8.7.9. Alternatively the sodium bisulfate preservation may be performed in the field. This is not recommended because of the many problems that can occur in the field setting. Ship at least two vials per sample. The field samplers must determine the weight of soil sampled. Each sample will require an additional bottle with no preservative for percent moisture determination, and an additional bottle preserved with methanol for the medium level procedure. Depending on the type of soil it may also be necessary to ship vials with no or extra preservative.

8.8. *Unpreserved soils*

8.8.1. *At specific client request unpreserved soils packed into glass jars or brass tubes may be accepted and subsampled in the lab. This is the old procedure based on method 5030A and method 8260A. It is no longer included in SW846 and is likely to generate results that are biased low, possibly be more than an order of magnitude.*


8.9. Aqueous samples are stored in glass containers with Teflon lined septa at 4°C +/- 2°C, with minimum headspace.

8.10. Medium level solid extracts are aliquoted into 2 - 5 mL glass vials with Teflon lined caps and stored at 4°C +/- 2°C. The extracts are stored with minimum headspace.


8.11. The maximum holding time is 14 days from sampling until the sample is analyzed. (Samples that are found to be unpreserved still have a 14 day holding time. However they should be analyzed as soon as possible. The lack of preservation should be addressed in the case narrative). Maximum holding time for the EnCore™ sampler (before the sample is added to methanol or sodium bisulfate) is 48 hours.

8.12. A holding blank is stored with the samples. This is analyzed and replaced if any of the trip blanks show any contamination. Otherwise it is replaced every 14 days.


EnCore procedure when low level is not required (field steps in gray)




EnCore procedure when low level is required



Field methanol extraction procedure (field steps in gray)



Field bisulfate preservation procedure (field steps in gray)



9. QUALITY CONTROL

9.1. Initial Demonstration of Capability

- 9.1.1. For the standard analyte list, the initial demonstration described in Section 13 and method detection limit (MDL) studies must be acceptable before analysis of samples may begin. MDLs should be analyzed for low and medium soils and aqueous samples.
- 9.1.2. For non-standard analytes, a MDL study must be performed and calibration curve generated before analyzing any samples, unless lesser requirements are previously agreed to with the client. In any event, the minimum initial demonstration required is analysis of a standard at the reporting limit and a single point calibration.

9.2. Control Limits

In-house historical control limits must be determined for surrogates, matrix spikes, and laboratory control samples (LCS). These limits must be determined at least annually. The recovery limits are mean recovery +/- 3 standard deviations for surrogates, matrix spikes and LCS. Precision limits for matrix spikes / matrix spike duplicates are 0 to mean relative percent difference + 3 standard deviations.

- 9.2.1. All surrogate, LCS, and MS recoveries (except for dilutions) must be entered into QuantIMS (when available) or other database so that accurate historical control limits can be generated. For tests without a separate extraction, surrogates and matrix spikes will be reported for all dilutions.
- 9.2.2. Refer to the QC Program document (QA-003) for further details of control limits.

9.3. Surrogates

Every sample, blank, and QC sample is spiked with surrogates. Surrogate recoveries in samples, blanks, and QC samples must be assessed to ensure that recoveries are within established limits. The compounds included in the surrogate spiking solutions are listed in Table 8. If any surrogates are outside limits, the following corrective actions must take place (except for dilutions):

- Check all calculations for error.
- Ensure that instrument performance is acceptable.
- Recalculate the data and/or reanalyze if either of the above checks reveal a problem.

-
- Reprepare and reanalyze the sample or flag the data as “Estimated Concentration” if neither of the above resolves the problem.

The decision to reanalyze or flag the data should be made in consultation with the client. It is only necessary to reprepare/reanalyze a sample once to demonstrate that poor surrogate recovery is due to matrix effect, unless the analyst believes that the repeated out of control results are not due to matrix effect.

- 9.3.1. If the surrogates are out of control for the sample, matrix spike, and matrix spike duplicate, then matrix effect has been demonstrated for that sample and reparation is not necessary. If the sample is out of control and the MS and/or MSD is in control, then reanalysis or flagging of the data is required.
- 9.3.2. Refer to the STL QC Program document (QA-003) for further details of the corrective actions.

9.4. Method Blanks

- 9.4.1. For each batch of samples, analyze a method blank. The method blank is analyzed after the calibration standards, normally before any samples. For low-level volatiles, the method blank consists of reagent water. For medium-level volatiles, the method blank consists of 25.0 mL of methanol. Surrogates are added and the method blank is carried through the entire analytical procedure. The method blank must not contain any analyte of interest at or above the reporting limit (except common laboratory contaminants, see below) or at or above 5% of the measured concentration of that analyte in the associated samples, whichever is higher.
 - If the analyte is a common laboratory contaminant (methylene chloride, acetone, 2-butanone) the data may be reported with qualifiers if the concentration of the analyte is less than five times the reporting limit. Such action must be taken in consultation with the client.
 - Reanalysis of samples associated with an unacceptable method blank is required when reportable concentrations are determined in the samples.
 - If there is no target analyte greater than the RL in the samples associated with an unacceptable method blank, the data may be reported with qualifiers. Such action should be done in consultation with the client.
- 9.4.2. The method blank must have acceptable surrogate recoveries. If surrogate recoveries are not acceptable, the data must be evaluated to determine if the method blank has served the purpose of demonstrating that the analysis is free of contamination. If surrogate recoveries are low and there are reportable analytes in the associated samples

re-extraction of the blank and affected samples will normally be required. Consultation with the client should take place.

9.4.3. If reanalysis of the batch is not possible due to limited sample volume or other constraints, the method blank is reported, all associated samples are flagged with a "B," and appropriate comments may be made in a narrative to provide further documentation.

9.4.4. Refer to the STL QC Program document (QA-003) for further details of the corrective actions.

9.5. Laboratory Control Samples (LCS)

9.5.1. For each batch of samples, analyze a LCS. The LCS is analyzed after the calibration standard, and normally before any samples. The LCS contains a representative subset of the analytes of interest (See Table 9), and must contain the same analytes as the matrix spike. If any analyte or surrogate is outside established control limits, the system is out of control and corrective action must occur. Corrective action will normally be reparation and reanalysis of the batch.

- If the batch is not re-extracted and reanalyzed, the reasons for accepting the batch must be clearly presented in the project records and the report. (Examples of acceptable reasons for not reanalyzing might be that the matrix spike and matrix spike duplicate are acceptable, and sample surrogate recoveries are good, demonstrating that the problem was confined to the LCS.)
- If re-extraction and reanalysis of the batch is not possible due to limited sample volume or other constraints, the LCS is reported, all associated samples are flagged, and appropriate comments are made in a narrative to provide further documentation.

9.5.2. Refer to the STL QC Program document (QA-003) for further details of the corrective action.

9.5.3. If full analyte spike lists are used at client request, it will be necessary to allow a percentage of the components to be outside control limits as this would be expected statistically. These requirements should be negotiated with the client. Refer to Section 17.5 for Ohio VAP specific analytes.

9.6. Matrix Spikes

9.6.1. For each QC batch, analyze a matrix spike and matrix spike duplicate. Spiking compounds and levels are given in Table 9. Compare the percent recovery and relative percent difference (RPD) to that in the laboratory specific historically generated limits. See Section 17.5 for Ohio VAP specific analytes.

- If any individual recovery or RPD falls outside the acceptable range, corrective action must occur. The initial corrective action will be to check the recovery of that analyte in the Laboratory Control Sample (LCS). Generally, if the recovery of the analyte in the LCS is within limits, then the laboratory operation is in control and analysis may proceed. The reasons for accepting the batch must be documented.
- If the recovery for any component is outside QC limits for both the matrix spike/spike duplicate and the LCS, the laboratory is out of control and corrective action must be taken. Corrective action will normally include reanalysis of the batch.
- If a MS/MSD is not possible due to limited sample, then a LCS duplicate should be analyzed. RPD of the LCS and LCSD are compared to the matrix spike limits.
- The matrix spike/duplicate must be analyzed at the same dilution as the unspiked sample, even if the matrix spike compounds will be diluted out.

9.7. Nonconformance and Corrective Action

9.7.1. Any deviations from QC procedures must be documented as a nonconformance, with applicable cause and corrective action approved by the facility QA Manager.

9.8. Quality Assurance Summaries

Certain clients may require specific project or program QC which may supersede these method requirements. Quality Assurance Summaries should be developed to address these requirements.

9.9. STL QC Program

Further details of QC and corrective action guidelines are presented in the STL QC Program document (QA-003). Refer to this document if in doubt regarding corrective actions.

10. CALIBRATION AND STANDARDIZATION

10.1. Summary

10.1.1. Prior to the analysis of samples and blanks, each GC/MS system must be tuned and calibrated. Hardware tuning is checked through the analysis of the 4-Bromofluorobenzene (BFB) to establish that a given GC/MS system meets the standard mass spectral abundance criteria. The GC/MS system must be calibrated initially at a minimum of five concentrations (analyzed under the same BFB tune), to determine the linearity of the response utilizing target calibration standards. Once the system has been calibrated, the calibration must be verified each twelve hour time period for each GC/MS system.

10.2.1. General

Electron Energy:	70 volts (nominal)
Mass Range:	35–300 AMU
Scan Time:	to give at least 5 scans/peak, but not to exceed 2 second/scan
Injector Temperature:	200–250°C
Source Temperature:	According to manufacturer's specifications
Transfer Line	Temperature: 250–300°C
Purge Flow:	40 mL/minute
Carrier Gas	Flow: 15 mL/minute
Make-up Gas Flow:	25–30 mL/minute

10.2.2. Gas chromatograph suggested temperature program

10.2.2.1. BFB Analysis

Isothermal:	170°C
-------------	-------

10.2.2.2. Sample Analysis

Initial Temperature:	40°C
Initial Hold Time:	4 minutes
Temperature Program:	8°C/minute
Final Temperature:	184°C
Second Temperature	Program: 40°C/minute
Final Temperature:	240°C
Final Hold Time:	2.6 minutes

10.3. Instrument Tuning

10.3.1. Each GC/MS system must be hardware-tuned to meet the abundance criteria listed in Table 10 for a maximum of a 50 ng injection or purging of BFB. Analysis must not begin until these criteria are met. These criteria must be met for each twelve-hour time period. The twelve-hour time period begins at the moment of injection of BFB.

10.4. Initial Calibration

10.4.1. A series of five initial calibration standards is prepared and analyzed for the target compounds and each surrogate compound. Six standards must be used for a quadratic least squares calibration. Suggested calibration levels for a 5 mL purge are: 5, 20, 50, 100, and 200 µg/L. Certain analytes are prepared at higher concentrations due to poor purge performance. Suggested calibration levels for a 25 mL purge are 1, 5, 10, 20, and 40 µg/L. Again, some analytes are prepared at higher levels. Tables 2 and 4 list the calibration levels for each analyte. Other calibration levels and purge volumes may be used depending on the capabilities of the specific instrument. (For example, adequate sensitivity can be obtained on the Agilent 5973 instruments to use a 5 mL purge volume to reach the same reporting limits that once required a 25 mL purge. The calibration levels will still be the same 1, 5, 10, 20, 40µg/L.) However, the same purge volume must be used for calibration and sample analysis, and the low level standard must be at or below the reporting limit.

10.4.2. It may be necessary to analyze more than one set of calibration standards to encompass all of the analytes required for same tests. For example, the Appendix IX list requires the Primary standard (Table 5) and the Appendix IX standard (Table 6). If acceptable analytical performance can be obtained the primary and appendix IX standards may be analyzed together.

10.4.3. Internal standard calibration is used. The internal standards are listed in Table 7. Target compounds should reference the nearest internal standard. Each calibration standard is analyzed and the response factor (RF) for each compound is calculated using the area response of the characteristic ions against the concentration for each compound and internal standard. See equation 1, Section 12, for calculation of response factor.

10.4.4. The % RSD of the calibration check compounds (CCC) must be less than 30%. Refer to Table 12 for the CCCs.

10.4.4.1. If none of the CCCs are required analytes, project specific calibration specifications must be agreed with the client.

10.4.5. The average RF must be calculated for each compound. A system performance check is made prior to using the calibration curve. The five system performance check compounds (SPCC) are checked for a minimum average response factor. Refer to Table 11 for the SPCC compounds and required minimum response factors.

10.4.6. If the average of all the %RSDs in the calibration is $\leq 15\%$, then all analytes may use average response factor for calibration.

10.4.6.1. If the software in use is capable of routinely reporting curve coefficients for data validation purposes, and the necessary calibration reports can be generated, then the analyst should evaluate analytes with %RSD $> 15\%$ for calibration on a curve. If it appears that substantially better accuracy would be obtained using quantitation from a curve then the appropriate curve should be used for quantitation. If Relative Standard Error (RSE) is used to evaluate the curve it must be better than 15%. Otherwise the correlation coefficient (coefficient of determination for non-linear curves) must be ≥ 0.990 .

10.4.6.2. If the average of all the %RSDs in the calibration is $> 15\%$ then calibration on a curve must be used for all analytes with %RSD $> 15\%$. Linear or quadratic curve fits may be used. The analyst should consider instrument maintenance to improve the linearity of response. If Relative Standard Error (RSE) is used to evaluate the curve it must be better than 15%. If the % RSD is $> 15\%$, the analyst may drop the low or high in the ICAL, as long as a minimum of 5 points are maintained and the quantitation range is adjusted accordingly. Otherwise the correlation coefficient, (coefficient of determination, r^2 for non-linear curves) must be ≥ 0.990 . If the correlation coefficient is < 0.990 , then any hit for these compounds must be flagged as estimated.

10.4.6.3. Refer to Section 17.5 for specific Ohio VAP criteria.

10.4.7. Weighting of data points

In a linear or quadratic calibration fit, the points at the lower end of the calibration curve have less weight in determining the curve generated than points at the high concentration end of the curve. However, in environmental analysis, accuracy at the low end of the curve is very important. For this reason it is preferable to increase the weighting of the lower concentration points. $1/\text{Concentration}^2$ weighting (often called $1/X^2$ weighting) will improve accuracy at the low end of the curve and should be used if the data system has this capability.

10.4.8. If time remains in the 12-hour period initiated by the BFB injection before the initial calibration, samples may be analyzed. Otherwise, proceed to continuing calibration.

10.4.9. The calibration standards for the initial 5-point calibration for low level soils that are not preserved in sodium bisulfate (i.e. are preserved by freezing, or not preserved) must be heated to 40°C for purging. Using this calibration curve for water samples is acceptable as long as all calibration, QC, and samples are also heated to 40°C . A separate five point calibration must be prepared for analysis of low level soils that are preserved with sodium bisulfate. Low level soils analysis requires the use of a closed vial autosampler

such as the Varian Archon, O.I. 4552 or Tekmar Precept. Each standard for analysis of sodium bisulfate preserved samples is prepared by spiking the methanolic standard solution through the septum of a VOA vial containing 5 mL of water and 1 g sodium bisulfate. The standards are heated to 40°C for purging. All low-level soil samples, standards, and blanks must also be heated to 40°C for purging. Medium soil extracts should be analyzed using the water (unheated or optionally heated) calibration curve as long as all calibration standards, samples, and QC samples are purged at the same temperature.

10.4.10. Non-standard analytes are sometimes requested. For these analytes, it is acceptable to analyze a single standard at the reporting limit with each continuing calibration rather than a five point initial calibration. If the analyte is detected in any of the samples, a five point initial calibration must be generated and the sample(s) reanalyzed for quantitation. However, if the analyte is not detected, the non-detect may be reported and no further action is necessary.

Note: This procedure is may not be used for Ohio VAP samples.

10.5. Continuing Calibration: The initial calibration must be verified every twelve hours.

10.5.1. Continuing calibration begins with analysis of BFB as described in Section 10.3. If the system tune is acceptable, the continuing calibration standard(s) are analyzed. The level 3 calibration standard is used as the continuing calibration.

10.5.2. The RF data from the standards are compared with the average RF from the initial five-point calibration to determine the percent drift of the CCC compounds. The calculation is given in equation 4, Section 12.3.4.

10.5.3. The % drift of the CCCs must be $\leq 20\%$ for the continuing calibration to be valid. The SPCCs are also monitored. The SPCCs must meet the criteria described in Table 11. In addition, the % drift of all analytes must be $\leq 50\%$ with allowance for up to six target analytes to have % drift $> 50\%$.

10.5.3.1. If none of the CCCs are required analytes, project specific calibration specifications must be agreed with the client.

10.5.3.2. Cyclohexanone, one of the components of the Appendix IX standard, is unstable in the calibration solution, forming 1,1-dimethoxycyclohexane. No calibration criteria are applied to cyclohexanone and quantitation is tentative. Cyclohexanone is included on the Universal Treatment Standard and FO-39 regulatory lists (but not on Appendix IX).

10.5.3.3. Refer to Table 12 for specific Ohio VAP analytes.

-
- 10.5.4. If the CCCs and or the SPCCs do not meet the criteria in Sections 10.5.3 and 10.5.4, the system must be evaluated and corrective action must be taken. The BFB tune and continuing calibration must be acceptable before analysis begins. Extensive corrective action such as a different type of column will require a new initial calibration.
- 10.5.5. Once the above criteria have been met, sample analysis may begin. **Initial calibration average RFs (or the calibration curve) will be used for sample quantitation, not the continuing calibration RFs.** Analysis may proceed until 12 hours from the injection of the BFB have passed. (A sample *desorbed* less than or equal to 12 hours after the BFB is acceptable.)

11. PROCEDURE

11.1. Procedural Variations

- 11.1.1. One time procedural variations are allowed only if deemed necessary in the professional judgment of supervision to accommodate variation in sample matrix, radioactivity, chemistry, sample size, or other parameters. Any variation shall be completely documented using a Nonconformance Memo and approved by a Supervisor or group leader and QA Manager. If contractually required, the client shall be notified. The Nonconformance Memo shall be filed in the project file.
- 11.1.2. Any unauthorized deviations from this procedure must also be documented as a nonconformance, with a cause and corrective action described.

11.2. Preliminary Evaluation

- 11.2.1. Where possible, samples are screened by headspace or GC/MS off-tune analysis to determine the correct aliquot for analysis. Alternatively, an appropriate aliquot can be determined from sample histories.
- 11.2.2. Dilutions should be done just prior to the GC/MS analysis of the sample. Dilutions are made in volumetric flasks or in a Luerlok syringe. Calculate the volume of reagent water required for the dilution. Fill the syringe with reagent water, compress the water to vent any residual air and adjust the water volume to the desired amount. Adjust the plunger to the mark and inject the proper aliquot of sample into the syringe. If the dilution required would use less than 1 μL of sample then serial dilutions must be made in volumetric flasks.
- 11.2.2.1. The diluted concentration is to be estimated to be in the upper half of the calibration range.

11.3. Sample Analysis Procedure

- 11.3.1. All analysis conditions for samples must be the same as for the continuing calibration standards (including purge time and flow, desorb time and temperature, column temperatures, multiplier setting etc.).
- 11.3.2. All samples must be analyzed as part of a batch. The batch is a set of up to 20 samples of the same matrix processed using the same procedures and reagents within the same time period. The batch also must contain a MS/MSD, a LCS, and a method blank.
 - 11.3.2.1. If there is insufficient time in the 12-hour tune period to analyze 20 samples, the batch may be continued into the next tune period. However, if any re-tuning of the instrument is necessary, or if a period of greater than 24 hours from the preceding BFB tune has passed, a new batch must be started. For medium level soils the batch is defined at the sample preparation stage.
 - 11.3.2.2. Laboratory generated QC samples (Blank, LCS, MS/MSD) do not count towards the maximum 20 samples in a batch. Field QC samples are included in the batch count.
 - 11.3.2.3. It is not necessary to reanalyze batch QC with reanalyses of samples. However, any reruns must be as part of a valid batch.

11.4. Water Samples

- 11.4.1. All samples and standard solutions must be at ambient temperature before analysis.
- 11.4.2. Fill a syringe with the sample. If a dilution is necessary it may be made in the syringe if the sample aliquot is $\geq 5 \mu\text{L}$. Check and document the pH of the remaining sample.
- 11.4.3. Add 250 ng of each internal and surrogate standard (10 μL of a 25 $\mu\text{g}/\text{mL}$ solution, refer to Tables 7, 8 and 16). The internal standards and the surrogate standards may be mixed and added as one spiking solution (this results in a 50 $\mu\text{g}/\text{L}$ solution for a 5 mL sample, and a 10 $\mu\text{g}/\text{L}$ solution for a 25 mL sample). Inject the sample into the purging chamber.
 - 11.4.3.1. For TCLP samples use 0.5 mL of TCLP leachate with 4.5 mL reagent water and spike with 10 μL of the 25 $\mu\text{g}/\text{mL}$ TCLP spiking solution. (Note that TCLP reporting limits will be 10 times higher than the corresponding aqueous limits).
- 11.4.4. Purge the sample for eleven minutes (the trap must be below 35°C).
- 11.4.5. After purging is complete, desorb the sample, start the GC temperature program, and begin data acquisition. After desorption, bake the trap for approximately 3-10 minutes

to condition it for the next analysis. When the trap is cool, it is ready for the next sample.

11.4.6. Desorb and bake time and temperature are optimized for the type of trap in use. The same conditions must be used for samples and standards.

11.5. Methanol Extract Soils

11.5.1. Rinse a gas-tight syringe with organic free water. Fill the syringe with the same volume of organic free water as used in the calibrations. Add no more than 2% (v/v) (100 μ L for a 5 mL purge) methanolic extract (from Section 8.5 or 8.6) to the syringe. Add internal standard (if used). Load the sample onto the purge and trap device and analyze as for aqueous samples. If less than 5 μ L of methanolic extract is to be added to the water, dilute the methanolic extract such that a volume greater than 5 μ L will be added to the water in the syringe. Refer to Section 17.8 for Michigan project requirements.

11.6. Liquid wastes that are soluble in methanol and insoluble in water.

11.6.1. Pipet 2 mL of the sample into a tared vial. Use a top-loading balance. Record the weight to the nearest 0.1 gram.

11.6.2. Quickly add 7 mL of methanol, then add 1 mL of surrogate spiking solution to bring the final volume to 10 mL. Cap the vial and shake for 2 minutes to mix thoroughly. For a MS/MSD or LCS, 6 mL of methanol, 1 mL of surrogate solution, and 1 mL of matrix spike solution is used.

11.6.3. Rinse a gas-tight syringe with organic free water. Fill the syringe with the same volume of organic free water as used in the calibrations. Add no more than 2% (v/v) (100 μ L for a 5 mL purge) methanolic extract (from Section 8.5 or 8.6) to the syringe. Add internal standard (if used). Load the sample onto the purge and trap device and analyze as for aqueous samples. If less than 5 μ L of methanolic extract is to be added to the water, dilute the methanolic extract such that a volume greater than 5 μ L will be added to the water in the syringe.

11.7. Aqueous and Low level Soil Sample Analysis (Purge and Trap units that sample directly from the VOA vial)

11.7.1. Units which sample from the VOA vial should be equipped with a module which automatically adds surrogate and internal standard solution to the sample prior to purging the sample.

-
- 11.7.2. If the autosampler uses automatic IS/SS injection, no further preparation of the VOA vial is needed. Otherwise the internal and surrogate standards must be added to the vial. *Note:* Aqueous samples with high amounts of sediment present in the vial may not be suitable for analysis on this instrumentation, or they may need to be analyzed as soils.
- 11.7.3. Soil samples must be quantitated against a curve prepared with standards containing about the same amount of sodium bisulfate as the samples (1 g in 5 mL).
- 11.7.4. Sample remaining in the vial after sampling with one of these mechanisms is no longer valid for further analysis. A fresh VOA vial must be used for further sample analysis.
- 11.7.5. For aqueous samples, check the pH of the sample remaining in the VOA vial after analysis is completed.

11.8. *Low-Level Solids Analysis using discrete autosamplers, Method 8260A, 5030A.*

Note: This technique may seriously underestimate analyte concentration and must not be used except at specific client request for the purpose of comparability with previous data. It is no longer part of SW-846.

This method is based on purging a heated soil/sediment sample mixed with reagent water containing the surrogates and internal standards. Analyze all reagent blanks and standards under the same conditions as the samples (e.g., heated). The calibration curve is also heated during analysis. Purge temperature is 40°C.

- 11.8.1. *Do not discard any supernatant liquids. Mix the contents of the container with a narrow metal spatula.*
- 11.8.2. *Weigh out 5 g (or other appropriate aliquot) of sample into a disposable culture tube or other purge vessel. Record the weight to the nearest 0.1 g. If method sensitivity is demonstrated, a smaller aliquot may be used. Do not use aliquots less than 1.0 g. If the sample is contaminated with analytes such that a purge amount less than 1.0 g is appropriate, use the medium level method. For the medium level method, add 4g soil to 10 mL methanol containing the surrogates, mix for two minutes, allow to settle then remove a portion of the methanol and store in a clean Teflon capped vial at 4°C until analysis. Analyze as described in section 11.5.*
- 11.8.3. *Connect the purge vessel to the purge and trap device.*
- 11.8.4. *Rinse a 5 mL gas-tight syringe with organic free water, and fill. Compress to 5 mL. Add surrogate/internal standard (and matrix spike solutions if required.). Add directly to the sample from 11.5.2.*

11.8.5. The above steps should be performed rapidly and without interruption to avoid loss of volatile organics.

11.8.6. Add the heater jacket or other heating device and start the purge and trap unit.

11.8.7. Soil samples that have low IS recovery when analyzed (<50%) should be reanalyzed once to confirm matrix effect.

11.9. Medium-Level Soil/Sediment and Waste Samples

11.9.5. Sediments/soils and waste that are insoluble in methanol.

11.9.5.1. Sediments/soils and waste that are insoluble in methanol.

11.9.5.1.1. Gently mix the contents of the sample container with a narrow metal or wood spatula. Weigh 4 g (wet weight) into a tared vial. Use a top-loading balance. Record the weight to 0.1 gram. Do not discard any supernatant liquids.

11.9.5.1.2. Quickly add 9 mL of methanol, and 1 mL of surrogate spiking solution to bring the final volume of methanol to 10 mL. For an LCS or MS/MSD sample add 8 mL of methanol, 1 mL of surrogate spike solution, and 1 mL of matrix spike solution. Cap the vial and vortex to mix thoroughly.

NOTE: Sections 11.9.5.1.1 and 11.9.5.1.2 must be performed rapidly and without interruption to avoid the loss of volatile organics.

11.10. Initial review and corrective actions

11.10.1. If the retention time for any internal standard in the continuing calibration changes by more than 0.5 minutes from the mid-level initial calibration standard, the chromatographic system must be inspected for malfunctions and corrected. Reanalysis of samples analyzed while the system was malfunctioning is required.

11.10.2. If the internal standard response in the continuing calibration is more than 200% or less than 50% of the response in the mid-level of the initial calibration standard, the chromatographic system must be inspected for malfunctions and corrected. Reanalysis of samples analyzed while the system was malfunctioning is required.

11.10.2.1. Any samples that do not meet the internal standard criteria for the continuing calibration must be evaluated for validity. If the change in sensitivity is a matrix effect confined to an individual sample reanalysis is not necessary. If the change

in sensitivity is due to instrumental problems all affected samples must be reanalyzed after the problem is corrected.

11.10.3. The surrogate standard recoveries are evaluated to ensure that they are within limits. Corrective action for surrogates out of control will normally be to reanalyze the affected samples. However, if the surrogate standard response is out high and there are no target analytes or tentatively identified compounds, reanalysis may not be necessary. Out of control surrogate standard response may be a matrix effect. It is only necessary to reanalyze a sample once to demonstrate matrix effect, but reanalysis at a dilution should be considered.

11.11. Dilutions

If the response for any compound exceeds the working range of the GC/MS system, a dilution of the extract is prepared and analyzed. An appropriate dilution should be in the upper half of the calibration range. Samples may be screened to determine the appropriate dilution for the initial run. If the initial diluted run has no hits or hits below 20% of the calibration range and the matrix allows for analysis at a lesser dilution, then the sample must be reanalyzed at a dilution targeted to bring the largest hit above 50% of the calibration range.

11.10.1. Guidance for Dilutions Due to Matrix

If the sample is initially run at a dilution and the baseline rise is less than half the height of the internal standards, or if individual non target peaks are less than twice the height of the internal standards, then the sample should be reanalyzed at a more concentrated dilution. This requirement is approximate and subject to analyst judgement.

11.10.2. Reporting Dilutions

The most concentrated dilution with no target compounds above the calibration range will be reported. Other dilutions will only be reported at client request.

12. DATA ANALYSIS AND CALCULATIONS

12.1. Qualitative identification

An analyte is identified by retention time and by comparison of the sample mass spectrum with the mass spectrum of a standard of the suspected compound (standard reference spectrum). Mass spectra for standard reference may be obtained on the user's GC/MS by analysis of the calibration standards or from the NIST Library. Two criteria must be satisfied to verify identification: (1) elution of sample component at the same GC retention time as the standard component; and (2) correspondence of the sample

component and the standard component characteristic ions. (Note: Care must be taken to ensure that spectral distortion due to co-elution is evaluated.)

- The sample component retention time must compare to within ± 0.2 min. of the retention time of the standard component. For reference, the standard must be run within the same twelve hours as the sample.
- All ions present in the standard mass spectra at a relative intensity greater than 10% (most abundant ion in the spectrum equals 100%) should be present in the sample spectrum.
- The relative intensities of ions should agree to within $\pm 30\%$ between the standard and sample spectra. (Example: For an ion with an abundance of 50% in the standard spectra, the corresponding sample abundance must be between 20 and 80 percent.)

12.1.1. If a compound cannot be verified by all the above criteria, but in the technical judgment of the analyst, the identification is correct, then the analyst shall report that identification and proceed with quantitation.

12.2. Tentatively Identified Compounds (TICs)

12.2.1. If the client requests components not associated with the calibration standards, a search of the NIST library may be made for the purpose of tentative identification. Guidelines are:

- 12.2.1.1. Relative intensities of major ions in the reference spectrum (ions $> 10\%$ of the most abundant ion) should be present in the sample spectrum.
- 12.2.1.2. The relative intensities of the major ions should agree to within 20%. (Example: If an ion shows an abundance of 50% in the standard spectrum, the corresponding sample ion abundance must be between 30% and 70%).
- 12.2.1.3. Molecular ions present in the reference spectrum should be present in the sample spectrum.
- 12.2.1.4. Ions present in the sample spectrum but not in the reference spectrum should be reviewed for possible background contamination or presence of coeluting compounds.
- 12.2.1.5. Ions present in the reference spectrum but not in the sample spectrum should be reviewed for possible subtraction from the spectrum because of background contamination or coeluting peaks. (Data system reduction programs can sometimes create these discrepancies.)

12.2.1.6. Computer-generated library search routines should not use normalization routines that would misrepresent the library or unknown spectra when compared to each other. Only after visual inspection of the sample with the nearest library searches should the analyst assign a tentative identification.

12.3. Calculations.

12.3.1. Response factor (RF):

Equation 1

$$RF = \frac{A_x C_{is}}{A_{is} C_x}$$

Where:

A_x = Area of the characteristic ion for the compound to be measured

A_{is} = Area of the characteristic ion for the specific internal standard

C_{is} = Concentration of the specific internal standard, ng

C_x = Concentration of the compound being measured, ng

12.3.2. Standard deviation (SD):

Equation 2

$$SD = \sqrt{\sum_{i=1}^N \frac{(X_i - X)^2}{N - 1}}$$

X_i = Value of X at i through N

N = Number of points

X = Average value of X_i

12.3.3. Percent relative standard deviation (%RSD):

Equation 3

$$\%RSD = \frac{\text{Standard Deviation}}{RF_i} \times 100$$

RF_i = Mean of RF values in the curve

12.3.4. Percent drift between the initial calibration and the continuing calibration:

Equation 4

$$\% \text{ Drift} = \frac{C_{\text{expected}} - C_{\text{found}}}{C_{\text{expected}}} \times 100$$

Where

C_{expected} = Known concentration in standard

C_{found} = Measured concentration using selected quantitation method

12.3.5. Target compound and surrogate concentrations:

Concentrations in the sample may be determined from linear or second order (quadratic) curve fitted to the initial calibration points, or from the average response factor of the initial calibration points. Average response factor may only be used when the % RSD of the response factors in the initial calibration is $\leq 15\%$.

12.3.5.1. Calculation of concentration using Average Response Factors

Equation 5

$$\text{Concentration } \mu\text{g} / \text{L} = \frac{x}{RF}$$

12.3.5.2. Calculation of concentration using Linear fit

Equation 6

$$\text{Concentration } \mu\text{g} / \text{L} = A + Bx$$

12.3.5.3. Calculation of concentration using Quadratic fit

Equation 7

$$\text{Concentration } \mu\text{g} / \text{L} = A + Bx + Cx^2$$

x is defined in equations 8, 9 and 10

A is a constant defined by the intercept

B is the slope of the curve

C is the curvature

12.3.5.4. Calculation of x for Water and water-miscible waste:

Equation 8

$$x = \frac{(A_x)(I_s)(D_f)}{(A_{is})(V_o)}$$

Where:

A_x = Area of characteristic ion for the compound being measured (secondary ion quantitation is allowed only when there are sample interferences with the primary ion)

A_{is} = Area of the characteristic ion for the internal standard

I_s = Amount of internal standard added in ng

$$\text{Dilution Factor} = D_f = \frac{\text{Total volume purged (mL)}}{\text{Volume of original sample used (mL)}}$$

V_o = Volume of water purged, mL

12.3.5.5. Calculation of x for Medium level soils:

Equation 9

$$x = \frac{(A_x)(I_s)(V_t)(1000)(D)}{(A_{is})(V_a)(W_s)(D)}$$

Where:

A_x , I_s , D_f , A_{is} , same as for water.

V_t = Volume of total extract, mL (Typically 25 mL)

V_a = Volume of extract added for purging, μL

W_s = Weight of sample extracted, g

$$D = \frac{100 - \% \text{moisture}}{100}$$

12.3.5.6. Calculation of x for Low level soils:

Equation 10

$$x = \frac{(A_x)(I_s)}{(A_{is})(W_s)(D)}$$

Where:

A_x , I_s , A_{is} , same as for water.

D is as for medium level soils

W_s = Weight of sample added to the purge vessel, g

12.3.5.7. Calculation of TICs: The calculation of TICs (tentatively identified compounds) is identical to the above calculations with the following exceptions:

A_x = Area in the total ion chromatogram for the compound being measured

A_{is} = Area of the total ion chromatogram for the nearest internal standard without interference

$RF = 1$

In other words, the concentration is equal to x as defined in equations 8, 9 and 10.

12.3.6. MS/MSD Recovery

Equation 11

$$\text{Matrix Spike Recovery, \%} = \frac{SSR - SR}{SA} \times 100$$

SSR = Spike sample result

SR = Sample result

SA = Spike added

12.3.7. Relative % Difference calculation for the MS/MSD

Equation 12

$$RPD = \frac{|MSR - MSDR|}{\frac{1}{2}(MSR + MSDR)} \times 100$$

Where:

RPD = Relative percent difference

MSR = Matrix spike result

MSDR = Matrix spike duplicate result

13. METHOD PERFORMANCE

13.1. Method Detection Limit

13.2. Generally, each laboratory must generate a valid method detection limit for each analyte of interest. The MDL must be below the reporting limit for each analyte. The procedure for determination of the method detection limit is given in 40 CFR Part 136, Appendix B, and further defined in QA Policy #: QA-005. When non-standard compounds are analyzed at client request, lesser requirements are possible with client agreement. At a minimum, a standard at the reporting limit must be analyzed to demonstrate the capability of the method.

13.3. Initial Demonstration

13.4. Each laboratory must make a one time initial demonstration of capability for each individual method. Demonstration of capability for both soil and water matrices is required. This requires analysis of QC check samples containing all of the standard analytes for the method. For some tests it may be necessary to use more than one QC check mix to cover all analytes of interest. The QC check sample is made up at 20 µg/L. (Some compounds will be at higher levels, refer to the calibration standard levels for guidance.)

13.4.1. Four aliquots of the QC check sample are analyzed using the same procedures used to analyze samples, including sample preparation.

13.4.2. Calculate the average recovery and standard deviation of the recovery for each analyte of interest. The %RSD should be $\leq 15\%$ for each analyte, and the % recovery should be within 80-120%.

13.4.3. If any analyte does not meet the acceptance criteria, check the acceptance limits in the reference methods (Table 6 of method 8240B, paragraph 8.3.5 of method 8260A). If the recovery or precision is outside the limits in the reference methods, the test must be repeated. Only those analytes that did not meet criteria in the first test need to be evaluated. Repeated failure for any analyte indicates the need for the laboratory to evaluate the analytical procedure and take corrective action.

13.4.4. Training Qualification

13.4.4.1. The group/team leader has the responsibility to ensure that this procedure is performed by an analyst who has been properly trained in its use and has the required experience.

14. POLLUTION PREVENTION

- 14.1. This method does not contain any specific modifications that serve to minimize or prevent pollution.

15. WASTE MANAGEMENT

- 15.1. Waste generated in the procedure must be segregated and disposed according to the facility hazardous waste procedures. The Health and Safety Director should be contacted if additional information is required.

16. REFERENCES

- 16.1. SW846, *Test Methods for Evaluating Solid Waste*, Third Edition, Gas Chromatography/Mass Spectrometry for Volatile Organics, Method 8260B, Update III, December 1996
- 16.2. SW846, *Test Methods for Evaluating Solid Waste*, Third Edition, Gas Chromatography/Mass Spectrometry for Volatile Organics, Method 8260A, Update II, September 1994.

17. MISCELLANEOUS

- 17.1. Modifications from the reference method
- 17.1.1. Ion 119 is used as the quantitation ion for chlorobenzene-d5 for 25 mL purge tests.
- 17.1.2. A retention time window of 0.2 minutes is used for all components, since some data systems do not have the capability of using the relative retention time units specified in the reference method.
- 17.1.3. The quantitation and qualifier ions for some compounds have been changed from those recommended in SW-846 in order to improve the reliability of qualitative identification.
- 17.1.4. Method 8260A recommends that the purge vessel is run through an additional purge cycle after 25 mL sample analysis to remove carryover. Instead, purge vessels are oven baked between analyses or disposable vessels are used one time only.
- 17.1.5. SW-846 recommends that a curve be used for any analytes with %RSD of the response factors > 15%. However, some industry standard data systems and forms generation software cannot report this data with the necessary information for data validation. In addition most software available does not allow weighting of the curve.

Unweighted curves may exhibit serious errors in quantitation at the low end, resulting in possible false positives or false negatives. Therefore, this SOP allows use of average response factors if the average %RSD for all compounds is $\leq 15\%$.

17.2. Modifications from previous revision

This SOP has been substantially revised to reflect the changes included in Update III to SW-846. Directions for method 524.2 and method 624 have also been added.

17.3. Facility specific SOPs

Each facility shall attach a list of facility-specific SOPs or approved attachments (if applicable) which are required to implement this SOP or which are used in conjunction with this SOP. If no facility specific SOPs or amendments are to be attached, a statement must be attached specifying that there are none.

17.4. Flow diagrams

17.4.1. Initial Demonstration and MDL

17.5. The following are protocols that must be followed when analyzing OhioVAP samples:

- Sections 9.5 and 9.6: n-Hexane must be spiked and reported for both the LCS and MS/MSD.
- Sections 10.4.6: All analytes must have a %RSDs $\leq 15\%$. Corrective action must be completed for any compounds failing the $<15\%$ requirement.
- Section 11.1 and 17.1.5 (Method deviations) are not to be performed.
- Section 11.9.2: For OhioVAP projects, the laboratory will reanalyze any sample where the internal standard fails and there is no evidence of matrix interference.

17.6. The following are protocols that must be followed when analyzing BP Oil – Lima Refinery RFI work plan.

- Section 8.1 STL will continue to follow the 14 day holding time specified in the Corporate SOP.
- Delete for this project Section 8.3 At specific client request, unpreserved soil samples may be accepted.
- Delete for this project Section 8.8.1 At specific client request unpreserved soils packed into glass jars or brass tubes may be accepted and subsampled in the lab. This is the old

procedure based on method 5030A. It is no longer included and is likely to generate results that are biased low, possibly by more than an order of magnitude.

- Modify Section 8.5.8 For the purpose of this project, the soil/methanol mixture may be stored for two days prior to analysis.
- Modify Section 8.6.12 For the purpose of this project, the soil/methanol mixture may be stored for two days prior to analysis.
- Modify (per discussion with Region V representative) to Section 10.4.6.2 Compounds with %RSD >15% are to be calibrated using an alternate calibration technique (e.g. linear or quadratic calibration curve). For poor responders, the alternate calibration technique requirements may not be met either. This sentence is added for those cases. If the correlation coefficient is < 0.990, then any hit for these compounds must be flagged as estimated.
- Modify Section 10.4.2 It is necessary to analyze the Appendix IX standard separately from the primary standard due to the presence of xylene solvent in the Appendix IX standard. Alternatively, STL will purchase the Appendix IX standard in a solvent other than xylene.
- Modify Section 10.4.9 For this project, this section will be modified to comply with the requirement of adding methanol to the calibration standards so that those standards contain the same amount of methanol as the diluted soil extracts.
- Modify Table 6
- For the project specific SOP, acetonitrile will be removed from table 6, page 49 and appended onto table 5, page 48. Acetonitrile will be calibrated as part of the STL primary standard, using a separate acetonitrile standard. This will ensure that the calibration curve for acetonitrile will be done free from any interference from allyl chloride.

17.7. The following are protocols that must be followed when analyzing South Carolina Projects only.

- **Delete** from Section 10.4.7 In a linear *or quadratic* calibration fit, the points at the lower end of the calibration curve have less weight in determining the curve generated than points at the high concentration end of the curve.
- **Delete** from Section 12.3.5 Concentrations in the sample may be determined from linear *or second order (quadratic)* curve fitted to the initial calibration points, or from the average response factor of the initial calibration points.
- **Delete** from Section 12.3.5.1

- Calculation of concentration using Quadratic fit

Equation 13

$$\text{Concentration } \mu\text{g} / \text{L} = A + Bx + Cx^2$$

x is defined in equations 8, 9 and 10

A is a constant defined by the intercept

B is the slope of the curve

C is the curvature

- **Change** Section 9.3 The compounds included in the surrogate spiking solutions are listed in Tables 8 and 9.

17.8. The following are protocols that must be followed to achieve the lower reporting limits required when analyzing Michigan projects.

17.8.1. Modify Section 8.5.4 and 8.6.8 (add 5 uL of 2500 ug/mL surrogate solution for a nominal 25 g sample).

17.8.2. Modify Section 8.5.5 and 8.6.9 (add 100 uL of 50 ug/mL spike solution for a nominal 25 g sample).

17.8.3. Modify Section 8.5.6 and 8.6.10 (add 100 uL of 50 ug/mL spike solution for a nominal 25g sample).

17.8.4. Michigan reporting limits for methanol preserved soils are achieved by injecting 100 uL of the methanol extract in a 5 mL purge. The instrument is calibrated using the recommended calibration levels in water of 1 ug/L, 2 ug/L (if a quadratic calibration is to be used), 5 ug/L, 10 ug/L, 20 ug/L and 40 ug/L. Some analytes are prepared at higher concentrations.

17.8.5. Samples for Michigan projects frequently require calibration for 2-Methylnaphthalene. Recommended calibration levels for this compounds are 2 ug/L, 10 ug/L, 20 ug/L, 40 ug/L and 80 ug/L.





Table 1
 STL Primary Standard and Reporting Limits

Compound	CAS Number	Reporting Limits ¹				
		5 mL Water µg/L	25 mL ³ water µg/L	Low soil µg/kg	8260B/ 5035 Soil ug/kg	8260A 5030A Med Level Soil µg/kg
Dichlorodifluoromethane	75-71-8	10	2	10	250	1200
Chloromethane	74-87-3	10	2	10	250	1200
Bromomethane	74-83-9	10	2	10	250	1200
Vinyl chloride	75-01-4	10	2	10	250	1200
Chloroethane	75-00-3	10	2	10	250	1200
Trichlorofluoromethane	75-69-4	10	2	10	250	1200
Acrolein	107-02-8	100	20	100	5000	12000
Acetone	67-64-1	20	10	20	1000	2500
Trichlorotrifluoroethane	76-13-1	5	1	5	250	620
Iodomethane	74-88-4	5	1	5	250	620
Carbon disulfide	75-15-0	5	1	5	250	620
Methylene chloride	75-09-2	5	1	5	250	620
tert-Butyl alcohol	75-65-0	200	50	200	10,000	25000
1,1-Dichloroethene	75-35-4	5	1	5	250	620
1,1-Dichloroethane	75-34-3	5	1	5	250	620
trans-1,2-Dichloroethene	156-60-5	5.0	1.0	5.0	250	310
Acrylonitrile	107-13-1	100	20	100	5000	12000
Methyl tert-butyl ether (MTBE)	1634-04-4	20	5	20	1000	2500
Hexane	110-54-3	5	1	5	250	620
cis-1,2-Dichloroethene	156-59-2	5.0	1.0	5.0	250	310
1,2-Dichloroethene (Total)	540-59-0	5	1	5	250	620
Tetrahydrofuran	109-99-9	20	5	20	1000	2500
Chloroform	67-66-3	5	1	5	250	620
1,2-Dichloroethane	107-06-2	5	1	5	250	620
Dibromomethane	74-95-3	5	1	5	250	620
2-Butanone	78-93-3	20	5	20	1000	2500
1,4-Dioxane	123-91-1	500	200	500	25000	62000
1,1,1-Trichloroethane	71-55-6	5	1	5	250	620
Carbon tetrachloride	56-23-5	5	1	5	250	620
Bromodichloromethane	75-27-4	5	1	5	250	620
1,2-Dichloropropane	78-87-5	5	1	5	250	620
cis-1,3-Dichloropropene	10061-01-5	5	1	5	250	620

Table 1
 STL Primary Standard and Reporting Limits

Compound	CAS Number	Reporting Limits ¹				
		5 mL Water µg/L	25 mL ³ water µg/L	Low soil µg/kg	8260B/ 5035 Soil ug/kg	8260A 5030A Med Level Soil µg/kg
Trichloroethene	79-01-6	5	1	5	250	620
Dibromochloromethane	124-48-1	5	1	5	250	620
1,2-Dibromoethane	106-93-4	5	1	5	250	620
1,2,3-Trichloropropane	96-18-4	5	1	5	250	620
1,1,2-Trichloroethane	79-00-5	5	1	5	250	620
Benzene	71-43-2	5	1	5	250	620
Ethylmethacrylate	97-63-2	5	1	5	250	620
trans-1,3-Dichloropropene	10061-02-6	5	1	5	250	620
Bromoform	75-25-2	5	1	5	250	620
4-Methyl-2-pentanone	108-10-1	20	5	20	1000	2500
2-Hexanone	591-78-6	20	5	20	1000	2500
Tetrachloroethene	127-18-4	5	1	5	250	620
Toluene	108-88-3	5	1	5	250	620
1,1,2,2-Tetrachloroethane	79-34-5	5	1	5	250	620
2-Chloroethyl vinyl ether	110-75-8	N/A ²	N/A	50	1000	6200
Vinyl acetate	108-05-4	10	2	10	500	1200
Chlorobenzene	108-90-7	5	1	5	250	620
Ethylbenzene	100-41-4	5	1	5	250	620
Styrene	100-42-5	5	1	5	250	620
t-1,4-Dichloro-2-butene	110-57-6	5	1	5	250	620
m and p Xylenes		5.0	0.5	2.5	125	310
o-xylene	95-47-6	5.0	0.5	2.5	125	310
Total xylenes	1330-20-7	5	1	5	250	620
1,3-Dichlorobenzene	541-73-1	5	1	5	250	620
1,4-Dichlorobenzene	106-46-7	5	1	5	250	620
1,2-Dichlorobenzene	95-50-1	5	1	5	250	620
2,2-Dichloropropane	590-20-7	5	1	5	250	
Bromochloromethane	74-97-5	5	1	5	250	
1,1-Dichloropropene	563-58-6	5	1	5	250	
Bromodichloromethane	75-27-4	5	1	5	250	
1,2-Dichloropropane	78-87-5	5	1	5	250	
1,3-Dichloropropane	142-28-9	5	1	5	250	

Table 1

STL Primary Standard and Reporting Limits

Compound	CAS Number	Reporting Limits ¹				
		5 mL Water µg/L	25 mL ³ water µg/L	Low soil µg/kg	8260B/ 5035 Soil ug/kg	8260A 5030A Med Level Soil µg/kg
Isopropylbenzene	98-82-8	5	1	5	250	
Bromobenzene	108-86-1	5	1	5	250	
n-Propylbenzene	103-65-1	5	1	5	250	
2-Chlorotoluene	95-49-8	5	1	5	250	
4-Chlorotoluene	106-43-4	5	1	5	250	
1,3,5-Trimethylbenzene	108-67-8	5	1	5	250	
tert-Butylbenzene	98-06-6	5	1	5	250	
1,2,4-Trimethylbenzene	95-63-6	5	1	5	250	
sec-butylbenzene	135-98-8	5	1	5	250	
4-Isopropyltoluene	99-87-6	5	1	5	250	
n-Butylbenzene	104-51-8	5	1	5	250	
1,2,4-Trichlorobenzene	120-82-1	5	1	5	250	
Napthalene	91-20-3	5	1	5	250	
Hexachlorobutadiene	87-68-3	5	1	5	250	
1,2,3-Trichlorobenzene	87-61-6	5	1	5	250	
Acetonitrile	75-05-8	100	20	100	5000	

¹ Reporting limits listed for soil/sediment are based on wet weight. The reporting limits calculated by the laboratory for soil/sediment, calculated on dry weight basis, will be higher.

² 2-Chloroethyl vinyl ether cannot be reliably recovered from acid preserved samples

³ Optionally, 5 mL purge volume if adequate sensitivity is obtained.

Table 2

STL Primary Standard Calibration Levels, 5 mL purge¹

Compound	Calibration Level ug/L				
	Level 1	Level 2	Level 3	Level 4	Level 5
1,2-Dichloroethane-d4 (Surrogate)	5	20	50	100	200
Toluene-d8 (Surrogate)	5	20	50	100	200
4-Bromofluorobenzene (Surrogate)	5	20	50	100	200
Dichlorodifluoromethane	5	20	50	100	200
Chloromethane	5	20	50	100	200
Bromomethane	5	20	50	100	200
Vinyl chloride	5	20	50	100	200
Chloroethane	5	20	50	100	200
Trichlorofluoromethane	5	20	50	100	200
Acrolein	50	200	500	1000	2000
Acetone	5	20	50	100	200
Trichlorotrifluoroethane	5	20	50	100	200
Iodomethane	5	20	50	100	200
Carbon disulfide	5	20	50	100	200
Methylene chloride	5	20	50	100	200
tert-Butyl alcohol	100	400	1,000	2,000	4,000
1,1-Dichloroethene	5	20	50	100	200
1,1-Dichloroethane	5	20	50	100	200
trans-1,2-Dichloroethene	5	20	50	100	200
Acrylonitrile	50	200	500	1,000	2,000
Methyl tert-butyl ether (MTBE)	5	20	50	100	200
Hexane	5	20	50	100	200
cis-1,2-Dichloroethene	5	20	50	100	200
Tetrahydrofuran	5	20	50	100	200
Chloroform	5	20	50	100	200
1,2-Dichloroethane	5	20	50	100	200
Dibromomethane	5	20	50	100	200
2-Butanone	5	20	50	100	200
1,4-Dioxane	250	1000	2,500	5,000	10,000
1,1,1-Trichloroethane	5	20	50	100	200
Carbon tetrachloride	5	20	50	100	200
Bromodichloromethane	5	20	50	100	200
1,2-Dichloropropane	5	20	50	100	200
cis-1,3-Dichloropropene	5	20	50	100	200
Trichloroethene	5	20	50	100	200
Dibromochloromethane	5	20	50	100	200

Table 2
 STL Primary Standard Calibration Levels, 5 mL purge¹

Compound	Calibration Level ug/L				
	Level 1	Level 2	Level 3	Level 4	Level 5
1,2-Dibromoethane	5	20	50	100	200
1,2,3-Trichloropropane	5	20	50	100	200
Acetonitrile	50	200	500	1000	2000
1,1,2-Trichloroethane	5	20	50	100	200
Benzene	5	20	50	100	200
Ethylmethacrylate	5	20	50	100	200
trans-1,3-Dichloropropene	5	20	50	100	200
Bromoform	5	20	50	100	200
4-Methyl-2-pentanone	5	20	50	100	200
2-Hexanone	5	20	50	100	200
Tetrachloroethene	5	20	50	100	200
Toluene	5	20	50	100	200
1,1,2,2-Tetrachloroethane	5	20	50	100	200
2-Chloroethyl vinyl ether	10	40	100	200	400
Vinyl acetate	5	20	50	100	200
Chlorobenzene	5	20	50	100	200
Ethylbenzene	5	20	50	100	200
Styrene	5	20	50	100	200
t-1,4-Dichloro-2-butene	5	20	50	100	200
m and p Xylenes	10	40	100	200	400
o-xylene	5	20	50	100	200
1,3-Dichlorobenzene	5	20	50	100	200
1,4-Dichlorobenzene	5	20	50	100	200
1,2-Dichlorobenzene	5	20	50	100	200
2,2-Dichloropropane	5	20	50	100	200
Bromochloromethane	5	20	50	100	200
1,1-Dichloropropene	5	20	50	100	200
Bromodichloromethane	5	20	50	100	200
1,2-Dichloropropane	5	20	50	100	200
1,3-Dichloropropane	5	20	50	100	200
Isopropylbenzene	5	20	50	100	200
Bromobenzene	5	20	50	100	200
n-Propylbenzene	5	20	50	100	200
2-Chlorotoluene	5	20	50	100	200
4-Chlorotoluene	5	20	50	100	200
1,3,5-Trimethylbenzene	5	20	50	100	200

Table 2

STL Primary Standard Calibration Levels, 5 mL purge¹

Compound	Calibration Level ug/L				
	Level 1	Level 2	Level 3	Level 4	Level 5
tert-Butylbenzene	5	20	50	100	200
1,2,4-Trimethylbenzene	5	20	50	100	200
sec-butylbenzene	5	20	50	100	200
4-Isopropyltoluene	5	20	50	100	200
n-Butylbenzene	5	20	50	100	200
1,2,4-Trichlorobenzene	5	20	50	100	200
Napthalene	5	20	50	100	200
Hexachlorobutadiene	5	20	50	100	200
1,2,3-Trichlorobenzene	5	20	50	100	200

¹ Levels for 25 mL purge are 5 times lower in all cases

Table 2A
 STL Primary Standard Calibration Levels, Low Level ¹

Compound	Calibration Level ug/L				
	Level 1	Level 2	Level 3	Level 4	Level 5
Dibromofluoromethane (Surrogate)	1	5	10	20	40
1,2-Dichloroethane-d4 (Surrogate)	1	5	10	20	40
Toluene-d8 (Surrogate)	1	5	10	20	40
Bromofluorobenzene (Surrogate)	1	5	10	20	40
Dichlorodifluoromethane	1	5	10	20	40
Chloromethane	1	5	10	20	40
Vinyl Chloride	1	5	10	20	40
Bromomethane	1	5	10	20	40
Chloroethane	1	5	10	20	40
Trichlorofluoromethane	1	5	10	20	40
Acrolein	10	50	100	200	400
Acetone	2	10	20	40	80
1,1-Dichloroethene	1	5	10	20	40
Trichlorotrifluoroethane	1	5	10	20	40
Iodomethane	1	5	10	20	40
Carbon Disulfide	1	5	10	20	40
Methylene Chloride	1	5	10	20	40
Acetonitrile	10	50	100	200	400
Acrylonitrile	10	50	100	200	400
Methyl tert-butyl ether	1	5	10	20	40
trans-1,2-Dichloroethene	1	5	10	20	40
Hexane	1	5	10	20	40
Vinyl acetate	1	5	10	20	40
1,1-Dichloroethane	1	5	10	20	40
tert-Butyl Alcohol	20	100	200	400	800
2-Butanone	2	10	20	40	80
cis-1,2-dichloroethene	1	5	10	20	40
2,2-Dichloropropane	1	5	10	20	40
Bromochloromethane	1	5	10	20	40
Chloroform	1	5	10	20	40
Tetrahydrofuran	1	5	10	20	40
1,1,1-Trichloroethane	1	5	10	20	40
1,1-Dichloropropene	1	5	10	20	40
Carbon Tetrachloride	1	5	10	20	40
1,2-Dichloroethane	1	5	10	20	40
Benzene	1	5	10	20	40
Trichloroethene	1	5	10	20	40
1,2-Dichloropropane	1	5	10	20	40

Table 2A

STL Primary Standard Calibration Levels, Low Level ¹

Compound	Calibration Level ug/L				
	Level 1	Level 2	Level 3	Level 4	Level 5
1,4-Dioxane	50	250	500	1000	2000
Dibromomethane	1	5	10	20	40
Bromodichloromethane	1	5	10	20	40
2-Chloroethyl vinyl ether	2	10	20	40	80
cis-1,3-Dichloropropene	1	5	10	20	40
4-Methyl-2-pentanone	2	10	20	40	80
Toluene	1	5	10	20	40
trans-1,3-Dichloropropene	1	5	10	20	40
Ethyl Methacrylate	1	5	10	20	40
1,1,2-Trichloroethane	1	5	10	20	40
1,3-Dichloropropane	1	5	10	20	40
Tetrachloroethene	1	5	10	20	40
2-Hexanone	2	10	20	40	80
Dibromochloromethane	1	5	10	20	40
1,2-Dibromoethane	1	5	10	20	40
Chlorobenzene	1	5	10	20	40
1,1,1,2-Tetrachloroethane	1	5	10	20	40
Ethylbenzene	1	5	10	20	40
m + p-Xylene	2	10	20	40	80
Xylene-o	1	5	10	20	40
Styrene	1	5	10	20	40
Bromoform	1	5	10	20	40
Isopropylbenzene	1	5	10	20	40
1,1,2,2-Tetrachloroethane	1	5	10	20	40
1,4-Dichloro-2-butene	1	5	10	20	40
1,2,3-Trichloropropane	1	5	10	20	40
Bromobenzene	1	5	10	20	40
n-Propylbenzene	1	5	10	20	40
2-Chlorotoluene	1	5	10	20	40
1,3,5-Trimethylbenzene	1	5	10	20	40
4-Chlorotoluene	1	5	10	20	40
tert-Butylbenzene	1	5	10	20	40
1,2,4-Trimethylbenzene	1	5	10	20	40
sec-Butylbenzene	1	5	10	20	40
4-Isopropyltoluene	1	5	10	20	40
1,3-Dichlorobenzene	1	5	10	20	40
1,4-Dichlorobenzene	1	5	10	20	40
n-Butylbenzene	1	5	10	20	40

Table 2A

STL Primary Standard Calibration Levels, Low Level ¹

Compound	Calibration Level ug/L				
	Level 1	Level 2	Level 3	Level 4	Level 5
1,2-Dichlorobenzene	1	5	10	20	40
1,2-Dibromo-3-chloropropane	1	5	10	20	40
1,2,4-Trichlorobenzene	1	5	10	20	40
Hexachlorobutadiene	1	5	10	20	40
Naphthalene	1	5	10	20	40
1,2,3-Trichlorobenzene	1	5	10	20	40
Cyclohexane	1	5	10	20	40
Methyl Acetate	2	10	20	40	80
Methylcyclohexane	1	5	10	20	40
1,3,5-Trichlorobenzene	1	5	10	20	40

¹ 25 mL purge samples analyzed at 5 mL purge on more sensitive equipment.

Table 3

STL Appendix IX Standard and Reporting Limits, 5 mL purge

Compound	CAS Number	Reporting Limits			
		5 mL Water µg/L	25mL ² water µg/L	Low Soil µg/kg	Medium Soil µg/mL
Allyl Chloride	107-05-1	10	2	10	500
Acetonitrile	75-05-8	100	20	100	5000
Dichlorofluoromethane		10	2	10	500
Isopropyl ether	108-20-3	10	2	10	500
Chloroprene	126-99-8	5	2	5	250
n-Butanol	71-36-3	200	50	200	10,000
Propionitrile	107-12-0	20	4	20	1000
Methacrylonitrile	126-98-7	5	2	5	250
Isobutanol	78-83-1	200	50	200	10,000
Methyl methacrylate	80-62-6	5	2	5	250
1,1,1,2-Tetrachloroethane	630-20-6	5	1	5	250
1,2-Dibromo-3-chloropropane	96-12-8	10	2	10	500
Ethyl ether	60-29-7	10	2	10	500
Ethyl Acetate	141-78-6	20	4	20	1,000
2-Nitropropane	79-46-9	10	4	10	500
Cyclohexanone	108-94-1	N/A ¹	N/A ¹	N/A ¹	N/A ¹
Isopropylbenzene	98-82-8	5	1	5	250
2-Methylnaphthalene (Michigan only)	91-57-6	NA	5	NA	330

¹ Cyclohexanone decomposes to 1,1-dimethoxycyclohexane in methanolic solution. Reporting limits cannot be accurately determined.

² Optionally, 5 mL purge volume if adequate sensitivity is obtained.

Table 4

Recommended/STL Appendix IX Standard Calibration Levels, µg/L

Compound	Level 1	Level 2	Level 3	Level 4	Level 5
Allyl Chloride	5	20	50	100	200
Acetonitrile	50	200	500	1,000	2,000
Dichlorofluoromethane	5	20	50	100	200
Isopropyl ether	5	20	50	100	200
Chloroprene	5	20	50	100	200
n-Butanol	100	400	1,000	2,000	4,000
Propionitrile	10	40	100	200	400
Methacrylonitrile	5	20	50	100	200
Isobutanol	100	400	1,000	2,000	4,000
Methyl methacrylate	5	20	50	100	200
1,1,1,2-Tetrachloroethane	5	20	50	100	200
1,2-Dibromo-3-chloropropane	10	40	100	200	400
Ethyl ether	5	20	50	100	200
Ethyl Acetate	10	40	100	200	400
2-Nitropropane	10	40	100	200	400
Cyclohexanone	50	200	500	1,000	2,000
2-Methylnaphthalene (Michigan only)	2	10	20	40	80

Table 5
 Reportable Analytes for STL Standard Tests, Primary Standard

Compound	CAS Number	STL Standard List	TCLP	TCL	Appendix IX	UTS
Dichlorodifluoromethane	75-71-8				X	X
Chloromethane	74-87-3	X		X	X	X
Bromomethane	74-83-9	X		X	X	X
Vinyl chloride	75-01-4	X	X	X	X	X
Chloroethane	75-00-3	X		X	X	X
Trichlorofluoromethane	75-69-4				X	X
Acrolein	107-02-8				X	X
Acetone	67-64-1	X		X	X	X
Trichlorotrifluoroethane	76-13-1					X
Ethanol	64-17-5					
Iodomethane	74-88-4				X	X
Carbon disulfide	75-15-0	X		X	X	X
Methylene chloride	75-09-2	X		X	X	X
tert-Butyl alcohol	75-65-0					
1,1-Dichloroethene	75-35-4	X	X	X	X	X
1,1-Dichloroethane	75-34-3	X		X	X	X
trans-1,2-Dichloroethene	156-60-5	X		X	X	X
Total dichloroethene		X		X	X	X
Acrylonitrile	107-13-1				X	X
Methyl tert-butyl ether (MTBE)	1634-04-4					
Hexane	110-54-3					
cis-1,2-Dichloroethene	156-59-2	X		X		
Tetrahydrofuran	109-99-9					
Chloroform	67-66-3	X	X	X	X	X
1,2-Dichloroethane	107-06-2	X	X	X	X	X
Dibromomethane	74-95-3				X	X
2-Butanone	78-93-3	X	X	X	X	X
1,4-Dioxane	123-91-1				X	X
1,1,1-Trichloroethane	71-55-6	X		X	X	X
Carbon tetrachloride	56-23-5	X	X	X	X	X
Bromodichloromethane	75-27-4	X		X	X	X
1,2-Dichloropropane	78-87-5	X		X	X	X
cis-1,3-Dichloropropene	10061-01-	X		X	X	X

Table 5

Reportable Analytes for STL Standard Tests, Primary Standard

Compound	CAS Number	STL Standard List	TCLP	TCL	Appendix IX	UTS
	5					
Trichloroethene	79-01-6	X	X	X	X	X
Dibromochloromethane	124-48-1	X		X	X	X
1,2-Dibromoethane	106-93-4				X	X
1,2,3-Trichloropropane	96-18-4				X	X
1,1,2-Trichloroethane	79-00-5	X		X	X	X
Benzene	71-43-2	X	X	X	X	X
Ethylmethacrylate	97-63-2				X	X
trans-1,3-Dichloropropene	10061-02-6	X		X	X	X
Bromoform	75-25-2	X		X	X	X
4-Methyl-2-pentanone	108-10-1	X		X	X	X
2-Hexanone	591-78-6	X		X	X	
Tetrachloroethene	127-18-4	X	X	X	X	X
Toluene	108-88-3	X		X	X	X
1,1,2,2-Tetrachloroethane	79-34-5	X		X	X	X
2-Chloroethyl vinyl ether	110-75-8					
Vinyl acetate	108-05-4				X	
Chlorobenzene	108-90-7	X	X	X	X	X
Ethylbenzene	100-41-4	X		X	X	X
Styrene	100-42-5	X		X	X	
t-1,4-Dichloro-2-butene	110-57-6				X	
m and p Xylenes		X		X	X	X
o-xylene	95-47-6	X		X	X	X
Total xylenes	1330-20-7	X		X	X	X
1,3-Dichlorobenzene	541-73-1					
1,4-Dichlorobenzene	106-46-7					
1,2-Dichlorobenzene	95-50-1					

Table 6

Reportable Analytes for STL Standard Tests, Appendix IX standard

Compound	Number	STL Standard List	TCLP	TCL	Appendix IX	UTS
Allyl Chloride	107-05-1				X	
Acetonitrile	75-05-8				X	X
Dichlorofluoromethane	75-43-4					
Isopropyl ether	108-20-3					
Chloroprene	126-99-8				X	
n-Butanol	71-36-3					
Propionitrile	107-12-0				X	
Methacrylonitrile	126-98-7				X	X
Isobutanol	78-83-1				X	X
Methyl methacrylate	80-62-6				X	X
1,1,1,2-Tetrachloroethane	630-20-6				X	X
1,2-Dibromo-3-chloropropane	96-12-8				X	X
Ethyl ether	60-29-7					X
Ethyl Acetate	141-78-6					X
2-Nitropropane	79-46-9					
Cyclohexanone	108-94-1					X
Isopropylbenzene	98-82-8					

Table 7
Internal Standards

	Standard Concentration µg/mL	Quantitation ion (5 mL purge)	Quantitation ion (25 mL purge)
Fluorobenzene	50	96	96
Chlorobenzene-d5	50	117	119
1,4-Dichlorobenzene-d4	50	152	152

Notes:

- 1) 5 µL of the internal standard is added to the sample. This results in a concentration of each internal in the sample of 50µg/L for a 5 mL purge or 10 µg/L for a 25 mL purge.
- 2) Except for medium level soils, the surrogate and internal standards may be combined in one solution.

Table 8
Surrogate Standards

Surrogate Compounds	Standard Concentration µg/mL
1,2-Dichloroethane-d ₄	50
Dibromofluoromethane	50
Toluene-d ₈	50
4-Bromofluorobenzene	50

Notes:

- 1) 5 µL of the surrogate standard is added to the sample. This results in a concentration of each surrogate in the sample of 50µg/L for a 5 mL purge or 10 µg/L for a 25 mL purge.
- 2) Except for medium level soils, the surrogate and internal standards may be combined in one solution.
- 3) Recovery limits for surrogates are generated from historical data and are maintained by the QA department.

Table 9

Matrix Spike / LCS Control Compounds

Compound	Standard Concentration $\mu\text{g}/\text{mL}$
1,1-Dichloroethene	50
Trichloroethene	50
Toluene	50
Benzene	50
Chlorobenzene	50
n-Hexane (Ohio VAP only)	50

Notes:

- 1) 5 μL of the standard is added to the LCS or matrix spiked sample. This results in a concentration of each spike analyte in the sample of 50 $\mu\text{g}/\text{L}$ for a 5 mL purge or 10 $\mu\text{g}/\text{L}$ for a 25 mL purge.
- 2) Recovery and precision limits for LCS and MS/MSD are generated from historical data and are maintained by the QA department.
- 3) Full analyte spikes may also be used at the laboratories option or at client request.

Table 10

BFB Key Ion Abundance Criteria

Mass	Ion Abundance Criteria
50	15% to 40% of Mass 95
75	30% to 60% of Mass 95
95	Base Peak, 100% Relative Abundance
96	5% to 9% of Mass 95
173	Less Than 2% of Mass 174
174	Greater Than 50% of Mass 95
175	5% to 9% of Mass 174
176	Greater Than 95%, But Less Than 101% of Mass 174
177	5% to 9% of Mass 176

Table 11
SPCC Compounds and Minimum Response Factors

Compound	8260B, 8260A Min. RF
Chloromethane	0.100
1,1-Dichloroethane	0.100
Bromoform	>0.100
1,1,2,2-Tetrachloroethane	0.300
Chlorobenzene	0.300

Table 12
CCC compounds

Compound	Max. %RSD from Initial Calibration	Max. %D for continuing calibration
Vinyl Chloride	<30.0	<20.0
1,1-Dichloroethene	<30.0	<20.0
Chloroform	<30.0	<20.0
1,2-Dichloropropane	<30.0	<20.0
Toluene	<30.0	<20.0
Ethylbenzene	<30.0	<20.0
n-Hexane (Ohio VAP only)	<30.0	<20.0

Table 13
Characteristic ions

Compound	Primary*	Secondary	Tertiary
1,2-Dichloroethane-d ₄ (Surrogate)	65	102	
Dichlorodifluoromethane	85	87	50, 101, 103
Chloromethane	50	52	49
Vinyl chloride	62	64	61
Bromomethane	94	96	79
Chloroethane	64	66	49
Trichlorofluoromethane	101	103	66
1,1-Dichloroethene	96	61	98

Table 13
 Characteristic ions

Compound	Primary*	Secondary	Tertiary
Acrolein	56	55	58
Iodomethane	142	127	141
Carbon disulfide	76	78	
Trichlorotrifluoroethane	151	101	153
Ethanol	45	46	
Acetone	43	58	
Methylene chloride	84	49	51, 86
tert-Butyl alcohol	59	74	
trans-1,2-Dichloroethene	96	61	98
Acrylonitrile	53	52	51
Methyl tert butyl ether	73		
Hexane	57	43	
1,1-Dichloroethane	63	65	83
cis-1,2-Dichloroethene	96	61	98
2-Butanone	43	72**	
Tetrahydrofuran	42	71	
Chloroform	83	85	47
1,2-Dichloroethane	62	64	98
Dibromomethane	93	174	95, 172, 176
1,4-Dioxane	88	58	
Vinyl acetate	43	86	
1,1,1-Trichloroethane	97	99	117
Carbon tetrachloride	117	119	121
Benzene	78	52	77
Trichloroethene	130	95	97, 132
1,2-Dichloropropane	63	65	41
Bromodichloromethane	83	85	129
2-Chloroethyl vinyl ether	63	65	106
cis-1,3-Dichloropropene	75	77	39
trans-1,3-Dichloropropene	75	77	39
1,1,2-Trichloroethane	97	83	85, 99
Chlorodibromomethane	129	127	131
Bromoform	173	171	175, 252
1,2,3-Trichloropropane	75	110	77, 112, 97
Toluene-d ₈ (Surrogate)	98	70	100
4-Bromofluorobenzene (Surrogate)	95	174	176
Toluene	91	92	65
4-Methyl-2-pentanone	43	58	57, 100
Tetrachloroethene	164	166	131

Table 13
 Characteristic ions

Compound	Primary*	Secondary	Tertiary
Ethyl methacrylate	69	41	99, 86, 114
2-Hexanone	43	58	57, 100
Chlorobenzene	112	114	77
Ethylbenzene	106	91	
Xylenes	106	91	
Styrene	104	103	78, 51, 77
Dichlorobenzene (all isomers)	146	148	111
trans 1,4-Dichloro-2-butene	53	75	89, 77, 124
1,1,2,2-Tetrachloroethane	83	85	131, 133
Allyl Chloride	76	41	78
Acetonitrile	40	41	
Dichlorofluoromethane	67	69	
Isopropyl ether	87	59	45
Chloroprene	53	88	90
n-Butanol	56	41	42
Propionitrile	54	52	55
Methacrylonitrile	41	67	52
Isobutanol	41	43	74
Methyl methacrylate	41	69	100
1,1,1,2-Tetrachloroethane	131	133	119
1,2-Dibromo-3-chloropropane	157	155	75
Ethyl ether	59	74	
Ethyl Acetate	43	88	61
2-Nitropropane	41	43	46
Cyclohexanone	55	42	98
Isopropylbenzene	105	120	

* The primary ion should be used for quantitation unless interferences are present, in which case a secondary ion may be used.

** m/z 43 may be used for quantitation of 2-Butanone, but m/z 72 must be present for positive identification.

1. REQUIREMENTS FOR EPA 624

- 1.1. Method 624 is required for demonstration of compliance with NPDES wastewater discharge permits. This method can be applied only to aqueous matrices. The standard analyte list and reporting limits are listed in Table B-1.
- 1.2. The tune period for this method is defined as 24 hours.
- 1.3. The initial calibration curve for this method requires at least three points.
- 1.4. Sample concentrations are calculated using the average RRF from the initial calibration curve.
- 1.5. Each target analyte is assigned to the closest eluting internal standard.
- 1.6. Initial demonstration of Proficiency
 - 1.6.1. The spiking level for the four replicate initial demonstration of proficiency is 20 µg/L. The acceptance criteria are listed in Table B-2
- 1.7. Initial calibration curve requirements:
 - 1.7.1. Target compounds must have RSD ≤ 35%.
 - 1.7.2. If this requirement can not be met, a regression curve must be constructed for the non-compliant compounds. There is no correlation coefficient requirement for the regression curve.
- 1.8. Continuing calibration verification requirements:
 - 1.8.1. The continuing calibration standard is from a different source than the initial calibration standard. The acceptance criteria are listed in Table B-2.
- 1.9. Matrix Spike and LCS requirements
 - 1.9.1. The matrix spike and LCS are spiked at 20 µg/L. A matrix spike duplicate is not necessary for this method. The recovery limits for matrix spike and LCS recovery are listed in Table C-2.
- 1.10. Method clarifications, modifications and additions

- 1.10.1. Section 5.2.2 of the source method describes the trap packing materials as Tenax GC, Methyl silicone, silica gel and coconut charcoal. STL routinely employs the Supelco VOCARB 3000, which consists of Carbopack B and Carboxen 1000 and 1001.
- 1.10.2. Section 5.3.2 of the source method describes a packed analytical column. STL routinely employs capillary columns when performing this method.
- 1.10.3. The source method provides a suggested list of compounds for internal and surrogate standards. STL uses the following two compounds which are not on the table: Chlorobenzene-d₅ (internal standard) and 1,2-Difluorobenzene-d₄ (surrogate).

Table A-1.

Method 624 Analytes and Reporting Limits

Analytes	µg/L
Benzene	5
Bromodichloromethane	5
Bromoform	5
Bromomethane	5
Carbon tetrachloride	5
Chlorobenzene	5
Chloroethane	5
2-Chloroethyl vinyl ether	5
Chloroform	5
Chloromethane	5
Dibromochloromethane	5
1,2-Dichlorobenzene	5
1,3-Dichlorobenzene	5
1,4-Dichlorobenzene	5
1,1-Dichloroethane	5
1,2-Dichloroethane	5
1,1-Dichloroethene	5
trans-1,2-Dichloroethene	5
1,2-Dichloropropane	5
cis-1,3-Dichloropropene	5
trans-1,3-Dichloropropene	5
Ethylbenzene	5
Methylene chloride	5
1,1,2,2-Tetrachloroethane	5
Tetrachloroethene	5
Toluene	5
1,1,1-Trichloroethane	5
1,1,2-Trichloroethane	5
Trichloroethene	5
Trichlorofluoromethane	5
Vinyl chloride	5

Table A-2.

Method 624 QC Acceptance Criteria

Analytes	Daily QC check acceptance criteria (20µg/L spike)	Mean recovery, 4 replicate initial demonstration acceptance criteria (20µg/L spike)	Standard deviation, 4 replicate initial demonstration acceptance criteria (20µg/L spike)	Matrix spike and LCS acceptance criteria (% recovery)
Benzene	12.8-27.2	15.2-26.0	6.9	37-151
Bromodichloromethane	13.1-26.9	10.1-28.0	6.4	35-155
Bromoform	14.2-25.8	11.4-31.1	5.4	45-169
Bromomethane	2.8-37.2	D-41.2	17.9	D-242
Carbon tetrachloride	14.6-25.4	17.2-23.5	5.2	70-140
Chlorobenzene	13.2-26.8	16.4-27.4	6.3	37-160
Chloroethane	7.6-32.4	8.4-40.4	11.4	14-230
2-Chloroethyl vinyl ether	D-44.8	D-50.4	25.9	D-305
Chloroform	13.5-26.5	13.7-24.2	6.1	51-138
Chloromethane	D-40.8	D-45.9	19.8	D-273
Dibromochloromethane	13.5-26.5	13.8-26.6	6.1	53-149
1,2-Dichlorobenzene	12.6-27.4	11.8-34.7	7.1	18-190
1,3-Dichlorobenzene	14.6-25.4	17.0-28.8	5.5	59-156
1,4-Dichlorobenzene	12.6-27.4	11.8-34.7	7.1	18-190
1,1-Dichloroethane	14.5-25.5	14.2-28.5	5.1	59-155
1,2-Dichloroethane	13.6-26.4	14.3-27.4	6.0	49-155
1,1-Dichloroethene	10.1-29.9	3.7-42.3	9.1	D-234
trans-1,2-Dichloroethene	13.9-26.1	13.6-28.5	5.7	54-156
1,2-Dichloropropane	6.8-33.2	3.8-36.2	13.8	D-210
cis-1,3-Dichloropropene	4.8-35.2	1.0-39.0	15.8	D-227
trans-1,3-Dichloropropene	10.0-30.0	7.6-32.4	10.4	17-183
Ethylbenzene	11.8-28.2	17.4-26.7	7.5	37-162
Methylene chloride	12.1-27.9	D-41.0	7.4	D-221
1,1,2,2-Tetrachloroethane	12.1-27.9	13.5-27.2	7.4	46-157
Tetrachloroethene	14.7-25.3	17.0-26.6	5.0	64-148
Toluene	14.9-25.1	16.6-26.7	4.8	47-150
1,1,1-Trichloroethane	15.0-25.0	13.7-30.1	4.6	52-162
1,1,2-Trichloroethane	14.2-25.8	14.3-27.1	5.5	52-150
Trichloroethene	13.3-26.7	18.6-27.6	6.6	71-157
Trichlorofluoromethane	9.6-30.4	8.9-31.5	10.0	17-181
Vinyl chloride	0.8-39.2	D-43.5	20.0	D-251

Controlled Copy
Copy No. _____
Implementation Date: _____

SOP No. CORP-IP-0003NC
Revision No. 1.3
Revision Date: 09/25/01
Page 1 of 37

STL STANDARD OPERATING PROCEDURE

**TITLE: ACID DIGESTION OF AQUEOUS SAMPLES BY SW846 AND MCAWW
200 SERIES METHODS**

(Supersedes: Revision 1.2 Dated 03/20/00)

Prepared by: _____ Date

Reviewed by: _____ Date
Technology Specialist

Reviewed by: _____ Date
Quality Assurance Manager

Reviewed by: _____ Date
Environmental Health and Safety

Reviewed by: _____ Date
Laboratory Director

Reviewed by: _____ Date
Corporate Technology

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use it for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: © 2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

APPENDIX A - TABLES

TABLE OF CONTENTS

1. SCOPE AND APPLICATION 3

2. SUMMARY OF METHOD..... 4

3. DEFINITIONS 5

4. INTERFERENCES..... 5

5. SAFETY 6

6. EQUIPMENT AND SUPPLIES 7

7. REAGENTS AND STANDARDS 8

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE..... 10

9. QUALITY CONTROL 10

10. CALIBRATION AND STANDARDIZATION..... 13

11. PROCEDURE 13

12. DATA ANALYSIS AND CALCULATIONS 19

13. METHOD PERFORMANCE..... 19

14. POLLUTION PREVENTION 19

15. WASTE MANAGEMENT 19

16. REFERENCES 20

17. MISCELLANEOUS (TABLES, APPENDICES, ETC. . .) 20

LIST OF APPENDICES:

APPENDIX A - TABLES 32

APPENDIX B - CONTAMINATION CONTROL GUIDELINES 38

APPENDIX A - TABLES

1. SCOPE AND APPLICATION

- 1.1. This procedure describes the preparation of aqueous samples for the analysis of certain metals by Graphite Furnace Atomic Absorption (GFAA), Flame Atomic Absorption (FLAA), Inductively Coupled Plasma Atomic Emission Spectroscopy (ICP), and Inductively Coupled Plasma-Mass Spectrometry (ICP/MS) using the MCAWW 200 series methods (NPDES) and SW846 Methods 3005A, 3010A, 3020A and 7060A/7740 (RCRA).
- 1.2. The applicability of each of these preparation protocols to specific analytes is detailed in Tables I and II (Appendix A) and the applicable determinative methods are illustrated by Figures 6 and 7 (Section 17). Additional elements may be analyzed following digestion by these protocols provided that the method performance criteria specified in Section 13.0 of this SOP are met.
- 1.3. This SOP provides procedures applicable to the preparation of dissolved, suspended, total recoverable and total elements in ground water, aqueous samples, certain aqueous sludges, wastes, and biological tissues, and leachates/extracts.
- 1.4. SW-846 Method 3005A is used to prepare surface and groundwater samples for total recoverable and dissolved metals determination by FLAA, ICP and GFAA (antimony only).
- 1.5. MCAWW Method 200.7 Section 9.4 is used to prepare surface water, domestic and industrial waste samples for total recoverable and dissolved metals determination by ICP.
- 1.6. SW-846 Method 3010A is used to prepare aqueous samples, EP and mobility-procedure extracts, and wastes that contain suspended solids for total metals analysis by FLAA or ICP.
- 1.7. MCAWW Method 200.7 Section 9.3 is used to prepare surface water and wastes that contain suspended solids for total metals analysis by ICP.
- 1.8. SW-846 Method 3020A is used to prepare aqueous samples, EP and mobility-procedure extracts, and wastes that contain suspended solids for total metals by GFAA, or ICP/MS.
- 1.9. MCAWW Method 200.0 Section 4.1.3 is used to prepare surface water and wastes that contain suspended solids for total metals analysis by GFAA.

APPENDIX A - TABLES

- 1.10. MCAWW Method 200.0 Section 4.1.4 is used to surface water, domestic and industrial waste samples for total recoverable and dissolved metals determination by GFAA.
- 1.11. SW-846 Methods 7060A and 7740, respectively, contain the procedure for the preparation of aqueous samples for arsenic and selenium.
- 1.12. MCAWW Methods 206.2 and 270.2, respectively, contain the procedure for the preparation of aqueous samples for arsenic and selenium.
- 1.13. All matrices require digestion prior to analysis with the exception of analyses for dissolved metals in filtered and acidified aqueous samples. Although digestion is not specifically required by the method, some clients and regulators do require digestion of dissolved samples and this must be clarified before project initiation.

2. SUMMARY OF METHOD

- 2.1. Method 3005A / Method 200.7 Section 9.4 - Preparation for Total Recoverable or Dissolved Metals Analysis by FLAA or ICP Spectroscopy
 - 2.1.1. A representative aliquot of sample is heated with nitric and hydrochloric acids and substantially reduced in volume. The digestate is filtered (if necessary) and diluted to volume.
- 2.2. Method 3010A / Method 200.7 Section 9.3 - Preparation for Total Metals Analysis by FLAA or ICP Spectroscopy
 - 2.2.1. A representative aliquot of sample is refluxed with nitric acid. After the digestate has been reduced to a low volume, it is refluxed with hydrochloric acid, filtered (if necessary) and brought up to volume.
- 2.3. Method 3020A / Method 200.0 Section 4.1.3 - Preparation for Total Metals for Analysis by GFAA Spectroscopy and ICP/MS.

APPENDIX A - TABLES

- 2.3.1. A representative aliquot of sample is refluxed with nitric acid. After the digestate has been reduced to a low volume, it is cooled, filtered (if necessary) and brought up to volume.
- 2.4. Methods 7060A/206.2 and Methods 7740/270.2 - Preparation for Arsenic/Selenium Analysis by GFAA
 - 2.4.1. A representative aliquot of sample is heated with nitric acid and peroxide until the digestate has been reduced to a low volume. The sample is cooled, filtered (if necessary) and brought up to volume.
- 2.5. Method 200.0 Section 4.1.4 - Total Recoverable GFAA Preparation (NPDES)
 - 2.5.1. A representative aliquot of sample is heated with nitric acid and until the digestate has been reduced to a low volume. The sample is cooled, filtered (if necessary) and brought up to volume.

3. **DEFINITIONS**

Additional definitions of terms used in this SOP may be found in the glossary of the QAMP.

- 3.1. Dissolved Metals: Those elements which pass through a 0.45 um membrane. (Sample is acidified after filtration).
- 3.2. Suspended Metals: Those elements which are retained by a 0.45 um membrane.
- 3.3. Total Metals: The concentration determined on an unfiltered sample following digestion.
- 3.4. Total Recoverable Metals: The concentration determined on an unfiltered sample following treatment with hot, dilute mineral acid.

4. **INTERFERENCES**

- 4.1. There are numerous routes by which samples may become contaminated. Potential sources of trace metals contamination include: metallic or metal-containing labware (e.g., talc gloves which contain high levels of zinc), containers, impure reagents, dirty glassware, improper sample transfers, dirty work areas, atmospheric inputs such as dirt and dust, etc. Be aware of potential sources of contamination and take appropriate measures to minimize or avoid them.

APPENDIX A - TABLES

- 4.2. The entire work area, including the bench top and fume hood, should be thoroughly cleaned on a routine schedule in order to minimize the potential for environmental contamination. Refer to Appendix D for additional contamination control guidelines.
- 4.3. Boron and silica from the glassware will migrate into the sample solution during and following sample processing. For critical low level determinations of boron and silica, only quartz and/or plastic labware should be used.
- 4.4. Physical interference effects may contribute to inaccuracies in the determinations of trace elements. Oils, solvents and other matrices may not be digested using these methods if they are not soluble with acids. If physical interferences are present, they should be documented.
- 4.5. Visual interferences or anomalies (such as foaming, emulsions, precipitates, etc.) must be documented.
- 4.6. Allowing samples to boil or go dry during digestion may result in the loss of volatile metals. If this occurs the sample must be reprepared. Antimony is easily lost by volatilization from hydrochloric acid media.
- 4.7. Precipitation of silver chloride (AgCl) may occur when chloride ions and high concentrations of silver (i.e., greater than 1 mg/L) are present in the sample.
- 4.8. Specific analytical interferences are discussed in each of the determinative methods.
5. **SAFETY**
 - 5.1. Procedures shall be carried out in a manner that protects the health and safety of all STL associates.
 - 5.2. Eye protection that satisfies ANSI Z87.1 (as per the Chemical Hygiene Plan), laboratory coat, and appropriate gloves must be worn while samples, standards, solvents, and reagents are being handled. Disposable gloves that have been contaminated will be removed and discarded; other gloves will be cleaned immediately.
 - 5.3. The health and safety hazards of many of the chemicals used in this procedure have not been fully defined. Additional health and safety information can be obtained from the Material Safety Data Sheets (MSDS) maintained in the laboratory. The following specific hazards are known:

APPENDIX A - TABLES

5.3.1. The following materials are known to be **corrosive**:

hydrochloric acid and nitric acid.

5.3.2. The following materials are known to be **oxidizing agents**:

nitric acid and hydrogen peroxide.

5.3.3. All sample digestions, including cooling of digestates, must be carried out in a fume hood.

5.4. The acidification of samples containing reactive materials may result in the release of toxic gases, such as cyanides or sulfides. Acidification of samples should be done in a fume hood. The analyst should also be aware of the potential for a vigorous reaction.

5.5. Exposure to chemicals must be maintained **as low as reasonably achievable**. Therefore, unless they are known to be non-hazardous, all samples should be opened, transferred and prepared in a fume hood, or under other means of mechanical ventilation. Solvent and waste containers will be kept closed unless transfers are being made.

5.6. All work must be stopped in the event of a known or potential compromise to the health and safety of a STL associate. The situation must be reported **immediately** to a laboratory supervisor.

5.7. Always carry bulk concentrated acid bottles in appropriate impact proof containers.

5.8. Acid/peroxide spills must be neutralized immediately, flushed with water and cleaned up using appropriate spill kits.

5.9. Discard chipped or broken beakers to prevent injury. Chipped glassware may be fire polished as an alternative to disposal.

5.10. Any and all accidents and spills must be reported to the lab supervisor or EH&S coordinator.

6. EQUIPMENT AND SUPPLIES

6.1. Hot plate, digestion block or other adjustable heating source capable of maintaining a temperature of 95°C (± 4).

APPENDIX A - TABLES

- 6.2. Calibrated thermometer that covers a temperature range of 0-200°C.
- 6.3. Griffin beakers of assorted sizes or equivalent.
- 6.4. Watch glasses, ribbed or equivalent.
- 6.5. Whatman No. 4 filter paper or equivalent.
- 6.6. Funnels or equivalent filtration apparatus.
- 6.7. Centrifugation equipment (if desired method of removing particulates is centrifugation).
- 6.8. Graduated cylinder or equivalent capable of measuring 50 mL within 3% accuracy.
- 6.9. Analytical balance capable of accurately weighing to the nearest 0.01 grams.
- 6.10. Repipetors or suitable reagent dispensers.
- 6.11. Calibrated automatic pipettes with corresponding pipette tips or Class A glass volumetric pipettes.
- 6.12. Class A volumetric flasks.
- 6.13. pH indicator strips (pH range 0 - 6).
- 6.14. Plastic digestate storage bottles.

7. REAGENTS AND STANDARDS

- 7.1. Reagent water must be produced by a Millipore DI system or equivalent. Reagent water must be free of the analytes of interest as demonstrated through the analysis of method blanks as defined in the determinative SOPs.
- 7.2. Laboratory Control Sample (LCS) and matrix spike (MS) solutions are purchased as custom STL solutions. All standards must be stored in FEP fluorocarbon or previously unused polyethylene or polypropylene bottles. Stock standard solutions must be replaced prior to the expiration date provided by the manufacturer. If no expiration date is provided, the stock solutions may be used for up to one year and must be replaced sooner if verification from an independent source indicates a problem.

APPENDIX A - TABLES

- 7.3. Working ICP LCS/MS spike solution: Prepare the ICP LCS/MS working spike solution from custom stock standards to the final concentration listed in Table III. The working spike must be prepared in a matrix of 5% HNO₃. This acid (5 mL of concentrated HNO₃ per 100 mL) must be added to the volumetric flask before the addition of the stock standard aliquot. The working ICP LCS solution must be made fresh every three months.
- 7.4. Working GFAA LCS/MS spike solution: Prepare the GFAA working LCS spike solution by diluting the custom stock solution (7.2) 200x. The working spike solution must be prepared in a matrix of 5% HNO₃. This acid (5 mL of concentrated HNO₃ per 100 mL) must be added to the volumetric flask before the addition of the stock standard aliquot. The working GFAA LCS solution must be made fresh every three months.
- 7.5. The TCLP MS working spike solution is provided directly by the vendor, no further standard preparation is necessary.
- 7.6. The LCS and MS samples must contain all the elements designated for analysis in each batch of samples. If a non-routine element is required that is not contained in the custom STL solution, the individual facility must purchase a solution from the designated vendor that will cover the additional analyte(s) of interest and provide for a final spike concentration that is appropriate to the determinative method.
- 7.7. Aqueous laboratory control samples (LCSW) and matrix spike samples are prepared as described in Sections 9.5 and 9.6. Refer to Tables III and IV(Appendix A) for details regarding the stock, working standard and final digestate spike concentrations for ICP and GFAA LCS and matrix spike preparations.
- 7.8. Nitric acid (HNO₃), concentrated, trace metal grade or better.
- 7.9. Nitric acid, 1:1 - dilute concentrated HNO₃ with an equal volume of reagent water.
- Note:** When preparing diluted acids always add acid to water. If the water is added to the acid a violent reaction may occur.
- 7.10. Hydrochloric acid (HCl), concentrated, trace metal grade or better.
- 7.11. Hydrochloric acid, 1:1 - dilute concentrated HCl with an equal volume of reagent water.
- Note:** When preparing diluted acids always add acid to water. If the water is added to the acid a violent reaction may occur.

7.12. 30% Hydrogen peroxide (H₂O₂), reagent grade.

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE

8.1. Sample holding time for metals included under the scope of this SOP is 180 days from the date of collection to the date of analysis.

8.2. Aqueous samples are preserved with nitric acid to a pH of <2 and may be stored in either plastic or glass. If boron or silica are to be determined, plastic containers are preferred. Refrigeration is not required. Preservation must be verified prior to analysis.

8.3. For dissolved metals analysis, the samples should be filtered through a 0.45 um filter prior to preservation. Filtration must be done in the field or within 24 hours of collection.

Note: If a sample being analyzed for dissolved metals is found to contain sediment the analyst should contact their supervisor or group leader. The client should be notified of the problem to decide how to treat the sample.

9. QUALITY CONTROL

Table VI (Appendix A) provides a summary of quality control requirements including type, frequency, acceptance criteria and corrective action.

9.1. Initial Demonstration of Capability

Prior to analysis of any analyte using any method contained within this SOP the following requirements must be met:

9.1.1. Method Detection Limit (MDL) - An MDL must be determined for each analyte/matrix prior to the analysis of any samples. The MDL is determined using seven replicates of reagent water, spiked with all the analytes of interest, that have been carried through the entire analytical procedure. MDL's must be redetermined on an annual basis in accordance with 40 CFR Part 136 Appendix B requirements as detailed in STL QA Policy QA-005. The spike level must be between the calculated MDL and 10X the MDL to be valid. The result of the MDL determination must be below the STL reporting limit.

9.1.2. Initial Demonstration Study - This requires the analysis of four QC check samples. The QC check sample is a well-characterized laboratory generated

APPENDIX A - TABLES

sample used to monitor method performance, which should contain all the analytes of interest. The results of the initial demonstration study must be acceptable before analysis of samples may begin. The results of the initial demonstration study may be used to extend a method for the analysis of other elements provided all acceptance criteria are met.

9.1.2.1. Four aliquots of the check sample (LCS) are prepared and analyzed using the procedures detailed in this SOP and the determinative SOPs.

9.1.2.2. Calculations and acceptance criteria for QC check samples are given in the determinative SOPs (CORP-MT-0001, CORP-MT-0003).

- 9.2. Preparation Batch - A group of up to 20 samples that are of the same matrix and are processed together using the same procedures and reagents. The preparation batch must contain a method blank, a LCS and a matrix spike/matrix spike duplicate. In some cases, at client request, it may be appropriate to process a matrix spike and sample duplicate in place of the MS/MSD. If clients specify specific samples for MS/MSD, the batch may contain multiple MS/MSD pairs.
- 9.3. Sample Count - Laboratory generated QC samples (method blanks, LCS) are not included in the sample count for determining the size of a preparation batch. MS/MSD are not included in the sample count unless there are multiple sets of MS/MSD per batch. In other words, the first MS/MSD are not counted; all additional MS and MSDs are counted as samples.
- 9.4. Method Blank (MB) - One method blank must be processed with each preparation batch. The method blank consists of reagent water containing all reagents specific to the method that is carried through the entire analytical procedure, including preparation and analysis. The method blank is used to identify any system and process interferences or contamination of the analytical system that may lead to the reporting of elevated analyte concentrations or false positive data. Criteria for the acceptance of blanks are contained within the individual analytical method SOP's. If the method blank does not meet the criteria contained within the analytical method SOPs; the blank and all associated samples in the batch must be redigested.
- 9.4.1. Aqueous method blanks are prepared by taking 50 mL or 50 g of reagent water through the appropriate procedure as described in Section 11.
- 9.4.2. TCLP method blanks are prepared by taking 50 mL or 50 g of leachate fluid through the appropriate procedure as described in Section 11.

9.5. Laboratory Control Sample (LCS) - One aqueous LCS must be processed with each preparation batch. The LCS must contain all analytes of interest and must be carried through the entire analytical procedure. The LCS is used to monitor the accuracy of the analytical process. On going monitoring of the LCS results provides evidence that the laboratory is performing the method within acceptable accuracy and precision guidelines. Criteria for the acceptance of LCS results are contained within the individual analytical method SOP's. Corrective action when LCS results fail to meet control limits will be re-preparation and reanalysis of the batch. Refer to Section 7.3 and 7.4 for instructions on preparation of the aqueous LCS spike solution.

9.5.1. The aqueous LCS is prepared by spiking a 50 mL aliquot of reagent water with 1.0 mL of the working LCS/MS spike solution (7.3 or 7.4). The LCS is then processed through the appropriate procedure as described in Section 11.

9.6. Matrix Spike/Matrix Spike Duplicate (MS/MSD) - One MS/MSD pair must be processed for each preparation batch. A matrix spike (MS) is a field sample to which known concentrations of target analytes have been added. A matrix spike duplicate (MSD) is a second aliquot of the same sample (spiked identically as the MS) prepared and analyzed along with the sample and matrix spike. Some client specific data quality objectives (DQO's) may require the use of sample duplicates in place of or in addition to MS/MSD's. The MS/MSD results are used to determine the effect of a matrix on the precision and accuracy of the analytical process. Samples identified as field blanks cannot be used for MS/MSD analysis. If any analyte recovery or RPD falls outside the acceptance range, the recovery of that analyte must be in control for the LCS. If the recovery of the LCS is outside limits, corrective action must be taken. Corrective action will include re-preparation and reanalysis of the batch. Corrective action when MS results fail to meet control limits does not include re-preparation of samples unless the results indicate that a spiking error may have occurred.

9.6.1. The aqueous matrix spike sample is prepared by spiking a 50 mL aliquot of a sample with 1.0 mL of the working LCS/MS spike solution (7.3 or 7.4). The matrix spike sample is then processed as described in Section 11.

9.6.2. The TCLP matrix spike sample is prepared by spiking a 50 mL aliquot of a leachate with 0.5 mL of the working TCLP spike solution (7.5). The matrix spike sample is then processed as described in Section 11.

NOTE: The TCLP matrix spike must be added prior to preservation of the leachate.

APPENDIX A - TABLES

9.6.3. If insufficient sample is available to process a MS/MSD, then a second LCS must be processed. The LCS pair is then evaluated according to the MS/MSD criteria.

9.7. Quality Assurance Summaries - Certain clients may require specific project or program QC that may supersede the SOP requirements. Quality Assurance Summaries (QAS) should be developed to address these requirements.

10. **CALIBRATION AND STANDARDIZATION**

10.1 Hotplate temperature must be verified daily for each hotplate used and must be recorded on either the metals preparation log or in a hotplate temperature logbook. The hotplate temperature should be verified by measuring the temperature of a beaker of reagent water placed on each hotplate.

11. **PROCEDURE**

- 11.1. One time procedural variations are allowed only if deemed necessary in the professional judgment of supervision to accommodate variation in sample matrix, radioactivity, chemistry, sample size, or other parameters. Any variation in procedure shall be completely documented using a Nonconformance Memo and is approved by a Technical Specialist and QA Manager. If contractually required, the client shall be notified. The Nonconformance Memo shall be filed in the project file.
- 11.2. Any unauthorized deviations from this procedure must also be documented as a nonconformance, with a cause and corrective action described.
- 11.3. All digestion procedures must be carried out in a properly functioning hood.
- 11.4. All samples are to be checked out of sample control with the chain of custody documentation filled out completely.
- 11.5. Proper sample identification is extremely important in any preparation procedure. Labeling of beakers and bottles must be done in a manner to ensure connection with the proper sample.
- 11.6. Samples are typically logged in as either waters or soils. Wastes such as organic liquids or sludges and tissues (animal/vegetable) are usually logged in with solid test codes. When initiating prep, examine the sample to see if the sample matches the matrix designation. If the sample is logged in as aqueous but it appears more like a waste (biphasic, sludge like, organic liquid, lots of sediment etc.) contact the lab

supervisor or project manager for further instructions. In some cases it may be more appropriate to process these samples as solids.

- 11.7. If possible prepare all the samples of a project at the same time to minimize the QC required and streamline the flow of the project through the lab and reporting group.
- 11.8. In most cases, both AA and ICP digests are required on each sample. It is recommended that both aliquots be measured out and processed at the same time.
- 11.9. Guidelines are provided in the appendices on procedures to minimize contamination of samples and standards.
- 11.10. The following procedure must be followed for all aqueous sample preparations:
 - 11.10.1. Measure sample pH with pH paper on a separate aliquot of sample.

Note: If the sample pH is > 2 pH units, the client must be notified of the anomaly.

Note: If sample pH has already been verified and documented in sample receipt this step may be omitted.
 - 11.10.2. Mix sample by shaking the container.
 - 11.10.3. Measure and transfer 50 mL or 50 g of the sample into a beaker.

Note: This SOP allows for samples to be weighed instead of measured volumetrically.
 - 11.10.4. Measure two extra aliquots of sample selected for the MS/MSD analysis. Spike each aliquot with the appropriate spiking solutions (7.3-7.5,9.6).
 - 11.10.5. Measure and transfer 50 mL of reagent water into a beaker for the method blank.
 - 11.10.6. Measure and transfer 50 mL of reagent water into a beaker for the LCS and add the appropriate spiking solutions (7.3-7.5,9.6).

APPENDIX A - TABLES

11.11. Proceed to the appropriate Section for the desired method as follows:

Method 3005A or Method 200.7 Section 9.4	11.12
Method 3010A or Method 200.7 Section 9.3	11.13
Method 3020A or Method 200.0 Section 4.1.3	11.14
Method 7060A/7740 or Method 206.2/270.2	11.15
Method 200.0 Section 4.1.4	11.16

11.12. **Method 3005A / Method 200.7 Section 9.4 - Preparation for Total Recoverable or Dissolved Metals Analysis by FLAA or ICP (See Figures 1, 6 and 7)**

11.12.1. To the sample beaker, add 1 mL of concentrated HNO₃ and 2.5 mL of concentrated HCl.

11.12.2. Cover with ribbed watch glass.

11.12.3. Heat at 95°C (± 4) until volume is reduced to between 15 and 20 mL.

NOTE: DO NOT ALLOW SAMPLE TO BOIL OR GO DRY. Doing so will result in the loss of analyte and the sample must be reprepared.

11.12.4. Cool the beaker in a fume hood.

11.12.5. Wash down beaker walls and watch glass with reagent water.

11.12.6. Filter sample, if insoluble materials are present, through Whatman 4 filter paper that has been pre-rinsed with dilute nitric acid.

Note: If any samples in a preparation batch are filtered, the method blank and LCS associated with that batch must also be filtered.

Note: In place of filtering, the samples, after dilution and mixing, may be centrifuged or allowed to settle by gravity overnight to remove insoluble material.

APPENDIX A - TABLES

11.12.7. Rinse beaker and filter paper with reagent water to ensure complete sample transfer.

11.12.8. Adjust the final volume/mass to 50 mL or 50 g with reagent water. The sample is now ready for analysis

11.13. Method 3010A / Method 200.7 Section 9.3 - Preparation for Total Metals Analysis by FLAA or ICP Spectroscopy (See Figures 2, 6 and 7)

11.13.1. To the sample beaker, add 3.0mL of concentrated HNO₃.

11.13.2. Cover with ribbed watch glass.

11.13.3. Place beaker on hotplate 95°C (± 4) and evaporate for 4-5 hours or to low volume of 15-20 mL while ensuring that no portion of the bottom of the beaker is allowed to go dry.

NOTE: DO NOT ALLOW SAMPLE TO BOIL OR GO DRY. Doing so will result in the loss of analyte and the sample must be reprepared.

11.13.4. If necessary, add another 1.5 ml portion of concentrated HNO₃ and re-cover the beaker. Reflux 15 minutes.

11.13.5. Add 5 mL of 1:1 HCl.

11.13.6. Cover and reflux for an additional 15 minutes to dissolve precipitate or residue. Cool in a fume hood.

11.13.7. Wash down beaker walls and watch glass with reagent water.

11.13.8. Filter sample, if insoluble materials are present, through Whatman 4 filter paper.

Note: If any samples in the QC batch are filtered the method blank and LCS associated with that batch must also be filtered.

Note: In place of filtering, the samples, after dilution and mixing, may be centrifuged or allowed to settle by gravity overnight to remove insoluble material.

APPENDIX A - TABLES

11.13.9. Rinse beaker and filter paper with reagent water to ensure complete sample transfer.

11.13.10. Adjust final volume/mass to 50 mL or 50 g with reagent water. The sample is now ready for analysis.

11.14. Method 3020A / Method 200.0 Section 4.1.3 - Preparation for Total Metals Analysis by GFAA or Total Recoverable Metals by ICPMS (See Figures 3, 6 and 7)

11.14.1. To the sample beaker, add 1.5 mL of concentrated HNO₃.

11.14.2. Cover with ribbed watch glass.

11.14.3. Place beaker on hotplate 95°C (± 4) and evaporate to low volume of 15-20 mL while ensuring that no portion of the bottom of the beaker is allowed to go dry.

NOTE: DO NOT ALLOW SAMPLE TO BOIL OR GO DRY. Doing so will result in the loss of analyte and the sample must be reprepared.

11.14.4. If necessary, add another 1.5 mL portion of concentrated HNO₃. Recover, and reflux 15 minutes. Cool the beaker in a fume hood.

11.14.5. Filter sample, if insoluble materials are present, through Whatman 4 filter paper.

Note: If any samples in the QC batch are filtered the method blank and LCS associated with that batch must also be filtered.

Note: In place of filtering, the samples, after dilution and mixing, may be centrifuged or allowed to settle by gravity overnight to remove insoluble material.

11.14.6. Rinse beaker and filter paper with reagent water to ensure complete sample transfer.

11.14.7. Adjust final volume to 50 mL with reagent water. The sample is now ready for analysis.

11.15. Method 7060A/7740 and Method 206.2/270.2 - Preparation for Arsenic and Selenium Analysis by GFAA (See Figures 4, 6 and 7)

APPENDIX A - TABLES

- 11.15.1. To the sample beaker, add 1 mL of 30 % H₂O₂ and 1.0 mL of 1:1 HNO₃.
- 11.15.2. Heat, until the digestion is complete, at 95°C (± 4) or until the volume has been reduced to 15-20 mL.
- 11.15.3. Cool beaker.
- 11.15.4. Filter sample, if insoluble materials are present, through Whatman 4 filter paper that has been pre-rinsed with dilute nitric acid.

Note: If any samples in the QC batch are filtered the method blank and LCS associated with that batch must also be filtered.

Note: In place of filtering, the samples, after dilution and mixing, may be centrifuged or allowed to settle by gravity overnight to remove insoluble material.
- 11.15.5. Rinse beaker and filter paper with reagent water to ensure complete sample transfer.
- 11.15.6. Adjust final volume to 50 mL with reagent water. The sample is now ready for analysis.

11.16. Method 200.0 Section 4.1.4 - Preparation for Total Recoverable GFAA Analyses. (See Figures 5 and 7)

- 11.16.1. To the sample beaker, add 1.0 mL of 1:1 HNO₃.
- 11.16.2. Heat, until the digestion is complete, at 95°C (± 4) or until the volume has been reduced to 15 - 20 mL.
- 11.16.3. Cool beaker.
- 11.16.4. Filter sample, if insoluble materials are present, though Whatman 4 filter paper.

Note: If any samples in the QC batch are filtered the method blank and LCS associated with that batch must also be filtered.

Note: In place of filtering, the samples, after dilution and mixing, may be centrifuged or allowed to settle by gravity overnight to remove insoluble material.

11.16.5. Rinse beaker and filter paper with reagent water to ensure complete sample transfer.

11.16.6. Adjust final volume to 50 mL with reagent water. The sample is now ready for analysis.

12. DATA ANALYSIS AND CALCULATIONS

Not Applicable.

13. METHOD PERFORMANCE

13.1. Method performance is determined by the analysis of matrix spike and matrix spike duplicate samples as well as method blanks and laboratory control samples. In general, the matrix spike recovery should fall within +/- 20 % and the matrix spike duplicates should compare within 20% RPD. Method blanks must meet the criteria specified in determinative SOPs. The laboratory control samples should recover within 20% of the true value until in house control limits are established. Acceptance criteria are given in the determinative SOPs.

13.2. The initial demonstration study as detailed in Section 9.1.2 must be acceptable before the analysis of field samples under this SOP may begin. The results of the initial demonstration study may be used to extend a method for the analysis of other elements provided all acceptance criteria are met.

13.3. Training Qualification:

The group/team leader has the responsibility to ensure that this procedure is performed by an associate who has been properly trained in its use and has the required experience.

14. POLLUTION PREVENTION

14.1. This method allows for the proportional reduction of sample and reagent volumes to decrease waste generation.

15. WASTE MANAGEMENT

15.1. Waste generated in the procedure must be segregated and disposed according to the facility hazardous waste procedures. The facility EH & S coordinator should be contacted if additional information is required.

APPENDIX A - TABLES

- 15.2. Standards should be purchased and prepared in volumes consistent with laboratory use to minimize the volume of expired standards to be disposed.

16. **REFERENCES**

- 16.1. Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW-846, 3rd Edition, Final Update I, Revision 1, July 1992. Methods 3005A, 3010A, 3020A, 7060A and 7740A.
- 16.2. Methods for the Chemical Analysis of Water and Waste (MCAWW), 1983.
- 16.3. CORP-MT-0001, Inductively Coupled Plasma-Atomic Emission Spectroscopy, Spectrometric Method for Trace Element Analysis of Water and Wastes, Method 6010A and Method 200.7.
- 16.4. CORP-MT-0003, Graphite Furnace Atomic Absorption Spectroscopy, SW846 Method 7000A and MCAWW 200 Series Methods.
- 16.5. QA-003, STL QC Program.
- 16.6. QA-004, Rounding and Significant Figures.
- 16.7. QA-005, Method Detection Limits.

17. **MISCELLANEOUS (TABLES, APPENDICES, ETC. . .)**

- 17.1. Modifications/Interpretations from reference methods.
- 17.1.1. Modifications applicable to SW-846 reference methods.
- 17.1.1.1. Chapter 1 of SW-846 states that the method blank should not contain any analyte of interest at or above the MDL. This SOP states that the method blank must not contain any analyte of interest at or above the reporting limit. Common lab contaminants are allowed up to two times the reporting limit in the blank following consultation with the client.
- 17.1.1.2. The referenced methods as well as Table 3-1 of SW-846 refer to the use of a 100 mL aliquot for digestion. This SOP requires the use of a 50 mL sample size to reduce waste generation. The use of reduced sample volumes are supported in EPA's document "Response to Public Comments Background Document,

APPENDIX A - TABLES

Promulgation of the Second Update to SW-846, Third Edition” dated November 3, 1994. This document stated “flexibility to alter digestion volumes is addressed and “allowed” by the table (3-1) and is also inherently allowed by specific digestion methods. Table 3-1 is only to be used as guidance when collecting samples...” EMSL-Ci has also taken the stance that “reduction in sample size and appropriate corresponding reduction in sample volume is not considered a significant change in the methodology.” Additionally, in written correspondence from the Office of Solid Waste, Olliver Fordham stated “ As a “representative sample” can be assured, scaling causes no loss of precision and accuracy in the analysis.”

17.1.2. Modifications Specific to Method 3010A

17.1.2.1. Section 11.13.7 of this SOP requires the sample be reduced to a volume of 15 - 20 mL. Section 7.2 of Method 3010A states the volume should be reduced to 3 mL but also states that no portion of the bottom of the beaker should go dry. The SOP required volume is a closer approximation of the volume required to provide an adequate covering of the beaker so as to prevent the loss of critical analytes through volatilization.

17.1.2.2. The scope of 3010A has been expanded to include silver based on comparison studies with 7760A. Method 3010A consistently demonstrated improved accuracy and precision over Method 7760A in the matrices tested (reagent water, surface water and TCLP leachate) up to a concentration of 1 ppm silver.

17.1.3. Modifications Specific to Method 3020A

17.1.3.1. Section 11.14.8 of this SOP requires the sample be reduced to a volume of 15 - 20 mL. Section 7.2 of Method 3010A states the volume should be reduced to 3 mL but also states that no portion of the bottom of the beaker should go dry. The SOP required volume is a closer approximation of the volume required to provide an adequate covering of the beaker so as to prevent the loss of critical analytes through volatilization.

17.1.4. Modifications Specific to Method 7060A/7740

17.1.4.1. Methods 7060A and 7740A incorporate the use of a two step dilution to accommodate the addition of a nickel nitrate modifier.

APPENDIX A - TABLES

This SOP performs the dilution directly in one step and omits the addition of the modifier. The modifier is added automatically at the instrument by direct injection into the furnace.

17.1.5. Modifications Specific to MCAWW Methods

It was determined by technical review that several of the MCAWW methods were equivalent to the SW-846 methods and therefore were combined under the scope of this SOP as described in Section 11.0. The nature of the differences were deemed insignificant in regards to the amount of acid added and the evaporative volume based on the flexibility allowed by the methods (i.e., add additional acid as required) and the subjective wording of the methods (i.e., evaporate to near dryness vs. an exact volume).

17.2. Modifications from previous SOP

17.2.1. Added ICP/MS to the digestion procedures.

17.3. Facility Specific SOPs

Each facility shall attach a list of facility specific SOPs or approved attachments (if applicable) which are required to implement this SOP or which are used in conjunction with this SOP. If no facility specific SOPs or amendments are to be attached, a statement must be attached specifying that there are none. Refer to the Appendices for any facility specific information required to support this SOP.


17.4. Documentation and Record Management

The preparation benchsheet should, at a minimum, include the following information:

- Preparation date, analyst name, matrix, prep type (ICP or GFAA), SOP reference.
- Sample ID; initial weight/volume and final weight/volume.
- Standards Documentation (source, lot, prep date, volume added).
- Analyst Signature.
- Reviewer's Signature and date.


APPENDIX A - TABLES

Figure 1. Method 3005A / Method 200.7 Section 9.4 (Section 11.12)



APPENDIX A - TABLES

Figure 2. Method 3010A / Method 200.7 Section 9.3 (Section 11.13)



APPENDIX A - TABLES

Figure 3. Method 3020A / Method 200.0 Section 4.1.3 (Section 11.14)





Figure 4. Method 7060A/7740A and Method 206.2/270.2 (Section 11.15)



APPENDIX A - TABLES

Figure 5. Method 200.0 Section 4.1.4 (Section 11.16)






Figure 6. Overview of SW846 Aqueous Preparation Methods by Determinative Method



TAL list: Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Ni, K, Se, Ag, Na, Tl, V, Zn

Figure 7. Overview of MCAWW Aqueous Preparation Methods by Determinative Technique



TAL list: Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Ni, K, Se, Ag, Na, Tl, V, Zn

APPENDIX A

TABLES

APPENDIX A - TABLES

TABLE I. Approved Preparation Method Analytes - SW846

ELEMENT	Symbol	CAS Number	3005A	3010 A	3020 A	7060 A 7740
Aluminum	Al	7429-90-5	X	X		
Antimony	Sb	7440-36-0	X			
Arsenic	As	7440-38-2	X	X		X
Barium	Ba	7440-39-3	X	X		
Beryllium	Be	7440-41-7	X	X	X	
Cadmium	Cd	7440-43-9	X	X	X	
Calcium	Ca	7440-70-2	X	X		
Chromium	Cr	7440-47-3	X	X	X	
Cobalt	Co	7440-48-4	X	X	X	
Copper	Cu	7440-50-8	X	X		
Iron	Fe	7439-89-6	X	X		
Lead	Pb	7439-92-1	X	X	X	
Magnesium	Mg	7439-95-4	X	X		
Manganese	Mn	7439-96-5	X	X		
Molybdenum	Mo	7439-98-7	X	X	X	
Nickel	Ni	7440-02-0	X	X		
Potassium	K	7440-09-7	X	X		
Selenium	Se	7782-49-2	X	X		X
Silver	Ag	7440-22-4	X	X		
Sodium	Na	7440-23-5	X	X		
Thallium	Tl	7440-28-0	X	X	X	
Vanadium	V	7440-62-2	X	X	X	
Zinc	Zn	7440-66-6	X	X		

X - Designates that the preparation method is approved for an element

Note: Additional elements may be analyzed following digestion by these protocols provided the method performance criteria specified in Section 13.0 of the SOP are met.

APPENDIX A - TABLES

TABLE II. Approved Preparation Method Analytes - NPDES

ELEMENT	Symbol	CAS Number	200.7 (9.4)	200.7 (9.3)	200.0 (4.1.4)	200.0 (4.1.3)	206.2 270.2
Aluminum	Al	7429-90-5	X	X			
Antimony	Sb	7440-36-0	X	X	X	X	
Arsenic	As	7440-38-2	X	X			X
Boron	B	7440-42-8	X	X			
Barium	Ba	7440-39-3	X	X	X	X	
Beryllium	Be	7440-41-7	X	X	X	X	
Cadmium	Cd	7440-43-9	X	X	X	X	
Calcium	Ca	7440-70-2	X	X			
Chromium	Cr	7440-47-3	X	X	X	X	
Cobalt	Co	7440-48-4	X	X	X	X	
Copper	Cu	7440-50-8	X	X	X	X	
Iron	Fe	7439-89-6	X	X	X	X	
Lead	Pb	7439-92-1	X	X	X	X	
Magnesium	Mg	7439-95-4	X	X			
Manganese	Mn	7439-96-5	X	X	X	X	
Molybdenum	Mo	7439-98-7	X	X	X	X	
Nickel	Ni	7440-02-0	X	X	X	X	
Potassium	K	7440-09-7	X	X			
Selenium	Se	7782-49-2	X	X			X
Silicon	Si	7631-86-9	X	X			
Silver	Ag	7440-22-4	X	X	X	X	
Sodium	Na	7440-23-5	X	X			
Thallium	Tl	7440-28-0	X	X	X	X	
Vanadium	V	7440-62-2	X	X	X	X	
Zinc	Zn	7440-66-6	X	X			

X - Designates that the preparation method is approved for an element

Note: Additional elements may be analyzed following digestion by these protocols provided the method performance criteria specified in Section 13.0 of the SOP are met.

APPENDIX A - TABLES

TABLE III. ICP and FLAA Matrix Spike and Aqueous Laboratory Control Sample Levels

ELEMENT	Working LCS/MS Standard (mg/L)	Aqueous LCS/ MS Level * (ug/l)
Aluminum	100	2000
Antimony	25	500
Arsenic	100	2000
Barium	100	2000
Beryllium	2.5	50
Cadmium	2.5	50
Calcium	2500	50000
Chromium	10	200
Cobalt	25	500
Copper	12.5	250
Iron	50	1000
Lead	50	500
Magnesium	2500	50000
Manganese	25	500
Molybdenum	50	1000
Nickel	25	500
Phosphorous	500	10000
Potassium	2500	50000
Selenium	100	2000
Silver	2.5	50
Sodium	2500	50000
Thallium	100	2000
Vanadium	25	500
Zinc	25	500
Boron	50	1000
Tin	100	2000
Titanium	50	1000

* Levels shown indicate the spike concentration in the final digestate of the aqueous LCS or matrix spike based on the addition of 1.0 mL working spike (7.3) to 50 mL of sample.

TABLE IV. GFAA Matrix Spike and Aqueous LCS Spike Levels

APPENDIX A - TABLES

ELEMENT	Stock LCS/MS Standard (mg/L)	Working LCS/MS Standard (ug/L)	Aqueous LCS/MS Level * (ug/l)
Arsenic	400	2000	40
Selenium	400	2000	40
Lead	400	2000	40
Thallium	400	2000	40
Antimony	400	2000	40
Cadmium	40	200	4
Chromium	100	500	10
Silver	50	250	5

* Levels shown indicate the spike concentration in the final digestate of the aqueous LCS or matrix spike based on the addition of 1.0 mL working spike (7.4) to 50 mL of sample.

TABLE V. TCLP Reporting Limits, Regulatory Limits and Matrix Spike Levels

ELEMENT	RL (ug/L)	Regulatory Limit (ug/L)	Spike Level (ug/L)*
Arsenic	500	5000	5000
Barium	10000	100000	50000
Cadmium	100	1000	1000
Chromium	500	5000	5000
Lead	500	5000	5000
Selenium	250	1000	1000
Silver	500	5000	1000

* Levels shown indicate the spike concentration in the final digestate of the aqueous LCS or matrix spike based on the addition of 0.5 mL working spike (7.4) to 50 mL of sample.

APPENDIX A - TABLES

TABLE VI. Summary of Quality Control Requirements

QC PARAMETER	FREQUENCY	ACCEPTANCE CRITERIA	CORRECTIVE ACTION
Method Blank	One per sample preparation batch of up to 20 samples.	Refer to determinative SOPs: CORP-MT-0001 CORP-MT-0003	Redigest and reanalyze samples associated with the method blank.
Laboratory Control Sample (LCS)	One per sample preparation batch of up to 20 samples.	Refer to determinative SOPs: CORP-MT-0001 CORP-MT-0003	Redigest and reanalyze all samples associated with the LCS.
Matrix Spike	One per sample preparation batch of up to 20 samples.	Refer to determinative SOPs: CORP-MT-0001 CORP-MT-0003	Reprep not required unless preparation error suspected.
Matrix Spike Duplicate	See Matrix Spike	Refer to determinative SOPs: CORP-MT-0001 CORP-MT-0003	See Corrective Action for Matrix Spike.

APPENDIX B
CONTAMINATION CONTROL GUIDELINES

APPENDIX B – CONTAMINATION CONTROL GUIDELINES

APPENDIX B. CONTAMINATION CONTROL GUIDELINES

The following procedures are strongly recommended to prevent contamination:

All work areas used to prepare standards and spikes should be cleaned before and after each use.

All glassware should be washed with detergent and tap water and rinsed with 1:1 nitric acid followed by deionized water.

Proper laboratory housekeeping is essential in the reduction of contamination in the metals laboratory. All work areas must be kept scrupulously clean.

Powdered or Latex Gloves must not be used in the metals laboratory since the powder contains silica and zinc, as well as other metallic analytes. Only vinyl or nitrile gloves should be used in the metals laboratory.

Glassware should be periodically checked for cracks and etches and discarded if found. Etched glassware can cause cross contamination of any metallic analytes.

Autosampler trays should be covered to reduce the possibility of contamination. Trace levels of elements being analyzed in the samples can be easily contaminated by dust particles in the laboratory.

The following are helpful hints in the identification of the source of contaminants:

Reagents or standards can contain contaminants or be contaminated with the improper use of a pipette.

Improper cleaning of glassware can cause contamination.

Separate glassware if an unusually high sample is analyzed and soak with sulfuric acid prior to routine cleaning.

Controlled Copy
Copy No. _____
Implementation Date: _____

SOP No. CORP-IP-0003NC
Revision No. 1.3
Revision Date: 09/25/01
Page 1 of 37

STL STANDARD OPERATING PROCEDURE

TITLE: ACID DIGESTION OF AQUEOUS SAMPLES BY SW846 AND MCAWW 200 SERIES METHODS

(Supersedes: Revision 1.2 Dated 03/20/00)

Prepared by: Patrick O'Malley 9/26/01
Date

Reviewed by: Lisa McGall 9-26-01
Technology Specialist Date

Reviewed by: Pauline Cohen 9/26/01
Quality Assurance Manager Date

Reviewed by: [Signature] 9-27-01
Environmental Health and Safety Date

Reviewed by: Christy R. Pignatelli 9-28-01
Laboratory Director Date

Reviewed by: _____
Corporate Technology
Date

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use it for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: © 2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

Controlled Copy
Copy No. _____
Implementation Date: _____

SOP No. CORP-IP-0003NC
Revision No. 1.3
Revision Date: 09/25/01
Page 1 of 37

STL STANDARD OPERATING PROCEDURE

**TITLE: ACID DIGESTION OF AQUEOUS SAMPLES BY SW846 AND MCAWW
200 SERIES METHODS**

(Supersedes: Revision 1.2 Dated 03/20/00)

Prepared by: _____ Date _____
Reviewed by: _____ Date _____
 Technology Specialist
Reviewed by: _____ Date _____
 Quality Assurance Manager
Reviewed by: _____ Date _____
 Environmental Health and Safety
Reviewed by: _____ Date _____
 Laboratory Director
Reviewed by: _____ Date 10/01/01
 Corporate Technology
Date _____

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use if for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: © 2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

Implementation Date: _____

STL NORTH CANTON STANDARD OPERATING PROCEDURE

TITLE: PREPARATION AND ANALYSIS OF MERCURY IN AQUEOUS SAMPLES BY COLD VAPOR ATOMIC ABSORPTION, SW846 7470A AND MCAWW 245.1

(SUPERSEDES: REVISION 2.2, DATED 02/05/01)

Prepared by: _____ Date _____
Intranet Controlled

Approved by: _____ Date _____
Technical Specialist

Approved by: _____ Date _____
Quality Assurance Manager

Approved by: _____ Date _____
Environmental, Health and Safety Coordinator

Approved by: _____ Date _____
Laboratory Director

Approved by: _____ Date 5/17/01
Corporate Technology

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use if for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

TABLE OF CONTENTS

1. SCOPE AND APPLICATION	3
2. SUMMARY OF METHOD.....	3
3. DEFINITIONS	4
4. INTERFERENCES.....	4
5. SAFETY	5
6. EQUIPMENT AND SUPPLIES.....	6
7. REAGENTS AND STANDARDS	8
8. SAMPLE COLLECTION, PRESERVATION AND STORAGE.....	9
9. QUALITY CONTROL.....	9
10. CALIBRATION AND STANDARDIZATION.....	13
11. PROCEDURE	13
12. DATA ANALYSIS AND CALCULATIONS	18
13. METHOD PERFORMANCE.....	19
14. POLLUTION PREVENTION.....	20
15. WASTE MANAGEMENT	20
16. REFERENCES	20
17. MISCELLANEOUS (TABLES, APPENDICES, ETC. . .).....	20

LIST OF APPENDICES:

APPENDIX A - TABLES	26
APPENDIX B - STL NORTH CANTON Hg DATA REVIEW CHECKLIST.....	30
APPENDIX C - MSA GUIDANCE	32
APPENDIX D - TROUBLESHOOTING GUIDE	35
APPENDIX E- CONTAMINATION CONTROL GUIDELINES	37
APPENDIX F - PREVENTATIVE MAINTENANCE.....	39
APPENDIX G – INSTRUMENT SET-UP.....	41

1. SCOPE AND APPLICATION

- 1.1. This procedure describes the preparation and analysis of mercury (Hg, CAS # 7439-97-6) by Cold Vapor Atomic Absorption Spectroscopy (CVAA) using SW-846 Method 7470A and MCAWW Method 245.1.
- 1.2. The associated LIMs method codes are BL (Method 245.1) and O8 (Method 7470A).
- 1.3. CVAA analysis provides for the determination of total mercury (organic and inorganic). The combination of the oxidants, potassium permanganate and potassium persulfate, has been found to give 100% recovery with both types of compounds. Detection limits, sensitivity and optimum concentration ranges for mercury analysis will vary with the matrices, instrumentation and volume of sample used.
- 1.4. Method 7470A is applicable to the preparation and analysis of mercury in ground water, aqueous samples, TCLP, and other leachates/extracts. Certain solid and sludge type wastes may also be analyzed, however Method 7471A (see CORP-MT-0007NC) is usually the method of choice. All matrices require sample preparation prior to analysis.
- 1.5. Method 245.1 is applicable to the determination of mercury in drinking, surface and saline waters and domestic and industrial wastes. All matrices require sample preparation prior to analysis.
- 1.6. The STL North Canton reporting limit for mercury in aqueous matrices is 0.0002 mg/L except for TCLP or SPLP leachates for which the reporting limit is 0.002 mg/L.

2. SUMMARY OF METHOD

- 2.1. This SOP describes a technique for the determination of mercury in solution. The procedure is a physical method based on the absorption of radiation at 253.7 nm by mercury vapor. A representative portion of the sample is digested in sulfuric and nitric acids. Organic mercury compounds are oxidized with potassium permanganate and potassium persulfate and the mercury reduced to its elemental state with stannous chloride and aerated from solution in a closed system. The mercury vapor passes through a cell positioned in the light path of an atomic absorption spectrophotometer. Absorbance is measured as a function of mercury concentration. Concentration of the analyte in the sample is determined by comparison of the sample absorbance to the calibration curve (absorbance vs. concentration).

3. DEFINITIONS

- 3.1. Dissolved Metals: Those elements which pass through a 0.45 um membrane. (Sample is acidified after filtration).
- 3.2. Suspended Metals: Those elements which are retained by a 0.45 um membrane.
- 3.3. Total Metals: The concentration determined on an unfiltered sample following digestion.

4. INTERFERENCES

Chemical and physical interferences may be encountered when analyzing samples using this method.

- 4.1. Potassium permanganate, which is used to breakdown organic mercury compounds also eliminates possible interferences from sulfide. Concentrations as high as 20 mg/L of sulfide as sodium sulfide do not interfere with the recovery of inorganic mercury from reagent water.
- 4.2. Copper has also been reported to interfere; however, copper concentrations as high as 10 mg/L had no effect on the recovery of mercury from spiked samples.
- 4.3. Chlorides can cause a positive interference. Sea waters, brines and industrial effluents high in chlorides require additional permanganate (as much as 25 mL) because, during the oxidation step, chlorides are converted to free chlorine, which also absorbs radiation at 253.7 nm. Care must be taken to ensure that free chlorine is absent before the mercury is reduced and swept into the cell. This is accomplished by adding excess hydroxylamine reagent (25 mL) and purging the sample headspace before stannous chloride is added. Both inorganic and organic mercury spikes have been quantitatively recovered from seawater using this technique.

Note: Sufficient addition of permanganate is apparent when the purple color persists at least 15 minutes. Some samples may require dilution prior to digestion due to extremely high concentrations of chloride.

- 4.4. Interference from certain volatile organic materials that absorb at this wavelength may also occur. If suspected, a preliminary run without stannous chloride can determine if this type of interference is present. While the possibility of absorption from certain organic substances present in the sample does exist, this problem is not routinely encountered. This is mentioned only to caution the analyst of the possibility. If this condition is found to exist, the mercury concentration in the sample

can be determined by subtracting the result of the sample run without the reducing reagent (stannous chloride) from that obtained with the reducing reagent.

- 4.5. Samples containing high concentrations of oxidizing organic materials, as evidenced by high COD levels, may not be completely oxidized by this procedure. When this occurs the recovery of mercury will be low. The problem can be eliminated by reducing the volume of original sample used.
- 4.6. The most common interference is laboratory contamination, which may arise from impure reagents, dirty glassware, improper sample transfers, dirty work areas, etc. Be aware of potential sources of contamination and take appropriate measures to minimize or avoid them.

5. SAFETY

- 5.1. Procedures shall be carried out in a manner that protects the health and safety of all STL North Canton associates.
- 5.2. Eye protection that satisfies ANSI Z87.1 (as per the Chemical Hygiene Plan), laboratory coat, and appropriate gloves must be worn while samples, standards, solvents, and reagents are being handled. Disposable gloves that have been contaminated will be removed and discarded; other gloves will be cleaned immediately.
- 5.3. The health and safety hazards of many of the chemicals used in this procedure have not been fully defined. Additional health and safety information can be obtained from the Material Safety Data Sheets (MSDS) maintained in the laboratory. The following specific hazards are known:
 - 5.3.1. The following materials are known to be **corrosive**:

hydrochloric acid, nitric acid and sulfuric acid.
 - 5.3.2. The following materials are known to be **oxidizing agents**:

nitric acid, potassium permanganate, potassium persulfate and magnesium perchlorate.
 - 5.3.3. Mercury is a highly toxic element that must be handled with care. The analyst must be aware of the handling and clean up techniques before working with mercury. Since mercury vapor is toxic, precaution must be taken to avoid its inhalation, ingestion or absorption through skin. All lines should be checked

for leakage and the mercury vapor must be vented into a hood or passed through a mercury absorbing media such as:

5.3.3.1. Equal volumes of 0.1 M KMnO_4 and 10% H_2SO_4 , or

5.3.3.2. Iodine, 0.25%, in a 3% KI solution.

5.3.4. Magnesium sulfate is known to be a reproductive toxin (mutagen).

- 5.4. Exposure to chemicals must be maintained **as low as reasonably achievable**. Therefore, unless they are known to be non-hazardous, all samples should be opened, transferred and prepared in a fume hood, or under other means of mechanical ventilation. Solvent and waste containers will be kept closed unless transfers are being made.
- 5.5. All work must be stopped in the event of a known or potential compromise to the health and safety of a STL North Canton associate. The situation must be reported **immediately** to a laboratory supervisor.
- 5.6. Do not look directly into the beam of the Hg lamp. The UV light that these lamps radiate is harmful to the eyes.
- 5.7. Cylinders of compressed gas must be handled with caution, in accordance with local regulations. It is recommended that, wherever possible, cylinders are located outside the laboratory and the gas led to the instrument through approved lines.
- 5.8. The CVAA apparatus must be properly vented to remove potentially harmful fumes generated during sample analysis.

6. EQUIPMENT AND SUPPLIES

- 6.1. Temperature controlled water bath (capable of maintaining a temperature of 90-95 °C) or autoclave that is able to obtain conditions of 15 lbs., 120 °C for 15 minutes.

6.2. Atomic Absorption Spectrophotometer equipped with:

- 6.2.1. Absorption Cell with quartz end windows perpendicular to the longitudinal axis. Dimensions of the cell must result in sufficient sensitivity to meet the SOP defined reporting limit. The quartz windows must be maintained to provide accurate measurements. Any scratches or fingerprints can alter the absorption of UV radiation.
- 6.2.2. Mercury specific hollow cathode lamp (HCL) or electrodeless discharge lamp (EDL).
- 6.2.3. Peristaltic pump, which can deliver 1 L/min, air.
- 6.2.4. Flowmeter capable of measuring an airflow of 1 L/min.
- 6.2.5. Recorder or Printer.
- 6.2.6. Aeration Tubing: A straight glass frit having a coarse porosity and Tygon tubing is used for the transfer of mercury vapor from the sample bottle to the absorption cell and return.
- 6.2.7. Drying device (a drying tube containing magnesium perchlorate or magnesium sulfate and/or a lamp with a 60 W bulb) to prevent condensation in cell. The lamp is positioned to shine on the absorption cell maintaining the air temperature in the cell about 10 °C above room temperature. Other drying devices that achieve the same purpose are also acceptable (i.e., Gortex filter).

NOTE: Instruments designed specifically for the measurement of mercury using the cold vapor technique may be substituted for the atomic absorption spectrophotometer.

- 6.3. BOD bottles or equivalent.
- 6.4. Nitrogen or argon gas supply, welding grade, or equivalent.
- 6.5. Calibrated automatic pipettes or Class A glass volumetric pipettes.
- 6.6. Class A volumetric flasks.
- 6.7. Thermometer (capable of accurate readings at 95 °C).
- 6.8. Disposable cups or tubes.

7. REAGENTS AND STANDARDS

- 7.1. Reagent water must be produced by a Millipore DI system or equivalent. Reagent water must be free of the analytes of interest as demonstrated through the analysis of method blanks.
- 7.2. Stock (10 ppm) mercury standards (in 10% HNO₃) are purchased as custom STL North Canton solutions. All standards must be stored in FEP fluorocarbon or previously unused polyethylene or polypropylene bottles. Stock standard solutions must be replaced prior to the expiration date provided by the manufacturer. If no expiration date is provided, the stock solutions may be used for up to one year and must be replaced sooner if verification from an independent source indicates a problem.
- 7.3. Working mercury standard (0.1 ppm): Take 1 mL of the stock mercury standard (7.2) and dilute to 100 mL with reagent water. The working mercury standard must be made daily and must be prepared in a matrix of 0.15% HNO₃. This acid (150 μ L of concentrated HNO₃) must be added to the flask/bottle before the addition of the stock standard aliquot.
- 7.4. The calibration standards listed in Table I must be prepared fresh daily from the working standard (7.3) by transferring 0, 0.2, 0.5, 1.0, 5.0 and 10.0 mL aliquots of the working mercury standard into 100 mL flasks and diluting to volume with reagent water.

Note: Alternate approaches to standard preparation may be taken and alternate volumes of standard may be prepared as long as the accuracy and final standard concentrations as detailed in Table I are maintained. For example, automated mercury systems do not require 100 mL of standard and therefore smaller volumes may be generated to reduce waste generation.
- 7.5. The initial calibration verification standard must be made from a different stock solution than that of the calibration standards.
- 7.6. Refer to Table I (Appendix A) for details regarding the working standard concentrations for calibration, calibration verification and spiking solutions. All standards must be processed through the entire analytical procedure including sample preparation.
- 7.7. Nitric acid (HNO₃), concentrated, trace metal grade or better.

Note: If a high reagent blank is obtained, it may be necessary to distill the nitric acid.

- 7.8. Sulfuric acid (H₂SO₄), concentrated, trace metal grade or better.
- 7.9. Stannous sulfate solution: Add 25 g of stannous sulfate to 250 mL of 0.5 N sulfuric acid. This mixture is a suspension and should appear cloudy. This solution should be made daily and should be stirred continuously during use.

Note: Stannous chloride may be used in place of stannous sulfate. Prepare the stannous chloride solution according to the recommendations provided by the instrument manufacturer.

- 7.10. Sodium chloride-hydroxylamine hydrochloride solution: Add 12 g of sodium chloride and 12 g of hydroxylamine hydrochloride to every 100 mL of reagent water.

Note: Hydroxylamine sulfate may be used in place of hydroxylamine hydrochloride.

- 7.11. Potassium permanganate, 5% solution (w/v): Dissolve 5 g of potassium permanganate for every 100 mL of reagent water.
- 7.12. Potassium persulfate, 5% solution (w/v): Dissolve 5 g of potassium persulfate for every 100 mL of reagent water.

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE

- 8.1. Sample holding time for mercury is 28 days from time of collection to the time of analysis.
- 8.2. Aqueous samples are preserved with nitric acid to a pH of <2 and may be stored in either plastic or glass. Refrigeration is not required. Preservation must be verified prior to analysis.

9. QUALITY CONTROL

Table II (Appendix A) provides a summary of quality control requirements including type, frequency, acceptance criteria and corrective action.

9.1. Initial Demonstration of Capability

Prior to the analysis of any analyte using 7470A or the 245.1, the following requirements must be met.

- 9.1.1. Method Detection Limit (MDL) - An MDL must be determined for each analyte/matrix prior to the analysis of any samples. The MDL is determined using seven replicates of reagent water, spiked with all the analytes of interest,

that have been carried through the entire analytical procedure. MDLs must be redetermined on an annual basis in accordance with 40 CFR Part 136 Appendix B requirements. The spike level must be between the calculated MDL and 10X the MDL to be valid. The result of the MDL determination must be below the STL North Canton reporting limit.

9.1.2. Initial Demonstration Study - This requires the analysis of four QC check samples. The QC check sample is a well-characterized laboratory generated sample used to monitor method performance. The results of the initial demonstration study must be acceptable before analysis of samples may begin.

9.1.2.1. Four aliquots of the check sample (LCS) are prepared and analyzed using the procedures detailed in this SOP and the determinative SOPs.

- 9.2. Preparation Batch - A group of up to 20 samples that are of the same matrix and are processed together using the same procedures and reagents. The preparation batch must contain a method blank, a LCS and a matrix spike/matrix spike duplicate. In some cases, at client request, it may be appropriate to process a matrix spike and sample duplicate in place of the MS/MSD. If clients specify specific samples for MS/MSD, the batch may contain multiple MS/MSD pairs.
- 9.3. Sample Count - Laboratory generated QC samples (Method Blanks, LCS, and MS/MSDs) are not included in the sample count for determining the size of a preparation batch.
- 9.4. Method Blank (MB) - One method blank must be processed with each preparation batch. The method blank consists of reagent water containing all reagents specific to the method that is carried through the entire analytical procedure, including preparation and analysis. The method blank is used to identify any system and process interferences or contamination of the analytical system that may lead to the reporting of elevated analyte concentrations or false positive data. The method blank should not contain any analyte of interest at or above the reporting limit or at or above 10% of the measured concentration of that analyte in associated samples, whichever is higher (sample result must be a minimum of 20 times higher than the blank contamination level).
- Repreparation and reanalysis of all samples associated with an unacceptable method blank is required when reportable concentrations are determined in the samples (see exception noted above).

-
- If there is no analyte greater than the RL in the samples associated with an unacceptable method blank, the data may be reported with qualifiers. **Such action must be taken in consultation with the client and must be addressed in the project narrative.**
 - If the above criteria are not met and reanalysis is not possible, then the sample data must be qualified. **This anomaly must be addressed in the project narrative and the client must be notified.**
- 9.5. Laboratory Control Sample (LCS) - One aqueous LCS must be processed with each preparation batch. The LCS is used to monitor the accuracy of the analytical process. On going monitoring of the LCS results provides evidence that the laboratory is performing the method within acceptable accuracy and precision guidelines. The LCS must be carried through the entire analytical procedure. If the LCS is outside established control limits the system is out of control and corrective action must occur. Until in-house control limits are established, a control limit of 80 - 120% recovery must be applied.
- In the instance where the LCS recovery is $> 120\%$ and the sample results are $< RL$, the data may be reported with qualifiers. Such action must be taken in consultation with the client and must be addressed in the case narrative.
 - In the event that an MS/MSD analysis is not possible, a Laboratory Control Sample Duplicate (LCSD) must be analyzed. The RPD of the LCS and LCSD must be compared to the matrix spike RPD limits.
 - Corrective action will be repreparation and reanalysis of the batch unless the client agrees that other corrective action is acceptable.
- 9.6. Matrix Spike/Matrix Spike Duplicate (MS/MSD) - One MS/MSD pair must be processed for each preparation batch. A matrix spike (MS) is a field sample to which known concentrations of target analytes have been added. A matrix spike duplicate (MSD) is a second aliquot of the same sample (spiked identically as the MS) prepared and analyzed along with the sample and matrix spike. Some client specific data quality objectives (DQO's) may require the use of sample duplicates in place of or in addition to MS/MSD's. The MS/MSD results are used to determine the effect of a matrix on the precision and accuracy of the analytical process. Due to the potential variability of the matrix of each sample, these results may have immediate bearing only on the specific sample spiked. Samples identified as field blanks cannot be used for MS/MSD analysis. Spiking levels are provided in Table I (Appendix A).
- If analyte recovery or RPD falls outside the acceptance range, the recovery of that analyte must be in control for the LCS. Until in-house control limits are

established, a control limit of 75 - 125 % recovery and 20% RPD must be applied to the MS/MSD. If the LCS recovery is within limits, then the laboratory operation is in control and the results may be accepted. If the recovery of the LCS is outside limits, corrective action must be taken. Corrective action will include repreparation and reanalysis of the batch. MS/MSD results, which fall outside the control limits, must be addressed in the narrative.

- If the native analyte concentration in the MS/MSD exceeds 4 times the spike level for that analyte, the recovery data are reported as NC (i.e., not calculated). If the reporting software does not have the ability to report NC then the actual recovery must be reported and narrated as follows: "Results outside of limits do not necessarily reflect poor method performance in the matrix due to high analyte concentrations in the sample relative to the spike level."
 - If an MS/MSD is not possible due to limited sample volume, then a laboratory control sample duplicate (LCSD) should be analyzed. The RPD of the LCS and LCSD must be compared to the matrix spike RPD limits.
- 9.7. Initial Calibration Verification (ICV/ICB) - Calibration accuracy is verified by analyzing a second source standard (ICV). The ICV result must fall within 10% of the true value for that solution. An ICB is analyzed immediately following the ICV to monitor low level accuracy and system cleanliness. The ICB result must fall within +/- the reporting limit (RL) from zero. If either the ICV or ICB fail to meet criteria, the analysis should be terminated, the problem corrected and the instrument recalibrated. (See Section 11.2.11 and Section 11.2.12 for required run sequence). If the cause of the ICV or ICB failure was not directly instrument related the corrective action will include repreparation of the ICV, ICB, CRA, CCV, and CCB with the calibration curve.
- 9.8. Continuing Calibration Verification (CCV/CCB) - Calibration accuracy is monitored throughout the analytical run through the analysis of a known standard after every 10 samples. The CCV must be a mid-range standard at a concentration other than that of the ICV. The CCV result must fall within 20% of the true value for that solution. A CCB is analyzed immediately following each CCV. (See Section 11.2.11 and 11.2.12 for required run sequence.) The CCB result must fall within +/- RL from zero. Each CCV and CCB analyzed must reflect the conditions of analysis of all associated samples. Sample results may only be reported when bracketed by valid ICV/CCV and ICB/CCB pairs.
- 9.9. Method of Standard Addition (MSA) -This technique involves adding known amounts of standard to one or more aliquots of the sample prior to preparation. This

technique compensates for a sample interferent that may enhance or depress the analyte signal, thus producing a different slope from that of the calibration standards. It will not correct for additive interferences, which cause a baseline shift. Refer to Section 11.2.13 for additional information on when full 4 point MSA is required as well as Appendix C for specific MSA requirements.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Calibration standards must be processed through the preparation procedure as described in Section 11.1.
- 10.2. Due to the differences in preparation protocols separate calibration and calibration verification standards must be prepared for aqueous and solid matrices.
- 10.3. Calibration must be performed daily (every 24 hours) and each time the instrument is set up. The instrument calibration date and time must be included in the raw data.
- 10.4. Set up the instrument with the operating parameters recommended by the manufacturer. Allow the instrument to become thermally stable before beginning calibration (approximately 30 minutes of warm-up is required). Refer to the facility specific instrument SOP and CVAA instrument manual for detailed setup and operation protocols.
- 10.5. Calibrate the instrument according to instrument manufacturer's instructions, using a minimum of five standards and a blank. One standard must be at the STL North Canton reporting limit. Analyze standards in ascending order beginning with the blank. Refer to Section 7.5 and Table I for additional information on preparing calibration standards and calibration levels.
- 10.6. The calibration curve must have a correlation coefficient of ≥ 0.995 or the instrument shall be stopped and recalibrated prior to running samples. Sample results can not be reported from a curve with an unacceptable correlation coefficient.
- 10.7. Refer to Section 9.0 for calibration verification procedures, acceptance criteria and corrective actions.

11. PROCEDURE

11.1. Sample Preparation:

- 11.1.1. All calibration and calibration verification standards (ICV, ICB, CCV, CCB) are processed through the digestion procedure as well as the field samples.

Transfer 0, 0.2, 0.5, 1.0, 5.0 and 10.0 mL aliquots of the working standard (7.3) into a series of 100 ml class A volumetrics, then dilute to volume. For the ICV, use a 2.5 ml aliquot of the working standard. The ICV working standard must be made from a source other than that used for the calibration standards.

- 11.1.2. Transfer 100 mL of well-mixed sample or standard to a clean sample digestion bottle.

Note: Reduced sample volumes can be used as long as a representative sample can be obtained and the reagent levels are adjusted to maintain the same sample to reagent ratio. All samples and standards must be processed similarly.

Note: Spiking is done before the addition of acids or reagents.

- 11.1.3. Add 5 mL of concentrated H_2SO_4 and 2.5 mL of concentrated HNO_3 mixing after each addition.

- 11.1.4. Add 15 mL of potassium permanganate solution. For samples high in organic materials or chlorides, additional permanganate may be added. Shake and add additional portions of permanganate solution until a purple color persists for at least 15 minutes. If after the addition of up to 25-mL additional permanganate the color does not persist, sample dilution prior to reanalysis may be required.

Note: When performing analyses using automated vs. manual techniques the sample dilution resultant from the addition of more than the original aliquot of permanganate solution must be compensated for by the addition of the same volume of permanganate to all other associated samples and standards in the run. In instances, where this is not feasible, the addition of excess reagent can be addressed through mathematical correction of the results to account for the resultant dilution effect.

- 11.1.5. Add 8 mL of potassium persulfate solution and heat for two hours in a water bath at 90 - 95 °C.

NOTE: Alternatively, for RCRA analyses using 7470A, samples may be digested using an autoclave for 15 minutes at 120 °C and 15 lbs.

11.1.6. Cool samples.

11.2. Sample Analysis:

- 11.2.1. Because of differences between various makes and models of CVAA instrumentation, no detailed operating instructions can be provided. Refer to the facility specific instrument operating SOP and the CVAA instrument manual for detailed setup and operation protocols.
- 11.2.2. All labs are required to detail the conditions/programs utilized for each instrument within the facility specific instrument operation SOP.
- 11.2.3. When ready to begin analysis, add 6 mL of sodium chloride-hydroxylamine hydrochloride solution to the samples to reduce the excess permanganate (the permanganate has been reduced when no purple color remains). Add this solution in 6-mL increments until the permanganate is completely reduced.
- 11.2.4. Manual determination:
- 11.2.4.1. Treating each sample individually, purge the headspace of the sample bottle for at least one minute.
- 11.2.4.2. Add 5 mL of stannous chloride solution and immediately attach the bottle to the aeration apparatus.
- 11.2.4.3. Allow the sample to stand quietly without manual agitation while the sample is aerated (1 L/min flow). Monitor the sample absorbance during aeration. When the absorbance reaches a maximum and the signal levels off, open the bypass valve and continue aeration until the absorbance returns to its baseline level. Close the bypass valve and remove the aeration device.
- 11.2.4.4. Place the aeration device into 100 mL of 1% HNO₃ and allow to bubble rinse until the next sample is analyzed.
- 11.2.5. Automated determination: Follow instructions provided by instrument manufacturer.
- 11.2.6. Perform a linear regression analysis of the calibration standards by plotting maximum response of the standards vs. concentration of mercury. Determine the mercury concentration in the samples from the linear

regression fit of the calibration curve. Calibration using computer or calculation based regression curve fitting techniques on concentration/response data is acceptable.

- 11.2.7. All measurements must fall within the defined calibration range to be valid. Dilute and reanalyze all samples for analytes that exceed the highest calibration standard.
- 11.2.8. If the sample results are negative and the absolute value of the negative result is greater than the reporting limit, the sample must be diluted and reanalyzed.
- 11.2.9. The samples must be allowed to cool to room temperature prior to analysis or a decrease in the response signal can occur.
- 11.2.10. Baseline correction is acceptable as long as it is performed after every sample or after the CCV and CCB; resloping is acceptable as long as it is immediately preceded and followed by a compliant CCV and CCB.
- 11.2.11. The following analytical sequence must be used with 7470A and 245.1:

Instrument Calibration

ICV

ICB

Maximum 10 samples

CCV

CCB

Repeat sequence of 10 samples between CCV/CCB pairs as required to complete run

CCV

CCB

Refer to Quality Control Section 9.0 and Table II (Appendix A) for quality control criteria to apply to Methods 7470A and 245.1.

Note: Samples include the method blank, LCS, MS, MSD, duplicate, field samples and sample dilutions.

- 11.2.12. The following run sequence is consistent with 7470A, CLP and 245.1 and may be used as an alternate to the sequence in 11.2.11. This run sequence is recommended if multiple method requirements must be accommodated

in one analytical run:

Instrument Calibration

ICV

ICB

CRA*

CCV

CCB

10 samples

CCV

CCB

Repeat sequence of 10 samples between CCV/CCB pairs as required to complete run.

CCV

CCB

Refer to the appropriate CLP SOP (CORP-MT-0006) for quality control requirements for QC samples.

* Refer to the CLP SOP for information on the CRA.

11.2.13. For TCLP samples, full four point MSA will be required if all of the following conditions are met:

- 1) recovery of the analyte in the matrix spike is not at least 50%,
- 2) the concentration of the analyte does not exceed the regulatory level, and,
- 3) the concentration of the analyte is within 20% of the regulatory level.

The reporting and matrix spike levels for TCLP analyses are detailed in Table I (Appendix A). Appendix E provides guidance on performing MSA analyses. For TCLP mercury determinations, MSA spikes must be added prior to sample preparation.

11.3. To facilitate the early identification of QC failures and samples requiring rerun it is strongly recommended that sample data are reviewed periodically throughout the run.

11.4. Guidelines are provided in the appendices on procedures to minimize contamination of samples and standards, preventive maintenance and troubleshooting.

- 11.5. One time procedural variations are allowed only if deemed necessary in the professional judgment of supervision to accommodate variation in sample matrix, radioactivity, chemistry, sample size, or other parameters. Any variation in procedure shall be completely documented using a Nonconformance Memo and is approved by a Technical Specialist and QA Manager. If contractually required, the client shall be notified. The Nonconformance Memo shall be filed in the project file.
- 11.6. Any unauthorized deviations from this procedure must also be documented as a nonconformance, with a cause and corrective action described.

12. DATA ANALYSIS AND CALCULATIONS

- 12.1. ICV percent recoveries are calculated according to the equation:

$$\%R = 100 \left(\frac{\text{Found}(ICV)}{\text{True}(ICV)} \right)$$

- 12.2. CCV percent recoveries are calculated according to the equation:

$$\%R = 100 \left(\frac{\text{Found}(CCV)}{\text{True}(CCV)} \right)$$

- 12.3. Matrix spike recoveries are calculated according to the following equation:

$$\%R = 100 \left(\frac{SSR - SR}{SA} \right)$$

Where:

SSR = Spike Sample Result

SR = Sample Result

SA = Spike Added

- 12.4. The relative percent difference (RPD) of matrix spike/matrix spike duplicates or sample duplicates are calculated according to the following equations:

$$RPD = 100 \left[\frac{|MSD - MS|}{\left(\frac{MSD + MS}{2} \right)} \right]$$

Where:

MS = determined spiked sample concentration

MSD = determined matrix spike duplicate concentration

$$RPD = 100 \left[\frac{|DU1 - DU2|}{\left(\frac{DU1 + DU2}{2} \right)} \right]$$

Where:

DU1 = Sample result

DU2 = Sample duplicate result

- 12.5. The final concentration for an aqueous sample is calculated as follows:

$$mg/L = C \times D$$

Where:

C = Concentration (mg/L) from instrument readout

D = Instrument dilution factor

- 12.6. The LCS percent recovery is calculated according to the following equation:

$$\%R = 100 \left(\frac{Found(LCS)}{True(LCS)} \right)$$

- 12.7. Appropriate factors must be applied to sample values if dilutions are performed.

- 12.8. Sample results should be reported with up to three significant figures in accordance with the STL North Canton significant figure policy.

13. METHOD PERFORMANCE

- 13.1. Each laboratory must have initial demonstration of performance data on file for each analyte of interest as described in Section 9.0.

- 13.2. Method performance is determined by the analysis of method blanks, laboratory control samples, matrix spike and matrix spike duplicate samples. The matrix spike recovery should fall within +/- 25 % and the matrix spike duplicates should compare within 20% RPD. The method blanks must meet the criteria in Section 9.3. The laboratory control sample should recover within 20% of the true value until in house limits are established.

13.3. Training Qualification:

The group/team leader has the responsibility to ensure that this procedure is performed by an associate who has been properly trained in its use and has the required experience.

14. **POLLUTION PREVENTION**

14.1. This method allows for the proportional reduction of sample and reagent volumes to decrease waste generation.

15. **WASTE MANAGEMENT**

15.1. Waste generated in the procedure must be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Director should be contacted if additional information is required.

16. **REFERENCES**

16.1. Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW-846, 3rd Edition, Final Update II, Revision I, September 1994, Method 7470A (Mercury).

16.2. "Methods for the Chemical Analysis of Water and Wastes", EPA-600/4-79-020, U.S.EPA, August 1983, Method 245.1.

16.3. U.S.EPA Statement of Work for Inorganics Analysis, ILMO3.0 and ILMO4.0.

16.4. QA-003, STL North Canton QC Program.

16.5. QA-004, Rounding and Significant Figures.

16.6. QA-005, Method Detection Limits.

17. **MISCELLANEOUS (TABLES, APPENDICES, ETC. . .)**

17.1. Modifications/Interpretations from reference method.

17.1.1. Modifications from both 7470A and 245.1.

17.1.1.1. The 200 series methods and Chapter 1 of SW846 specify the use of reagent water with a purity equivalent to ASTM Type II water. This SOP specifies the use of a Millipore DI system or equivalent

to produce reagent water. This SOP requires that reagent water must be free of the analytes of interest as demonstrated through the analysis of method blanks.

- 17.1.1.2. This SOP allows for the use of reduced sample volumes to decrease waste generation. Reagent levels are adjusted to maintain the same ratios as stated in the source methods. According to a letter from Robert Booth of EPA EMSL-Cinn to David Payne of EPA Region V, "Reduction in sample size and appropriate corresponding reduction in sample volume is not considered a significant change in the methodology."
- 17.1.1.3. The alternate run sequence presented in Section 11.2.12 is consistent with method requirements. An additional QC analysis (CRA) was added to accommodate the CLP protocol requirements.

17.1.2. Modifications from Method 7470A

- 17.1.2.1. Chapter 1 of SW-846 states that the method blank should not contain any analyte of interest at or above the MDL. This SOP states that the method blank must not contain any analyte of interest at or above the reporting limit.
- 17.1.2.2. Documentation is on file from EPA's Office of Solid Waste (Olliver Fordham 11/28/95) regarding the acceptance of the autoclave as an equivalent heating device to the water bath. In his letter, Mr. Fordham cited the CLP water protocol 245.1 CLP-M and therefore the operating parameters from that method were adopted for 7470A (15 minutes at 120 °C and 15 lbs.).

17.1.3. Modifications from 245.1

- 17.1.3.1. Method 245.1 Section 9.3 states concentrations should be reported as follows: Between 1 and 10 ug/L, one decimal; above 10 ug/L, to the nearest whole number. STL North Canton reports all Hg results under this SOP to two significant figures.

17.2. Modifications from previous SOP

17.2.1. Chapter 1 of SW-846 states that the method blank should not contain any analyte of interest at or above the MDL. This SOP states that the method blank must not contain any analyte of interest at or above the reporting limit.

17.2.2. Documentation is on file from EPA's Office of Solid Waste (Olliver Fordham 11/28/95) regarding the acceptance of the autoclave as an equivalent heating device to the water bath. In his letter, Mr. Fordham cited the CLP water protocol 245.1 CLP-M and therefore the operating parameters from the method were adopted for 7470A (15 minutes at 120° C and 15 lbs.).

17.3. Facility Specific SOPs

Each facility shall attach a list of facility specific SOPs or approved attachments (if applicable) which are required to implement this SOP or which are used in conjunction with this SOP. If no facility specific SOPs or amendments are to be attached, a statement must be attached specifying that there are none. Refer to the Appendices for any facility specific information required to support this SOP.

17.4. Documentation and Record Management

The following documentation comprises a complete CVAA raw data package:

- Raw data (direct instrument printout)
- Run log printout from instrument software where this option is available or manually generated run log. (A bench sheet may be substituted for the run log as long as it contains an accurate representation of the analytical sequence).
- Data review checklist - See Appendix B
- Standards Documentation (source, lot, date).
- Copy of digestion log.
- Non-conformance summary (if applicable).

Figure 1. Aqueous Sample Preparation - Mercury





Figure 2. CVAA Mercury Analysis



APPENDIX A

TABLES

**TABLE I. MERCURY REPORTING LIMITS, CALIBRATION STANDARD*, QC
 STANDARD AND SPIKING LEVELS (MG/L)**

Standard Aqueous RL	0.0002
TCLP RL	0.002
Std 0	0
Std 1	0.0002
Std 2	0.0005
Std 3	0.001
Std 4	0.005
Std 5 **	0.010
ICV	0.001 or 0.0025 ***
LCS/CCV	0.0025 or 0.005 ***
Aqueous MS	0.001
TCLP MS	0.005

- * SOP specified calibration levels must be used unless prevented by the instrument configuration or client specific requirements. Deviations from specified calibration levels must be documented in the facility specific instrument operation SOP and must be approved by the facility technical manager and Quality Assurance Manager.
- ** Optional standard which may be used to extend the calibration range as allowed by the instrument configuration. If the instrument configuration prevents the use of 6 standards, the 2-ppb standard may be eliminated in favor of the 10 ppb standard.
- *** Concentration level dependent on high calibration standard used. CCV must be 50% of high standard concentration and ICV must be 20-25% of high standard concentration.

TABLE II. Summary Of Quality Control Requirements

QC PARAMETER	FREQUENCY *	ACCEPTANCE CRITERIA	CORRECTIVE ACTION
ICV	Beginning of every analytical run.	90-110 % recovery.	Terminate analysis; Correct the problem; Recalibrate or reprep with calibration curve. (See Section 9.7).
ICB	Beginning of every analytical run, immediately following the ICV.	The result must be within +/- RL from zero.	Terminate analysis; Correct the problem; Recalibrate or reprep with calibration curve. (See Section 9.7).
CCV	Every 10 samples and at the end of the run.	80 - 120 % recovery.	Terminate analysis; Correct the problem; Recalibrate and rerun all samples not bracketed by acceptable CCV or reprep with calibration curve. (See Section 9.8).
CCB	Immediately following each CCV.	The result must be within +/- RL from zero.	Terminate analysis; Correct the problem; Recalibrate and rerun all samples not bracketed by acceptable CCB or reprep with calibration curve. (See Section 9.8).
Method Blank	One per sample preparation batch of up to 20 samples.	The result must be less than or equal to the RL. Sample results greater than 20x the blank concentration are acceptable. Samples for which the contaminant is < RL do not require redigestion (See Section 9.4).	Redigest and reanalyze samples. Note exceptions under criteria section. See Section 9.4 for additional requirements.

*See Sections 11.2.11 and 11.2.12 for exact run sequence to be followed.

TABLE II. Summary of Quality Control Requirements (Continued)

QC PARAMETER	FREQUENCY	ACCEPTANCE CRITERIA	CORRECTIVE ACTION
Laboratory Control Sample (LCS)	One per sample preparation batch of up to 20 samples.	Aqueous LCS must be within 80 - 120% recovery or in-house control limits.	Terminate analysis; Correct the problem; Redigest and reanalyze all samples associated with the LCS (see Section 9.5).
Matrix Spike	One per sample preparation batch of up to 20 samples.	75 - 125 % recovery or in-house control limits. If the MS/MSD is out for an analyte, it must be in control in the LCS.	In the absence of client specific requirements, flag the data; no flag required if the sample level is > 4x the spike added. (see Section 9.6) For TCLP see Section 11.2.13
Matrix Spike Duplicate	See Matrix Spike	75 - 125 % recovery or in-house control limits; RPD ≤ 20%. (See MS)	See Corrective Action for Matrix Spike.

APPENDIX B
EXAMPLE
STL NORTH CANTON Hg DATA REVIEW CHECKLIST

Example
STL North Canton Hg Data Review Checklist

Run/Project Information

Run Date: _____ Analyst: _____ Instrument: _____
 Prep Batches Run: _____

Circle Methods used: 7470A / 245.1 : CORP-MT-0005 Rev 1 7471 / 245.5 : CORP-MT-0007 Rev 1
 CLP - AQ : CORP-MT-0006 Rev 0 CLP - SOL : CORP-MT-0008 Rev 0

Review Items

A. Calibration/Instrument Run QC	Yes	No	N/A	2ndLevel
1. Instrument calibrated per manufacturer's instructions and at SOP specified levels?				
2. ICV/CCV analyzed at appropriate frequency and within control limits?				
3. ICB/CCB analyzed at appropriate frequency and within +/- RL or +/- CRDL (CLP)?				
4. CRA run (CLP only)?				
B. Sample Results				
1. Were samples with concentrations > the high calibration standard diluted and reanalyzed?				
2. All reported results bracketed by in control QC?				
3. Sample analyses done within holding time?				
C. Preparation/Matrix QC				
1. LCS done per prep batch and within QC limits?				
2. Method blank done per prep batch and < RL or CRDL (CLP)?				
3. MS run at required frequency and within limits?				
4. MSD or DU run at required frequency and RPD within SOP limits?				
D. Other				
1. Are all nonconformances documented appropriately?				
2. Current IDL/MDL data on file?				
3. Calculations and Transcriptions checked for error?				
4. All client/ project specific requirements met?				
5. Date of analysis verified as correct?				

Analyst: _____ Date: _____

Comments:

2nd Level Reviewer : _____ Date: _____

APPENDIX C
MSA GUIDANCE


APPENDIX C. MSA GUIDANCE

Method of Standard Addition

Four equal volume aliquots of sample are measured and known amounts of standards are added to three aliquots. The fourth aliquot is the unknown and no standard is added to it. The concentration of standard added to the first aliquot should be 50% of the expected concentration. The concentration of standard added to the second aliquot should be 100% of the expected concentration and the concentration of standard added to the third aliquot should be 150% of the expected concentration. The volume of the unspiked and spiked aliquots should be the same (i.e., the volume of the spike added should be negligible in relation to the volume of sample).

To determine the concentration of analyte in the sample, the absorbance (or response) of each solution is determined and a linear regression performed. On the vertical axis the absorbance (or response) is plotted versus the concentrations of the standards on the horizontal axis using 0 as the concentration of the unspiked aliquot. An example plot is shown in Figure 1. When the resulting line is extrapolated back to zero absorbance, the point of interception of the horizontal axis is the concentration of the unknown. Calculate the correlation coefficient (r) and the x-intercept (where $y=0$) of the curve. The concentration in the digestate is equal to the negative x-intercept.

Figure 1



- For the method of standard additions to be correctly applied, the following limitations must be taken into consideration.
- The plot of the sample and standards must be linear over the concentration range of concern. For best results, the slope of the curve should be similar to that of a plot of the aqueous standard curve.
- The effect of the interference should not vary as the ratio of the standard added to the sample matrix changes.

APPENDIX D
TROUBLESHOOTING GUIDE

APPENDIX D. TROUBLESHOOTING GUIDE

Problem	Possible Cause
Poor or No Absorbance or Sensitivity Check failed	Incorrect wavelength Dirty windows Window loose Etched or dirty optics Wrong lamp Bad lamp Not enough or no sample introduced Empty sample cup Incorrectly made standards Gas leak EDL power supply set on "Continuous"
Erratic Readings	Source lamp not aligned properly Lamp not prewarmed Injection tip partially clogged Contaminated reagents Contaminated glassware Drying tube saturated Bad lamp Injection tip hitting outside of tube Injection tip coated or not set properly Leak in sample tubing Power fluctuations Air bubbles in tubing
EDL Won't Light	Lamp cable not plugged in Lamp power set at 0 Lamp is dead Power supply fuse is blown Short in cord
Standards reading twice or half normal absorbance or concentration	Incorrect standard used Incorrect dilution performed Dirty cell
Background Correction Light Blinking	Background screen or attenuator faulty

APPENDIX E
CONTAMINATION CONTROL GUIDELINES

APPENDIX E. CONTAMINATION CONTROL GUIDELINES

The following procedures are strongly recommended to prevent contamination:

All work areas used to prepare standards and spikes should be cleaned before and after each use.

All glassware should be washed with detergent and tap water and rinsed with 1:1 nitric acid followed by deionized water.

Proper laboratory housekeeping is essential in the reduction of contamination in the metals laboratory. All work areas must be kept scrupulously clean.

Powdered or Latex Gloves must not be used in the metals laboratory since the powder contains silica and zinc, as well as other metallic analytes. Only vinyl or nitrile gloves should be used in the metals laboratory.

Glassware should be periodically checked for cracks and etches and discarded if found. Etched glassware can cause cross contamination of any metallic analytes.

Autosampler trays should be covered to reduce the possibility of contamination. Trace levels of elements being analyzed in the samples can be easily contaminated by dust particles in the laboratory.

The following are helpful hints in the identification of the source of contaminants:

Reagents or standards can contain contaminants or be contaminated with the improper use of a pipette.

Improper cleaning of glassware can cause contamination.

Separate glassware if an unusually high sample is analyzed and soak with sulfuric acid prior to routine cleaning.

APPENDIX F
PREVENTIVE MAINTENANCE

APPENDIX F. PREVENTIVE MAINTENANCE

A maintenance log is used to record when maintenance is performed on instruments. When an instrument problem occurs indicate the date, time and instrument number, then identify the problem and corrective action in the maintenance log.

The following procedures are required to ensure that that the instrument is fully operational.

Cold Vapor Atomic Absorption (Leeman PS 200) ⁽¹⁾

Daily	Semi-annually	Annually
Clean lens.	Check Hg lamp intensity.	Change Hg lamp.
Check aperture.		Check liquid/gas separator.
Check argon flow.		
Check tubing.		
Check drain.		
Replace drying tube.		

Cold Vapor Atomic Absorption (PE 5000) ⁽¹⁾

Daily	Monthly
Clean aspirator by flushing with DI water.	Clean cell in aqua regia.
Check tubing and replace if needed.	Clean aspirator in aqua regia.
Clean windows with methanol.	
Change silica gel in drying tube.	
Check argon gas supply.	
Adjust lamp.	

APPENDIX G
INSTRUMENT SET UP

Hg Analysis (Leeman PS200II)

SYSTEM INITIALIZATION AND WARM UP

1. F1 Menu
2. Instrument
 - a. TASKMASTER
 - b. #4 Wake System Up Enter

The warming up period takes approximately 10 minutes.

TO SET UP INSTRUMENT FOR ANALYSIS






1. F1 Menu
2. Autosampler
 - A. Rack Entry
 - B. Edit (ex. Rack 1), Enter
 - C. Cup ID - Enter (clears sample #'s)
 - D. Extended ID- type in matrix of sample (water or solid), Enter.
 - E. Press Insert Key and move cursor with arrows to cup ID and begin typing labels.
 - F. F3 Print Screen
3. Press F2 Macro key and type in analyst's first name - Enter
 - A. Enter folder name - ex. HG0306, Enter. If folder does not exist, type Y - Enter.
 - B. Type in - "Rack 1", "Rack 2" etc. , Enter.

C. Type in : FROM CUP TO CUP

Ex. = 1 30

Do the same for position 2 if needed. If not needed, you must
press Enter 3 times to begin analysis.

STL Austin
SOP REVIEW FORM

SOP NUMBER: AUS-GC-0002, Rev. 2	
SOP TITLE: ANALYSIS OF DISSOLVED GASES IN WATER	
<p>The subject SOP has undergone the required peer/management review. No modifications are necessary at this point. The SOP will be reviewed one year from the date indicated below.</p>	
REVIEWED BY/DATE:	 8/16/01
*APPROVED BY:	
 Technical Director	8/16/01 Date
 Environmental Health and Safety Coordinator	8/16/01 Date
 Quality Assurance Manager	8/16/01 Date
 Laboratory Director	8/16/01 Date

*Must be same signature authorities of SOP being reviewed.

Controlled Copy
Copy No. **UNCONTROLLED COPY**

SOP No. AUS-GC-0002
Revision No. 2
Revision Date: 07/01/99
Page 1 of 17


Implementation Date 7/12/99


OPERATION-SPECIFIC STANDARD OPERATING PROCEDURE

TITLE: ANALYSIS OF DISSOLVED GASES IN WATER


(SUPERSEDES: AUS-GC-0002, Rev. 1))

Prepared by: Brook Derenzy

Reviewed by: 
Technical Specialist, Carl Hogberg

Approved by: 
Quality Assurance Manager, Alice Wusterhausen

Approved by: 
Environmental, Health and Safety Coordinator, Stuart Bosio

Approved by: 
Laboratory Manager, Neal Salcher

Proprietary Information Statement:

This document has been prepared by and remains the sole property of Quanterra[®] Incorporated. It is submitted to a client or government agency solely for its use in evaluating Quanterra's[®] qualifications in connection with the particular project, certification, or approval for which it was prepared and is to be held proprietary to Quanterra[®].

The user agrees by its acceptance or use of this document to return it upon Quanterra's[®] request and not to reproduce, copy, lend, or otherwise disclose or dispose of the contents, directly or indirectly, and not to use it for any purpose other than that for which it was specifically furnished. The user also agrees that where consultants or others outside of the user's organization are involved in the evaluation process, access to these documents shall not be given to those parties unless those parties also specifically agree to these conditions.

TABLE OF CONTENTS

1. SCOPE AND APPLICATION	3
2. SUMMARY OF METHOD	3
3. DEFINITIONS	3
4. INTERFERENCES	3
5. SAFETY	4
6. EQUIPMENT AND SUPPLIES	4
7. REAGENTS AND STANDARDS	5
8. SAMPLE COLLECTION, PRESERVATION AND STORAGE	5
9. QUALITY CONTROL	5
10. CALIBRATION AND STANDARDIZATION	6
11. PROCEDURE	7
12. DATA ANALYSIS AND CALCULATIONS	9
13. METHOD PERFORMANCE	11
14. POLLUTION PREVENTION	11
15. WASTE MANAGEMENT	11
16. REFERENCES	11
17. MISCELLANEOUS (TABLES, APPENDICES, ETC...)	11
TABLE 1 - GC CONDITIONS	12
TABLE 2 - COMPOUND LIST, REPORTING LIMITS AND METHOD DETECTION LIMITS	12
APPENDIX A - EXAMPLE RSK-175 WORKSHEET	13
APPENDIX B - EXAMPLE HENRY'S LAW DATA	14
APPENDIX C - EXAMPLE SPIKING LEVEL CALCULATION WORKSHEET	14
APPENDIX D - EXAMPLE FINAL RESULT WORKSHEETS	15
APPENDIX E - DATA REVIEW CHECKLIST	16

1.0 SCOPE AND APPLICATION

- 1.1. This analysis is for the determination of the concentration of aliphatic and olefinic hydrocarbons, normally found in the gas phase at room temperature, in water. These include methane, ethane and ethene; specific information is found in Table 2 of this SOP. Other gases are amenable to this method. The method is a modification of RSKSOP-175 from the U.S.E.P.A. R.S. Kerr laboratory.

2.0 SUMMARY OF METHOD

- 2.1. A water sample is collected in a 40-mL vial with no headspace. A percentage of the headspace is displaced with helium. The sample in the vial is shaken to equilibrate the hydrocarbons between the water and helium headspace. A headspace aliquot is withdrawn and analyzed with a GC FID system to determine the concentration of gaseous hydrocarbons present in the headspace. The headspace concentration is related to the starting water concentration through the use of Henry's Law.

3.0 DEFINITIONS

- 3.1. Henry's Law: Henry's law states that the ratio of the partial pressure of a gas in a closed system and molar concentration in solution is a constant. This constant varies with temperature. The constant is compound specific. (See Appendix B.)

4.0 INTERFERENCES

- 4.1. Background atmospheric methane is commonly present at levels of 5-10 ppm in the atmosphere. This causes small amounts of methane to be present in laboratory blanks (typically less than 0.25 µg/L).
- 4.2. Before any analysis can be performed on samples, all systems related hardware must be shown to be free of interferences by the analysis of a method blank.
- 4.3. Cross-over contamination is routinely not a problem with this analysis due to the volatile nature of the gases tested, the characteristics of the column, and the GC temperature program that is hot enough and long enough to prevent carry over. Each sample has its own new VOA vial and is directly injected on the GC. The syringe is flushed between samples. The analyst must be familiar with the characteristics of the system to determine when cross over may have occurred. Reanalysis of suspected cross over samples must be done as soon as possible.
- 4.4. Helium and other gases used in this procedure must be of known purity and shown to be free of interferences by running a clean daily lab blank.
- 4.5. Samples do not gain contamination from volatiles diffusing through the septa during handling or transport.

5.0 SAFETY

5.1. Normal office-dependent safety precautions must be taken in performing this SOP. If personnel are required to perform any portion of the procedure in laboratory areas, appropriate personal protective equipment and precautions must be utilized. Procedures shall be carried out in a manner that protects the health and safety of all Quanterra® associates.

5.1.1. Eye protection that satisfies ANSI Z87.1 (as per the Chemical Hygiene Plan), laboratory coat, and appropriate gloves must be worn while samples, standards, solvents, and reagents are being handled. Disposable gloves that have been contaminated will be removed and discarded; other gloves will be cleaned immediately. Viton gloves may be worn when halogenated solvents are used for extractions or sample preparation. Nitrile gloves may be used when other solvents are handled. [Note: Viton is readily degraded by acetone; all solvents will readily pass through disposable latex rubber gloves.]

5.1.2. The health and safety hazards of many of the chemicals used in this procedure have not been fully defined. Additional health and safety information can be obtained from the MSDS files maintained in the laboratory. The following specific hazards are known:

Chemicals known to be **flammable** are: hydrogen, aliphatic and olefinic hydrocarbons from C1-C7. These include methane, ethane, ethene, etc.

5.2. Exposure to chemicals must be maintained as low as reasonably achievable, therefore, unless they are known to be non-hazardous, all samples must be opened, transferred and prepared in a fume hood, or under other means of mechanical ventilation. Solvent and waste containers will be kept closed unless transfers are being made.

5.3. All work must be stopped in the event of a known or potential compromise to the health and safety of a Quanterra® associate. The situation must be reported immediately to a laboratory supervisor.

6.0 EQUIPMENT AND SUPPLIES

6.1. 1-mL and 5-mL gas tight syringes.

6.2. 40-mL glass screw-cap VOA vials with Teflon™-faced silicone septa.

6.3. Gas tight syringes with valves, 0.1 mL, 0.5 mL, 1.0 mL, and 5.0 mL.

6.4. Hot water bath.

6.5. Boiled water.

- 6.6. HP-Plot column 0.53 mm X 30 m Al₂O₃.
- 6.7. GC system with flame ionization detector such as a HP model 5890.
- 6.8. Data acquisition system such as PE Nelson Turbochrom.

7.0 REAGENTS AND STANDARDS

- 7.1. 100 ppmv methane, ethane, and ethene gas mix (Quality Standards and Research Gases Inc. or equivalent)
- 7.2. 1 % gas standard of methane, ethane, and ethene in nitrogen or helium
- 7.2. Gas mixture from a second source at 100 ppmv for the LCS, MS/MSD

8.0 SAMPLE COLLECTION, PRESERVATION AND STORAGE

- 8.1 Water samples are collected in 40-mL VOA vials with zero headspace and stored at 4 °C ± 2°C. Samples should be preserved with 1:1 hydrochloric acid to a pH of less than 2 before they are capped. The suggested holding time for this analysis is 14 days; however there is evidence that samples may be stable for as long as 28 days. Any samples analyzed outside the suggested 14 day hold time will be documented using a nonconformance memo.

9.0 QUALITY CONTROL

- 9.1. A method blank must be analyzed with each batch. The method blank must yield concentrations of less than the reporting limit before all associated samples may be analyzed. Methane is commonly present in the atmosphere at levels of 5-10 ppm. Care must be taken to ensure that blank water used in this method is stored in a well sealed container to prevent adsorption of atmospheric methane. All samples which have levels of methane within 5 times the daily blank will be footnoted.
- 9.2. An LCS must be analyzed with each batch and must meet 70-130% of the theoretical value. [In-house historical control limits will be determined when a sufficient number of data points are accumulated (~30)]. If the LCS does not meet the acceptance criteria, a new LCS will be prepared and analyzed. If the second LCS passes, samples may be analyzed. Continued failure of the LCS is an indication that the system may need maintenance and that a new calibration curve will be necessary.
- 9.3. One MS/MSD pair must be analyzed per 20 samples or every 24 hours. MS/MSD recovery limits are 70-130% of the theoretical value and the RPD acceptance limit is 30%. [In-house historical control limits will be determined when a sufficient number of data points are accumulated (~30)]. If the MS/MSD

does not meet the acceptance criteria, the data may be reported based on the LCS as per Quanterra® Quality Assurance Policy, QA-003.

- 9.4. For each sample, and QC sample, the retention times of the analytes should be within the determined retention time window.

10.0 CALIBRATION AND STANDARDIZATION

10.1. Initial calibration

10.1.1. At a minimum, a three-point curve directly injected on the GC from 2 to 800 ppmv standards is run. The curve is prepared by making serial dilutions of the 10000 ppmv standard in a gas dilution vessel. The curve is acceptable if it meets the appropriate criteria for either average calibration factor or correlation coefficient curve fit. The average calibration factor may be used if the average %RSD of the calibration factors is $\leq 15\%$; or the correlation coefficient must be ≥ 0.995 . If neither of these two criteria is met, one point may be rerun. If after rerunning one point, the %RSD is still $> 15\%$ or the correlation coefficient is still < 0.995 , the entire curve must be rerun. Calibration curves are historically stable for weeks to several months. Failure to have linear response may be an indication of instrument problems which should be evaluated and corrected.

10.2. Retention Time Window Determination.

10.2.1. Retention time (RT) windows for each target analyte should be established before interpreting sample data. This is a manual injection method, and as such, retention times and retention time windows of each analyst are expected to be slightly different due to analyst technique. New retention time windows should be reestablished whenever a new GC column is installed, a new analyst is performing this method, or the operating parameters are significantly changed.

10.2.2. Retention time windows should be established by the evaluation of the same level standard three times over the course of three days.

10.2.3. The standard deviation of the three retention times for each compound is calculated. The retention time window is defined as plus or minus three times this standard deviation.

10.2.4. If the standard deviation of the retention time for any analyte is zero, the standard deviation of a closely eluting peak may be substituted for determination of the retention time window. In some cases analytical judgment may be used where RT windows are unrealistically narrow. Approval by the Team Leader or Group Leader must be obtained where

analytical judgment is used. Until specific windows have been established, a nominal window of 0.10 minute may be used.

10.2.5. The retention time windows established in this manner are used with the retention time of each peak in the daily calibration verification for identification of the target analytes in the samples.

10.3. Daily Calibration Verification

10.3.1. A check of the initial calibration curve must be made at the start of every 24 hour sequence of analysis runs.

10.3.2. The calibration check consists of the injection of a midpoint standard at 400 ppm. The check standard must meet $\pm 15\%$ of its theoretical value. This standard is prepared by serially diluting the 10000 ppmv gas standard in a gas dilution vessel.

10.3.3 If the conditions of 10.3.2 are not met, corrective action must be taken. The standard may be remade and reinjected once, but failure to meet the conditions twice requires a new initial calibration. Historically it is found that the initial calibration for this method is stable for months.

10.4. Continuing Calibration

10.4.1. A continuing calibration standard must be run after every 20 samples and at the end of all analytical sequences. The continuing calibration standard is usually the midpoint standard but other levels may be run to verify the entire curve range.

10.4.2. The analysis of samples may continue after the continuing calibration if the calibration standard has met the criteria of 10.3.2. Samples analyzed after a failed calibration check must be reanalyzed.

11.0 PROCEDURE

11.1. Any unauthorized deviations from this procedure must also be documented as a nonconformance, with a cause and corrective action described.

11.2. The GC instrument is set up using the conditions listed in Table 1.

11.3. The GC septum is changed daily or after 50 samples. The column flow is turned off and the GC cooled below 40 °C before the septum is removed. After the septum has been changed, the column flow is turned on and the GC brought up to initial temperature.

- 11.4. The calibration curve is run at nominal 2,10,40,200,400 and 800 ppm(v/v) by injecting 300 μ L of prepared mixes or dilutions of prepared mixes. Alternatively a reduced volume of a higher standard may be used to prepare a lower calibration point. No injections greater than 300 μ L will be made and no injections smaller than 30 μ L should be made. Standards may be prepared by making serial dilutions of a more concentrated standard into a gas dilution vessel.
- 11.5. Samples are brought to room temperature before analysis. The temperature is recorded in the RSK-175 worksheet. (See Appendix A.) This temperature is used for calculation of the extrapolated Henry's Law constant and the molar density for each analyte. The headspace is prepared by displacement of the water from the sample. This is done by injecting helium into the vial through a needle at approximately 1-2 psi while simultaneously collecting the water by forcing it through a second needle into a collecting syringe (attach a 20 gauge needle to a flow controlled source of helium). Prepare a 5- or 10-mL syringe with a 20 gauge needle. Clamp the sample in a 40-mL VOA vial upside down. Insert the collecting syringe needle. Insert the helium needle and displace 4 mL of water from the VOA vial into the collecting syringe. Record the displaced volume. The volume of the water remaining is determined after the analysis is complete by weight, assuming 1 g = 1 mL.
- 11.6. Shake the samples upside down (septa side down) at a setting of 5 on a Vortex Genie for 5 minutes to equilibrate the gases between the gaseous and liquid phases.
- 11.7. Using a 500- μ L syringe equipped with an on/off valve, remove 300 μ L of headspace and close the valve. Inject the 300 μ L aliquot into the GC smoothly and start the GC temperature program and the data system.
- 11.8. Dilutions are prepared by adding a known amount of headspace gas into a gas tight syringe containing helium/nitrogen as a diluent. Alternatively a reduced volume may be injected onto the system. (For example a 30- μ L injection would represent a 10X dilution.) Higher dilutions will be made by using gas sampling cylinders. Dilutions may also be prepared by making a serial dilution of a second unanalyzed sample vial into a volumetric flask using organic free water as a diluent; then transferring, without agitation, the resulting dilution into a clean 40-mL vial with a Teflon/rubber seal. (Note: A single inversion is sufficient to mix the contents of the volumetric flask; excessive mixing will result in the loss of the target gases.)
- 11.9. The LCS is prepared by injecting 0.2 mL of a 10000 ppmv gas mix into the headspace of a water blank. The sample is then shaken as noted in section 11.7. The spiking gases must be in equilibrium with the aqueous phase prior to analysis.
- 11.10. MS/MSD samples are prepared in a same way as in 11.9 using a sample instead of lab blank water.

11.11. Analytical Documentation

- 11.11.1. Record all analytical information in the analytical logbook/benchsheet (Appendix A), including the analytical data from standards, blanks, LCSs, MS/MSDs, and any corrective actions or modifications to the method. Submit a copy of the logbook/benchsheet along with the worksheets (Appendix C and D) and the data review checklist (Appendix E) for second-level data review.
- 11.11.2. All standards are logged into a standards inventory database, run_std. All standards are assigned a unique number for identification. Entries should be reviewed by the supervisor or designee.
- 11.11.3. Documentation such as all associated instrument printouts (final runs, screens, reruns, QC samples, etc.) daily calibration data corresponding to all final runs, and the data review checklist is available for each data file.
- 11.11.4. Sample results and associated QC are entered into the LIMS after final technical review.
- 11.11.5. Second-level review is performed by the supervisor or designee. After completion of the final review, the data is released for reporting.

12.0 DATA ANALYSIS AND CALCULATIONS

12.1. Calculated Results

- 12.1.1. Calculation of the Henry's Law Constant at a given temperature is performed using an Excel 5.0 linear fit algorithm on data tables found in Perry's Chemical Engineer's Handbook (see Appendices B and C) over a range of temperatures which are inclusive of the sample temperature.
- 12.1.2. Calculation of the molar density at a given temperature is given below:

$$MD = \frac{mw}{22.4 \left(\frac{T}{273} \right)}$$

where: MD = molar density, g/L
mw = molecular weight
T = daily temperature in degrees centigrade + 273

- 12.1.3. Calculation of the concentration of an analyte in solution, $\mu\text{g/L}$:

$$A = \frac{(\alpha)(0.000001)(55.5)(1000000)(mw)}{H}$$

where: a = concentration from cal curve in units of ppmv/v
 0.000001 = the number of atmospheres per ppmv/v
 55.5 = the number of moles per liter of water
 1000000 = μg per gram conversion
 mw = molecular weight
 H = Henry's Law Constant at temperature T, in units of moles

12.1.4. Calculation of the concentration of the analyte in the headspace, $\mu\text{g/L}$:

$$B = \frac{(\alpha)(0.000001)(MD)(V_{(h)})(1000000)}{V_{(s)}}$$

where: a = concentration of analyte in ppmv/v in headspace
 MD = molar density, g/L
 $V_{(h)}$ = volume of the headspace (mL)
 $V_{(s)}$ = volume of the water (mL)
 0.000001 = conversion of ppmv/v to atmospheres
 1000000 = μg per gram conversion

12.1.5. Total concentration of gases, $\mu\text{g/L}$:

$$A + B = C$$

where: A = analyte concentration in the liquid phase ($\mu\text{g/L}$), from 12.1.3
 B = analyte concentration in the gas phase ($\mu\text{g/L}$), from 12.1.4
 C = sample concentration ($\mu\text{g/L}$)

12.2. LCS and Matrix Spike Calculation

12.2.1. Calculation of the spike level (see Appendix C)

$$S = \frac{(V_{(std)})(C_{(std)})(0.000001)(MD)(1000000)}{V_{(smp)}}$$

where: S = spike concentration in $\mu\text{g/L}$
 $V_{(std)}$ = volume of standard added (mL)
 $C_{(std)}$ = concentration of standard in ppmv
 MD = molar density in g/L at temperature T
 1000000 = μg per gram conversion
 0.000001 = converts ppmv to mL
 $V_{(smp)}$ = volume of sample (mL)

12.2.2. Calculation of LCS recovery:

$$R = \frac{C_{(Found)}}{C_{(Spike)}} \times 100$$

where: R = recovery

$C_{(found)}$ = concentration in mg/L found in LCS

$C_{(spike)}$ = concentration in mg/L spiked

12.2.3. Calculation of MS recovery

$$R = \frac{C_{(Found)} - C_{(Unspiked)}}{C_{(Spike)}} \times 100$$

where: R = recovery

$C_{(found)}$ = concentration in mg/L found

$C_{(unspiked)}$ = concentration in mg/L found in sample

$C_{(spike)}$ = concentration spiked in mg/L

12.2.4 Calculation of %RPD

$$\%RPD = \frac{|SR - SDR|}{\left(\frac{SR + SDR}{2}\right)} \times 100$$

where: SR = Spike Result

SDR = Spike Duplicate Result

13.0 METHOD PERFORMANCE

- 13.1. Training Qualification: The group/team leader has the responsibility to ensure that this procedure is performed by an analyst who has been properly trained in its use and has the required experience. To qualify on this method, an analyst will run a LCS study with 4 spikes at the curve mid point and 7 spikes at the low point of the curve. The LCS spikes must meet $\pm 30\%$ of the theoretical value, the low spikes may be used to determine a new MDL and are compared with the existing MDL.

14.0 POLLUTION PREVENTION

- 14.1. This method does not contain any specific modifications that serve to minimize or prevent pollution

15.0 WASTE MANAGEMENT

- 15.1. Waste generated in this procedure will be segregated and disposed according to the facility hazardous waste procedures. The Environmental Health and Safety Coordinator should be contacted if additional information is required.

16.0 REFERENCES

- 16.1. Perry's Chemical Engineers Handbook, sixth edition
- 16.2. RSKSOP-175, (U.S.) Robert S. Kerr Environmental Research Lab, Ada, OK

17.0 MISCELLANEOUS (TABLES, APPENDICES, ETC.)

- 17.1. Modifications from RSKSOP-175: The Austin laboratory uses 40-mL glass screw-cap VOA vials with Teflon™-faced silicone septa for sample collection instead of 60-mL serum bottles with butyl rubber Teflon™-faced septa and aluminum crimp caps.
- 17.2. Table 1: GC Conditions
- 17.3. Table 2: Compound List, Reporting Limits and Method Detection Limits
- 17.4. Appendix A: Example RSK-175 Worksheet
- 17.5. Appendix B: Example Henry's Law Data
- 17.6. Appendix C: Example Spiking Level Calculation Worksheet
- 17.7. Appendix D: Example Final Result Worksheets
- 17.8. Appendix E: Data Review Checklist

TABLE 1

GC Conditions

Column:	HP Plot 0.53 mm X 30 meters Al ₂ O ₃
Carrier gas:	He
Flow:	20 psi
Make up gas:	N ₂
Flow:	20 mL/min
Initial temperature:	40 °C
Initial time:	2.5 min
Rate:	50 °C/min
Final temperature:	180 °C
Final time:	1.2 min

TABLE 2

Compound List, Reporting Limits and Method Detection

Limits in µg/L

(based on 1 ppm (v/v) reporting level)

Compound	CAS Number	MDL (µg/L)	RL (µg/L)
Methane	74-82-8	0.1	0.5
Ethane	74-84-0	0.1	0.5
Ethene	74-85-1	0.1	0.5

APPENDIX A - EXAMPLE RSK-175 WORKSHEET

RSK-175 Worksheet							
Method	RSK-175	Analyte	Methane	Ethane	Ethylene		
Matrix:	Water	Mol. Wt.	16	30	28		
Prep Date:	8/19/97	H (T)	38629.00	27551.00	10563.50		
Instrument:	GCK1	M.D. (T)	0.662139	1.241511	1.158744		
		g/L					
Analyst	twc						
Temp (C):	21.5						
	Sample #	Headspace volume (mL)	Water volume (mL)	Dilution Factor	Raw Amount		
					Methane	Ethane	Ethene
1	BLANK	4.1	38.3	1	2.4577	0.0000	0.0000
2	LCS	4.3	38.6	1	304.4413	281.3018	188.9934
3	LCSD	4.5	38.9	1	309.9261	298.7994	205.4615
4	Sample #1	4.6	38.0	1	3.5800	2.6900	1.3600
5	Sample #2	4.2	39.0	1	3.8200	3.0300	1.5000
6	Sample #3	4.0	39.5	1	4.2400	3.2400	1.7100
7							
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							

APPENDIX B - HENRY'S LAW DATA

Henry's Law					
Methane		Ethane		Ethene	
degrees C	H solubility	degrees C	H solubility	degrees C	H solubility
15	33700.00	15	2.26E+04	15	8.95E+03
20	37600.00	20	2.63E+04	20	1.02E+04
25	41300.00	25	3.02E+04	25	1.14E+04
30	44900.00	30	3.42E+04	30	1.27E+04

APPENDIX C - EXAMPLE SPIKING LEVEL CALCULATION WORKSHEET

Calculations:						
Conc. in H2O = Raw amt * 1E-6 / H (T) * 55.5 * Mol. wt. * 1E6						
Conc. in HS = (Vol (hs) * Raw amt * 1E-6) / (Vol (water) * 1E-3) * MD (T) * 1E3						
Final conc. = DF * (Conc (water) + Conc (hs))						
Analyte	Methane	Ethane	Ethylene			
Mol. Wt.	16	30	28			
H (T)	38629.00	27551.00	10563.50			
M.D. (T) g/L	0.662139	1.241511	1.158744			
Spiking level calculation:						
Spike conc. = DF * Vol (std) * Conc (std) * 1E-6 * MD / (Vol (samp) * 1E-3) * 1E3						
PM2929	Compound	Conc. ppmv				
	Methane	10000				
	Ethane	9900				
	Ethene	9900				
	water volume	mls of spike	Dil Factor			
			Methane ug/L			
			Ethane ug/L			
			Ethylene ug/L			
LCS	38.6	0.2	1	34.3077	63.6837	59.4381
LCSD	38.9	0.2	1	34.0431	63.1926	58.9798
MS	39.1	0.2	20	677.3803	1257.3871	1173.5613
MSD	39.1	0.2	20	677.3803	1257.3871	1173.5613
MS	39.3	0.2	1	33.6967	62.5494	58.3795
MS	39.1	0.1	3	50.8035	94.3040	88.0171

APPENDIX D - EXAMPLE FINAL RESULT WORKSHEETS

RSK-175 Report for		BLANK				
Prep Date:	8/19/97	Analyte	Methane	Ethane	Ethylene	
Instrument:	GCK1	Mol. Wt.	16	30	28	
Analyst	twc	H (T)	38629	27551	10563.5	
Temp (C):	21.5	M.D. (T) g/L	0.662139	1.241511	1.158744	
Sample #	Headspace volume (mL)	Water volume (mL)	Dilution Factor			
BLANK	4.1	38.3	1			
Compound	Raw Amt. (ppmv)	Dilution Factor	Conc. in H2O ug/L	Conc. in HS ug/L	Compound	Final Conc. ug/L
Methane	2.4577	1	0.0565	0.1742	Methane	0.231
Ethane	0	1	0.0000	0.0000	Ethane	0.000
Ethene	0	1	0.0000	0.0000	Ethene	0.000

RSK-175 Report for		LCS				
Prep Date:	8/19/97	Analyte	Methane	Ethane	Ethylene	
Instrument:	GCK1	Mol. Wt.	16	30	28	
Analyst	twc	H (T)	38629	27551	10563.5	
Temp (C):	21.5	M.D. (T) g/L	0.662139	1.241511	1.158744	
Sample #	Headspace volume (mL)	Water volume (mL)	Dilution Factor			
LCS	4.3	38.6	1			
Compound	Raw Amt. (ppmv)	Dilution Factor	Conc. in H2O ug/L	Conc. in HS ug/L	Compound	Final Conc. ug/L
Methane	304.4413	1	6.9985	22.4561	Methane	29.455
Ethane	281.3018	1	17.0000	38.9049	Ethane	55.905
Ethene	188.9934	1	27.8029	24.3958	Ethene	52.199

APPENDIX E - DATA REVIEW CHECKLIST

Quanterra®, Austin

QC Check List, RSK-175

WO # / Lot # / Batch #	Instrument ID:
	Analysis Date:
	SOP #: AUS-GC-0002, Rev.0

Initial Calibration:

	Yes	No	N/A
Minimum 3 point calibration performed?			
Date the curve was run if different than this date:			
Percent RSD +/- 15% for all target analytes?			

Continuing Calibration:

	Yes	No	N/A
Calib. verification for all target analytes within +/- 15% of true value?			
Frequency of continuing calibration is 1 per every 20 samples?			
Continuing calibration meets +/- 15% acceptance criteria?			
Ending calibration check standard run?			

Quality Control Criteria:

	Yes	No	N/A
Method blank less than reporting limits?			
LCS meets the % recovery criteria?			
MS/MSD meets the recovery criteria?			
MS/MSD meets the % RPD criteria?			

Sample Analysis:

	Yes	No	N/A
Name and header information correct?			
Runs checked for saturation?			
The reported results are within the calibration range?			
Calculations have been checked?			
Results account for dilutions, etc.?			
Units correctly reported?			
Sample hold times met?			

Other:

	Yes	No	N/A
Transcriptions checked for accuracy?			
Calculations checked at minimum frequency?			
All manual integrations checked by 2nd level reviewer?			
All nonconformances included and noted?			

Analyst: _____

Date: _____

Comments:

2nd Level Reviewer: _____

Date: _____

Comments:

Controlled Copy
Copy No: _____
Implementation Date 06/05/01

SOP No: CORP-GC-0001NC
Revision No: 5.6
Revision Date: 05/25/01
Page 1 of 92

STL NORTH CANTON STANDARD OPERATING PROCEDURE

**TITLE: GAS CHROMATOGRAPHIC ANALYSIS BASED ON METHOD 8000B,
8021B, 8081A, 8082, 8151A, 8310, 8141A, 8015B, 608, 610 and Wisconsin DNR Modified
DRO Method**

(SUPERSEDES: Revision 5.5 Dated 03/16/01)

Prepared by:	<u>Raymond W. Baskin</u>	<u>5-25-01</u>
Approved by:	<u>Wally L. Sabrowski</u>	<u>5-25-01</u>
	Technology Specialist	Date
Approved by:	<u>Pauline Cob</u>	<u>5/25/01</u>
	Quality Assurance Manager	Date
Approved by:	<u>[Signature]</u>	<u>5-25-01</u>
	Environmental Health and Safety	Date
Approved by:	<u>Cathy Reprandi</u>	<u>5-25-01</u>
	Laboratory Director	Date
Approved by:	<u>[Signature]</u>	<u>6-4-01</u>
	Corporate Technology	Date
		<u>06-05-01</u>

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use it for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

Controlled Copy
Copy No. _____

SOP No. CORP-MT-0001NC
Revision No. 3.2
Revision Date: 01/19/01
Page 1 of 55

Implementation Date: 2/12/01

STL NORTH CANTON STANDARD OPERATING PROCEDURE

**TITLE: INDUCTIVELY COUPLED PLASMA-ATOMIC EMISSION SPECTROSCOPY,
SPECTROMETRIC METHOD FOR TRACE ELEMENT ANALYSES,
SW-846 METHOD 6010B AND EPA METHOD 200.7**

(SUPERSEDES: REVISION 3.1, REVISION DATE 10/04/00)

Approved by: Karen S. Counts 2-05-01
Date

Approved by: Patrick O'My 2/15/01
Technical Specialist Date

Approved by: Paul Ai. Cuba 2/6/01
Quality Assurance Manager Date

Approved by: [Signature] 2-8-01
Environmental, Health and Safety Coordinator Date

Approved by: Christy McQuarrie 2-10-01
Laboratory Manager Date

Approved by: [Signature] 1/30/01
Corporate Technology Date

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use if for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

Controlled Copy
Copy No. _____

SOP No. CORP-MT-0001NC
Revision No. 3.2
Revision Date: 01/19/01
Page 1 of 57

Implementation Date: _____

STL NORTH CANTON STANDARD OPERATING PROCEDURE

**TITLE: INDUCTIVELY COUPLED PLASMA-ATOMIC EMISSION SPECTROSCOPY,
SPECTROMETRIC METHOD FOR TRACE ELEMENT ANALYSES,
SW-846 METHOD 6010B AND EPA METHOD 200.7**

(SUPERSEDES: REVISION 3.1, REVISION DATE 10/04/00)

Approved by: _____ Date

Approved by: _____ Date
Technical Specialist

Approved by: _____ Date
Quality Assurance Manager

Approved by: _____ Date
Environmental, Health and Safety Coordinator

Approved by: _____ Date
Laboratory Manager

Approved by: _____ Date
Corporate Technology *R. B.* 1/30/01

Proprietary Information Statement:

This documentation has been prepared by Severn Trent Laboratories (STL) solely for STL's own use and the use of STL's customers in evaluating its qualifications and capabilities in connection with a particular project. The user of this document agrees by its acceptance to return it to STL upon request and not to reproduce, copy, lend, or otherwise disclose its contents, directly or indirectly, and not to use if for any other purpose other than that for which it was specifically provided. The user also agrees that where consultants or other outside parties are involved in the evaluation process, access to these documents shall not be given to said parties unless those parties also specifically agree to these conditions.

THIS DOCUMENT CONTAINS VALUABLE CONFIDENTIAL AND PROPRIETARY INFORMATION. DISCLOSURE, USE OR REPRODUCTION OF THESE MATERIALS WITHOUT THE WRITTEN AUTHORIZATION OF SEVERN TRENT LABORATORIES IS STRICTLY PROHIBITED. THIS UNPUBLISHED WORK BY STL IS PROTECTED BY STATE AND FEDERAL LAW OF THE UNITED STATES. IF PUBLICATION OF THIS WORK SHOULD OCCUR THE FOLLOWING NOTICE SHALL APPLY: ©2000 SEVERN TRENT LABORATORIES, INC. ALL RIGHTS RESERVED.

TABLE OF CONTENTS

1. SCOPE AND APPLICATION	4
2. SUMMARY OF METHOD.....	5
3. DEFINITIONS	5
4. INTERFERENCES	6
5. SAFETY	7
6. EQUIPMENT AND SUPPLIES.....	8
7. REAGENTS AND STANDARDS	8
8. SAMPLE COLLECTION, PRESERVATION AND STORAGE	9
9. QUALITY CONTROL.....	9
10. CALIBRATION AND STANDARDIZATION	17
11. PROCEDURE	17
12. DATA ANALYSIS AND CALCULATIONS	23
13. METHOD PERFORMANCE.....	25
14. POLLUTION PREVENTION.....	25
15. WASTE MANAGEMENT.....	25
16. REFERENCES	25
17. MISCELLANEOUS (TABLES, APPENDICES, ETC.).....	26

LIST OF APPENDICES:

APPENDIX A - TABLES	30
APPENDIX B – EXAMPLE STL NORTH CANTON ICP DATA REVIEW CHECKLIST	41
APPENDIX C - CROSS REFERENCE OF TERMS USED IN METHODS AND IN SOP	43
APPENDIX D - MSA GUIDANCE	45
APPENDIX E - TROUBLESHOOTING GUIDE.....	47
APPENDIX F - CONTAMINATION CONTROL GUIDELINES	49
APPENDIX G - PREVENTATIVE MAINTENANCE.....	51
APPENDIX H – ICP OPERATING INSTRUCTIONS.....	54

1. SCOPE AND APPLICATION

- 1.1. This procedure describes the analysis of trace elements including metals in solution by Inductively Coupled Plasma -Atomic Emission Spectroscopy (ICP-AES) using SW-846 Method 6010B and EPA Method 200.7 . Table I of Appendix A lists the elements appropriate for analysis by Methods 6010B and 200.7. Additional elements may be analyzed under Methods 6010B and 200.7 provided that the method performance criteria presented in Section 13.0 are met.
- 1.2. ICP analysis provides for the determination of metal concentrations over several orders of magnitude. Detection limits, sensitivity and optimum concentration ranges of the metals will vary with the matrices and instrumentation used. For instance, in comparison to conventional ICP technique, ICP-Trace can achieve detection levels comparable to those determined using the graphite furnace atomic absorption spectroscopy (GFAAS) technique.
- 1.3. Method 6010B is applicable to the determination of dissolved, suspended, total recoverable and total elements in ground water, aqueous samples, soils, sludges, wastes, sediments, and TCLP, EP and other leachates/extracts. All matrices require digestion prior to analysis with the exception of analyses for dissolved metals in filtered and acidified aqueous samples. Although digestion is not specifically required by the method, some clients and regulators may require digestion of **dissolved samples** and this must be clarified and documented before project initiation. Silver concentrations must be below 2.0 mg/L in aqueous samples and 100 mg/kg in solid matrix samples. Precipitation may occur in samples where silver concentrations exceed these levels and lead to the generation of erroneous data.
- 1.4. Method 200.7 is applicable to the determination of dissolved, suspended, total recoverable, and total elements in water, waste water, and solid wastes. All matrices require digestion prior to analysis with the exception of analyses for dissolved metals in filtered and acidified aqueous samples if the criteria in Section 11.1 are met. Silver concentrations must be below 0.1 mg/L in aqueous samples and 50 mg/kg in solid matrix samples.
- 1.5. State-specific requirements may take precedence over this SOP for drinking water sample analyses.
- 1.6. The applicable LIMS method codes are QO (6010B), QM (6010B Trace), AS (200.7), JI (200.7 Trace).

2. SUMMARY OF METHOD

- 2.1. This method describes a technique for the determination of multi elements in solution using sequential or simultaneous optical systems and axial or radial viewing of the plasma. The basis of the method is the measurement of atomic emission by an optical spectroscopic technique. Samples are nebulized and the aerosol that is produced is transported to the plasma torch where excitation occurs. Characteristic atomic-line emission spectra are produced by a radio frequency inductively coupled plasma (ICP). The spectra are dispersed by a grating spectrometer and the intensities of the emission lines are monitored by photomultiplier tubes. The photocurrents from the photomultiplier tubes are processed and controlled by a computer system. A background correction technique is required to compensate for variable background contribution to the determination of trace elements. Background must be measured adjacent to analyte lines during analysis. The position selected for the background intensity measurement, on either or both sides of the analytical line, will be determined by the complexity of the spectrum adjacent to the analyte line. The position used must be free of spectral interferences and reflect the same change in background intensity as occurs at the analyte wavelength measured. Background correction is not required in cases of line broadening where a background correction measurement would actually degrade the analytical result. The possibility of additional interferences should also be recognized and appropriate actions taken. Alternatively, multivariate calibration methods may be chosen for which point selection for background correction is superfluous since whole spectral regions are processed.
- 2.2. Refer to the appropriate SOPs for details on sample preparation methods.

3. DEFINITIONS

- 3.1. Dissolved Metals: Those elements which pass through a 0.45 um membrane. (Sample is acidified after filtration).
- 3.2. Suspended Metals: Those elements which are retained by a 0.45 um membrane.
- 3.3. Total Metals: The concentration determined on an unfiltered sample following vigorous digestion.
- 3.4. Total Recoverable Metals: The concentration determined on an unfiltered sample following treatment with hot, dilute mineral acid.

4. INTERFERENCES

- 4.1. Spectral, physical and chemical interference effects may contribute to inaccuracies in the determinations of trace elements by ICP. Spectral interferences are caused by:
- Overlap of a spectral line from another element.
 - Unresolved overlap of molecular band spectra.
 - Background contribution from continuous or recombination phenomena.
 - Stray light from the line emission of high concentration elements.
- 4.1.1. A background correction technique is required to compensate for variable background contribution to the determination of trace elements. Background correction is not required in cases where a background corrective measurement would actually degrade the analytical result.
- 4.1.2. Inter-element correction factors (IECs) are necessary to compensate for spectral overlap. Inter-element interferences occur when elements in the sample emit radiation at wavelengths so close to that of the analyte that they contribute significant intensity to the analyte channel. If such conditions exist, the intensity contributed by the matrix elements will cause an excessively high (or sometimes low) concentration to be reported for the analyte. Inter-element corrections IECs must be applied to the analyte to remove the effects of these unwanted emissions.

- 4.1.3. Physical interferences are generally considered to be effects associated with sample transport, nebulization and conversion within the plasma. These interferences may result in differences between instrument responses for the sample and the calibration standards. Physical interferences may occur in the transfer of solution to the nebulizer (e.g., viscosity effects), at the point of aerosol formation and transport to the plasma (e.g., surface tension) or during excitation and ionization processes within the plasma itself. Changes in viscosity and surface tension can cause significant inaccuracies, especially in samples containing high dissolved solids or high acid concentrations. If physical interferences are present, dilution of the sample, use of a peristaltic pump, mass flow controller, use of an internal standard and/or use of a high solids nebulizer can reduce the effect.
- 4.1.4. Chemical interferences are characterized by molecular compound formation, ionization effects and solute vaporization effects. Normally these effects are not significant with the ICP technique but if observed can be minimized by buffering the sample, matrix matching or standard addition procedures.

5. SAFETY

- 5.1. Procedures shall be carried out in a manner that protects the health and safety of all STL North Canton associates.
- 5.2. Eye protection that satisfies ANSI Z87.1 (as per the Chemical Hygiene Plan), laboratory coat, and appropriate gloves must be worn while samples, standards, solvents, and reagents are being handled. Disposable gloves that have been contaminated will be removed and discarded; other gloves will be cleaned immediately.
- 5.3. The health and safety hazards of many of the chemicals used in this procedure have not been fully defined. Additional health and safety information can be obtained from the Material Safety Data Sheets (MSDS) maintained in the laboratory.
- 5.3.1. The following materials are known to be **corrosive**:
sulfuric acid, hydrochloric acid, nitric acid and hydrofluoric acid. (NOTE: sulfuric and hydrofluoric acids are used in cleaning the ICP torch and hydrofluoric acid is also commonly used in air toxics preparations.)
- 5.3.2. The following materials are known to be **oxidizing agents**:
nitric acid and hydrogen peroxide.

- 5.3.3. The plasma emits strong UV light and is harmful to vision. **NOTE: AVOID looking directly at the plasma.**
- 5.3.4. The RF generator produces strong radio frequency waves, most of which are unshielded. People with pacemakers should not go near the instrument while in operation.
- 5.4. Exposure to chemicals must be maintained **as low as reasonably achievable**, therefore, unless they are known to be non-hazardous, all samples should be opened, transferred and prepared in a fume hood, or under other means of mechanical ventilation. Metals digestates can be processed outside of a fume hood. Solvent and waste containers will be kept closed unless transfers are being made.
- 5.5. All work must be stopped in the event of a known or potential compromise to the health and safety of a STL North Canton associate. The situation must be reported **immediately** to a laboratory supervisor.
- 5.6. The use of hydrofluoric acid requires special safety precautions. Consult the facility EH&S Manager and laboratory supervisor for guidance.

6. EQUIPMENT AND SUPPLIES

- 6.1. Inductively Coupled Plasma Atomic Emission Spectrometer equipped with autosampler and background correction.
- 6.2. Radio Frequency Generator.
- 6.3. Argon gas supply, welding grade or equivalent.
- 6.4. Coolflow or appropriate water cooling device.
- 6.5. Peristaltic Pump.
- 6.6. Calibrated automatic pipettes or Class A glass volumetric pipettes.
- 6.7. Class A volumetric flasks.
- 6.8. Autosampler tubes.

7. REAGENTS AND STANDARDS

- 7.1. Intermediate standards are purchased as custom STL North Canton multi-element mixes or as single-element solutions. All standards must be stored in FEP

fluorocarbon or unused polyethylene or polypropylene bottles. Intermediate standard solutions must be replaced prior to the expiration date provided by the manufacturer. If no expiration date is provided, the intermediate solutions may be used for up to one year and must be replaced sooner if verification from an independent source indicates a problem. Expiration dates can be extended provided that the acceptance criteria described in laboratory-specific SOPs are met.

- 7.2. Working calibration and calibration verification solutions may be used for up to 3 months and must be replaced sooner if verification from an independent source indicates a problem. Standards should be prepared in a matrix of 5% hydrochloric and 5% nitric acids. An exception to this is in the event the Trace ICP is utilized without the internal standard. In this case, the standard acid matrix must be matched to the final preparation matrix as listed in Section 11.9.
- 7.3. Refer to Tables III, IV, IVA, V and VI (Appendix A) for details regarding the working standard concentrations for calibration, calibration verification, interference correction and spiking solutions.
- 7.4. Concentrated nitric acid (HNO₃), trace metal grade or better.
- 7.5. Concentrated hydrochloric acid (HCl), trace metal grade or better.
- 7.6. Reagent water must be produced by a Millipore DI system or equivalent. Reagent water must be free of the analytes of interest as demonstrated through the analysis of method blanks.

8. SAMPLE COLLECTION, PRESERVATION AND STORAGE

- 8.1. Sample holding times for metals are six months from time of collection to the time of analysis.
- 8.2. Aqueous samples are preserved with nitric acid to a pH of <2 and may be stored in either plastic or glass. If boron or silica are to be determined, plastic containers are preferred. Refrigeration is not required. Preservation must be verified prior to analysis.
- 8.3. Soil samples do not require preservation but must be stored at 4°C ± 2° until the time of preparation .

9. QUALITY CONTROL

Table VII (Appendix A) provides a summary of quality control requirements including type, frequency, acceptance criteria and corrective action.

- 9.1. Initial Demonstration of Capability

Prior to analysis of any analyte using either Method 200.7 or Method 6010B, the following requirements must be met.

Instrument Detection Limit (IDL) - The IDL for each analyte must be determined for each analyte wavelength used least each instrument. The IDL must be determined annually. If the instrument is adjusted in anyway that may affect the IDL, the IDL for that instrument must be redetermined. The IDL shall be determined by multiplying by 3, the standard deviation obtained from the analysis of a standard solution (each analyte in reagent water) at a concentration 3x - 5x the previously determined IDL, with seven consecutive measurements. Each measurement must be performed as though it were a separate analytical sample (i.e., each measurement must be followed by a rinse and/or any other procedure performed between the analysis of separate samples). The result of the IDL determination must be below the STL North Canton reporting limit. The CLP IDL procedure can be used for this method.

9.1.1. Method Detection Limit (MDL) - An MDL must be determined for each analyte prior to the analysis of any client samples. Refer to STL North Canton SOP NC-QA-0021 for details on MDL analysis and criteria.

9.1.2. Linear Range Verification (LR) - The linear range must be determined on at least an annual basis for each analyte wavelength used on each instrument. The linear range is the concentration above which results cannot be reported without dilution of the sample. The standards used to verify the linear range limit must be analyzed during a routine analytical run and must read within 5% of the expected value.

For the initial determination of the upper limit of the linear dynamic range (LDR) for each wavelength, determine the signal responses from a minimum of three to five different concentration standards across the estimated range. One standard should be near the upper limit of the estimated range. The concentration measured at the LDR must be no more than 10% less than the expected level extrapolated from lower standards. If the instrument is adjusted in any way that may affect the LRs, new dynamic ranges must be determined. The LR data must be documented and kept on file.

9.1.3. Background Correction Points - To determine the appropriate location for off-line background correction when establishing methods, the user must scan the area on either side adjacent to the wavelength and record the apparent emission intensity from all other method analytes. This spectral information must be documented and kept on file. The location selected for background correction must be either free of off-line interelement spectral interference or a computer routine must be used for automatic correction on all determinations. Tests to determine spectral interference must be done using analyte concentrations that will adequately describe the interference. Background

correction points must be set prior to determining IECs. Refer to the facility-specific instrument operation SOP and ICP instrument manual for specific procedures to be used in setting background correction points.

- 9.1.4. Inter-element Corrections (IECs) - ICP interelement correction factors must be determined prior to the analysis of samples and every six months thereafter. If the instrument is adjusted in any way that may affect the IECs, the IECs must be redetermined. When initially determining IECs for an instrument, wavelength scans must be performed to ensure that solutions in use are free from contaminants. If an IEC varies significantly from the previously determined IEC then the possibility of contamination should be investigated. The purity of the IEC check solution can be verified by using a standard from a second source or an alternate method (i.e., GFAA or ICP-MS). Published wavelength tables (e.g. MIT tables, Inductively Coupled Plasma-Atomic Spectroscopy: Prominent Lines) can also be consulted to evaluate the validity of the IECs. Refer to the facility specific instrument operation SOP and instrument manufacturer's recommendations for specific procedures to be used in setting IECs. An IEC must be established to compensate for any interelement interference which results in a false analyte signal greater than \pm the RL as defined in Tables I, IA or II. To determine IECs, run a single element standard at the established linear range. To calculate an IEC, divide the observed concentration of the analyte by the actual concentration of the "interfering element."

Note: Trace ICP IECs are more sensitive to small changes in the plasma and instrument setup conditions. Adjustments in the IECs will be required on a more frequent basis for the Trace as reflected by the ICESA response. Additional spectral interference is present from easily ionizable elements such as potassium and sodium in axial viewing instruments.

- 9.1.5. Rinse Time Determination - Rinse times must be determined upon initial set-up of an ICP instrument. To determine the appropriate rinse time for a particular ICP system, the linear range verification standard (see 9.1.3) should be aspirated as a regular sample followed by the analysis of a series of rinse blanks. The length of time required to reduce the analyte signals to $<$ RL will define the rinse time for a particular ICP system. For some analytes it may be impractical to set the rinse time based on the linear range standard result (i.e., analyte not typically detected in environmental samples at that level and an excessive rinse time would be required at the linear range level). Until the required rinse time is established, the method recommends a rinse period of at least 60 seconds between samples and standards. If a memory effect is suspected, the sample must be reanalyzed after a rinse period of sufficient length. Rinse time studies can be conducted at additional concentration levels.

These additional studies must be documented and kept on file, if a concentration other than the linear range level is used to set the rinse time. The concentration levels used to establish the rinse time must be taken into consideration when reviewing the data.

- 9.2. Method Blank (MB) - One method blank must be processed with each preparation batch. The method blank consists of reagent water containing all reagents specific to the method that is carried through the entire analytical procedure, including preparation and analysis. The method blank is used to identify any system and process interferences or contamination of the analytical system that may lead to the reporting of elevated analyte concentrations or false positive data. The method blank should not contain any analyte of interest at or above the reporting limit (exception: common laboratory contaminants, see below) or at or above 5% of the measured concentration of that analyte in associated samples, whichever is higher (sample result must be a minimum of 20x higher than the blank contamination level).
- If the analyte is a common laboratory contaminant (copper, iron, lead (Trace only) or zinc) the data may be reported with qualifiers if the concentration of the analyte in the method blank is less than two times the RL. **Such action must be taken in consultation with the client and must be addressed in the project narrative.**
 - Repreparation and reanalysis of all samples associated with an unacceptable method blank is required when reportable concentrations are determined in the samples (see exception noted above).
 - If there is no analyte greater than the RL in the samples associated with an unacceptable method blank, the data may be reported with qualifiers. **Such action must be taken in consultation with the client and must be addressed in the project narrative.**
 - If the above criteria are not met and reanalysis is not possible, then the sample data must be qualified. **This anomaly must be addressed in the project narrative and the client must be notified.**
 - For dissolved metals samples which have not been digested, a CCB result is reported as the method blank. The CCB run immediately prior to the start of the dissolved sample analyses must be used for this purpose. No more than 20 samples can be associated with one CCB.
- 9.3. Laboratory Control Sample (LCS) - One aqueous LCS must be processed with each preparation batch. The LCS must contain all analytes of interest and must be carried through the entire analytical procedure. Aqueous LCS spike levels are provided in Table III (Appendix A). The LCS is used to monitor the accuracy of the analytical process. On-going monitoring of the LCS results provides evidence that the

laboratory is performing the method within acceptable accuracy and precision guidelines.

- If any analyte is outside established control limits the system is out of control and corrective action must occur. Unless in-house control limits are established, a control limit of 80 - 120% recovery must be applied.
- In the event that an MS/MSD analysis is not possible a Laboratory Control Sample Duplicate (LCSD) must be analyzed. The RPD of the LCS and LCSD must be compared to the matrix spike RPD limits.
- In the instance where the LCS recovery is greater than 120% and the sample results are < RL, the data may be reported with qualifiers. **Such action must be taken in consultation with the client and must be addressed in the report narrative.**
- Corrective action will be repreparation and reanalysis of the batch unless the client agrees that other corrective action is acceptable.
- For dissolved metals samples which have not been digested, a CCV result is reported as the LCS. The CCV run immediately prior to the start of the dissolved sample analyses must be used for this purpose. No more than 20 samples can be associated with one CCV.

9.4. Matrix Spike/Matrix Spike Duplicate (MS/MSD) - One MS/MSD pair must be processed for each preparation batch. A matrix spike (MS) is a field sample to which known concentrations of target analytes have been added. A matrix spike duplicate (MSD) is a second aliquot of the same sample (spiked identically as the MS) prepared and analyzed along with the sample and matrix spike. Some client specific data quality objectives (DQO's) may require the use of sample duplicates in place of or in addition to MS/MSDs. The MS/MSD results are used to determine the effect of a matrix on the precision and accuracy of the analytical process. Due to the potential variability of the matrix of each sample, these results may have immediate bearing only on the specific sample spiked. Samples identified as field blanks cannot be used for MS/MSD analysis. Spiking levels are provided in Tables III and VI (Appendix A).

- If any analyte recovery or RPD falls outside the acceptance range, the recovery of that analyte must be in control for the LCS. For both methods 200.7 and 6010B, control limits of 75-125% recovery and 20% RPD or historical acceptance criteria must be applied to the MS/MSD. If the LCS recovery is within limits, then the laboratory operation is in control and the results may be accepted. If the recovery of the LCS is outside limits corrective action must be taken. Corrective

action will include reparation and reanalysis of the batch. MS/MSD results which fall outside the control limits must be addressed in the narrative.

- If the native analyte concentration in the MS/MSD exceeds 4x the spike level for that analyte, the recovery data are reported as NC, MSB (i.e., not calculated). If the reporting software does not have the ability to report NC, MSB then the actual recovery must be reported and narrated as follows: “Results outside of limits do not necessarily reflect poor method performance in the matrix due to high analyte concentrations in the sample relative to the spike level.” Two other narrative notes for metals analyses: Matrix spike/spike duplicate spike recovery/recoveries was/were outside the acceptance limits of some analytes. The acceptable LCS analysis data indicated that the analytical system was operating within control and this condition is most likely due to matrix interference. See the Matrix Spike Report for the affected analytes which will be flagged with N. Matrix spike/spike duplicate relative percent difference (RPD) exceeded the acceptance limits for some analytes. The imprecision may be attributed to sample heterogeneity. See the Matrix Spike Report for the affected analytes, which will be flagged with *.
 - If an MS/MSD is not possible due to limited sample volume then a laboratory control sample duplicate (LCSD) should be analyzed. The RPD of the LCS and LCSD must be compared to the matrix spike RPD limits.
 - For dissolved metals samples which have not been digested, a MS/MSD must be performed per batch of up to 20 samples by spiking two aliquots of the sample at the levels specified in Table III (Appendix A).
- 9.5. Dilution test – A dilution test is performed to determine whether significant physical or chemical interferences exist due to the sample matrix. One sample per preparation batch must be processed as a dilution test. The test is performed by running a sample at a 5x (1:4) dilution. Samples identified as field blanks cannot be used for dilution tests. The results of the diluted sample, after correction for dilution, should agree within 10% of the original sample determination when the original sample concentration is greater than 50x the IDL. If the results are not within 10%, the possibility of chemical or physical interference exists.
- 9.6. Initial Calibration Verification (ICV/ICB) - Calibration accuracy is verified by analyzing a second source standard (ICV). For analyses conducted under Method 200.7, the ICV result must fall within 5% of the true value for that solution with relative standard deviation <3% from replicate (minimum of two) exposures. For Method 6010B, the ICV must fall within 10% of the true value for that solution with relative standard deviation <5% from replicate (minimum of two) exposures. An ICB is analyzed immediately following the ICV to monitor low level accuracy and system cleanliness. The ICB result must fall within +/- the RL from zero. If either the ICV or ICB fail to meet criteria, the analysis should be terminated, the problem corrected,

the instrument recalibrated and the calibration reverified. (See Section 11.10 or 11.13 for required run sequence).

- 9.7. Continuing Calibration Verification (CCV/CCB) - Calibration accuracy is monitored throughout the analytical run through the analysis of a known standard after every 10 samples. The CCV is be a mid-range standard made from a dilution of the calibration standard. The CCV for both methods must fall within 10% of the true value for that solution with relative standard deviation <5% from replicate (minimum of two) exposures. A CCB is analyzed immediately following each CCV. (See Section 11.10 or 11.13 for required run sequence.) The CCB result must fall within +/- RL from zero. If the blank is less than 1/10 the concentration of the action level of interest, and no sample is within 10% of the action limit, reanalysis and recalibration are not required before continuation of the run. Sample results may only be reported when bracketed by valid CCV/CCB pairs. If a mid-run CCV or CCB fails, the analysis for the affected element must be terminated, the problem corrected, the instrument recalibrated, the calibration verified and the affected samples reanalyzed. (Refer to Section 11.13 for an illustration of the appropriate rerun sequence).
- 9.8. Interference Check Analysis (ICSA/ICSAB) - The validity of the interelement correction factors is demonstrated through the successful analysis of interference check solutions. The ICSA contains only interfering elements, the ICSAB contains analytes and interferents. Refer to Table V (Appendix A) for the details of ICSA and ICSAB composition. Custom STL North Canton multielement ICS solutions must be used. All analytes should be spiked into the ICSAB solution, therefore, if a non-routine analyte is required then it should be manually spiked into the ICSAB using a certified ultra high purity single element solution or custom lab-specific mix. If the ICP will display overcorrection as a negative number then the non-routine elements can be controlled from the ICSA as described in section 9.8.3. Elements known to be interferents on a required analyte must be included in the ICP run when that analyte is determined. Aluminum, iron, calcium and magnesium must always be included in all ICP runs.
- 9.8.1. The ICSA and ICSAB solutions must be run at the beginning of the run. (See Section 11.10 or 11.13 for required run sequence.)
- 9.8.2. The ICSAB results for the interferents must fall within 80 - 120% of the true value. If any ICSAB interferent result fails criteria, the analysis should be terminated, the problem corrected, the instrument recalibrated and the samples rerun.
- 9.8.3. ICSA results for the non-interfering elements with reporting limits $\leq 10 \mu\text{g/L}$ must fall within the STL North Canton guidelines of $\pm 2x$ RL from zero. ICSA results for the non-interfering elements with RLs $> 10 \mu\text{g/L}$ must fall within the STL North Canton guidelines of $\pm 1x$ RL from zero. If the ICSA results

for the non-interfering elements do not fall within $\pm 2x$ RL (RL ≤ 10) or $\pm 1x$ RL (RL > 10) from zero the field sample data must be evaluated as follows:

- If the non-interfering element concentration in the ICSA is the result of contamination versus a spectral interference, and this reason is documented, the field sample data can be accepted.
 - If the affected element was not required then the sample data can be accepted.
 - If the interfering elements are not present in the field sample at a concentration which would result in a false positive or negative result greater than $\pm 2x$ RL from zero then the field sample data can be accepted.
 - If the interfering element is present in the field sample at a level which would result in a false analyte signal greater than $\pm 2x$ RL from zero, the data can be accepted only if the concentration of the affected analyte in the field sample is more than 10x the analyte signal in the ICSA.
 - If the data does not meet the above conditions then the IECs must be re-evaluated and corrected if necessary and the affected samples reanalyzed or the sample results manually corrected through application of the new IEC to the raw results. If the results are recalculated manually the calculations must be clearly documented on the raw data.
- 9.9. CRI - To verify linearity near the CRDL for ICP analysis, a CRI standard is run at the beginning of each sample analysis run but not before the ICV. (See Section 11.10 or 11.13 for required run sequence).
- Note: The custom STL North Canton CRI mix contains most analytes at a level two times the standard lab reporting limit.
- 9.10. Method of Standard Addition (MSA) -This technique involves adding known amounts of standard to one or more aliquots of the processed sample solution. This technique compensates for a sample interferent that may enhance or depress the analyte signal, thus producing a different slope from that of the calibration standards. It will not correct for additive interferences which cause a baseline shift. Refer to Section 11.17 for additional information on when MSA is required as well as Appendix D for specific MSA requirements.
- 9.11. Quality Assurance/Project Summaries - Certain clients may require project- or program-specific QC which may supersede this SOP requirements. Quality Assurance Summaries (QASs) or equivalent documents providing project-specific

requirements should be developed so that project staff clearly understand the special project requirements.

10. CALIBRATION AND STANDARDIZATION

- 10.1. Set up the instrument with the operating parameters recommended by the manufacturer. Allow the instrument to become thermally stable before beginning calibration (approximately 30 minutes of warm-up is required).
- 10.2. Profile and calibrate the instrument according to the instrument manufacturer's recommended procedures. Flush the system with the calibration blank between each standard or as the manufacturer recommends. The calibration curve must consist of a minimum of a blank and a standard. Refer to the facility-specific instrument SOP or ICP instrument manual for a detailed set up and operation protocols.
- 10.3. Calibration must be performed daily and each time the instrument is set up. Instrument runs may be continued over periods exceeding 24 hours as long as all calibration verification (CCV) and interference check QC criteria are met. The instrument standardization date and time must be included in the raw data.
- 10.4. Refer to Section 9.0 for calibration verification procedures, acceptance criteria and corresponding corrective actions.

11. PROCEDURE

- 11.1. For 200.7 analyses, dissolved (preserved) samples must be digested unless it can be documented that the sample meets all of the following criteria:
 - A. Visibly transparent with a turbidity measurement of 1 NTU or less.
 - B. Is of one liquid phase and free of particulate or suspended matter following acidification.
 - C. Is NOT being analyzed for silver.
- 11.2. A minimum of two exposures for each standard, field sample and QC sample is required. The average of the exposures is reported. For Trace ICP analyses, the results of the sum channel must be used for reporting.
- 11.3. Prior to calibration and between each sample/standard the system is rinsed with the calibration blank solution. The minimum rinse time between analytical samples is 60 seconds unless following the protocol outlined in 9.1.6 it can be demonstrated that a shorter rinse time may be used. Triton-X can be added to the rinse solution to facilitate the rinse process.

- 11.4. The use of an autosampler for all runs is strongly recommended.
- 11.5. The use of automated QC checks through the instrument software is highly recommended for all calibration verification samples (ICV,CCV), blanks (ICB,CCB,PB), interference checks (ICSA,ICSAB) and field samples (linear range) to improve the data review process.
- 11.6. To facilitate the early identification of QC failures and samples requiring rerun it is strongly recommended that sample data be reviewed periodically throughout the run.
- 11.7. To facilitate the data review and reporting processes it is strongly recommended that all necessary dilutions be performed before closing out the instrument run.
- 11.8. The use of an internal standard is recommended on the conventional, non-Trace ICPs as an alternative to using the method of standard additions. This technique is useful in overcoming matrix interferences especially in high solids matrices. However, for conventional ICP techniques, internal standards may not be necessary provided that one of the following is performed to minimize physical interferences: (1) peristaltic pump is used, (2) high solids nebulizer is used, or (3) high solids samples are diluted and reanalyzed.
- 11.9. The use of an internal standard is **required** on the Trace ICP unless the calibration and QC standards are matrix matched to each digestion procedure used as follows:

Preparation Method	% HNO ₃	% HCl
CLP Aqueous	1	5
CLP Soil	5	2.5
SW846 3050	10	10
SW846 3005	2	5
SW846 3010	6	5

The following procedural guidelines must be followed when using an internal standard:

- 11.9.1. Typically used internal standards are: yttrium or scandium. (Note: Any element can be used that is not typically found in environmental samples at a high rate of occurrence.)

- 11.9.2. The internal standard (IS) must be added to every sample and standard at the same concentration. It is recommended that the IS be added to each analytical sample automatically through use of a third pump channel and mixing coil. Internal standards should be added to blanks, samples and standards in a like manner, so that dilution effects resulting from the addition may be disregarded.
- 11.9.3. The concentration of the internal standard should be sufficiently high to obtain good precision in the measurement of the IS analyte used for data correction and to minimize the possibility of correction errors if the IS analyte is naturally present in the sample.
- 11.9.4. The internal standard raw intensity counts must be printed on the raw data.
- 11.9.5. The analyst must monitor the response of the internal standard throughout the sample analysis run. This information is used to detect potential problems and identify possible background contributions from the sample (i.e., natural occurrence of IS analyte). The instrument automatically adjusts sample results based on comparison of the internal standard intensity in the sample to the internal standard intensity at calibration.
 - 11.9.5.1. If the internal standard counts fall within $\pm 30\%$ of the counts observed in the ICB then the data is acceptable.
 - 11.9.5.2. If the internal standard counts in the field samples are more than $\pm 30\%$ higher than the expected level, the field samples must then be:
 - (1) Diluted and reanalyzed;
 - (2) The IS concentrations must be raised; or
 - (3) A different internal standard must be used.

11.10. The following analytical sequence must be used for Methods 6010B and 200.7:

Instrument Calibration

ICV

ICB

CRI

ICSA

ICSAB

7 samples

CCV

CCB

10 samples

CCV

CCB

Repeat sequence of up to 10 samples between CCV/CCB pairs as required to
complete run

CCV

CCB

Refer to Quality Control Section 9.0 and Table VII (Appendix A) for Method 6010B
and 200.7 quality control criteria.

- 11.11. Additional quality control analyses are necessary for analysis under the Contract Laboratory Program (CLP). If these are included then CLP, 6010 and 200.7 samples can be included in the same sequence. Refer to CORP-MT-0002NC for details.
- 11.12. Full method required QC must be available for each wavelength used in determining reported analyte results.
- 11.13. The following run sequence provides an illustration of a mid-run CCV or CCB failure and the appropriate corrective action run sequence as described in Section 9.7:

Original Run: Instrument Calibration

ICV

ICB

CRI

ICSA

ICSAB

7 samples

CCV1

CCB1

10 samples

CCV2

CCB2

10 samples **

CCV3 * * Failure occurs at CCV3/CCB3

CCB3 * **Samples requiring rerun for affected analytes

10 samples **

CCV4

CCB4

10 samples

CCV5

CCB5

Reanalysis: Recalibrate

ICV

ICB

CRI

ICSA

ICSAB

CCV2

CCB2

10 samples

CCV3

CCB3

10 samples

CCV4

CCB4

Notes:

Samples between CCV4 and CCV5 do not require reanalysis as they were bracketed by compliant QC samples.

See CORP-MT-0002NC for the appropriate reanalysis sequence if CLP requirements must also be met.

- 11.14. Guidelines are provided in the appendices on procedures to minimize contamination of samples and standards, preventive maintenance and troubleshooting.
- 11.15. All measurements must fall within the defined linear range where spectral interference correction factors are valid. Dilute and reanalyze all samples for required analytes that exceed the linear range or use an alternate wavelength for which QC data are established. If an interelement correction exists for an analyte which exceeds the linear range, the IEC may be inaccurately applied. Therefore, even if an overrange analyte may not be required to be reported for a sample, if that analyte is a interferent for any requested analyte in that sample, the sample must be diluted. Acid strength must be maintained in the dilution of samples.
- 11.16. For TCLP samples, full four-point MSA will be required if all of the following conditions are met:
- 1) recovery of the analyte in the matrix spike is not at least 50%,
 - 2) the concentration of the analyte does not exceed the regulatory level, and,
 - 3) the concentration of the analyte is within 20% of the regulatory level.
- The reporting and regulatory limits for TCLP analyses as well as matrix spike levels are detailed in Table VI (Appendix A). Appendix E provides guidance on performing MSA analyses.
- 11.17. Any variation in procedure shall be completely documented using instrument run logs, maintenance logs, report narratives, a Nonconformance Memo, or an anomaly report and is approved by a Supervisor/Group Leader and QA Manager. If contractually required, the client shall be notified by the Project Manager.
- 11.18. Nonconformance documentation shall be filed in the project file.
- 11.19. Any unauthorized deviations from this procedure must also be documented as a nonconformance, with a cause and corrective action described.

12. DATA ANALYSIS AND CALCULATIONS

12.1. ICV percent recoveries are calculated according to the equation:

$$\%R = 100 \left(\frac{\text{Found}(ICV)}{\text{True}(ICV)} \right)$$

12.2. CCV percent recoveries are calculated according to the equation:

$$\%R = 100 \left(\frac{\text{Found}(CCV)}{\text{True}(CCV)} \right)$$

12.3. Matrix Spike Recoveries are calculated according to the following equation:

$$\%R = 100 \left(\frac{SSR - SR}{SA} \right)$$

Where:

SSR = Spike Sample Result

SR = Sample Result

SA = Spike Added

12.4. The relative percent difference (RPD) of matrix spike/matrix spike duplicates are calculated according to the following equations:

$$RPD = 100 \left[\frac{|MSD - MS|}{\left(\frac{MSD + MS}{2} \right)} \right]$$

Where:

MS = determined spiked sample concentration

MSD = determined matrix spike duplicate concentration

12.5. The final concentration for a digested aqueous sample is calculated as follows:

$$mg / L = \frac{C \times V1 \times D}{V2}$$

Where:

C = Concentration (mg/L) from instrument readout

D = Instrument dilution factor
V1 = Final volume in liters after sample preparation
V2 = Initial volume of sample digested in liters

- 12.6. The final concentration determined in digested solid samples when reported on a dry weight basis is calculated as follows:

$$mg / Kg, dry weight = \frac{C \times V \times D}{W \times S}$$

Where:

C = Concentration (mg/L) from instrument readout
D = Instrument dilution factor
V = Final volume in liters after sample preparation
W = Weight in Kg of wet sample digested
S = Percent solids/100

Note: A Percent Solids determination must be performed on a separate aliquot when dry weight concentrations are to be reported. If the results are to be reported on wet weight basis the "S" factor should be omitted from the above equation.

- 12.7. The LCS percent recovery is calculated according to the following equation:

$$\%R = 100 \left(\frac{Found(LCS)}{True(LCS)} \right)$$

- 12.8. The dilution test percent difference for each component is calculated as follows:

$$\%Difference = \frac{|I - S|}{I} \times 100$$

Where:

I = Sample result (Instrument reading)
S = Dilution test result (Instrument reading × 5)

- 12.9. Appropriate factors must be applied to sample values if dilutions are performed.
- 12.10. Sample results should be reported with up to three significant figures in accordance with the STL North Canton significant figure policy.

13. METHOD PERFORMANCE

- 13.1. Each laboratory must have initial demonstration of performance data on file for each analyte of interest as described in Section 9.0.
- 13.2. Refer to Tables I, IA & II in Appendix A for the list of Method 6010B and 200.7 analytes as well as additional analytes that may be analyzed using this SOP.
- 13.3. Method performance is determined by the analysis of MS and MSD samples as well as method blanks and laboratory control samples. The MS or MSD recovery should fall within +/- 25 % and the MS/MSD should compare within 20% RPD or within the laboratory's historical acceptance limits. These criteria apply to analyte concentrations greater than or equal to 10xIDL. Method blanks must meet the criteria specified in Section 9.2. The laboratory control samples should recover within 20% of the true value or within the laboratory's historical acceptance limits.
- 13.4. Training Qualification:

The group/team leader or the supervisor has the responsibility to ensure that this procedure is performed by an associate who has been properly trained in its use and has the required experience.

14. POLLUTION PREVENTION

- 14.1. This method does not contain any specific modifications that serve to minimize or prevent pollution.

15. WASTE MANAGEMENT

- 15.1. Waste generated in the procedure must be segregated and disposed of according to the facility hazardous waste procedures and per the local, state, and federal regulations. The Environmental Health and Safety Director should be contacted, if additional information is required.
- 15.2. Standards should be purchased and prepared in volumes consistent with laboratory use to minimize the volume of expired standards to be disposed.

16. REFERENCES

- 16.1. 40 CFR Part 136, Appendix B, 7-5-95, Determination of Method Detection Limits.
- 16.2. Test Methods for Evaluating Solid Waste , Physical/Chemical Methods, SW-846, 3rd Edition, Final Update III, Revision 2, December 1996. Method 6010B.

- 16.3. Determination of Metals and Trace Elements in Water and Wastes by Inductively Coupled Plasma-Atomic Emission Spectrometry, Revision 4.4, May 1994. Method 200.7.
 - 16.4. CORP-MT-0002NC, Inductively Coupled Plasma-Atomic Emission Spectroscopy, Method 200.7 & CLP-M, SOW ILMO3.0 and ILMO4.0).
 - 16.5. QA-003, STL North Canton QC Program.
 - 16.6. QA-004, Rounding and Significant Figures.
 - 16.7. QA-005, Method Detection Limits.
- 17. MISCELLANEOUS (TABLES, APPENDICES, ETC.)**
- 17.1. Modifications/Interpretations from reference method
 - 17.1.1. Modifications/interpretations from both Methods 6010B and 200.7.
 - 17.1.1.1. STL North Canton laboratories use mixed calibration standard solutions purchased from approved vendors instead of using individual mixes prepared in house as recommended by the subject methods.
 - 17.1.1.2. Methods 200.7 and 6010B state that if the correction routine is operating properly, the determined apparent analyte(s) concentration from analysis of each interference solution should fall within a specific concentration range around the calibration blank. In determining IECs, because of lack of definition clarification for "concentration range around the calibration blank," STL North Canton has adopted the procedure in EPA CLP ILMO4.0.
 - 17.1.1.3. Section 8.5 of Method 6010B and Section 9.5 of Method 200.7 recommend that whenever a new or unusual matrix is encountered, a series of tests be performed prior to reporting concentration data for that analyte. The dilution test helps determine if a chemical or physical interference exists. Because STL North Canton laboratories receive no prior information from clients regarding when to expect a new or unusual matrix, STL North Canton may select to perform a dilution test on one sample in each prep batch. According to the method, the post digestion spike (PDS) determines any

potential matrix interferences. At STL North Canton labs, matrix interference is determined by evaluating data for the LCS and MS/MSD. STL North Canton requires documented, clear guidance when a new or unusual matrix will be received for a project and a request to perform the dilution test or PDS on a client-identified sample.

17.1.2. Modifications from Method 200.7.

- 17.1.2.1. Method 200.7 defines the IDL as the concentration equivalent to a signal, due to the analyte, which is equal to three times the standard deviation of a series of ten replicate measurements of the calibration blank signal at the same wavelength. STL North Canton labs utilize the CLP IDL definition as defined in Section 9.1.1 of this SOP.
- 17.1.2.2. The calibration blank is prepared in an acid matrix of 5% HNO₃/5% HCl instead of the specified 2% HNO₃/10% HCl matrix as the former matrix provides for improved performance relative to the wide variety of digestate acid matrices which result from the various EPA preparation protocols applied.
- 17.1.2.3. Method section 9.3.4 specifies that "Analysis of the IPC (ICSA/AB) solution immediately following calibration must verify that the instrument is within $\pm 5\%$ of calibration with a relative standard deviation $<3\%$ from replicate integrations ≥ 4 ." STL North Canton uses a minimum of two exposures.
- 17.1.2.4. Section 7.12 of 200.7 indicates that the QCS (ICV) should be prepared at a concentration near 1 ppm. The ICV specified in this SOP accommodates the 1 ppm criteria for the majority of analytes. For the remaining analytes, this SOP specifies ICV concentrations which are appropriate to the range of calibration. The intent of the ICV, verification of calibration standard accuracy, is independent of the ICV concentration used.
- 17.1.2.5. The ICS criteria applied by this SOP differ from those stated in the method. Method 200.7 section 10.4 states that results should fall within the established control limits of 3 times the standard deviation of the calibration blank for that analyte. The control limits listed in this SOP are those applicable to the EPA designed solution.
- 17.1.2.6. Method 200.7 section 9.3.4 states the CCB should be less than the IDL, but $>$ the lower 3-sigma control limit of the calibration blank.

The intent of this requirement is to ensure that the calibration is not drifting at the low end. STL North Canton has adopted an absolute control limit of +/- RL from zero for calibration blank criteria. SOP section 9.7 provides the detailed corrective action criteria that must be followed.

17.1.3. Modifications from Method 6010B.

17.1.3.1. Chapter 1 of SW-846 states that the method blank should not contain any analyte of interest at or above the MDL. This SOP states that the method blank must not contain any analyte of interest at or above the reporting limit. Common lab contaminants are allowed up to two times the reporting limit in the blank following consultation with the client.

17.1.3.2. Method 6010B section 8.6.1.3 states that the results of the calibration blank are to agree within 3x the IDL. If not, repeat the analysis two or more times and average the results. If the average is not within three standard deviations of the background mean, terminate the analysis, correct the problem, recalibrate, and reanalyze the previous 10 samples. The intent of this requirement is to ensure that the calibration is not drifting at the low end. STL North Canton has adopted an absolute control limit of +/- RL from zero for calibration blank criteria. See SOP Section 9.7 for a detailed description of the required corrective action procedures.

17.2. Modifications from previous SOP

Refer to revision 2 of this SOP.

17.3. Facility-Specific SOPs

Each facility shall review and revise as appropriate this SOP to reflect any facility-specific requirements. If no facility-specific amendments are required, the SOP can be adopted as is.


17.4. Documentation and Record Management

The following documentation comprises a complete ICP raw data package:

- Raw data (direct instrument printout).
- Relevant sample preparation benchsheets.

- Run log printout from instrument software where this option is available (TJA) or manually generated run log (i.e., Ward WSL printout).
- Data review checklist - See Appendix B.
- Standards documentation (including prep and expiration dates, source, and lot #).
- Nonconformance/anomaly documentation (if applicable).

17.5. Flow Diagram



APPENDIX A
TABLES

TABLE I. Method 200.7 and 6010B Target Analyte List

ELEMENT	Symbol	CAS #	6010B analyte	200.7 analyte	Reporting Limit (ug/L) Water	Reporting Limit (mg/kg) Soil
Aluminum	Al	7429-90-5	X	X	200	20
Antimony	Sb	7440-36-0	X	X	60	6
Arsenic	As	7440-38-2	X	X	300	30
Barium	Ba	7440-39-3	X	X	200	20
Beryllium	Be	7440-41-7	X	X	5.0	0.5
Boron	B	7440-42-8		X	200	20
Cadmium	Cd	7440-43-9	X	X	5.0	0.5
Calcium	Ca	7440-70-2	X	X	5000	500
Chromium	Cr	7440-47-3	X	X	10	1
Cobalt	Co	7440-48-4	X	X	50	5
Copper	Cu	7440-50-8	X	X	25	2.5
Iron	Fe	7439-89-6	X	X	100	10
Lead	Pb	7439-92-1	X	X	100	10
Magnesium	Mg	7439-95-4	X	X	5000	500
Manganese	Mn	7439-96-5	X	X	15	1.5
Molybdenum	Mo	7439-98-7	X	X	40	4
Nickel	Ni	7440-02-0	X	X	40	4
Potassium	K	7440-09-7	X	X	5000	500
Selenium	Se	7782-49-2	X	X	250	25
Silver	Ag	7440-22-4	X	X	10	1
Sodium	Na	7440-23-5	X	X	5000	500
Thallium	Tl	7440-28-0	X	X	2000	200
Vanadium	V	7440-62-2	X	X	50	5
Zinc	Zn	7440-66-6	X	X	20	2

TABLE IA. Method 200.7 and 6010B Trace ICP Target Analyte List

ELEMENT	Symbol	CAS #	Reporting Limit (ug/L) Water	Reporting Limit (mg/kg) Soil
Arsenic	As	7440-38-2	10	1.0
Lead	Pb	7439-92-1	3.0	0.3
Selenium	Se	7782-49-2	5.0	0.5
Thallium	Tl	7440-28-0	10	1.0
Antimony	Sb	7440-36-0	10	1.0
Cadmium	Cd	7440-43-9	2.0	0.2
Silver	Ag	7440-22-4	5.0	0.5
Chromium	Cr	7440-47-3	5.0	0.5

TABLE II. Non-Routine Analyte List

ELEMENT	Symbol	CAS #	Reporting Limit (ug/L) Water	Reporting Limit (mg/kg) Soil
Tin	Sn	7440-31-5	100	10
Titanium	Ti	7440-32-6	50	5

TABLE III. Matrix Spike and Aqueous Laboratory Control Sample Levels

ELEMENT	LCS Level (ug/L)	Matrix Spike Level (ug/L)
Aluminum	2000	2000
Antimony	500	500
Arsenic	2000	2000
Barium	2000	2000
Beryllium	50	50
Cadmium	50	50
Calcium	50000	50000
Chromium	200	200
Cobalt	500	500
Copper	250	250
Iron	1000	1000
Lead	500	500
Magnesium	50000	50000
Manganese	500	500
Molybdenum	1000	1000
Nickel	500	500
Potassium	50000	50000
Selenium	2000	2000
Silver	50	50
Sodium	50000	50000
Thallium	2000	2000
Vanadium	500	500
Zinc	500	500
Boron	1000	1000
Tin	2000	2000
Titanium	1000	1000