



Wisconsin State Laboratory of Hygiene
 2601 Agriculture Drive, PO Box 7996
 Madison, WI 53707-7996
 (800)442-4618 - FAX (608)224-6213
<http://www.slh.wisc.edu>

Laboratory Report

Environmental Health Division

WSLH Sample: 536912001

Report To:
 CASSY HANAN
 TOWN OF CAMPBELL
 2219 BAINBRIDGE ST
 LA CROSSE, WI 54603

Invoice To:
 CASSY HANAN
 TOWN OF CAMPBELL
 2219 BAINBRIDGE ST
 LA CROSSE, WI 54603
 Customer ID: 268768

Field #: TRIZMA-SLED7801 COMMUNITY
 Project No:
 Collection End: 11/17/2020 2:00:00 PM
 Collection Start:
 Collected By: LENNY HASZ
 Date Received: 11/18/2020
 Date Reported: 2/16/2021
 Sample Reason:

ID#:
 Sample Location:
 Sample Description: KITCHEN SINK FAUCET
 Sample Type: PO-PRIVATE WELL
 Waterbody:
 Point or Outfall:
 Sample Depth:
 Program Code:
 Region Code:
 County:

PFAS in Water

Analyte	Analysis Method	Result	Units	LOD	LOQ
Prep Date: 12/10/20 08:45		Analysis Date: 12/28/20 12:53			
PFPeA (2706-90-3)	Modified ISO 21675	1.39	ng/L	0.339	0.379
PFBS (375-73-5)	Modified ISO 21675	5.90	ng/L	0.419	0.947
4:2 FTSA (757124-72-4)	Modified ISO 21675	<0.432	ng/L	0.432	0.947
PFHxA (307-24-4)	Modified ISO 21675	1.69	ng/L	0.400	0.947
PFPeS (2706-91-4)	Modified ISO 21675	0.488	ng/L	0.259	0.379
HFPO-DA (13252-13-6)	Modified ISO 21675	<0.505	ng/L	0.505	0.947
PFHpA (375-85-9)	Modified ISO 21675	0.489F	ng/L	0.451	0.947
PFHxS (355-46-4)	Modified ISO 21675	4.41	ng/L	0.392	0.947
DONA (919005-14-4)	Modified ISO 21675	<0.402	ng/L	0.402	0.947
6:2 FTSA (27619-97-2)	Modified ISO 21675	<0.489	ng/L	0.489	0.947
PFOA (335-67-1)	Modified ISO 21675	4.89	ng/L	0.439	0.947
PFHpS (375-92-8)	Modified ISO 21675	<0.383	ng/L	0.383	0.947
PFOS (1763-23-1)	Modified ISO 21675	5.96	ng/L	0.325	0.379
PFNA (375-95-1)	Modified ISO 21675	<0.413	ng/L	0.413	0.947
9CI-PF3ONS (756426-58-1)	Modified ISO 21675	<0.401	ng/L	0.401	0.947
8:2 FTSA (39108-34-4)	Modified ISO 21675	<0.427	ng/L	0.427	0.947

Environmental Health Division

WSLH Sample: 536912001

PFAS in Water

Analyte	Analysis Method	Result	Units	LOD	LOQ
Prep Date: 12/10/20 08:45		Analysis Date: 12/28/20 12:53			
PFDA (335-76-2)	Modified ISO 21675	<0.376	ng/L	0.376	0.947
PFNS (68259-12-1)	Modified ISO 21675	<0.478	ng/L	0.478	0.947
N-MeFOSAA (2355-31-9)	Modified ISO 21675	<0.513	ng/L	0.513	0.947
N-EtFOSAA (2991-50-6)	Modified ISO 21675	<0.410	ng/L	0.410	0.947
FOSA (754-91-6)	Modified ISO 21675	<3.89	ng/L	3.89	4.73
PFUnA (2058-94-8)	Modified ISO 21675	<0.389	ng/L	0.389	0.947
PFDS (335-77-3)	Modified ISO 21675	<0.436	ng/L	0.436	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
11CI-PF3OUdS (763051-92-9)	Modified ISO 21675	<0.377	ng/L	0.377	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
PFDoA (307-55-1)	Modified ISO 21675	<0.367	ng/L	0.367	0.947
10:2 FTSA (120226-60-0)	Modified ISO 21675	<0.415	ng/L	0.415	0.947
PFDoS (79780-39-5)	Modified ISO 21675	<0.495	ng/L	0.495	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
PFTrDA (72629-94-8)	Modified ISO 21675	<0.383	ng/L	0.383	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-MeFOSA (31506-32-8)	Modified ISO 21675	<0.769	ng/L	0.769	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-MeFOSE (24448-09-7)	Modified ISO 21675	<0.387	ng/L	0.387	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-EtFOSA (4151-50-2)	Modified ISO 21675	<0.630	ng/L	0.630	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-EtFOSE (1691-99-2)	Modified ISO 21675	<0.395	ng/L	0.395	0.947
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
PFTeDA (376-06-7)	Modified ISO 21675	<0.339	ng/L	0.339	0.379



Wisconsin State Laboratory of Hygiene
2601 Agriculture Drive, PO Box 7996
Madison, WI 53707-7996
(800)442-4618 - FAX (608)224-6213
<http://www.slh.wisc.edu>

Laboratory Report

Environmental Health Division

WSLH Sample: 536912001

WDNR LAB ID:113133790 NELAP LAB ID:2091 EPA LAB ID:WI00007, WI00008 WI DATCP ID:105-415

List of Abbreviations:

LOD = Level of detection
LOQ = Level of quantification
ND = None detected. Results are less than the LOD
F next to result = Result is between LOD and LOQ
Z next to result = Result is between 0 (zero) and LOD
if LOD=LOQ, Limits were not statistically derived

Test results for NELAP accredited tests are certified to meet the requirements of the NELAC standards. For a list of accredited analytes

see <http://www.slh.wisc.edu/about/compliance/nelac-laboratory-accreditation>

Results, LOD and LOQ values have been adjusted for analytical dilutions and percent moisture where applicable.

Results relate only to the items tested.

This Laboratory Report shall not be reproduced except in full, without written approval of the laboratory.

The water microbiology unit analyzes samples as received and not all samples are tested for preservation before analysis is performed.

Responsible Party

Inorganic Chemistry: Graham Anderson, Supervisor 608-224-6281

Metals: Graham Anderson, Supervisor 608-224-6281

Organics: Erin Mani, Supervisor 608-224-6269

Environmental Toxicology: Dawn Perkins, Supervisor 608-224-6230

Water Microbiology: Martin Collins, Supervisor 608-224-6239

Radiochemistry: David Webb, Division Director 608-224-6227



Wisconsin State Laboratory of Hygiene
 2601 Agriculture Drive, PO Box 7996
 Madison, WI 53707-7996
 (800)442-4618 - FAX (608)224-6213
<http://www.slh.wisc.edu>

Laboratory Report

Environmental Health Division

WSLH Sample: 536912003

Report To:
 CASSY HANAN
 TOWN OF CAMPBELL
 2219 BAINBRIDGE ST
 LA CROSSE, WI 54603

Invoice To:
 CASSY HANAN
 TOWN OF CAMPBELL
 2219 BAINBRIDGE ST
 LA CROSSE, WI 54603
 Customer ID: 268768

Field #: TRIZMA SLB26597 LIBRARY
 Project No:
 Collection End: 11/17/2020 1:30:00 PM

ID#:
 Sample Location:
 Sample Description: WOMENS BATHROOM SINK
 FAUCET

Collection Start:
 Collected By: LENNY HASZ
 Date Received: 11/18/2020
 Date Reported: 2/16/2021
 Sample Reason:

Sample Type: PO-PRIVATE WELL
 Waterbody:
 Point or Outfall:
 Sample Depth:
 Program Code:
 Region Code:
 County:

PFAS in Water

Analyte	Analysis Method	Result	Units	LOD	LOQ
Prep Date: 12/14/20 12:45		Analysis Date: 12/22/20 09:31			
PFBA (375-22-4)	Modified ISO 21675	142	ng/L	1.84	3.68
Interference					
PFPeA (2706-90-3)	Modified ISO 21675	3.63	ng/L	0.329	0.368
PFBS (375-73-5)	Modified ISO 21675	3.02	ng/L	0.407	0.919
4:2 FTSA (757124-72-4)	Modified ISO 21675	<0.419	ng/L	0.419	0.919
PFHxA (307-24-4)	Modified ISO 21675	3.53	ng/L	0.389	0.919
PFPeS (2706-91-4)	Modified ISO 21675	1.04	ng/L	0.252	0.368
HFPO-DA (13252-13-6)	Modified ISO 21675	<0.490	ng/L	0.490	0.919
PFHpA (375-85-9)	Modified ISO 21675	1.12	ng/L	0.438	0.919
PFHxS (355-46-4)	Modified ISO 21675	6.25	ng/L	0.381	0.919
DONA (919005-14-4)	Modified ISO 21675	<0.391	ng/L	0.391	0.919
6:2 FTSA (27619-97-2)	Modified ISO 21675	<0.475	ng/L	0.475	0.919
PFOA (335-67-1)	Modified ISO 21675	12.8	ng/L	0.427	0.919
PFHpS (375-92-8)	Modified ISO 21675	<0.371	ng/L	0.371	0.919
PFOS (1763-23-1)	Modified ISO 21675	13.3	ng/L	0.315	0.368

Environmental Health Division

WSLH Sample: 536912003

PFAS in Water

Analyte	Analysis Method	Result	Units	LOD	LOQ
Prep Date: 12/14/20 12:45		Analysis Date: 12/22/20 09:31			
PFNA (375-95-1)	Modified ISO 21675	<0.401	ng/L	0.401	0.919
9CI-PF3ONS (756426-58-1)	Modified ISO 21675	<0.390	ng/L	0.390	0.919
8:2 FTSA (39108-34-4)	Modified ISO 21675	<0.415	ng/L	0.415	0.919
PFDA (335-76-2)	Modified ISO 21675	<0.365	ng/L	0.365	0.919
PFNS (68259-12-1)	Modified ISO 21675	<0.464	ng/L	0.464	0.919
N-MeFOSAA (2355-31-9)	Modified ISO 21675	<0.498	ng/L	0.498	0.919
N-EtFOSAA (2991-50-6)	Modified ISO 21675	<0.398	ng/L	0.398	0.919
FOSA (754-91-6)	Modified ISO 21675	<3.78	ng/L	3.78	4.60
PFOA (2058-94-8)	Modified ISO 21675	<0.378	ng/L	0.378	0.919
PFDS (335-77-3)	Modified ISO 21675	<0.423	ng/L	0.423	0.919
11CI-PF3OUdS (763051-92-9)	Modified ISO 21675	<0.366	ng/L	0.366	0.919
PFDoA (307-55-1)	Modified ISO 21675	<0.357	ng/L	0.357	0.919
10:2 FTSA (120226-60-0)	Modified ISO 21675	<0.403	ng/L	0.403	0.919
PFDoS (79780-39-5)	Modified ISO 21675	<0.481	ng/L	0.481	0.919
PFOTrDA (72629-94-8)	Modified ISO 21675	<0.372	ng/L	0.372	0.919
N-MeFOSA (31506-32-8)	Modified ISO 21675	<0.746	ng/L	0.746	0.919
N-MeFOSE (24448-09-7)	Modified ISO 21675	0.715F	ng/L	0.376	0.919
N-EtFOSA (4151-50-2)	Modified ISO 21675	<0.611	ng/L	0.611	0.919
N-EtFOSE (1691-99-2)	Modified ISO 21675	<0.383	ng/L	0.383	0.919
PFOTrDA (376-06-7)	Modified ISO 21675	<0.329	ng/L	0.329	0.368



Wisconsin State Laboratory of Hygiene
2601 Agriculture Drive, PO Box 7996
Madison, WI 53707-7996
(800)442-4618 - FAX (608)224-6213
<http://www.slh.wisc.edu>

Laboratory Report

Environmental Health Division

WSLH Sample: 536912003

WDNR LAB ID:113133790 NELAP LAB ID:2091 EPA LAB ID:WI00007, WI00008 WI DATCP ID:105-415

List of Abbreviations:

LOD = Level of detection
LOQ = Level of quantification
ND = None detected. Results are less than the LOD
F next to result = Result is between LOD and LOQ
Z next to result = Result is between 0 (zero) and LOD
if LOD=LOQ, Limits were not statistically derived

Test results for NELAP accredited tests are certified to meet the requirements of the NELAC standards. For a list of accredited analytes

see <http://www.slh.wisc.edu/about/compliance/nelac-laboratory-accreditation>

Results, LOD and LOQ values have been adjusted for analytical dilutions and percent moisture where applicable.

Results relate only to the items tested.

This Laboratory Report shall not be reproduced except in full, without written approval of the laboratory.

The water microbiology unit analyzes samples as received and not all samples are tested for preservation before analysis is performed.

Responsible Party

Inorganic Chemistry: Graham Anderson, Supervisor 608-224-6281
Metals: Graham Anderson, Supervisor 608-224-6281
Organics: Erin Mani, Supervisor 608-224-6269
Environmental Toxicology: Dawn Perkins, Supervisor 608-224-6230
Water Microbiology: Martin Collins, Supervisor 608-224-6239
Radiochemistry: David Webb, Division Director 608-224-6227



Wisconsin State Laboratory of Hygiene
 2601 Agriculture Drive, PO Box 7996
 Madison, WI 53707-7996
 (800)442-4618 - FAX (608)224-6213
<http://www.slh.wisc.edu>

Laboratory Report

Environmental Health Division

WSLH Sample: 536912002

Report To:
 CASSY HANAN
 TOWN OF CAMPBELL
 2219 BAINBRIDGE ST
 LA CROSSE, WI 54603

Invoice To:
 CASSY HANAN
 TOWN OF CAMPBELL
 2219 BAINBRIDGE ST
 LA CROSSE, WI 54603
 Customer ID: 268768

Field #: TRISMA SLCD7801 TOWN HALL
 Project No:
 Collection End: 11/17/2020 1:00:00 PM
 Collection Start:
 Collected By: LENNY HASZ
 Date Received: 11/18/2020
 Date Reported: 2/16/2021
 Sample Reason:

ID#:
 Sample Location:
 Sample Description: MENS BATHROOM SINK FAUCET
 Sample Type: PO-PRIVATE WELL
 Waterbody:
 Point or Outfall:
 Sample Depth:
 Program Code:
 Region Code:
 County:

PFAS in Water

Analyte	Analysis Method	Result	Units	LOD	LOQ
Prep Date: 12/10/20 08:45		Analysis Date: 12/28/20 13:21			
PFPeA (2706-90-3)	Modified ISO 21675	0.537	ng/L	0.337	0.377
PFBS (375-73-5)	Modified ISO 21675	0.833F	ng/L	0.417	0.942
4:2 FTSA (757124-72-4)	Modified ISO 21675	<0.430	ng/L	0.430	0.942
PFHxA (307-24-4)	Modified ISO 21675	0.673F	ng/L	0.399	0.942
PFPeS (2706-91-4)	Modified ISO 21675	<0.258	ng/L	0.258	0.377
HFPO-DA (13252-13-6)	Modified ISO 21675	<0.502	ng/L	0.502	0.942
PFHpA (375-85-9)	Modified ISO 21675	<0.448	ng/L	0.448	0.942
PFHxS (355-46-4)	Modified ISO 21675	1.32	ng/L	0.390	0.942
DONA (919005-14-4)	Modified ISO 21675	<0.400	ng/L	0.400	0.942
6:2 FTSA (27619-97-2)	Modified ISO 21675	<0.487	ng/L	0.487	0.942
PFOA (335-67-1)	Modified ISO 21675	1.96	ng/L	0.437	0.942
PFHpS (375-92-8)	Modified ISO 21675	<0.381	ng/L	0.381	0.942
PFOS (1763-23-1)	Modified ISO 21675	1.14	ng/L	0.323	0.377
PFNA (375-95-1)	Modified ISO 21675	<0.411	ng/L	0.411	0.942
9CI-PF3ONS (756426-58-1)	Modified ISO 21675	<0.400	ng/L	0.400	0.942
8:2 FTSA (39108-34-4)	Modified ISO 21675	<0.425	ng/L	0.425	0.942

Environmental Health Division

WSLH Sample: 536912002

PFAS in Water

Analyte	Analysis Method	Result	Units	LOD	LOQ
Prep Date: 12/10/20 08:45		Analysis Date: 12/28/20 13:21			
PFDA (335-76-2)	Modified ISO 21675	<0.374	ng/L	0.374	0.942
PFNS (68259-12-1)	Modified ISO 21675	<0.476	ng/L	0.476	0.942
N-MeFOSAA (2355-31-9)	Modified ISO 21675	<0.511	ng/L	0.511	0.942
N-EtFOSAA (2991-50-6)	Modified ISO 21675	<0.408	ng/L	0.408	0.942
FOSA (754-91-6)	Modified ISO 21675	<3.87	ng/L	3.87	4.71
PFUnA (2058-94-8)	Modified ISO 21675	<0.387	ng/L	0.387	0.942
PFDS (335-77-3)	Modified ISO 21675	<0.433	ng/L	0.433	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
11CI-PF3OUdS (763051-92-9)	Modified ISO 21675	<0.375	ng/L	0.375	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
PFDoA (307-55-1)	Modified ISO 21675	<0.366	ng/L	0.366	0.942
10:2 FTSA (120226-60-0)	Modified ISO 21675	<0.413	ng/L	0.413	0.942
PFDoS (79780-39-5)	Modified ISO 21675	<0.493	ng/L	0.493	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
PFTrDA (72629-94-8)	Modified ISO 21675	<0.382	ng/L	0.382	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-MeFOSA (31506-32-8)	Modified ISO 21675	<0.765	ng/L	0.765	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-MeFOSE (24448-09-7)	Modified ISO 21675	<0.385	ng/L	0.385	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-EtFOSA (4151-50-2)	Modified ISO 21675	<0.627	ng/L	0.627	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
N-EtFOSE (1691-99-2)	Modified ISO 21675	<0.393	ng/L	0.393	0.942
The Laboratory Control Spike (LCS) does not meet the lower QC limit.					
PFTeDA (376-06-7)	Modified ISO 21675	<0.337	ng/L	0.337	0.377



Wisconsin State Laboratory of Hygiene
2601 Agriculture Drive, PO Box 7996
Madison, WI 53707-7996
(800)442-4618 - FAX (608)224-6213
<http://www.slh.wisc.edu>

Laboratory Report

Environmental Health Division

WSLH Sample: 536912002

WDNR LAB ID:113133790 NELAP LAB ID:2091 EPA LAB ID:WI00007, WI00008 WI DATCP ID:105-415

List of Abbreviations:

LOD = Level of detection
LOQ = Level of quantification
ND = None detected. Results are less than the LOD
F next to result = Result is between LOD and LOQ
Z next to result = Result is between 0 (zero) and LOD
if LOD=LOQ, Limits were not statistically derived

Test results for NELAP accredited tests are certified to meet the requirements of the NELAC standards. For a list of accredited analytes

see <http://www.slh.wisc.edu/about/compliance/nelac-laboratory-accreditation>

Results, LOD and LOQ values have been adjusted for analytical dilutions and percent moisture where applicable.

Results relate only to the items tested.

This Laboratory Report shall not be reproduced except in full, without written approval of the laboratory.

The water microbiology unit analyzes samples as received and not all samples are tested for preservation before analysis is performed.

Responsible Party

Inorganic Chemistry: Graham Anderson, Supervisor 608-224-6281

Metals: Graham Anderson, Supervisor 608-224-6281

Organics: Erin Mani, Supervisor 608-224-6269

Environmental Toxicology: Dawn Perkins, Supervisor 608-224-6230

Water Microbiology: Martin Collins, Supervisor 608-224-6239

Radiochemistry: David Webb, Division Director 608-224-6227