



Northern Lake Service, Inc • 400 N Lake Ave • Crandon, WI 54520
800-278-1254 • www.nlslab.com

August 31, 2023

Mark Meske
3661 Rominski Road
Rhineland, WI 54501

Project: 2023 On-line Ordering
Project Number: [none]
Work Order: CB09971
Received: 08/21/23

Enclosed are the results of analyses for samples received by our laboratory on 8/21/2023. If you have any questions concerning this report, please feel free to contact a client service representative at clientservices@nlslab.com.

Sincerely,

A handwritten signature in black ink, appearing to read "Tom Priebe".

Tom Priebe For Kristin Tienor
Northern Lake Service, Inc.



Mark Meske
3661 Rominski Road
Rhineland, WI 54501

Project: 2023 On-line Ordering
Project Number:
Project Manager: Mark Meske

Reported:
8/31/23 13:00

Work Order:
CB09971

Sample Summary

Descriptions of all qualifiers listed throughout this report can be found on the Qualifiers and Definitions Page.

Lab ID	Sample	Matrix	Sample Type	Qualifiers	Date Sampled	Date Received
CB09971-01	Basement Faucet	DW			8/21/23 10:30	8/21/23 14:48



Mark Meske
3661 Rominski Road
Rhineland, WI 54501

Project: 2023 On-line Ordering
Project Number:
Project Manager: Mark Meske

Reported:
8/31/23 13:00

Work Order:
CB09971

Sample Results

Sample: Basement Faucet
CB09971-01 (DW) Sampled: 08/21/23 10:30

Analyte	Result	Qualifier	LOD	LOQ	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
Semi-Volatiles											
11-chloroeicosafluoro-3-oxaundecane-1-sulfonic acid (11Cl-PF3OUdS)	ND		0.32	1.0		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9Cl-PF3ONS)	ND		0.35	1.1		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND		0.38	1.2		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
hexafluoropropylene oxide dimer acid (HFPO DA)	ND		0.42	1.4		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
N-ethyl perfluorooctanesulfonamidoacetic acid (NEtFOSAA)	ND		0.48	1.6		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
n-methyl perfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		0.41	1.3		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorobutanesulfonic acid (PFBS)	ND		0.31	1.0		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorodecanoic acid (PFDA)	ND		0.34	1.1		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorododecanoic acid (PFDoA)	ND		0.23	0.79		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluoroheptanoic acid (PFHpA)	ND		0.45	1.5		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanoic acid (PFHxA)	ND		0.48	1.6		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanesulfonic acid (PFHxS)	ND		0.35	1.1		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorononanoic acid (PFNA)	ND		0.47	1.5		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanoic acid (PFOA)	ND		0.50	1.6		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanesulfonic acid (PFOS)	ND		0.32	1.0		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorotetradecanoic acid (PFTA)	ND		0.35	1.1		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluorotridecanoic acid (PFTrDA)	ND		0.44	1.4		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
perfluoroundecanoic acid (PFUnA)	ND		0.31	1.0		ng/L	8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFHxA	95%		Limits: 70-130%				8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-HFPODA	91%		Limits: 70-130%				8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFDA	94%		Limits: 70-130%				8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) d5-NEtFOSAA	80%		Limits: 70-130%				8/28/23 5:32	8/29/23 12:41	RAW	EPA 537.1, Rev 2.0	2



Mark Meske
3661 Rominski Road
Rhineland, WI 54501

Project: 2023 On-line Ordering
Project Number:
Project Manager: Mark Meske

Reported:
8/31/23 13:00

Work Order:
CB09971

List of Certifications

Code	Description	Number	Expires
2	NLS (Crandon) WDNR Laboratory ID No.	721026460	8/31/24



Mark Meske
3661 Rominski Road
Rhineland, WI 54501

Project: 2023 On-line Ordering
Project Number:
Project Manager: Mark Meske

Reported:
8/31/23 13:00

Work Order:
CB09971

Qualifiers and Definitions

Item	Definition
ND	Analyte NOT DETECTED at or above the LOD or MRL.
LOD	Limit of Detection.
LOQ	Limit of Quantitation.
NA	Not Applicable.
Dry	Dry Weight Basis.
Wet	Wet Weight Basis.
% Dry	Equal to: (mg/kg dry) / 10000.
1000 ug/L	Equal to: 1 mg/L.
MCL	Maximum Contaminant Levels for Drinking Water Samples. Shaded results indicate >MCL.
RPD	Relative Percent Difference.
%REC	Percent Recovery.
Source	Sample that was matrix spiked or duplicated.

All LOD/LOQs adjusted to reflect preparation volumes, dilutions, and/or solids content.

PFAS by EPA Method 537.1 Test Kit
18 PFAS compounds



CB09971

Return your sample as soon as possible to one of our two locations:

Northern Lake Service
400 N Lake Ave
Crandon, WI 54520

OR

Northern Lake Service
2420 N Grandview Blvd
Waukesha, WI 53188

Please provide the following information:

Name: MARK MESKE
Address: 3661 ROMINSKY RD
City/State/Zip: RHINELAUDER / WI / ~~53~~ 54501
Phone: 414-791-0126 Email: MESKEMARK@GMAIL.COM
Sample Collection Date: 8/21/2023 Sample Collection Time: 10:30 AM/PM
Sample Collection Location (ex. Pressure Tank) BASEMENT FAUCET
Sample Collected By: MARK MESKE
Report by email only Report by mail only (\$5 fee)

Per EPA 537.1, each sample set **must be accompanied by a field blank. The purpose of the field blank is to allow for the identification of potential contamination during sample collection and handling. If you choose not to have the field blank analyzed, the PFAS data may be viewed as a screening and may not be suitable for compliance use.

Please initial in **ONE** of the check boxes below that apply:

Cost for PFAS Kit Without Field Blank - \$367.50

OR

Cost for PFAS Kit Including Field Blank - \$552.50

Your credit card will be charged according to the option you have chosen above

8/21/23 14:48
Aly Pearce

Hand Del
3-20C #12