



James Finnan  
4122 Camp Bryn Afon Rd  
Rhinelander, WI 54501

Project: 2023 On-line Ordering  
Project Number: From well test valve before pressure tank  
Project Manager: James Finnan

Reported:  
10/16/23 12:39

Work Order:  
CB11647

Sample Results

Sample: Sample #1

CB11647-01 (DW) Sampled: 09/25/23 12:00

Analyte	Result	Qualifier	LOD	LOQ	MCL	Units	Date Prepared	Date Analyzed	Analyst	Method	Lab Cert Code
<b>Semi-Volatiles</b>											
11-chloroicosafluoro-3-oxaundecane-1-sulfonic acid (11C-PF30UdS)	ND		0.32	1.0		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
9-chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9C-PF30NS)	ND		0.35	1.1		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
4,8-dioxa-3H-perfluorononanoic acid (ADONA)	ND		0.39	1.2		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
hexafluoropropylene oxide dimer acid (HFPO DA)	ND		0.43	1.5		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
N-ethyl perfluorooctanesulfonamidoacetic acid (NEFOSAA)	ND		0.49	1.7		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
n-methyl perfluorooctanesulfonamidoacetic acid (NMeFOSAA)	ND		0.42	1.4		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorobutanesulfonic acid (PFBS)	ND		0.31	1.0		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorodecanoic acid (PFDA)	ND		0.34	1.1		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorododecanoic acid (PFDoA)	ND		0.24	0.80		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluoroheptanoic acid (PFHpA)	ND		0.46	1.6		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanoic acid (PFHxA)	ND		0.49	1.7		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorohexanesulfonic acid (PFHxS)	ND		0.35	1.1		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorononanoic acid (PFNA)	ND		0.48	1.6		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanoic acid (PFOA)	ND		0.51	1.7		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorooctanesulfonic acid (PFOS)	ND		0.32	1.0		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorotetradecanoic acid (PFTA)	ND		0.35	1.1		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluorotridecanoic acid (PFTrDA)	ND		0.45	1.5		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
perfluoroundecanoic acid (PFUnA)	ND		0.31	1.0		ng/L	9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFHxA	103%		Limits: 70-130%				9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-HFPoDA	91%		Limits: 70-130%				9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) C13-PFDoA	90%		Limits: 70-130%				9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2
Surrogate: (SURR) d5-NEFOSAA	84%		Limits: 70-130%				9/28/23 5:13	9/28/23 18:28	RAW	EPA 537.1, Rev 2.0	2

The contents of this report apply to the sample(s) analyzed in accordance with the chain of custody document. No duplication of this report is allowed, except in its entirety.